

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041823\
 Data File : BN024962.D
 Acq On : 17 Apr 2023 19:08
 Operator : CG/JU
 Sample : PB152135BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB152135BL

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 04/18/2023
 Supervised By :Jagrut Upadhyay 04/18/2023

Quant Time: Apr 18 05:32:21 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN040523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 18 05:29:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.940	152	7587	0.400	ng	0.00
7) Naphthalene-d8	10.750	136	20924	0.400	ng	0.00
13) Acenaphthene-d10	14.576	164	11486	0.400	ng	0.00
19) Phenanthrene-d10	17.336	188	25794	0.400	ng	# 0.01
29) Chrysene-d12	21.531	240	19635	0.400	ng	# 0.00
35) Perylene-d12	23.970	264	16452	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.491	112	6872	0.392	ng	0.00
5) Phenol-d6	7.102	99	7820	0.354	ng	0.00
8) Nitrobenzene-d5	9.116	82	5889	0.349	ng	0.01
11) 2-Methylnaphthalene-d10	12.343	152	10945	0.395	ng	0.00
14) 2,4,6-Tribromophenol	16.094	330	1380m	0.267	ng	0.02
15) 2-Fluorobiphenyl	13.208	172	16133	0.368	ng	0.00
27) Fluoranthene-d10	19.363	212	24269	0.415	ng	0.00
31) Terphenyl-d14	19.953	244	21800	0.490	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041823\
Data File : BN024962.D
Acq On : 17 Apr 2023 19:08
Operator : CG/JU
Sample : PB152135BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB152135BL

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 04/18/2023
Supervised By :Jagrut Upadhyay 04/18/2023

Quant Time: Apr 18 05:32:21 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN040523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 18 05:29:26 2023
Response via : Initial Calibration

