

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041823\
 Data File : BN024979.D
 Acq On : 18 Apr 2023 15:31
 Operator : CG/JU
 Sample : 02354-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SP-100-20230413

Quant Time: Apr 18 16:28:11 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN040523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 18 05:29:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.933	152	9734	0.400	ng	0.00
7) Naphthalene-d8	10.750	136	29142	0.400	ng	0.00
13) Acenaphthene-d10	14.576	164	15483	0.400	ng	0.00
19) Phenanthrene-d10	17.323	188	33552	0.400	ng	0.00
29) Chrysene-d12	21.522	240	27651	0.400	ng	0.00
35) Perylene-d12	23.967	264	24070	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.492	112	4260	0.190	ng	0.00
5) Phenol-d6	7.102	99	2888	0.102	ng	0.00
8) Nitrobenzene-d5	9.106	82	9730	0.414	ng	0.00
11) 2-Methylnaphthalene-d10	12.332	152	17234	0.447	ng	0.00
14) 2,4,6-Tribromophenol	16.070	330	2666	0.383	ng	0.00
15) 2-Fluorobiphenyl	13.197	172	26757	0.452	ng	-0.01
27) Fluoranthene-d10	19.355	212	37759	0.496	ng	0.00
31) Terphenyl-d14	19.949	244	33144	0.529	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.339	88	60884	5.692	ng	97
6) bis(2-Chloroethyl)ether	7.355	93	1243	0.043	ng	99
34) Bis(2-ethylhexyl)phtha...	21.414	149	5056	0.101	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041823\
Data File : BN024979.D
Acq On : 18 Apr 2023 15:31
Operator : CG/JU
Sample : 02354-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SP-100-20230413

Quant Time: Apr 18 16:28:11 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN040523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 18 05:29:26 2023
Response via : Initial Calibration

