

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041823\
 Data File : BN024988.D
 Acq On : 18 Apr 2023 22:05
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 19 05:47:32 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN040523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 19 05:45:45 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	111	0.00
2	1,4-Dioxane	0.400	0.444	-11.0	104	0.00
3	n-Nitrosodimethylamine	0.400	0.432	-8.0	105	0.00
4 S	2-Fluorophenol	0.400	0.443	-10.7	109	0.00
5 S	Phenol-d6	0.400	0.463	-15.8	115	0.00
6	bis(2-Chloroethyl)ether	0.400	0.435	-8.7	105	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	121	0.00
8 S	Nitrobenzene-d5	0.400	0.437	-9.2	118	0.00
9	Naphthalene	0.400	0.457	-14.2	120	0.00
10	Hexachlorobutadiene	0.400	0.460	-15.0	117	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.446	-11.5	118	0.00
12	2-Methylnaphthalene	0.400	0.442	-10.5	117	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	116	0.00
14 S	2,4,6-Tribromophenol	0.400	0.442	-10.5	120	0.00
15 S	2-Fluorobiphenyl	0.400	0.465	-16.3	113	0.00
16	Acenaphthylene	0.400	0.449	-12.2	117	0.00
17	Acenaphthene	0.400	0.455	-13.7	114	0.00
18	Fluorene	0.400	0.440	-10.0	109	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	114	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.329	17.8	70	0.00
21	4-Bromophenyl-phenylether	0.400	0.427	-6.7	108	0.00
22	Hexachlorobenzene	0.400	0.454	-13.5	112	0.00
23	Atrazine	0.400	0.400	0.0	106	0.00
24	Pentachlorophenol	0.400	0.581	-45.2#	215	0.00
25	Phenanthrene	0.400	0.433	-8.2	109	0.00
26	Anthracene	0.400	0.419	-4.7	110	0.00
27 SURR	Fluoranthene-d10	0.400	0.436	-9.0	111	0.00
28	Fluoranthene	0.400	0.421	-5.2	108	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	98	0.00
30	Pyrene	0.400	0.521	-30.3#	108	0.00
31 S	Terphenyl-d14	0.400	0.506	-26.5#	103	0.00
32	Benzo(a)anthracene	0.400	0.436	-9.0	91	0.00
33	Chrysene	0.400	0.469	-17.2	99	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.411	-2.7	96	0.00
35 I	Perylene-d12	0.400	0.400	0.0	84	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.368	8.0	68	0.00
37	Benzo(b)fluoranthene	0.400	0.439	-9.7	81	0.00
38	Benzo(k)fluoranthene	0.400	0.447	-11.7	81	0.00
39 C	Benzo(a)pyrene	0.400	0.421	-5.2	78	0.00
40	Dibenzo(a,h)anthracene	0.400	0.372	7.0	69	0.00
41	Benzo(g,h,i)perylene	0.400	0.369	7.8	68	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4