

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041823\
 Data File : BN024992.D
 Acq On : 19 Apr 2023 01:04
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 19 05:48:33 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN040523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 19 05:45:45 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	114	0.00
2	1,4-Dioxane	0.400	0.449	-12.2	107	0.00
3	n-Nitrosodimethylamine	0.400	0.451	-12.7	112	0.00
4 S	2-Fluorophenol	0.400	0.427	-6.7	107	0.00
5 S	Phenol-d6	0.400	0.444	-11.0	113	0.00
6	bis(2-Chloroethyl)ether	0.400	0.435	-8.7	107	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	122	0.00
8 S	Nitrobenzene-d5	0.400	0.440	-10.0	119	0.00
9	Naphthalene	0.400	0.457	-14.2	120	0.00
10	Hexachlorobutadiene	0.400	0.479	-19.7	123	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.445	-11.2	118	0.00
12	2-Methylnaphthalene	0.400	0.440	-10.0	117	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	117	0.00
14 S	2,4,6-Tribromophenol	0.400	0.422	-5.5	116	0.00
15 S	2-Fluorobiphenyl	0.400	0.463	-15.8	114	0.00
16	Acenaphthylene	0.400	0.439	-9.7	116	0.00
17	Acenaphthene	0.400	0.452	-13.0	114	0.00
18	Fluorene	0.400	0.436	-9.0	109	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	112	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.332	17.0	71	0.00
21	4-Bromophenyl-phenylether	0.400	0.441	-10.2	110	0.00
22	Hexachlorobenzene	0.400	0.470	-17.5	114	0.00
23	Atrazine	0.400	0.390	2.5	102	0.00
24	Pentachlorophenol	0.400	0.551	-37.8#	194	0.00
25	Phenanthrene	0.400	0.436	-9.0	108	0.00
26	Anthracene	0.400	0.417	-4.2	108	0.00
27 SURR	Fluoranthene-d10	0.400	0.443	-10.7	111	0.00
28	Fluoranthene	0.400	0.433	-8.2	109	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	96	0.00
30	Pyrene	0.400	0.530	-32.5#	107	0.00
31 S	Terphenyl-d14	0.400	0.514	-28.5#	102	0.00
32	Benzo(a)anthracene	0.400	0.431	-7.7	88	0.00
33	Chrysene	0.400	0.462	-15.5	95	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.380	5.0	87	0.00
35 I	Perylene-d12	0.400	0.400	0.0	78	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.380	5.0	65	-0.01
37	Benzo(b)fluoranthene	0.400	0.460	-15.0	79	0.00
38	Benzo(k)fluoranthene	0.400	0.462	-15.5	78	0.00
39 C	Benzo(a)pyrene	0.400	0.410	-2.5	71	0.00
40	Dibenzo(a,h)anthracene	0.400	0.382	4.5	66	-0.01
41	Benzo(g,h,i)perylene	0.400	0.392	2.0	67	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4