



Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN041823\  
Data File : BN024969.D  
Acq On : 18 Apr 2023 09:12  
Operator : CG/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270-SIM-BN040523.M  
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
Last Update : Wed Apr 19 05:45:45 2023






AutoFind: Scans 2477, 2478, 2479; Background Corrected with Scan 2469

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 56.1         | 306849     | PASS                |
| 68             | 69              | 0.00            | 2               | 1.8          | 4546       | PASS                |
| 69             | 198             | 0.00            | 100             | 45.2         | 247104     | PASS                |
| 70             | 69              | 0.00            | 2               | 0.6          | 1461       | PASS                |
| 127            | 198             | 10              | 80              | 48.6         | 265963     | PASS                |
| 197            | 198             | 0.00            | 2               | 0.4          | 2369       | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 547285     | PASS                |
| 199            | 198             | 5               | 9               | 6.8          | 37472      | PASS                |
| 275            | 198             | 10              | 60              | 25.4         | 138816     | PASS                |
| 365            | 198             | 1               | 100             | 3.4          | 18489      | PASS                |
| 441            | 198             | 0.01            | 100             | 14.8         | 80963      | PASS                |
| 442            | 442             | 50              | 100             | 100.0        | 429419     | PASS                |
| 443            | 442             | 15              | 24              | 21.6         | 92923      | PASS                |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN041823\  
 Data File : BN024969.D  
 Acq On : 18 Apr 2023 09:12  
 Operator : CG/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 DFTPP

Quant Time: Apr 19 07:28:39 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Apr 19 07:28:34 2023  
 Response via : Initial Calibration



TIC: BN024969.D\data.ms

**(70) Pentachlorophenol (C)**

16.975min ( 0.000) 22837.95 ng




response 697191

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 62.20  | 63.82  |
| 264.00 | 61.60  | 63.46  |
| 0.00   | 0.00   | 0.00   |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN041823\  
Data File : BN024969.D  
Acq On : 18 Apr 2023 09:12  
Operator : CG/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP

Quant Time: Apr 19 07:28:39 2023  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Apr 19 07:28:34 2023  
Response via : Initial Calibration



TIC: BN024969.D\data.ms

(77) Benzidine

19.575min ( 0.000) 1784831.23 ng

response 3191049

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 15.50  | 13.87  |
| 183.00 | 13.20  | 11.59  |
| 0.00   | 0.00   | 0.00   |

# DDT Breakdown

| Date          | Instrument Name  | DFTPP Data File    |
|---------------|------------------|--------------------|
| 4/18/2023     | BNA_N            | <u>BN024969.D</u>  |
|               |                  |                    |
|               |                  |                    |
| Compound Name | Response         | Retention Time     |
|               |                  |                    |
| DDT           | 1854807          | 20.816             |
| DDD           | 48479            | 20.369             |
| DDE           | 8318             | 19.857             |
|               |                  |                    |
|               |                  |                    |
| SUM(DDD+DDE)  | SUM(DDT+DDD+DDE) | % Breakdown Of DDT |
|               |                  |                    |
| 56797         | 1911604          | 2.97               |
|               |                  |                    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP