



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041823\
Data File : BN024987.D
Acq On : 18 Apr 2023 21:29
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN040523.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Apr 19 05:45:45 2023






AutoFind: Scans 2477, 2478, 2479; Background Corrected with Scan 2469

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	52.7	328094	PASS
68	69	0.00	2	1.7	4560	PASS
69	198	0.00	100	43.3	269205	PASS
70	69	0.00	2	0.5	1420	PASS
127	198	10	80	47.6	295915	PASS
197	198	0.00	2	0.6	3897	PASS
198	198	100	100	100.0	622272	PASS
199	198	5	9	6.4	39952	PASS
275	198	10	60	25.1	156213	PASS
365	198	1	100	3.8	23533	PASS
441	198	0.01	100	13.9	86637	PASS
442	442	50	100	100.0	542549	PASS
443	442	15	24	19.1	103683	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041823\
Data File : BN024987.D
Acq On : 18 Apr 2023 21:29
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Apr 19 07:29:40 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 19 07:28:34 2023
Response via : Initial Calibration



TIC: BN024987.D\data.ms

(70) Pentachlorophenol (C)

16.975min (+ 0.000) 24148.56 ng




response 851287

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	62.26
264.00	61.60	66.64
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041823\
 Data File : BN024987.D
 Acq On : 18 Apr 2023 21:29
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Quant Time: Apr 19 07:29:40 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 19 07:28:34 2023
 Response via : Initial Calibration



TIC: BN024987.D\data.ms

(77) Benzidine

19.575min (+ 0.000) 0.00 ng

response 4579985

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.06
183.00	13.20	11.67
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
4/18/2023	BNA_N	<u>BN024987.D</u>
Compound Name	Response	Retention Time
DDT	2029053	20.804
DDD	245924	20.363
DDE	4200	19.851
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
250124	2279177	10.97

Instrument :
BNA_N
ClientSampleId :
DFTPP