

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN042021\
 Data File : BN014329.D
 Acq On : 19 Apr 2021 17:52
 Operator : CG/JU
 Sample : PB135369TB
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLEB369

Quant Time: Apr 20 02:00:03 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-BN041521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Apr 19 15:27:04 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.69	152	127073	20.00	ng/ul	0.00
20) Naphthalene-d8	10.47	136	500747	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.33	164	281061	20.00	ng/ul	0.00
64) Phenanthrene-d10	17.07	188	568365	20.00	ng/ul	0.00
79) Chrysene-d12	21.27	240	581707	20.00	ng/ul	0.00
88) Perylene-d12	23.50	264	604432	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.22	96	29196	8.74	ng/uL	0.00
4) Pyridine-d5	3.62	84	372911	42.16	ng/ul	0.00
7) Phenol-d5	6.86	99	445701	41.95	ng/ul	0.00
9) Bis-(2-Chloroethyl)ether-d	7.03	67	280981	44.17	ng/ul	0.00
11) 2-Chlorophenol-d4	7.23	132	373528	44.70	ng/ul	0.00
15) 4-Methylphenol-d8	8.39	113	337673	41.93	ng/ul	0.00
21) Nitrobenzene-d5	8.84	128	167776	46.36	ng/ul	0.00
24) 2-Nitrophenol-d4	9.56	143	158915	44.85	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.09	165	321825	42.16	ng/ul	0.00
31) 4-Chloroaniline-d4	10.60	131	467989	41.32	ng/ul	0.00
46) Dimethylphthalate-d6	13.74	166	972799	47.78	ng/ul	0.00
49) Acenaphthylene-d8	14.02	160	1232527	50.32	ng/ul	0.00
54) 4-Nitrophenol-d4	14.52	143	141847	35.92	ng/ul	0.00
60) Fluorene-d10	15.32	176	867394	48.19	ng/ul	0.00
65) 4,6-Dinitro-2-methylphenol	15.44	200	102127	33.58	ng/ul	0.00
73) Anthracene-d10	17.17	188	1360405	51.28	ng/ul	0.00
81) Pyrene-d10	19.47	212	1558271	55.18	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.36	264	1726814	53.85	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN042021\
Data File : BN014329.D
Acq On : 19 Apr 2021 17:52
Operator : CG/JU
Sample : PB135369TB
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SLEB369

Quant Time: Apr 20 02:00:03 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-BN041521.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Apr 19 15:27:04 2021
Response via : Initial Calibration

