

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN042021\
 Data File : BN014330.D
 Acq On : 19 Apr 2021 19:08
 Operator : CG/JU
 Sample : PB135369BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleID :
 SLCS369

Manual Integrations
 APPROVED

mohammad
 4/21/2021 8:38:12 AM

Quant Time: Apr 20 02:02:10 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-BN041521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Apr 19 15:27:04 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.69	152	151843	20.00	ng/ul	0.00
20) Naphthalene-d8	10.47	136	608294	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.33	164	345153	20.00	ng/ul	0.00
64) Phenanthrene-d10	17.07	188	704251	20.00	ng/ul	0.00
79) Chrysene-d12	21.27	240	744537	20.00	ng/ul	0.00
88) Perylene-d12	23.50	264	715129	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.22	96	23861	5.98	ng/uL	0.00
4) Pyridine-d5	3.62	84	307937	29.13	ng/ul	0.00
7) Phenol-d5	6.86	99	388086	30.57	ng/ul	0.00
9) Bis-(2-Chloroethyl)ether-d	7.03	67	235590	30.99	ng/ul	0.00
11) 2-Chlorophenol-d4	7.23	132	316344	31.68	ng/ul	0.00
15) 4-Methylphenol-d8	8.39	113	294463	30.60	ng/ul	0.00
21) Nitrobenzene-d5	8.84	128	146507	33.33	ng/ul	0.00
24) 2-Nitrophenol-d4	9.56	143	142838	33.18	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.09	165	295464	31.86	ng/ul	0.00
31) 4-Chloroaniline-d4	10.60	131	381061	27.70	ng/ul	0.00
46) Dimethylphthalate-d6	13.74	166	852789	34.11	ng/ul	0.00
49) Acenaphthylene-d8	14.02	160	1121930	37.30	ng/ul	0.00
54) 4-Nitrophenol-d4	14.53	143	158989	32.78	ng/ul	0.00
60) Fluorene-d10	15.33	176	754294	34.13	ng/ul	0.00
65) 4,6-Dinitro-2-methylphenol	15.44	200	122240	32.43	ng/ul	0.00
73) Anthracene-d10	17.17	188	1146596	34.88	ng/ul	0.00
81) Pyrene-d10	19.47	212	1288961	35.66	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.36	264	1375665	36.26	ng/ul	0.00

Target Compounds

					Ovalue
2) 1,4-Dioxane	3.26	88	48053	11.824	ng/uL# 96
5) Pyridine	3.64	79	335711	30.906	ng/ul 95
6) Benzaldehyde	6.84	77	225480	35.181	ng/ul 99
8) Phenol	6.89	94	429760	31.495	ng/ul 99
10) Bis(2-Chloroethyl)ether	7.12	93	337883	31.485	ng/ul 99
12) 2-Chlorophenol	7.26	128	343240	32.153	ng/ul 97
13) 2-Methylphenol	8.13	108	307867	31.115	ng/ul 99
14) 2,2'-oxybis(1-Chloropropan	8.23	45	367957	31.018	ng/ul 99
16) Acetophenone	8.51	105	514529	31.813	ng/ul 99
17) N-Nitroso-di-n-propylamine	8.50	70	251735	33.476	ng/ul 99
18) 4-Methylphenol	8.46	108	341148	31.836	ng/ul 99
19) Hexachloroethane	8.76	117	146591	32.668	ng/ul 98
22) Nitrobenzene	8.89	77	383527	32.983	ng/ul 99
23) Isophorone	9.40	82	659193	33.708	ng/ul 100
25) 2-Nitrophenol	9.59	139	170504	33.523	ng/ul 97
26) 2,4-Dimethylphenol	9.65	107	358673	31.676	ng/ul 100
27) Bis(2-Chloroethoxy)methane	9.89	93	428951	32.196	ng/ul 100
29) 2,4-Dichlorophenol	10.12	162	310466	32.625	ng/ul 98
30) Naphthalene	10.52	128	1067053	31.650	ng/ul 99
32) 4-Chloroaniline	10.63	127	400340	28.282	ng/ul 99
33) Hexachlorobutadiene	10.81	225	197427	31.776	ng/ul 94
34) Caprolactam	11.39	113	89006m	33.556	ng/ul

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN042021\
 Data File : BN014330.D
 Acq On : 19 Apr 2021 19:08
 Operator : CG/JU
 Sample : PB135369BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS369

Manual Integrations
 APPROVED

mohammad
 4/21/2021 8:38:12 AM

Quant Time: Apr 20 02:02:10 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-BN041521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Apr 19 15:27:04 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
35) 4-Chloro-3-methylphenol	11.76	107	317272	33.730	ng/ul	98
36) 2-Methylnaphthalene	12.14	142	745049	32.433	ng/ul	100
37) 1-Methylnaphthalene	12.36	142	715073	31.260	ng/ul	98
39) 1,2,4,5-Tetrachlorobenzene	12.51	216	381214	32.632	ng/ul	100
40) Hexachlorocyclopentadiene	12.49	237	214210	29.527	ng/ul	98
41) 2,4,6-Trichlorophenol	12.75	196	223493	33.100	ng/ul	94
42) 2,4,5-Trichlorophenol	12.82	196	254794	34.004	ng/ul	99
43) 1,1'-Biphenyl	13.16	154	935844	32.967	ng/ul	99
44) 2-Chloronaphthalene	13.20	162	731794	32.978	ng/ul	98
45) 2-Nitroaniline	13.40	65	194306	37.771	ng/ul	95
47) Dimethylphthalate	13.79	163	918730	35.393	ng/ul	100
48) 2,6-Dinitrotoluene	13.90	165	178020	39.106	ng/ul	100
50) Acenaphthylene	14.05	152	1094220	34.189	ng/ul	100
51) 3-Nitroaniline	14.23	138	180896	32.250	ng/ul	96
52) Acenaphthene	14.39	153	785750	33.657	ng/ul	98
53) 2,4-Dinitrophenol	14.44	184	80359	30.242	ng/ul	95
55) 4-Nitrophenol	14.54	109	142006	34.174	ng/ul	99
56) Dibenzofuran	14.73	168	1098825	33.821	ng/ul	99
57) 2,4-Dinitrotoluene	14.69	165	258017	39.235	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	14.96	232	211114	35.812	ng/ul	95
59) Diethylphthalate	15.16	149	910026	34.272	ng/ul	99
61) Fluorene	15.38	166	913595	34.689	ng/ul	99
62) 4-Chlorophenyl-phenylether	15.38	204	453682	34.981	ng/ul	99
63) 4-Nitroaniline	15.40	138	192671	35.288	ng/ul	98
66) 4,6-Dinitro-2-methylphenol	15.46	198	130703	32.749	ng/ul#	96
67) N-Nitrosodiphenylamine	15.59	169	770473	35.161	ng/ul	99
68) 4-Bromophenyl-phenylether	16.27	248	258844	34.742	ng/ul	94
69) Hexachlorobenzene	16.38	284	317954	35.752	ng/ul	99
70) Atrazine	16.55	200	265367	34.077	ng/ul	97
71) Pentachlorophenol	16.73	266	166502	32.200	ng/ul	90
72) Phenanthrene	17.12	178	1428396	34.739	ng/ul	98
74) Anthracene	17.21	178	1463958	35.052	ng/ul	99
75) 1,2,3,4-Tetrachlorobenzene	13.12	216	380430	32.804	ng/uL	97
76) Pentachlorobenzene	14.65	250	360356	31.263	ng/uL	93
77) Carbazole	17.47	167	1299047	34.384	ng/ul	99
78) Di-n-butylphthalate	18.05	149	1480434	36.215	ng/ul	99
80) Fluoranthene	19.14	202	1628091	35.649	ng/ul	98
82) Pyrene	19.50	202	1722137	35.413	ng/ul	98
83) Butylbenzylphthalate	20.42	149	616820	38.630	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.19	252	464010	34.846	ng/ul	94
85) Benzo(a)anthracene	21.26	228	1707749	35.498	ng/ul	99
86) Bis(2-ethylhexyl)phthalate	21.20	149	998772	38.378	ng/ul	99
87) Chrysene	21.31	228	1683194	36.096	ng/ul	95
89) Di-n-octyl phthalate	22.07	149	1584215	35.481	ng/ul	100
90) Benzo(b)fluoranthene	22.83	252	1752636	36.634	ng/ul	99
91) Benzo(k)fluoranthene	22.88	252	1687513	35.432	ng/ul	99
93) Benzo(a)pyrene	23.40	252	1558908	36.829	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	25.75	276	2053878	36.610	ng/ul	99
95) Dibenzo(a,h)anthracene	25.76	278	1736552	37.495	ng/ul	97
96) Benzo(g,h,i)perylene	26.43	276	1642095	36.219	ng/ul	99

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN042021\
Data File : BN014330.D
Acq On : 19 Apr 2021 19:08
Operator : CG/JU
Sample : PB135369BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SLCS369


Manual Integrations
APPROVED

mohammad
4/21/2021 8:38:12 AM

Quant Time: Apr 20 02:02:10 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-BN041521.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Apr 19 15:27:04 2021
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
--------------------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2

1,4-Dioxane

Concen: 11.824 ng/uL

RT: 3.26 min Scan# 29

Delta R.T. 0.00 min

Lab File: BN014330.D

Acq: 19 Apr 2021 19:08

Instrument :

BNA_N





ClientSampled :

SLCS369

Tot Ion:	88	Resp:	48053
Ion	Ratio	Lower	Upper
88	100		
43	29.7	19.6	29.4#
58	63.5	51.8	77.6

Manual Integrations
APPROVED

mohammad
4/21/2021 8:38:12 AM



#5

Pyridine

Concen: 30.906 ng/uL




RT: 3.64 min Scan# 94


Delta R.T. 0.00 min

Lab File: BN014330.D

Acq: 19 Apr 2021 19:08

Tgt Ion:	79	Resp:	335711
Ion	Ratio	Lower	Upper
79	100		
52	63.3	54.3	81.5
51	30.4	25.9	38.9





#6

Benzaldehyde

Concen: 35.181 ng/ul

RT: 6.84 min Scan# 638

Delta R.T. 0.00 min

Lab File: BN014330.D

Acq: 19 Apr 2021 19:08

Instrument :

BNA_N





ClientSampleId :

SLCS369

Tot Ion:	77	Resp:	225480
Ion	Ratio	Lower	Upper
77	100		
105	98.1	78.2	117.4
106	94.4	74.5	111.7

Manual Integrations
APPROVED

mohammad
4/21/2021 8:38:12 AM



#8

Phenol

Concen: 31.495 ng/ul




RT: 6.89 min Scan# 647


Delta R.T. 0.01 min

Lab File: BN014330.D

Acq: 19 Apr 2021 19:08

Tgt Ion:	94	Resp:	429760
Ion	Ratio	Lower	Upper
94	100		
65	27.3	22.5	33.7
66	36.1	29.0	43.6









#10
Bis(2-Chloroethyl)ether
Concen: 31.485 ng/ul
RT: 7.12 min Scan# 686
Delta R.T. 0.00 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Instrument :
BNA_N
ClientSampled :
SLCS369

Tot Ion:	93	Resp:	337883
Ion	Ratio	Lower	Upper
93	100		
63	71.7	57.1	85.7
95	32.0	26.2	39.2




Manual Integrations
APPROVED


mohammad
4/21/2021 8:38:12 AM



#12
2-Chlorophenol
Concen: 32.153 ng/ul
RT: 7.26 min Scan# 710
Delta R.T. 0.01 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Tgt Ion:	128	Resp:	343240
Ion	Ratio	Lower	Upper
128	100		
64	41.7	34.3	51.5
130	31.7	27.6	41.4





#13

2-Methylphenol

Concen: 31.115 ng/ul

RT: 8.13 min Scan# 858

Delta R.T. 0.01 min

Lab File: BN014330.D

Acq: 19 Apr 2021 19:08

Instrument :

BNA_N

ClientSampled :

SLCS369

Tot Ion: 108 Resp: 307867

Ion Ratio Lower Upper

108 100


107 90.5 71.6 107.4

Manual Integrations

APPROVED

mohammad

4/21/2021 8:38:12 AM



Abundance Ion 108.00 (107.70 to 108.70): E

Ion 107.00 (106.70 to 107.70): E

8.13

200000



150000

100000

50000

0

Time--> 8.05 8.10 8.15 8.20



#14

2,2'-oxybis(1-Chloropropane)

Concen: 31.018 ng/ul

RT: 8.23 min Scan# 874

Delta R.T. 0.01 min

Lab File: BN014330.D

Acq: 19 Apr 2021 19:08


Tgt Ion: 45 Resp: 367957

Ion Ratio Lower Upper

45 100

77 18.2 14.2 21.2

79 14.1 11.1 16.7



Abundance Ion 45.00 (44.70 to 45.70): BN0

Ion 77.00 (76.70 to 77.70): BN0

Ion 79.00 (78.70 to 79.70): BN0

8.23

200000


150000


100000

50000

0

Time--> 8.15 8.20 8.25









#16
Acetophenone
Concen: 31.813 ng/ul
RT: 8.51 min Scan# 922
Delta R.T. 0.01 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Instrument :
BNA_N
ClientSampleId :
SLCS369

Tot Ion:	105	Resp:	514529
Ion Ratio	Lower	Upper	
105	100		
77	77.6	62.0	93.0
51	22.5	19.4	29.0




Manual Integrations
APPROVED


mohammad
4/21/2021 8:38:12 AM



#17
N-Nitroso-di-n-propylamine
Concen: 33.476 ng/ul
RT: 8.50 min Scan# 920
Delta R.T. 0.01 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Tgt Ion:	70	Resp:	251735
Ion Ratio	Lower	Upper	
70	100		
42	51.0	41.4	62.0
101	10.0	8.1	12.1
130	21.7	17.1	25.7





#18

4-Methylphenol

Concen: 31.836 ng/ul

RT: 8.46 min Scan# 913

Delta R.T. 0.01 min

Lab File: BN014330.D

Acq: 19 Apr 2021 19:08

Instrument :

BNA_N

ClientSampled :

SLCS369

Tot Ion:108 Resp: 341148

Ion Ratio Lower Upper

108 100


107 111.7 88.2 132.4

Manual Integrations

APPROVED

mohammad

4/21/2021 8:38:12 AM



Abundance Ion 108.00 (107.70 to 108.70): E

Ion 107.00 (106.70 to 107.70): E

250000

200000

150000



100000

50000

0

8.40 8.45 8.50 8.55

Time-->



#19

Hexachloroethane

Concen: 32.668 ng/ul

RT: 8.76 min Scan# 965

Delta R.T. 0.00 min

Lab File: BN014330.D

Acq: 19 Apr 2021 19:08


Tgt Ion:117 Resp: 146591

Ion Ratio Lower Upper

117 100

201 95.5 78.0 117.0

199 59.3 49.2 73.8



Abundance Ion 117.00 (116.70 to 117.70): E

Ion 201.00 (200.70 to 201.70): E

Ion 199.00 (198.70 to 199.70): E

120000

100000

80000

60000


40000


20000

0

8.70 8.75 8.80

Time-->









#22
Nitrobenzene
Concen: 32.983 ng/ul
RT: 8.89 min Scan# 986
Delta R.T. 0.01 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Instrument :
BNA_N
ClientSampleId :
SLCS369

Tot Ion	Ratio	Lower	Upper
77	100		
123	45.5	35.6	53.4
65	12.6	10.2	15.4




Manual Integrations
APPROVED


mohammad
4/21/2021 8:38:12 AM



#23
Isophorone
Concen: 33.708 ng/ul
RT: 9.40 min Scan# 1074
Delta R.T. 0.00 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Tgt Ion	Ratio	Lower	Upper
82	100		
95	6.9	5.5	8.3
138	17.1	13.5	20.3





#25

2-Nitrophenol

Concen: 33.523 ng/ul

RT: 9.59 min Scan# 1106

Delta R.T. 0.01 min

Lab File: BN014330.D

Acq: 19 Apr 2021 19:08

Instrument :

BNA_N





ClientSampled :

SLCS369

Tot Ion:	139	Resp:	170504
Ion Ratio	Lower	Upper	
139	100		
65	48.8	41.1	61.7
109	29.8	23.7	35.5

Manual Integrations
APPROVED

mohammad
4/21/2021 8:38:12 AM



#26

2,4-Dimethylphenol

Concen: 31.676 ng/ul




RT: 9.65 min Scan# 1116


Delta R.T. 0.00 min

Lab File: BN014330.D

Acq: 19 Apr 2021 19:08

Tgt Ion:	107	Resp:	358673
Ion Ratio	Lower	Upper	
107	100		
121	50.9	40.7	61.1
122	86.1	69.1	103.7





#27

Bis(2-Chloroethoxy)methane

Concen: 32.196 ng/ul

RT: 9.89 min Scan# 1157

Delta R.T. 0.01 min

Lab File: BN014330.D

Acq: 19 Apr 2021 19:08

Instrument :

BNA_N





ClientSampleId :

SLCS369

Tot Ion:	93	Resp:	428951
Ion Ratio	Lower	Upper	
93	100		
95	31.9	25.6	38.4
123	11.4	9.0	13.4

Manual Integrations
APPROVED

mohammad
4/21/2021 8:38:12 AM



#29

2,4-Dichlorophenol

Concen: 32.625 ng/ul




RT: 10.12 min Scan# 1196


Delta R.T. 0.01 min

Lab File: BN014330.D

Acq: 19 Apr 2021 19:08

Tgt Ion:	162	Resp:	310466
Ion Ratio	Lower	Upper	
162	100		
164	63.2	51.4	77.2
98	35.0	29.1	43.7









#30
Naphthalene
Concen: 31.650 ng/ul
RT: 10.52 min Scan# 1264
Delta R.T. 0.00 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Instrument :
BNA_N
ClientSampleId :
SLCS369

Tot Ion	Ratio	Lower	Upper
128	100		
129	10.7	8.7	13.1
127	12.9	10.6	15.8




Manual Integrations
APPROVED


mohammad
4/21/2021 8:38:12 AM



#32
4-Chloroaniline
Concen: 28.282 ng/ul
RT: 10.63 min Scan# 1282
Delta R.T. 0.00 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Tgt Ion	Ratio	Lower	Upper
127	100		
129	32.7	25.8	38.6









#33
Hexachlorobutadiene
Concen: 31.776 ng/ul
RT: 10.81 min Scan# 1313
Delta R.T. 0.00 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Instrument :
BNA_N
ClientSampleId :
SLCS369

Tot Ion	Ratio	Lower	Upper
225	100		
223	63.9	55.0	82.6
227	66.7	50.0	75.0




Manual Integrations
APPROVED


mohammad
4/21/2021 8:38:12 AM



#34
Caprolactam
Concen: 33.556 ng/ul m
RT: 11.39 min Scan# 1411
Delta R.T. 0.01 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Tgt Ion	Ratio	Lower	Upper
113	100		
55	140.5	115.4	173.2
56	110.5	87.2	130.8









#35
4-Chloro-3-methylphenol
Concen: 33.730 ng/ul
RT: 11.76 min Scan# 1474
Delta R.T. 0.01 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Instrument :
BNA_N
ClientSampled :
SLCS369

Tot Ion	Ratio	Lower	Upper
107	100		
144	26.8	21.0	31.4
142	81.4	66.4	99.6




Manual Integrations
APPROVED


mohammad
4/21/2021 8:38:12 AM



#36
2-Methylnaphthalene
Concen: 32.433 ng/ul
RT: 12.14 min Scan# 1539
Delta R.T. 0.01 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Tgt Ion	Ratio	Lower	Upper
142	100		
141	88.6	71.0	106.6









#37
1-Methylnaphthalene
Concen: 31.260 ng/ul
RT: 12.36 min Scan# 1576
Delta R.T. 0.00 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Instrument :
BNA_N
ClientSampleId :
SLCS369

Tot Ion	Ratio	Lower	Upper
142	100		
141	91.7	74.8	112.2
116	4.0	3.3	4.9




Manual Integrations
APPROVED


mohammad
4/21/2021 8:38:12 AM



#39
1,2,4,5-Tetrachlorobenzene
Concen: 32.632 ng/ul
RT: 12.51 min Scan# 1602
Delta R.T. 0.00 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Tgt Ion	Ratio	Lower	Upper
216	100		
214	78.1	62.6	93.8
179	21.3	16.8	25.2
108	18.2	14.4	21.6









#40
Hexachlorocyclopentadiene
Concen: 29.527 ng/ul
RT: 12.49 min Scan# 1599
Delta R.T. 0.00 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Instrument :
BNA_N
ClientSampleId :
SLCS369

Tot Ion:	237	Resp:	214210
Ion Ratio	Lower	Upper	
237	100		
235	61.1	50.2	75.4
272	12.6	9.7	14.5




Manual Integrations
APPROVED


mohammad
4/21/2021 8:38:12 AM



#41
2,4,6-Trichlorophenol
Concen: 33.100 ng/ul
RT: 12.75 min Scan# 1643
Delta R.T. 0.00 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Tgt Ion:	196	Resp:	223493
Ion Ratio	Lower	Upper	
196	100		
198	99.3	73.8	110.6
200	31.8	24.7	37.1









#42
 2,4,5-Trichlorophenol
 Concen: 34.004 ng/ul
 RT: 12.82 min Scan# 1655
 Delta R.T. 0.00 min
 Lab File: BN014330.D
 Acq: 19 Apr 2021 19:08

Instrument :
 BNA_N
 ClientSampled :
 SLCS369

Tot Ion	Ratio	Resp	Lower	Upper
196	100	254794		
198	97.1	78.6	117.8	
200	30.9	24.7	37.1	




Manual Integrations
 APPROVED


mohammad
 4/21/2021 8:38:12 AM



#43
 1,1'-Biphenyl
 Concen: 32.967 ng/ul
 RT: 13.16 min Scan# 1713
 Delta R.T. 0.01 min
 Lab File: BN014330.D
 Acq: 19 Apr 2021 19:08

Tgt Ion	Ratio	Resp	Lower	Upper
154	100	935844		
153	38.3	31.4	47.0	
76	12.3	9.9	14.9	









#44
2-Chloronaphthalene
Concen: 32.978 ng/ul
RT: 13.20 min Scan# 1719
Delta R.T. 0.00 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Instrument :
BNA_N
ClientSampleId :
SLCS369

Tot Ion	Ion	Ratio	Lower	Upper
162	100			
164	34.0	26.2	39.4	
127	38.9	30.6	46.0	




Manual Integrations
APPROVED


mohammad
4/21/2021 8:38:12 AM



#45
2-Nitroaniline
Concen: 37.771 ng/ul
RT: 13.40 min Scan# 1754
Delta R.T. 0.01 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Tgt Ion	Ion	Ratio	Lower	Upper
65	100			
92	71.5	56.7	85.1	
138	115.1	84.9	127.3	





#47

Dimethylphthalate

Concen: 35.393 ng/ul

RT: 13.79 min Scan# 1820

Delta R.T. 0.01 min

Lab File: BN014330.D

Acq: 19 Apr 2021 19:08

Instrument :

BNA_N


ClientSampleId :

SLCS369




Tot Ion:	163	Resp:	918730
Ion Ratio	Lower	Upper	
163	100		
194	4.2	3.3	4.9
164	9.7	7.8	11.8

Manual Integrations
APPROVED

mohammad
4/21/2021 8:38:12 AM



Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E



#48

2,6-Dinitrotoluene

Concen: 39.106 ng/ul


RT: 13.90 min Scan# 1839

Delta R.T. 0.00 min



Lab File: BN014330.D


Acq: 19 Apr 2021 19:08

Tgt Ion:	165	Resp:	178020
Ion Ratio	Lower	Upper	
165	100		
89	62.5	49.8	74.6
121	22.9	18.2	27.2



Abundance Ion 165.00 (164.70 to 165.70): E
Ion 89.00 (88.70 to 89.70): BNO
Ion 121.00 (120.70 to 121.70): E





#50

Acenaphthylene

Concen: 34.189 ng/ul

RT: 14.05 min Scan# 1864

Delta R.T. 0.01 min

Lab File: BN014330.D

Acq: 19 Apr 2021 19:08

Instrument :

BNA_N

ClientSampled :

SLCS369

Tot Ion:152 Resp: 1094220

Ion Ratio Lower Upper


152 100

151 19.6 15.5 23.3

153 13.1 10.5 15.7

Manual Integrations
APPROVED




mohammad
4/21/2021 8:38:12 AM



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#51

3-Nitroaniline

Concen: 32.250 ng/ul

RT: 14.23 min Scan# 1895

Delta R.T. 0.01 min

Lab File: BN014330.D

Acq: 19 Apr 2021 19:08


Tgt Ion:138 Resp: 180896

Ion Ratio Lower Upper

138 100

108 9.5 8.3 12.5



92 112.1 93.7 140.5




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BNC





#52

Acenaphthene

Concen: 33.657 ng/ul

RT: 14.39 min Scan# 1922

Delta R.T. 0.00 min

Lab File: BN014330.D

Acq: 19 Apr 2021 19:08

Instrument :

BNA_N


ClientSampleId :

SLCS369




Tot Ion:	153	Resp:	785750
Ion Ratio	Lower	Upper	
153	100		
152	48.2	39.4	59.2
154	87.0	70.9	106.3

Manual Integrations
APPROVED

mohammad
4/21/2021 8:38:12 AM



Abundance Ion 153.00 (152.70 to 153.70): E
Ion 152.00 (151.70 to 152.70): E
Ion 154.00 (153.70 to 154.70): E



#53

2,4-Dinitrophenol

Concen: 30.242 ng/ul


RT: 14.44 min Scan# 1930

Delta R.T. 0.01 min



Lab File: BN014330.D


Acq: 19 Apr 2021 19:08

Tgt Ion:	184	Resp:	80359
Ion Ratio	Lower	Upper	
184	100		
63	65.5	55.8	83.6
154	63.5	53.5	80.3



Abundance Ion 184.00 (183.70 to 184.70): E
Ion 63.00 (62.70 to 63.70): BNO
Ion 154.00 (153.70 to 154.70): E









#55
4-Nitrophenol
Concen: 34.174 ng/ul
RT: 14.54 min Scan# 1947
Delta R.T. 0.01 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Instrument :
BNA_N
ClientSampleId :
SLCS369

Tot Ion	Ion	Ratio	Resp	Lower	Upper
109	100				
139	128.0	102.8	154.2		
65	127.3	99.8	149.6		




Manual Integrations
APPROVED


mohammad
4/21/2021 8:38:12 AM



#56
Dibenzofuran
Concen: 33.821 ng/ul
RT: 14.73 min Scan# 1979
Delta R.T. 0.00 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
168	100				
139	38.5	30.2	45.4		









#57
 2,4-Dinitrotoluene
 Concen: 39.235 ng/ul
 RT: 14.69 min Scan# 1973
 Delta R.T. 0.00 min
 Lab File: BN014330.D
 Acq: 19 Apr 2021 19:08

Instrument :
 BNA_N
 ClientSampled :
 SLCS369

Tot Ion	Ratio	Resp	Lower	Upper
165	100	258017		
63	44.3		37.4	56.2
182	2.8		2.4	3.6




Manual Integrations
 APPROVED


mohammad
 4/21/2021 8:38:12 AM



#58
 2,3,4,6-Tetrachlorophenol
 Concen: 35.812 ng/ul
 RT: 14.96 min Scan# 2018
 Delta R.T. 0.00 min
 Lab File: BN014330.D
 Acq: 19 Apr 2021 19:08

Tgt Ion	Ratio	Resp	Lower	Upper
232	100	211114		
131	45.0		39.4	59.0
130	2.3		2.3	3.5
166	30.2		25.4	38.2









#59
Diethylphthalate
Concen: 34.272 ng/ul
RT: 15.16 min Scan# 2053
Delta R.T. 0.01 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Instrument :
BNA_N
ClientSampleId :
SLCS369

Tot Ion	Ratio	Lower	Upper
149	100		
177	21.1	17.3	25.9
150	12.1	10.0	15.0




Manual Integrations
APPROVED


mohammad
4/21/2021 8:38:12 AM



#61
Fluorene
Concen: 34.689 ng/ul
RT: 15.38 min Scan# 2090
Delta R.T. 0.00 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Tgt Ion	Ratio	Lower	Upper
166	100		
165	97.5	77.6	116.4
167	13.3	10.6	15.8









#62
 4-Chlorophenyl-phenylether
 Concen: 34.981 ng/ul
 RT: 15.38 min Scan# 2090
 Delta R.T. 0.01 min
 Lab File: BN014330.D
 Acq: 19 Apr 2021 19:08

Instrument :
 BNA_N
 ClientSampled :
 SLCS369

Tot Ion:	204	Resp:	453682
Ion Ratio	Lower	Upper	
204	100		
206	35.4	26.6	40.0
141	60.2	48.2	72.4




Manual Integrations
 APPROVED


mohammad
 4/21/2021 8:38:12 AM



#63
 4-Nitroaniline
 Concen: 35.288 ng/ul
 RT: 15.40 min Scan# 2093
 Delta R.T. 0.01 min
 Lab File: BN014330.D
 Acq: 19 Apr 2021 19:08

Tgt Ion:	138	Resp:	192671
Ion Ratio	Lower	Upper	
138	100		
92	50.6	40.9	61.3
108	79.2	65.8	98.8









#66
 4,6-Dinitro-2-methylphenol
 Concen: 32.749 ng/ul
 RT: 15.46 min Scan# 2103
 Delta R.T. 0.01 min
 Lab File: BN014330.D
 Acq: 19 Apr 2021 19:08

Instrument :
 BNA_N
 ClientSampled :
 SLCS369

Tot Ion	Ratio	Resp	Lower	Upper
198	100	130703		
182	4.1	4.5	6.7#	
77	26.0	22.6	33.8	




Manual Integrations
 APPROVED


mohammad
 4/21/2021 8:38:12 AM



#67
 N-Nitrosodiphenylamine
 Concen: 35.161 ng/ul
 RT: 15.59 min Scan# 2126
 Delta R.T. 0.00 min
 Lab File: BN014330.D
 Acq: 19 Apr 2021 19:08

Tgt Ion	Ratio	Resp	Lower	Upper
169	100	770473		
168	66.5	53.8	80.8	
167	37.2	29.7	44.5	









#68
4-Bromophenyl-phenylether
Concen: 34.742 ng/ul
RT: 16.27 min Scan# 2242
Delta R.T. 0.00 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Instrument :
BNA_N
ClientSampled :
SLCS369

Tot Ion	Ratio	Lower	Upper
248	100		
250	89.7	76.4	114.6
141	71.1	60.4	90.6




Manual Integrations
APPROVED


mohammad
4/21/2021 8:38:12 AM



#69
Hexachlorobenzene
Concen: 35.752 ng/ul
RT: 16.38 min Scan# 2260
Delta R.T. -0.01 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Tgt Ion	Ratio	Lower	Upper
284	100		
142	37.6	29.8	44.6
249	29.8	22.9	34.3





#70

Atrazine

Concen: 34.077 ng/ul

RT: 16.55 min Scan# 2288

Delta R.T. 0.00 min

Lab File: BN014330.D

Acq: 19 Apr 2021 19:08

Instrument :

BNA_N





ClientSampled :

SLCS369

Tot Ion:	200	Resp:	265367
Ion Ratio	Lower	Upper	
200	100		
173	25.5	21.2	31.8
215	47.1	39.5	59.3

Manual Integrations
APPROVED

mohammad
4/21/2021 8:38:12 AM



#71

Pentachlorophenol

Concen: 32.200 ng/ul




RT: 16.73 min Scan# 2319


Delta R.T. 0.00 min

Lab File: BN014330.D

Acq: 19 Apr 2021 19:08

Tgt Ion:	266	Resp:	166502
Ion Ratio	Lower	Upper	
266	100		
264	60.9	55.4	83.2
268	60.3	54.7	82.1





#72

Phenanthrene

Concen: 34.739 ng/ul

RT: 17.12 min Scan# 2385

Delta R.T. 0.00 min

Lab File: BN014330.D

Acq: 19 Apr 2021 19:08

Instrument :

BNA_N

ClientSampleId :

SLCS369

Tot Ion:178 Resp: 1428396

Ion Ratio Lower Upper

178 100


179 16.0 12.2 18.4

176 19.7 15.0 22.4

Manual Integrations
APPROVED

mohammad




4/21/2021 8:38:12 AM



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E



#74

Anthracene

Concen: 35.052 ng/ul

RT: 17.21 min Scan# 2401

Delta R.T. 0.01 min

Lab File: BN014330.D

Acq: 19 Apr 2021 19:08


Tgt Ion:178 Resp: 1463958

Ion Ratio Lower Upper

178 100

179 16.3 12.3 18.5



176 18.6 15.1 22.7




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E





#75

1,2,3,4-Tetrachlorobenzene

Concen: 32.804 ng/uL

RT: 13.12 min Scan# 1706

Delta R.T. 0.00 min

Lab File: BN014330.D

Acq: 19 Apr 2021 19:08

Instrument :

BNA_N


ClientSampleId :

SLCS369




Tot Ion:	216	Resp:	380430
Ion Ratio	Lower	Upper	
216	100		
214	81.7	62.2	93.2
218	49.6	40.2	60.2

Manual Integrations
APPROVED

mohammad
4/21/2021 8:38:12 AM



Abundance Ion 216.00 (215.70 to 216.70): E
Ion 214.00 (213.70 to 214.70): E
Ion 218.00 (217.70 to 218.70): E



#76

Pentachlorobenzene

Concen: 31.263 ng/uL


RT: 14.65 min Scan# 1966

Delta R.T. 0.00 min



Lab File: BN014330.D


Acq: 19 Apr 2021 19:08

Tgt Ion:	250	Resp:	360356
Ion Ratio	Lower	Upper	
250	100		
248	59.6	52.3	78.5
252	61.7	53.0	79.6



Abundance Ion 250.00 (249.70 to 250.70): E
Ion 248.00 (247.70 to 248.70): E
Ion 252.00 (251.70 to 252.70): E









#77
 Carbazole
 Concen: 34.384 ng/ul
 RT: 17.47 min Scan# 2446
 Delta R.T. 0.00 min
 Lab File: BN014330.D
 Acq: 19 Apr 2021 19:08

Instrument :
 BNA_N
 ClientSampleId :
 SLCS369

Tot Ion	Ratio	Lower	Upper
167	100		
166	21.6	16.8	25.2
139	13.0	10.5	15.7




Manual Integrations
 APPROVED


mohammad
 4/21/2021 8:38:12 AM



#78
 Di-n-butylphthalate
 Concen: 36.215 ng/ul
 RT: 18.05 min Scan# 2544
 Delta R.T. 0.00 min
 Lab File: BN014330.D
 Acq: 19 Apr 2021 19:08

Tgt Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.3	10.9
104	5.4	4.6	6.8









#80
Fluoranthene
Concen: 35.649 ng/ul
RT: 19.14 min Scan# 2729
Delta R.T. 0.00 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Instrument :
BNA_N
ClientSampleId :
SLCS369

Tot Ion	Ion	Ratio	Lower	Upper
202	100			
101	10.3	9.0	13.6	
100	8.0	7.0	10.4	




Manual Integrations
APPROVED


mohammad
4/21/2021 8:38:12 AM



#82
Pyrene
Concen: 35.413 ng/ul
RT: 19.50 min Scan# 2791
Delta R.T. 0.00 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Tgt Ion	Ion	Ratio	Lower	Upper
202	100			
101	12.4	10.5	15.7	
100	10.5	8.7	13.1	





#83

Butylbenzylphthalate

Concen: 38.630 ng/ul

RT: 20.42 min Scan# 2946

Delta R.T. 0.00 min

Lab File: BN014330.D

Acq: 19 Apr 2021 19:08

Instrument :

BNA_N





ClientSampleId :

SLCS369

Tot Ion:	149	Resp:	616820
Ion Ratio	Lower	Upper	
149	100		
91	71.3	58.2	87.2
206	21.5	16.5	24.7

Manual Integrations
APPROVED

mohammad
4/21/2021 8:38:12 AM



#84

3,3'-Dichlorobenzidine

Concen: 34.846 ng/ul




RT: 21.19 min Scan# 3078


Delta R.T. 0.00 min

Lab File: BN014330.D

Acq: 19 Apr 2021 19:08

Tgt Ion:	252	Resp:	464010
Ion Ratio	Lower	Upper	
252	100		
254	63.2	55.2	82.8
126	11.0	9.4	14.2









#85
Benzo(a)anthracene
Concen: 35.498 ng/ul
RT: 21.26 min Scan# 3089
Delta R.T. 0.00 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Instrument :
BNA_N
ClientSampleId :
SLCS369

Tot Ion	Ratio	Lower	Upper
228	100		
229	19.6	15.4	23.2
226	28.0	21.8	32.8




Manual Integrations
APPROVED


mohammad
4/21/2021 8:38:12 AM



#86
Bis(2-ethylhexyl)phthalate
Concen: 38.378 ng/ul
RT: 21.20 min Scan# 3079
Delta R.T. 0.00 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Tgt Ion	Ratio	Lower	Upper
149	100		
167	26.9	22.0	33.0
279	4.0	3.0	4.6









#87
Chrysene
Concen: 36.096 ng/ul
RT: 21.31 min Scan# 3098
Delta R.T. 0.00 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Instrument :
BNA_N
ClientSampleId :
SLCS369

Tot Ion	Ratio	Lower	Upper
228	100		
226	28.6	26.0	39.0
229	20.2	16.4	24.6




Manual Integrations
APPROVED


mohammad
4/21/2021 8:38:12 AM



#89
Di-n-octyl phthalate
Concen: 35.481 ng/ul
RT: 22.07 min Scan# 3228
Delta R.T. -0.01 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Tgt Ion:149 Resp: 1584215





#90

Benzo(b)fluoranthene

Concen: 36.634 ng/ul

RT: 22.83 min Scan# 3357

Delta R.T. 0.00 min

Lab File: BN014330.D

Acq: 19 Apr 2021 19:08

Instrument :

BNA_N

ClientSampleId :

SLCS369

Tot Ion:252 Resp: 1752636

Ion Ratio Lower Upper


252 100

253 22.1 17.3 25.9

125 9.7 8.1 12.1

Manual Integrations
APPROVED

mohammad
4/21/2021 8:38:12 AM






Abundance

Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#91

Benzo(k)fluoranthene

Concen: 35.432 ng/ul

RT: 22.88 min Scan# 3365

Delta R.T. 0.01 min

Lab File: BN014330.D

Acq: 19 Apr 2021 19:08


Tgt Ion:252 Resp: 1687513

Ion Ratio Lower Upper

252 100

253 22.6 17.8 26.8

125 9.1 8.2 12.4






Abundance

Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E









#93
Benzo(a)pyrene
Concen: 36.829 ng/u1
RT: 23.40 min Scan# 3454
Delta R.T. 0.00 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Instrument :
BNA_N
ClientSampled :
SLCS369

Tot Ion	Ratio	Lower	Upper
252	100		
253	23.3	17.4	26.2
125	11.2	9.3	13.9




Manual Integrations
APPROVED


mohammad
4/21/2021 8:38:12 AM



#94
Indeno(1,2,3-cd)pyrene
Concen: 36.610 ng/u1
RT: 25.75 min Scan# 3853
Delta R.T. 0.01 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Tgt Ion	Ratio	Lower	Upper
276	100		
138	24.6	19.0	28.4
277	23.5	19.3	28.9









#95
Dibenzo(a,h)anthracene
Concen: 37.495 ng/ul
RT: 25.76 min Scan# 3855
Delta R.T. 0.01 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Instrument :
BNA_N
ClientSampleId :
SLCS369

Tot Ion	Ratio	Lower	Upper
278	100		
139	16.3	12.1	18.1
279	24.3	18.2	27.2

Manual Integrations
APPROVED

mohammad
4/21/2021 8:38:12 AM



#96
Benzo(g,h,i)perylene
Concen: 36.219 ng/ul
RT: 26.43 min Scan# 3969
Delta R.T. 0.01 min
Lab File: BN014330.D
Acq: 19 Apr 2021 19:08

Tgt Ion	Ratio	Lower	Upper
276	100		
138	22.0	17.0	25.6
277	24.2	19.7	29.5

