


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042122\
 Data File : BN019509.D
 Acq On : 20 Apr 2022 12:33
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042122.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Apr 20 17:50:10 2022






AutoFind: Scans 2491, 2492, 2493; Background Corrected with Scan 2484

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.2	188069	PASS
68	69	0.00	2	1.9	3471	PASS
69	198	0.00	100	39.9	186522	PASS
70	69	0.00	2	0.6	1079	PASS
127	198	10	80	46.4	217152	PASS
197	198	0.00	2	0.2	702	PASS
198	198	100	100	100.0	467861	PASS
199	198	5	9	6.9	32211	PASS
275	198	10	60	27.4	128219	PASS
365	198	1	100	3.7	17190	PASS
441	198	0.01	100	12.5	58333	PASS
442	442	50	100	100.0	349760	PASS
443	442	15	24	19.3	67621	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042122\
Data File : BN019509.D
Acq On : 20 Apr 2022 12:33
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Apr 20 15:51:00 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 23 16:56:32 2022
Response via : Initial Calibration



TIC: BN019509.D\data.ms

(70) Pentachlorophenol (C)

16.845min (-0.027) 21693.28 ng




response 572869

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	63.26
264.00	61.60	64.15
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042122\
Data File : BN019509.D
Acq On : 20 Apr 2022 12:33
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Apr 20 15:51:00 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 23 16:56:32 2022
Response via : Initial Calibration



TIC: BN019509.D\data.ms

(77) Benzidine

19.433min (-0.033) 0.00 ng

response 2673541

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.23
183.00	13.20	12.40
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
4/21/2022	BNA_N	<u>BN019509.D</u>
Compound Name	Response	Retention Time
DDT	1556517	20.67
DDD	1987	20.23
DDE	0	19.9
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
1987	1558504	0.13