




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042222\
Data File : BN019519.D
Acq On : 22 Apr 2022 13:08
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 23 07:36:40 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 23 16:56:32 2022
Response via : Initial Calibration



TIC: BN019519.D\data.ms

(70) Pentachlorophenol (C)




16.845min (-0.027) 17585.42 ng

response 419952

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	68.13
264.00	61.60	62.79
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042222\
Data File : BN019519.D
Acq On : 22 Apr 2022 13:08
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 23 07:36:40 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 23 16:56:32 2022
Response via : Initial Calibration



TIC: BN019519.D\data.ms

(77) Benzidine

19.433min (-0.033) 0.00 ng

response 2684062

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.88
183.00	13.20	12.05
0.00	0.00	0.00