

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042222\
 Data File : BN019520.D
 Acq On : 22 Apr 2022 13:45
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 23 05:30:48 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 20 17:50:10 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	130	0.00
2	1,4-Dioxane	0.400	0.410	-2.5	127	0.00
3	n-Nitrosodimethylamine	0.400	0.393	1.8	132	0.00
4 S	2-Fluorophenol	0.400	0.390	2.5	129	0.00
5 S	Phenol-d6	0.400	0.348	13.0	133	0.00
6	bis(2-Chloroethyl)ether	0.400	0.415	-3.7	138	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	134	0.00
8 S	Nitrobenzene-d5	0.400	0.358	10.5	137	-0.01
9	Naphthalene	0.400	0.404	-1.0	134	0.00
10	Hexachlorobutadiene	0.400	0.406	-1.5	132	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.400	0.0	135	0.00
12	2-Methylnaphthalene	0.400	0.399	0.3	135	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	134	0.00
14 S	2,4,6-Tribromophenol	0.400	0.326	18.5	127	0.00
15 S	2-Fluorobiphenyl	0.400	0.402	-0.5	137	0.00
16	Acenaphthylene	0.400	0.372	7.0	135	0.00
17	Acenaphthene	0.400	0.398	0.5	133	0.00
18	Fluorene	0.400	0.393	1.8	131	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	133	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.338	15.5	132	0.00
21	4-Bromophenyl-phenylether	0.400	0.386	3.5	130	0.00
22	Hexachlorobenzene	0.400	0.400	0.0	131	0.00
23	Atrazine	0.400	0.333	16.8	128	0.00
24	Pentachlorophenol	0.400	0.345	13.8	134	0.00
25	Phenanthrene	0.400	0.392	2.0	131	0.00
26	Anthracene	0.400	0.368	8.0	130	0.00
27 SURR	Fluoranthene-d10	0.400	0.392	2.0	135	0.00
28	Fluoranthene	0.400	0.390	2.5	135	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	146	0.00
30	Pyrene	0.400	0.404	-1.0	135	0.00
31 S	Terphenyl-d14	0.400	0.402	-0.5	141	0.00
32	Benzo(a)anthracene	0.400	0.360	10.0	155	0.00
33	Chrysene	0.400	0.406	-1.5	146	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.345	13.8	141	0.00
35 I	Perylene-d12	0.400	0.400	0.0	163	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.377	5.8	170	0.00
37	Benzo(b)fluoranthene	0.400	0.359	10.3	158	0.00
38	Benzo(k)fluoranthene	0.400	0.374	6.5	165	0.00
39 C	Benzo(a)pyrene	0.400	0.392	2.0	169	0.00
40	Dibenzo(a,h)anthracene	0.400	0.379	5.3	181	0.00
41	Benzo(g,h,i)perylene	0.400	0.384	4.0	167	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4