

Data Path : Z:\HPCHEM1\BNA N\DATA\BN042618\
 Data File : BN000954.D
 Acq On : 27 Apr 2018 04:03
 Operator : JU/SJ
 Sample : J2526-06
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :

Quant Time: Apr 27 09:29:26 2018
 Quant Method : Z:\HPCHEM1\BNA N\METHODS\SOM-EPA-BN042418MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 27 09:25:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.62	152	156196	20.00	ng/ul	0.00
2) Naphthalene-d8	10.39	136	655362	20.00	ng/ul	0.00
6) Acenaphthene-d10	14.26	164	315203	20.00	ng/ul	0.00
12) Phenanthrene-d10	17.00	188	571428	20.00	ng/ul	0.00
17) Chrysene-d12	21.22	240	502974	20.00	ng/ul	0.00
22) Perylene-d12	23.41	264	527451	20.00	ng/ul	0.00
System Monitoring Compounds						
7) Acenaphthylene-d8	13.95	160	1007553	33.28	ng/ul	0.00
10) Fluorene-d10	15.25	176	668455	33.08	ng/ul	0.00
14) Anthracene-d10	17.10	188	877881	33.59	ng/ul	0.00
18) Pyrene-d10	19.41	212	872730	33.86	ng/ul	0.00
25) Benzo(a)pyrene-d12	23.27	264	924237	37.09	ng/ul	0.00
Target Compounds				Ovalue		
13) Phenanthrene	17.05	178	71852	2.305	ng/ul	99
16) Fluoranthene	19.07	202	121056	3.828	ng/ul#	94
19) Pyrene	19.44	202	104383	3.231	ng/ul#	91
20) Benzo(a)anthracene	21.20	228	60380	2.008	ng/ul	95
21) Chrysene	21.25	228	58029	2.053	ng/ul	97
23) Benzo(b)fluoranthene	22.76	252	88536m	2.954	ng/ul	
26) Benzo(a)pyrene	23.32	252	58822	2.053	ng/ul#	91
27) Indeno(1,2,3-cd)pyrene	25.59	276	43666	1.262	ng/ul#	85
29) Benzo(g,h,i)perylene	26.26	276	37253	1.287	ng/ul#	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA N\DATA\BN042618\
Data File : BN000954.D
Acq On : 27 Apr 2018 04:03
Operator : JU/SJ
Sample : J2526-06
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :

Quant Time: Apr 27 09:29:26 2018
Quant Method : Z:\HPCHEM1\BNA N\METHODS\SOM-EPA-BN042418MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Apr 27 09:25:35 2018
Response via : Initial Calibration

