

Data Path : Z:\HPCHEM1\BNA_N\Data\BN042618\
 Data File : BN000928.D
 Acq On : 26 Apr 2018 12:48
 Operator : JU/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD02064

Quant Time: Apr 27 07:45:09 2018
 Quant Method : Z:\HPCHEM1\BNA_N\Methods\SOM-EPA-BN042418MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 26 08:36:26 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
3	Naphthalene	1.008	1.025	-1.7	103	0.00
4	2-Methylnaphthalene	0.673	0.667	0.9	100	0.00
5	1-Methylnaphthalene	0.660	0.650	1.5	100	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
7 S	Acenaphthylene-d8	1.921	2.012	-4.7	100	0.00
8	Acenaphthylene	1.901	1.949	-2.5	99	0.00
9 C	Acenaphthene	1.314	1.336	-1.7	99	0.00
10 S	Fluorene-d10	1.282	1.307	-2.0	99	0.00
11	Fluorene	1.398	1.417	-1.4	97	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
13	Phenanthrene	1.091	1.104	-1.2	99	0.00
14 S	Anthracene-d10	0.915	0.959	-4.8	102	0.00
15	Anthracene	1.094	1.119	-2.3	100	0.00
16 C	Fluoranthene	1.107	1.232	-11.3	107	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
18 S	Pyrene-d10	1.025	1.052	-2.6	106	0.00
19	Pyrene	1.285	1.302	-1.3	105	0.00
20	Benzo(a)anthracene	1.196	1.239	-3.6	105	0.00
21	Chrysene	1.124	1.144	-1.8	102	0.00
22 I	Perylene-d12	1.000	1.000	0.0	103	0.01
23	Benzo(b)fluoranthene	1.137	1.183	-4.0	108	0.00
24	Benzo(k)fluoranthene	1.112	1.100	1.1	100	0.01
25 S	Benzo(a)pyrene-d12	0.945	0.996	-5.4	106	0.00
26 C	Benzo(a)pyrene	1.086	1.124	-3.5	104	0.01
27	Indeno(1,2,3-cd)pyrene	1.312	1.363	-3.9	105	0.02
28	Dibenzo(a,h)anthracene	1.093	1.145	-4.8	105	0.02
29	Benzo(g,h,i)perylene	1.097	1.146	-4.5	105	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0