

Data Path : Z:\HPCHEM1\BNA_N\DATA\BN042618\
 Data File : BN000936.D
 Acq On : 26 Apr 2018 17:28
 Operator : JU/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD02065

Quant Time: Apr 27 08:21:02 2018
 Quant Method : Z:\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN042418MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 27 07:47:48 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	93	0.00
3	Naphthalene	1.008	1.029	-2.1	93	0.00
4	2-Methylnaphthalene	0.673	0.675	-0.3	91	0.00
5	1-Methylnaphthalene	0.660	0.663	-0.5	92	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	91	0.00
7 S	Acenaphthylene-d8	1.921	2.050	-6.7	94	0.00
8	Acenaphthylene	1.901	1.986	-4.5	94	0.00
9 C	Acenaphthene	1.314	1.345	-2.4	92	0.00
10 S	Fluorene-d10	1.282	1.321	-3.0	93	0.00
11	Fluorene	1.398	1.456	-4.1	93	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.00
13	Phenanthrene	1.091	1.096	-0.5	95	0.00
14 S	Anthracene-d10	0.915	0.957	-4.6	97	0.00
15	Anthracene	1.094	1.127	-3.0	96	0.00
16 C	Fluoranthene	1.107	1.237	-11.7	103	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
18 S	Pyrene-d10	1.025	1.040	-1.5	103	0.00
19	Pyrene	1.285	1.270	1.2	100	0.00
20	Benzo(a)anthracene	1.196	1.254	-4.8	103	0.00
21	Chrysene	1.124	1.157	-2.9	101	0.00
22 I	Perylene-d12	1.000	1.000	0.0	100	0.00
23	Benzo(b)fluoranthene	1.137	1.159	-1.9	102	0.00
24	Benzo(k)fluoranthene	1.112	1.135	-2.1	99	0.00
25 S	Benzo(a)pyrene-d12	0.945	0.995	-5.3	103	0.00
26 C	Benzo(a)pyrene	1.086	1.127	-3.8	101	0.00
27	Indeno(1,2,3-cd)pyrene	1.312	1.366	-4.1	101	0.00
28	Dibenzo(a,h)anthracene	1.093	1.135	-3.8	101	0.01
29	Benzo(g,h,i)perylene	1.097	1.148	-4.6	102	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0