

Data Path : Z:\HPCHEM1\BNA_N\DATA\BN042618\
 Data File : BN000953.D
 Acq On : 27 Apr 2018 03:19
 Operator : JU/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC066

Quant Time: Apr 27 09:11:21 2018
 Quant Method : Z:\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN042418MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 27 08:22:19 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	139	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	138	0.00
3	Naphthalene	1.008	1.029	-2.1	139	0.00
4	2-Methylnaphthalene	0.673	0.687	-2.1	139	0.00
5	1-Methylnaphthalene	0.660	0.675	-2.3	140	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	138	0.00
7 S	Acenaphthylene-d8	1.921	2.028	-5.6	141	0.00
8	Acenaphthylene	1.901	1.966	-3.4	140	0.00
9 C	Acenaphthene	1.314	1.334	-1.5	138	0.00
10 S	Fluorene-d10	1.282	1.286	-0.3	137	0.00
11	Fluorene	1.398	1.391	0.5	134	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	127	0.00
13	Phenanthrene	1.091	1.104	-1.2	126	0.00
14 S	Anthracene-d10	0.915	0.951	-3.9	128	0.00
15	Anthracene	1.094	1.116	-2.0	126	0.00
16 C	Fluoranthene	1.107	1.067	3.6	118	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
18 S	Pyrene-d10	1.025	1.092	-6.5	115	0.00
19	Pyrene	1.285	1.351	-5.1	113	0.00
20	Benzo(a)anthracene	1.196	1.235	-3.3	108	0.00
21	Chrysene	1.124	1.148	-2.1	107	0.00
22 I	Perylene-d12	1.000	1.000	0.0	105	0.00
23	Benzo(b)fluoranthene	1.137	1.149	-1.1	106	0.00
24	Benzo(k)fluoranthene	1.112	1.095	1.5	101	0.00
25 S	Benzo(a)pyrene-d12	0.945	1.001	-5.9	108	0.00
26 C	Benzo(a)pyrene	1.086	1.130	-4.1	106	0.00
27	Indeno(1,2,3-cd)pyrene	1.312	1.366	-4.1	106	0.00
28	Dibenzo(a,h)anthracene	1.093	1.145	-4.8	107	0.00
29	Benzo(g,h,i)perylene	1.097	1.153	-5.1	107	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0