

Data Path : Z:\HPCHEM1\BNA_N\DATA\BN042618\
 Data File : BN000965.D
 Acq On : 27 Apr 2018 13:40
 Operator : JU/SJ
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC068

Quant Time: Apr 27 16:45:46 2018
 Quant Method : Z:\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN042418MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 27 09:25:35 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	89	0.00
3	Naphthalene	1.008	1.005	0.3	88	0.00
4	2-Methylnaphthalene	0.673	0.652	3.1	85	0.00
5	1-Methylnaphthalene	0.660	0.643	2.6	86	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	81	0.00
7 S	Acenaphthylene-d8	1.921	2.016	-4.9	82	0.00
8	Acenaphthylene	1.901	1.966	-3.4	82	0.00
9 C	Acenaphthene	1.314	1.331	-1.3	81	0.00
10 S	Fluorene-d10	1.282	1.255	2.1	79	0.00
11	Fluorene	1.398	1.377	1.5	78	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	76	0.00
13	Phenanthrene	1.091	1.088	0.3	75	0.00
14 S	Anthracene-d10	0.915	0.940	-2.7	76	0.00
15	Anthracene	1.094	1.103	-0.8	75	0.00
16 C	Fluoranthene	1.107	1.152	-4.1	76	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	78	0.00
18 S	Pyrene-d10	1.025	0.998	2.6	77	0.00
19	Pyrene	1.285	1.231	4.2	76	0.00
20	Benzo(a)anthracene	1.196	1.218	-1.8	78	0.00
21	Chrysene	1.124	1.123	0.1	77	0.00
22 I	Perylene-d12	1.000	1.000	0.0	78	0.00
23	Benzo(b)fluoranthene	1.137	1.122	1.3	78	0.00
24	Benzo(k)fluoranthene	1.112	1.054	5.2	73	0.00
25 S	Benzo(a)pyrene-d12	0.945	0.976	-3.3	79	0.00
26 C	Benzo(a)pyrene	1.086	1.099	-1.2	78	0.00
27	Indeno(1,2,3-cd)pyrene	1.312	1.341	-2.2	78	0.00
28	Dibenzo(a,h)anthracene	1.093	1.099	-0.5	77	0.00
29	Benzo(g,h,i)perylene	1.097	1.114	-1.5	78	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0