

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042721\
 Data File : BN014452.D
 Acq On : 27 Apr 2021 10:23
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Apr 27 11:43:34 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 27 11:43:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.684	152	10742	0.40	ng	0.00
7) Naphthalene-d8	10.454	136	40707	0.40	ng	0.00
13) Acenaphthene-d10	14.308	164	22168	0.40	ng	0.00
19) Phenanthrene-d10	17.055	188	46001	0.40	ng	# 0.00
29) Chrysene-d12	21.250	240	41765	0.40	ng	0.00
36) Perylene-d12	23.475	264	38996	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.284	112	9750	0.44	ng	0.00
5) Phenol-d6	6.847	99	12525	0.46	ng	0.00
8) Nitrobenzene-d5	8.819	82	9066	0.27	ng	0.00
11) 2-Methylnaphthalene-d10	12.049	152	26343	0.43	ng	0.00
14) 2,4,6-Tribromophenol	15.805	330	2754	0.39	ng	0.00
15) 2-Fluorobiphenyl	12.925	172	34877	0.40	ng	0.00
27) Fluoranthene-d10	19.094	212	48952	0.40	ng	0.00
31) Terphenyl-d14	19.700	244	38044	0.43	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.247	88	4802	0.37	ng	98
3) n-Nitrosodimethylamine	3.593	42	2312	0.41	ng	# 95
6) bis(2-Chloroethyl)ether	7.112	93	11293	0.43	ng	97
9) Naphthalene	10.505	128	43149	0.43	ng	100
10) Hexachlorobutadiene	10.796	225	8086	0.37	ng	# 99
12) 2-Methylnaphthalene	12.120	142	28506	0.43	ng	99
16) Acenaphthylene	14.029	152	32602	0.35	ng	99
17) Acenaphthene	14.372	154	24900	0.40	ng	97
18) Fluorene	15.361	166	30450	0.40	ng	99
20) 4,6-Dinitro-2-methylph...	15.437	198	1542	0.30	ng	# 82
21) 4-Bromophenyl-phenylether	16.252	248	9582	0.37	ng	# 89
22) Hexachlorobenzene	16.374	284	12681	0.36	ng	98
23) Atrazine	16.520	200	5315	0.35	ng	96
24) Pentachlorophenol	16.715	266	2549	0.39	ng	99
25) Phenanthrene	17.092	178	51607	0.40	ng	100
26) Anthracene	17.189	178	39956	0.39	ng	99
28) Fluoranthene	19.124	202	55022	0.40	ng	99
30) Pyrene	19.491	202	55831	0.45	ng	100
32) Benzo(a)anthracene	21.239	228	46118	0.42	ng	100
33) Chrysene	21.292	228	55626	0.43	ng	100
34) Bis(2-ethylhexyl)phtha...	21.176	149	13392	0.34	ng	99
35) Indeno(1,2,3-cd)pyrene	25.707	276	62653	0.43	ng	99
37) Benzo(b)fluoranthene	22.856	252	60199	0.42	ng	99
38) Benzo(k)fluoranthene	22.856	252	60524	0.41	ng	99
39) Benzo(a)pyrene	23.380	252	53514	0.44	ng	# 88
40) Dibenzo(a,h)anthracene	25.724	278	51717	0.38	ng	99
41) Benzo(g,h,i)perylene	26.385	276	53390	0.37	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042721\
Data File : BN014452.D
Acq On : 27 Apr 2021 10:23
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Apr 27 11:43:34 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 27 11:43:13 2021
Response via : Initial Calibration






Abundance Scan 610 (7.692 min): BN014350.D\data.ms (-603) (#1)




1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.684 min Scan# 609
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	152	Resp:	10742
Ion	Ratio	Lower	Upper
152	100		
150	152.3	123.4	185.2
115	55.5	47.5	71.3







Abundance Scan 59 (3.255 min): BN014350.D\data.ms (-56) (#2)



1,4-Dioxane
Concen: 0.37 ng
RT: 3.247 min Scan# 58
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23

Tgt Ion:	88	Resp:	4802
Ion	Ratio	Lower	Upper
88	100		
58	66.2	52.2	78.2
43	24.2	20.6	30.8









n-Nitrosodimethylamine
 Concen: 0.41 ng
 RT: 3.593 min Scan# 101
 Delta R.T. 0.000 min
 Lab File: BN014452.D
 Acq: 27 Apr 2021 10:23




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion: 42 Resp: 2312
 Ion Ratio Lower Upper
 42 100
 74 140.6 108.0 162.0
 44 10.6 0.0 0.0#



2-Fluorophenol
 Concen: 0.44 ng
 RT: 5.284 min Scan# 311
 Delta R.T. 0.000 min
 Lab File: BN014452.D
 Acq: 27 Apr 2021 10:23

Tgt Ion: 112 Resp: 9750
 Ion Ratio Lower Upper
 112 100
 64 52.0 43.0 64.6
 63 26.7 23.4 35.0









Phenol-d6
 Concen: 0.46 ng
 RT: 6.847 min Scan# 505
 Delta R.T. 0.000 min
 Lab File: BN014452.D
 Acq: 27 Apr 2021 10:23




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Tgt Ion:	99	Resp:	12525
Ion	Ratio	Lower	Upper
99	100		
42	16.3	16.0	24.0
71	32.1	27.8	41.8




bis(2-Chloroethyl)ether
 Concen: 0.43 ng
 RT: 7.112 min Scan# 538
 Delta R.T. 0.000 min
 Lab File: BN014452.D
 Acq: 27 Apr 2021 10:23

Tgt Ion:	93	Resp:	11293
Ion	Ratio	Lower	Upper
93	100		
63	70.3	59.4	89.2
95	33.3	27.0	40.4






Abundance Scan 857 (10.467 min): BN014350.D\data.ms (-857) (-)




Naphthalene-d8
Concen: 0.40 ng
RT: 10.454 min Scan# 856
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	136	Resp:	40707
Ion	Ratio	Lower	Upper
136	100		
137	11.0	9.0	13.6
54	5.0	4.4	6.6
68	4.8	3.9	5.9





Abundance Scan 728 (8.832 min): BN014350.D\data.ms (-725) (-)




Nitrobenzene-d5
Concen: 0.27 ng
RT: 8.819 min Scan# 727
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23


Tgt Ion:	82	Resp:	9066
Ion	Ratio	Lower	Upper
82	100		
128	36.8	30.3	45.5
54	45.2	35.8	53.8



Abundance Scan 727 (8.819 min): BN014452.D\data.ms (-711) (-)






Abundance Scan 861 (10.518 min): BN014350.D\data.ms (-855) (-)




Naphthalene
Concen: 0.43 ng
RT: 10.505 min Scan# 860
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	128	Resp:	43149
Ion	Ratio	Lower	Upper
128	100		
129	11.1	9.0	13.6
127	12.7	10.3	15.5






Abundance Scan 884 (10.809 min): BN014350.D\data.ms (-875) (-)




Hexachlorobutadiene
Concen: 0.37 ng
RT: 10.796 min Scan# 883
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23

Tgt Ion:	225	Resp:	8086
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.0	50.6	75.8






Abundance Scan 1063 (12.058 min): BN014350.D\data.ms (-1011) (-)




2-Methylnaphthalene-d10
Concen: 0.43 ng
RT: 12.049 min Scan# 1061
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:152 Resp: 26343
Ion Ratio Lower Upper
152 100
151 20.2 16.3 24.5






Abundance Scan 1080 (12.133 min): BN014350.D\data.ms (-1010) (-)



2-Methylnaphthalene
Concen: 0.43 ng
RT: 12.120 min Scan# 1077
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23


Tgt Ion:142 Resp: 28506
Ion Ratio Lower Upper
142 100
141 88.0 70.6 105.8
115 34.0 29.0 43.6





Abundance Scan 1307 (14.321 min): BN014350.D\data.ms (-1299) (-)

Acenaphthene-d10
Concen: 0.40 ng
RT: 14.308 min Scan# 1306
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:	164	Resp:	22168
Ion	Ratio	Lower	Upper
164	100		
162	101.6	81.0	121.6
160	47.3	40.7	61.1





Abundance Scan 1306 (14.308 min): BN014452.D\data.ms (-1299) (-)




Abundance Scan 1425 (15.818 min): BN014350.D\data.ms (-1416) (-)

2,4,6-Tribromophenol
Concen: 0.39 ng
RT: 15.805 min Scan# 1424
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23


Tgt Ion:	330	Resp:	2754
Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	35.1	26.2	39.2



Abundance Scan 1424 (15.805 min): BN014452.D\data.ms (-1416) (-)





Abundance Scan 1198 (12.938 min): BN014350.D\data.ms (-1198) (-)




2-Fluorobiphenyl
Concen: 0.40 ng
RT: 12.925 min Scan# 1197
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	172	Resp:	34877
Ion	Ratio	Lower	Upper
172	100		
171	35.1	28.2	42.2
170	24.2	19.6	29.4



Abundance Scan 1197 (12.925 min): BN014452.D\data.ms (-1189) (-)





Abundance Scan 1285 (14.042 min): BN014350.D\data.ms (-1215) (-)




Acenaphthylene
Concen: 0.35 ng
RT: 14.029 min Scan# 1284
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23


Tgt Ion:	152	Resp:	32602
Ion	Ratio	Lower	Upper
152	100		
151	19.1	15.6	23.4
153	13.0	10.5	15.7



Abundance Scan 1284 (14.029 min): BN014452.D\data.ms (-1268) (-)






Abundance Scan 1312 (14.384 min): BN014350.D\data.ms (-1303) (-)




Acenaphthene
Concen: 0.40 ng
RT: 14.372 min Scan# 1311
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:154 Resp: 24900
Ion Ratio Lower Upper
154 100
153 110.7 89.2 133.8
152 56.6 49.0 73.4






Abundance Scan 1390 (15.374 min): BN014350.D\data.ms (-1315) (-)



Fluorene
Concen: 0.40 ng
RT: 15.361 min Scan# 1389
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23


Tgt Ion:166 Resp: 30450
Ion Ratio Lower Upper
166 100
165 97.7 78.6 118.0
167 13.4 10.9 16.3






Abundance Scan 1527 (17.067 min): BN014350.D\data.ms (-1519) (-)

Phenanthrene-d10
Concen: 0.40 ng
RT: 17.055 min Scan# 1526
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







Tgt Ion:	188	Resp:	46001
Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	9.5	6.2	9.4#




Abundance Scan 1396 (15.450 min): BN014350.D\data.ms (-1520) (-)

4,6-Dinitro-2-methylphenol
Concen: 0.30 ng
RT: 15.437 min Scan# 1395
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23

Tgt Ion:	198	Resp:	1542
Ion	Ratio	Lower	Upper
198	100		
51	86.8	60.0	90.0
105	54.6	31.5	47.3#






Abundance Scan 1461 (16.264 min): BN014350.D\data.ms (-1421) (-)




4-Bromophenyl-phenylether
Concen: 0.37 ng
RT: 16.252 min Scan# 1460
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	248	Resp:	9582
Ion	Ratio	Lower	Upper
248	100		
250	93.6	77.4	116.0
141	70.5	41.4	62.0#






Abundance Scan 1470 (16.373 min): BN014350.D\data.ms (-1425) (-)



Hexachlorobenzene
Concen: 0.36 ng
RT: 16.374 min Scan# 1470
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23


Tgt Ion:	284	Resp:	12681
Ion	Ratio	Lower	Upper
284	100		
142	35.1	26.5	39.7
249	29.2	24.0	36.0





Abundance Scan 1483 (16.532 min): BN014350.D\data.ms (-1429) (-)

Atrazine
Concen: 0.35 ng
RT: 16.520 min Scan# 1482
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:	200	Resp:	5315
Ion	Ratio	Lower	Upper
200	100		
173	29.8	21.0	31.6
215	46.7	38.7	58.1






Abundance Scan 1482 (16.520 min): BN014452.D\data.ms (-1441) (-)




Abundance Scan 1498 (16.714 min): BN014350.D\data.ms (-1429) (-)

Pentachlorophenol
Concen: 0.39 ng
RT: 16.715 min Scan# 1498
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23


Tgt Ion:	266	Resp:	2549
Ion	Ratio	Lower	Upper
266	100		
264	63.9	50.6	76.0
268	64.8	51.4	77.0



Abundance Scan 1498 (16.715 min): BN014452.D\data.ms (-1490) (-)





Abundance Scan 1530 (17.104 min): BN014350.D\data.ms (-1525) (-)




Phenanthrene
Concen: 0.40 ng
RT: 17.092 min Scan# 1529
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	178	Resp:	51607
Ion	Ratio	Lower	Upper
178	100		
176	18.9	15.2	22.8
179	15.4	12.2	18.4



Abundance Scan 1529 (17.092 min): BN014452.D\data.ms (-1504) (-)





Abundance Scan 1538 (17.201 min): BN014350.D\data.ms (-1525) (-)




Anthracene
Concen: 0.39 ng
RT: 17.189 min Scan# 1537
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23


Tgt Ion:	178	Resp:	39956
Ion	Ratio	Lower	Upper
178	100		
176	18.4	15.0	22.4
179	15.4	12.2	18.2



Abundance Scan 1537 (17.189 min): BN014452.D\data.ms (-1529) (-)





Abundance Scan 1825 (19.101 min): BN014350.D\data.ms (-1827) (-)




Fluoranthene-d10
Concen: 0.40 ng
RT: 19.094 min Scan# 1822
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:212 Resp: 48952
Ion Ratio Lower Upper
212 100
106 12.8 9.7 14.5
104 7.0 5.4 8.0



Abundance Scan 1822 (19.094 min): BN014452.D\data.ms (-1782) (-)





Abundance Scan 1831 (19.131 min): BN014350.D\data.ms (-1828) (-)




Fluoranthene
Concen: 0.40 ng
RT: 19.124 min Scan# 1828
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23


Tgt Ion:202 Resp: 55022
Ion Ratio Lower Upper
202 100
101 10.9 8.2 12.4
203 17.2 13.8 20.6



Abundance Scan 1828 (19.124 min): BN014452.D\data.ms (-1788) (-)






Abundance Scan 2125 (21.258 min): BN014350.D\data.ms (-2125) (-)




Chrysene-d12
Concen: 0.40 ng
RT: 21.250 min Scan# 2123
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	240	Resp:	41765
Ion	Ratio	Lower	Upper
240	100		
120	15.4	10.7	16.1
236	27.1	21.5	32.3






Abundance Scan 1905 (19.498 min): BN014350.D\data.ms (-1905) (-)



Pyrene
Concen: 0.45 ng
RT: 19.491 min Scan# 1902
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23


Tgt Ion:	202	Resp:	55831
Ion	Ratio	Lower	Upper
202	100		
200	20.5	16.6	24.8
203	17.4	14.1	21.1






Abundance Scan 1947 (19.707 min): BN014350.D\data.ms (-1934) (-)

Terphenyl-d14
Concen: 0.43 ng
RT: 19.700 min Scan# 1944
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion	Ratio	Lower	Upper
244	100		
212	7.4	6.2	9.4
122	12.1	9.3	13.9






Abundance Scan 2124 (21.247 min): BN014350.D\data.ms (-2132) (-)


Benzo(a)anthracene
Concen: 0.42 ng
RT: 21.239 min Scan# 2122
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23



Tgt Ion	Ratio	Lower	Upper
228	100		
226	26.7	21.3	31.9
229	19.6	15.7	23.5






Abundance Scan 2129 (21.300 min): BN014350.D\data.ms (-2129) (-)




Chrysene
Concen: 0.43 ng
RT: 21.292 min Scan# 2127
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	228	Resp:	55626
Ion	Ratio	Lower	Upper
228	100		
226	29.0	23.1	34.7
229	19.7	15.7	23.5






Abundance Scan 2118 (21.184 min): BN014350.D\data.ms (-2118) (-)




Bis(2-ethylhexyl)phthalate
Concen: 0.34 ng
RT: 21.176 min Scan# 2116
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23

Tgt Ion:	149	Resp:	13392
Ion	Ratio	Lower	Upper
149	100		
167	27.3	21.3	31.9
279	3.7	3.1	4.7




Abundance Scan 3203 (25.725 min): BN014350.D\data.ms (-3175) (-)





Indeno(1,2,3-cd)pyrene
Concen: 0.43 ng
RT: 25.707 min Scan# 3196
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion:	276	Resp:	62653
Ion	Ratio	Lower	Upper
276	100		
138	24.2	18.9	28.3
277	24.9	19.8	29.8





Abundance Scan 2550 (23.490 min): BN014350.D\data.ms (-2535) (-)




Perylene-d12
Concen: 0.40 ng
RT: 23.475 min Scan# 2544
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23



Tgt Ion:	264	Resp:	38996
Ion	Ratio	Lower	Upper
264	100		
260	24.5	19.8	29.6
265	34.5	34.4	51.6






Abundance Scan 2354 (22.819 min): BN014350.D\data.ms (-2354) (-)




Benzo(b)fluoranthene
Concen: 0.42 ng
RT: 22.856 min Scan# 2363
Delta R.T. 0.044 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23

Tgt Ion:	252	Resp:	60199
Ion	Ratio	Lower	Upper
252	100		
253	23.3	19.0	28.6
125	11.5	9.7	14.5






Abundance Scan 2367 (22.864 min): BN014350.D\data.ms (-2367) (-)




Benzo(k)fluoranthene
Concen: 0.41 ng
RT: 22.856 min Scan# 2363
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23

Tgt Ion:	252	Resp:	60524
Ion	Ratio	Lower	Upper
252	100		
253	23.3	19.3	28.9
125	11.5	9.4	14.2



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Abundance Scan 2521 (23.391 min): BN014350.D\data.ms (-2520) (-)




Benzo(a)pyrene
Concen: 0.44 ng
RT: 23.380 min Scan# 2516
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	252	Resp:	53514
Ion	Ratio	Lower	Upper
252	100		
253	25.6	20.7	31.1
125	28.2	11.8	17.6#



Abundance Scan 2516 (23.380 min): BN014452.D\data.ms (-2458) (-)





Abundance Scan 3207 (25.739 min): BN014350.D\data.ms (-3146) (-)




Dibenzo(a,h)anthracene
Concen: 0.38 ng
RT: 25.724 min Scan# 3201
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23


Tgt Ion:	278	Resp:	51717
Ion	Ratio	Lower	Upper
278	100		
139	17.2	13.3	19.9
279	24.6	20.2	30.2



Abundance Scan 3201 (25.724 min): BN014452.D\data.ms (-3178) (-)




Abundance Scan 3403 (26.410 min): BN014350.D\data.ms (-3343) (-)




Benzo(g,h,i)perylene
Concen: 0.37 ng
RT: 26.385 min Scan# 3394
Delta R.T. 0.000 min
Lab File: BN014452.D
Acq: 27 Apr 2021 10:23

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Abundance Scan 3394 (26.385 min): BN014452.D\data.ms



Tgt Ion:276 Resp: 53390
Ion Ratio Lower Upper
276 100
277 23.9 19.1 28.7
138 22.2 16.7 25.1



Abundance Scan 3394 (26.385 min): BN014452.D\data.ms (-3336) (-)

