

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042721\
 Data File : BN014455.D
 Acq On : 27 Apr 2021 12:45
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 27 13:19:09 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 27 11:43:13 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	105	0.00
2	1,4-Dioxane	0.400	0.358	10.5	95	0.00
3	n-Nitrosodimethylamine	0.400	0.356	11.0	69	0.02
4 S	2-Fluorophenol	0.400	0.381	4.8	98	0.00
5 S	Phenol-d6	0.400	0.390	2.5	103	0.00
6	bis(2-Chloroethyl)ether	0.400	0.392	2.0	101	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	115	0.00
8 S	Nitrobenzene-d5	0.400	0.257	35.8#	75	0.00
9	Naphthalene	0.400	0.401	-0.3	115	0.00
10	Hexachlorobutadiene	0.400	0.350	12.5	99	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.403	-0.8	116	0.00
12	2-Methylnaphthalene	0.400	0.403	-0.8	117	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	119	0.00
14 S	2,4,6-Tribromophenol	0.400	0.351	12.3	115	0.00
15 S	2-Fluorobiphenyl	0.400	0.376	6.0	116	0.01
16	Acenaphthylene	0.400	0.325	18.8	102	0.00
17	Acenaphthene	0.400	0.375	6.3	116	0.00
18	Fluorene	0.400	0.379	5.3	119	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	123	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.277	30.8#	91	0.00
21	4-Bromophenyl-phenylether	0.400	0.346	13.5	110	0.00
22	Hexachlorobenzene	0.400	0.336	16.0	106	0.00
23	Atrazine	0.400	0.313	21.8	107	0.00
24	Pentachlorophenol	0.400	0.317	20.8	114	0.00
25	Phenanthrene	0.400	0.374	6.5	119	0.01
26	Anthracene	0.400	0.364	9.0	120	0.00
27 SURR	Fluoranthene-d10	0.400	0.365	8.8	117	0.00
28	Fluoranthene	0.400	0.368	8.0	118	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	115	0.00
30	Pyrene	0.400	0.418	-4.5	122	0.00
31 S	Terphenyl-d14	0.400	0.399	0.3	114	0.00
32	Benzo(a)anthracene	0.400	0.382	4.5	113	0.00
33	Chrysene	0.400	0.398	0.5	112	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.303	24.3	97	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.386	3.5	118	0.00
36 I	Perylene-d12	0.400	0.400	0.0	121	0.00
37	Benzo(b)fluoranthene	0.400	0.385	3.8	119	0.00
38	Benzo(k)fluoranthene	0.400	0.387	3.3	121	0.00
39 C	Benzo(a)pyrene	0.400	0.404	-1.0	135	0.00
40	Dibenzo(a,h)anthracene	0.400	0.363	9.3	116	0.00
41	Benzo(g,h,i)perylene	0.400	0.364	9.0	111	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4