

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042722\
 Data File : BN019540.D
 Acq On : 27 Apr 2022 11:27
 Operator : CG/JU
 Sample : SSTD04013
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD040213

Quant Time: Apr 27 12:51:20 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN042722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 27 12:47:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.881	152	249955	20.000	ng/u1	0.00
20) Naphthalene-d8	10.675	136	1029671	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.498	164	579239	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.239	188	1086029	20.000	ng/u1	0.00
79) Chrysene-d12	21.422	240	963292	20.000	ng/u1	# 0.00
88) Perylene-d12	23.780	264	1023010	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.346	96	103465	17.257	ng/uL	0.00
4) Pyridine-d5	3.758	84	670807	42.085	ng/u1	0.00
7) Phenol-d5	7.040	99	862417	40.247	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.210	67	500854	37.627	ng/u1	0.00
11) 2-Chlorophenol-d4	7.416	132	707693	43.316	ng/u1	0.00
15) 4-Methylphenol-d8	8.587	113	667652	38.851	ng/u1	0.00
21) Nitrobenzene-d5	9.034	128	325653	41.376	ng/u1	0.00
24) 2-Nitrophenol-d4	9.752	143	339128	38.792	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.293	165	616413	38.792	ng/u1	0.00
31) 4-Chloroaniline-d4	10.804	131	894413	40.835	ng/u1	0.00
46) Dimethylphthalate-d6	13.916	166	1630909	38.163	ng/u1	0.00
49) Acenaphthylene-d8	14.192	160	2154788	40.856	ng/u1	0.00
54) 4-Nitrophenol-d4	14.687	143	325320	41.469	ng/u1	0.00
60) Fluorene-d10	15.492	176	1380613	38.614	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.604	200	245156	36.780	ng/u1	0.00
73) Anthracene-d10	17.339	188	2027437	40.283	ng/u1	0.00
81) Pyrene-d10	19.628	212	2200526	36.641	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.627	264	2094059	40.134	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.381	88	105655	17.565	ng/uL	94
5) Pyridine	3.775	79	706702	42.223	ng/u1	98
6) Benzaldehyde	7.016	77	503758	43.883	ng/u1	93
8) Phenol	7.069	94	865752	40.160	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.305	93	675156	39.160	ng/u1	99
12) 2-Chlorophenol	7.446	128	713958	42.699	ng/u1	99
13) 2-Methylphenol	8.316	108	651086	40.259	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.416	45	966328	38.366	ng/u1	99
16) Acetophenone	8.699	105	1001554	39.066	ng/u1	96
17) N-Nitroso-di-n-propyla...	8.699	70	486899	35.792	ng/u1	98
18) 4-Methylphenol	8.652	108	697297	39.683	ng/u1	97
19) Hexachloroethane	8.963	117	287688	41.284	ng/u1	87
22) Nitrobenzene	9.075	77	727136	37.457	ng/u1	96
23) Isophorone	9.610	82	1349062	36.590	ng/u1	98
25) 2-Nitrophenol	9.787	139	369751	40.724	ng/u1	95
26) 2,4-Dimethylphenol	9.852	107	751596	39.096	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.087	93	843544	37.157	ng/u1	97
29) 2,4-Dichlorophenol	10.316	162	605218	38.668	ng/u1	97
30) Naphthalene	10.722	128	2156986	41.264	ng/u1	99
32) 4-Chloroaniline	10.828	127	893107	40.884	ng/u1	99
33) Hexachlorobutadiene	11.022	225	349104	33.606	ng/u1	97
34) Caprolactam	11.599	113	89655	17.233	ng/u1	95
35) 4-Chloro-3-methylphenol	11.951	107	663602	37.630	ng/u1	98
36) 2-Methylnaphthalene	12.334	142	1400827	39.025	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042722\
 Data File : BN019540.D
 Acq On : 27 Apr 2022 11:27
 Operator : CG/JU
 Sample : SST04013
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST040213

Quant Time: Apr 27 12:51:20 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN042722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 27 12:47:24 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.551	142	1396257	38.437	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.698	216	633458	36.935	ng/ul	97
40) Hexachlorocyclopentadiene	12.687	237	398517	40.356	ng/ul	94
41) 2,4,6-Trichlorophenol	12.934	196	430972	39.792	ng/ul	94
42) 2,4,5-Trichlorophenol	13.004	196	464241	38.745	ng/ul	97
43) 1,1'-Biphenyl	13.340	154	1804622	42.130	ng/ul	100
44) 2-Chloronaphthalene	13.381	162	1357035	41.071	ng/ul	96
45) 2-Nitroaniline	13.575	65	397258	39.599	ng/ul	93
47) Dimethylphthalate	13.963	163	1585134	37.964	ng/ul	99
48) 2,6-Dinitrotoluene	14.075	165	323578	39.243	ng/ul	96
50) Acenaphthylene	14.222	152	2203332	41.906	ng/ul	100
51) 3-Nitroaniline	14.398	138	354111	41.681	ng/ul	99
52) Acenaphthene	14.563	153	1420528	41.854	ng/ul	98
53) 2,4-Dinitrophenol	14.598	184	161815	30.635	ng/ul	100
55) 4-Nitrophenol	14.704	109	255851	38.751	ng/ul	95
56) Dibenzofuran	14.898	168	1945552	39.686	ng/ul	100
57) 2,4-Dinitrotoluene	14.857	165	464218	38.637	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.122	232	345925	35.807	ng/ul	92
59) Diethylphthalate	15.328	149	1627750	38.724	ng/ul	98
61) Fluorene	15.545	166	1511777	39.523	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.545	204	678966	34.975	ng/ul	95
63) 4-Nitroaniline	15.557	138	339444	41.664	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.616	198	245690	37.528	ng/ul	92
67) N-Nitrosodiphenylamine	15.751	169	1334622	42.932	ng/ul	99
68) 4-Bromophenyl-phenylether	16.434	248	406460	37.559	ng/ul	90
69) Hexachlorobenzene	16.551	284	454384	36.839	ng/ul#	91
70) Atrazine	16.704	200	446880	38.445	ng/ul	96
71) Pentachlorophenol	16.886	266	293689	39.205	ng/ul	98
72) Phenanthrene	17.281	178	2367717	41.553	ng/ul	99
74) Anthracene	17.375	178	2397859	41.995	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.304	216	657066	40.215	ng/uL	97
76) Pentachlorobenzene	14.822	250	576218	38.660	ng/uL	91
77) Carbazole	17.639	167	2195509	44.310	ng/ul	99
78) Di-n-butylphthalate	18.216	149	2654095	42.828	ng/ul	100
80) Fluoranthene	19.292	202	2616618	37.622	ng/ul#	94
82) Pyrene	19.657	202	2631135	37.328	ng/ul#	93
83) Butylbenzylphthalate	20.569	149	1172343	41.724	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.339	252	859642	45.448	ng/ul#	98
85) Benzo(a)anthracene	21.404	228	2426299	39.379	ng/ul	96
86) Bis(2-ethylhexyl)phtha...	21.351	149	1725606	42.767	ng/ul#	97
87) Chrysene	21.463	228	2363737	39.654	ng/ul	99
89) Di-n-octyl phthalate	22.269	149	2999663	36.275	ng/ul	100
90) Benzo(b)fluoranthene	23.063	252	2646624	38.946	ng/ul#	98
91) Benzo(k)fluoranthene	23.116	252	2353267	37.276	ng/ul	96
93) Benzo(a)pyrene	23.680	252	2486958	39.709	ng/ul	96
94) Indeno(1,2,3-cd)pyrene	26.209	276	2909319	44.996	ng/ul	93
95) Dibenzo(a,h)anthracene	26.233	278	2472251	44.348	ng/ul	98
96) Benzo(g,h,i)perylene	26.957	276	2592646	47.306	ng/ul#	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042722\
 Data File : BN019540.D
 Acq On : 27 Apr 2022 11:27
 Operator : CG/JU
 Sample : SST04013
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST040213

Quant Time: Apr 27 12:51:20 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN042722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 27 12:47:24 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 17.565 ng/uL
RT: 3.381 min Scan# 50
Delta R.T. 0.000 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion: 88 Resp: 105655
Ion Ratio Lower Upper
88 100
43 30.1 22.4 33.6
58 82.3 61.0 91.4



#5
Pyridine
Concen: 42.223 ng/uL
RT: 3.775 min Scan# 117
Delta R.T. 0.000 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27

Tgt Ion: 79 Resp: 706702
Ion Ratio Lower Upper
79 100
52 79.6 65.0 97.4
51 38.0 31.6 47.4









#6
Benzaldehyde
Concen: 43.883 ng/ul
RT: 7.016 min Scan# 668
Delta R.T. 0.000 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion: 77 Resp: 503758
Ion Ratio Lower Upper
77 100
105 99.4 74.7 112.1
106 95.8 70.2 105.4



#8
Phenol
Concen: 40.160 ng/ul
RT: 7.069 min Scan# 677
Delta R.T. 0.006 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27

Tgt Ion: 94 Resp: 865752
Ion Ratio Lower Upper
94 100
65 28.4 23.6 35.4
66 37.9 31.2 46.8





#10

Bis(2-Chloroethyl)ether

Concen: 39.160 ng/ul

RT: 7.305 min Scan# 716

Delta R.T. 0.006 min

Lab File: BN019540.D

Acq: 27 Apr 2022 11:27

Instrument :

BNA_N

ClientSampleId :

SSTD040213





Tgt Ion: 93 Resp: 675156

Ion Ratio Lower Upper

93 100

63 77.5 63.4 95.0

95 32.6 26.2 39.2



#12

2-Chlorophenol

Concen: 42.699 ng/ul

RT: 7.446 min Scan# 741

Delta R.T. 0.006 min

Lab File: BN019540.D

Acq: 27 Apr 2022 11:27




Tgt Ion:128 Resp: 713958


Ion Ratio Lower Upper

128 100

64 47.8 38.6 57.8

130 32.0 25.0 37.4





#13

2-Methylphenol

Concen: 40.259 ng/ul

RT: 8.316 min Scan# 889

Delta R.T. 0.000 min

Lab File: BN019540.D

Acq: 27 Apr 2022 11:27

Instrument :

BNA_N

ClientSampleId :





SSTD040213

Tgt Ion:108 Resp: 651086

Ion Ratio Lower Upper

108 100

107 88.2 71.5 107.3



#14

2,2'-oxybis(1-Chloropropane)

Concen: 38.366 ng/ul

RT: 8.416 min Scan# 906

Delta R.T. 0.000 min

Lab File: BN019540.D

Acq: 27 Apr 2022 11:27




Tgt Ion: 45 Resp: 966328


Ion Ratio Lower Upper

45 100

77 13.2 10.9 16.3

79 10.2 8.6 13.0









#16
Acetophenone
Concen: 39.066 ng/ul
RT: 8.699 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:105 Resp: 1001554
Ion Ratio Lower Upper
105 100
77 74.7 63.5 95.3
51 28.4 23.5 35.3



#17
N-Nitroso-di-n-propylamine
Concen: 35.792 ng/ul
RT: 8.699 min Scan# 954
Delta R.T. 0.012 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27

Tgt Ion: 70 Resp: 486899
Ion Ratio Lower Upper
70 100
42 59.7 47.8 71.6
101 10.5 7.8 11.6
130 23.4 15.8 23.8









#18
4-Methylphenol
Concen: 39.683 ng/ul
RT: 8.652 min Scan# 946
Delta R.T. 0.012 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:108 Resp: 697297
Ion Ratio Lower Upper
108 100
107 112.9 88.0 132.0



#19
Hexachloroethane
Concen: 41.284 ng/ul
RT: 8.963 min Scan# 999
Delta R.T. 0.000 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27

Tgt Ion:117 Resp: 287688
Ion Ratio Lower Upper
117 100
201 88.5 83.2 124.8
199 55.2 50.2 75.4









#22
Nitrobenzene
Concen: 37.457 ng/ul
RT: 9.075 min Scan# 1017
Delta R.T. 0.006 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion: 77 Resp: 727136
Ion Ratio Lower Upper
77 100
123 46.1 34.5 51.7
65 14.1 11.1 16.7



#23
Isophorone
Concen: 36.590 ng/ul
RT: 9.610 min Scan# 1109
Delta R.T. 0.006 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27

Tgt Ion: 82 Resp: 1349062
Ion Ratio Lower Upper
82 100
95 7.3 5.8 8.8
138 18.5 13.8 20.8









#25
2-Nitrophenol
Concen: 40.724 ng/ul
RT: 9.787 min Scan# 1138
Delta R.T. 0.006 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:139 Resp: 369751
Ion Ratio Lower Upper
139 100
65 50.0 43.7 65.5
109 30.7 25.8 38.6



#26
2,4-Dimethylphenol
Concen: 39.096 ng/ul
RT: 9.852 min Scan# 1150
Delta R.T. 0.006 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27

Tgt Ion:107 Resp: 751596
Ion Ratio Lower Upper
107 100
121 50.5 38.3 57.5
122 87.2 66.3 99.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 37.157 ng/ul
 RT: 10.087 min Scan# 1190
 Delta R.T. 0.000 min
 Lab File: BN019540.D
 Acq: 27 Apr 2022 11:27




Instrument :
 BNA_N
 ClientSampleId :
 SST040213


Tgt Ion: 93 Resp: 843544
 Ion Ratio Lower Upper
 93 100
 95 33.2 25.4 38.0
 123 11.7 8.3 12.5



#29
 2,4-Dichlorophenol
 Concen: 38.668 ng/ul
 RT: 10.316 min Scan# 1229
 Delta R.T. 0.000 min
 Lab File: BN019540.D
 Acq: 27 Apr 2022 11:27

Tgt Ion: 162 Resp: 605218
 Ion Ratio Lower Upper
 162 100
 164 67.3 52.9 79.3
 98 40.4 29.8 44.8









#30
Naphthalene
Concen: 41.264 ng/ul
RT: 10.722 min Scan# 1298
Delta R.T. 0.000 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:128 Resp: 2156986
Ion Ratio Lower Upper
128 100
129 11.2 8.7 13.1
127 12.9 10.6 16.0



#32
4-Chloroaniline
Concen: 40.884 ng/ul
RT: 10.828 min Scan# 1316
Delta R.T. 0.006 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27

Tgt Ion:127 Resp: 893107
Ion Ratio Lower Upper
127 100
129 33.0 26.1 39.1









#33
Hexachlorobutadiene
Concen: 33.606 ng/ul
RT: 11.022 min Scan# 1349
Delta R.T. 0.006 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:225 Resp: 349104
Ion Ratio Lower Upper
225 100
223 60.3 51.8 77.6
227 63.9 50.6 76.0



#34
Caprolactam
Concen: 17.233 ng/ul
RT: 11.599 min Scan# 1447
Delta R.T. 0.012 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27

Tgt Ion:113 Resp: 89655
Ion Ratio Lower Upper
113 100
55 169.6 145.5 218.3
56 131.6 106.5 159.7









#35
4-Chloro-3-methylphenol
Concen: 37.630 ng/ul
RT: 11.951 min Scan# 1507
Delta R.T. 0.006 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:107 Resp: 663602
Ion Ratio Lower Upper
107 100
144 25.3 20.6 31.0
142 80.1 62.2 93.4



#36
2-Methylnaphthalene
Concen: 39.025 ng/ul
RT: 12.334 min Scan# 1572
Delta R.T. 0.006 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27

Tgt Ion:142 Resp: 1400827
Ion Ratio Lower Upper
142 100
141 89.3 70.6 105.8





#37

1-Methylnaphthalene

Concen: 38.437 ng/ul

RT: 12.551 min Scan# 1

Delta R.T. 0.006 min

Lab File: BN019540.D

Acq: 27 Apr 2022 11:27

Instrument :

BNA_N

ClientSampleId :

SSTD040213





Tgt Ion:142 Resp: 1396257

Ion Ratio Lower Upper

142 100

141 91.6 70.9 106.3

116 3.8 3.1 4.7



#39

1,2,4,5-Tetrachlorobenzene

Concen: 36.935 ng/ul

RT: 12.698 min Scan# 1634

Delta R.T. 0.000 min

Lab File: BN019540.D

Acq: 27 Apr 2022 11:27

Tgt Ion:216 Resp: 633458




Ion Ratio Lower Upper


216 100

214 75.6 63.4 95.0

179 21.6 17.8 26.8

108 20.2 15.1 22.7









#40
Hexachlorocyclopentadiene
Concen: 40.356 ng/ul
RT: 12.687 min Scan# 1632
Delta R.T. 0.000 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:237 Resp: 398517
Ion Ratio Lower Upper
237 100
235 66.1 48.8 73.2
272 12.4 9.3 13.9



#41
2,4,6-Trichlorophenol
Concen: 39.792 ng/ul
RT: 12.934 min Scan# 1674
Delta R.T. 0.000 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27

Tgt Ion:196 Resp: 430972
Ion Ratio Lower Upper
196 100
198 92.4 79.8 119.8
200 29.9 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 38.745 ng/ul
RT: 13.004 min Scan# 1686
Delta R.T. 0.006 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:196 Resp: 464241
Ion Ratio Lower Upper
196 100
198 93.7 77.6 116.4
200 29.1 24.4 36.6



#43
1,1'-Biphenyl
Concen: 42.130 ng/ul
RT: 13.340 min Scan# 1743
Delta R.T. 0.000 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27

Tgt Ion:154 Resp: 1804622
Ion Ratio Lower Upper
154 100
153 39.2 31.5 47.3
76 13.2 10.5 15.7









#44
2-Chloronaphthalene
Concen: 41.071 ng/ul
RT: 13.381 min Scan# 1750
Delta R.T. 0.006 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:162 Resp: 1357035
Ion Ratio Lower Upper
162 100
164 32.4 25.8 38.6
127 43.1 31.4 47.2



#45
2-Nitroaniline
Concen: 39.599 ng/ul
RT: 13.575 min Scan# 1783
Delta R.T. 0.006 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27

Tgt Ion: 65 Resp: 397258
Ion Ratio Lower Upper
65 100
92 68.9 54.1 81.1
138 106.6 77.0 115.4









#47
Dimethylphthalate
Concen: 37.964 ng/ul
RT: 13.963 min Scan# 1849
Delta R.T. 0.006 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:163 Resp: 1585134
Ion Ratio Lower Upper
163 100
194 4.0 3.4 5.0
164 10.1 7.7 11.5



#48
2,6-Dinitrotoluene
Concen: 39.243 ng/ul
RT: 14.075 min Scan# 1868
Delta R.T. 0.006 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27

Tgt Ion:165 Resp: 323578
Ion Ratio Lower Upper
165 100
89 58.6 50.6 75.8
121 22.3 17.9 26.9









#50
Acenaphthylene
Concen: 41.906 ng/ul
RT: 14.222 min Scan# 1893
Delta R.T. 0.000 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:152 Resp: 2203332
Ion Ratio Lower Upper
152 100
151 20.3 16.0 24.0
153 13.4 10.7 16.1



#51
3-Nitroaniline
Concen: 41.681 ng/ul
RT: 14.398 min Scan# 1923
Delta R.T. 0.006 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27

Tgt Ion:138 Resp: 354111
Ion Ratio Lower Upper
138 100
108 12.1 8.6 12.8
92 117.1 92.9 139.3









#52
Acenaphthene
Concen: 41.854 ng/ul
RT: 14.563 min Scan# 1951
Delta R.T. 0.000 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:153 Resp: 1420528
Ion Ratio Lower Upper
153 100
152 46.7 38.4 57.6
154 88.5 69.7 104.5



#53
2,4-Dinitrophenol
Concen: 30.635 ng/ul
RT: 14.598 min Scan# 1957
Delta R.T. 0.000 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27

Tgt Ion:184 Resp: 161815
Ion Ratio Lower Upper
184 100
63 70.5 55.9 83.9
154 66.4 53.1 79.7









#55
4-Nitrophenol
Concen: 38.751 ng/ul
RT: 14.704 min Scan# 1973
Delta R.T. 0.012 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:109 Resp: 255851
Ion Ratio Lower Upper
109 100
139 135.9 100.1 150.1
65 132.4 104.6 157.0



#56
Dibenzofuran
Concen: 39.686 ng/ul
RT: 14.898 min Scan# 2008
Delta R.T. 0.000 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27

Tgt Ion:168 Resp: 1945552
Ion Ratio Lower Upper
168 100
139 39.2 31.1 46.7









#57
2,4-Dinitrotoluene
Concen: 38.637 ng/ul
RT: 14.857 min Scan# 2001
Delta R.T. 0.006 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:165 Resp: 464218
Ion Ratio Lower Upper
165 100
63 45.4 36.3 54.5
182 2.9 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 35.807 ng/ul
RT: 15.122 min Scan# 2046
Delta R.T. 0.000 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27

Tgt Ion:232 Resp: 345925
Ion Ratio Lower Upper
232 100
131 56.1 38.5 57.7
130 2.7 2.2 3.2
166 34.1 25.8 38.8









#59
Diethylphthalate
Concen: 38.724 ng/ul
RT: 15.328 min Scan# 2080
Delta R.T. 0.006 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:149 Resp: 1627750
Ion Ratio Lower Upper
149 100
177 21.4 17.8 26.8
150 12.4 10.5 15.7



#61
Fluorene
Concen: 39.523 ng/ul
RT: 15.545 min Scan# 2118
Delta R.T. 0.000 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27

Tgt Ion:166 Resp: 1511777
Ion Ratio Lower Upper
166 100
165 95.6 77.7 116.5
167 13.1 11.1 16.7









#62
4-Chlorophenyl-phenylether
Concen: 34.975 ng/ul
RT: 15.545 min Scan# 2117
Delta R.T. 0.006 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:204 Resp: 678966
Ion Ratio Lower Upper
204 100
206 32.5 26.7 40.1
141 67.4 50.2 75.2



#63
4-Nitroaniline
Concen: 41.664 ng/ul
RT: 15.557 min Scan# 2120
Delta R.T. 0.006 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27

Tgt Ion:138 Resp: 339444
Ion Ratio Lower Upper
138 100
92 49.0 41.0 61.6
108 75.9 65.3 97.9









#66
4,6-Dinitro-2-methylphenol
Concen: 37.528 ng/ul
RT: 15.616 min Scan# 2129
Delta R.T. 0.006 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:198 Resp: 245690
Ion Ratio Lower Upper
198 100
182 4.4 3.0 4.4
77 29.4 20.1 30.1



#67
N-Nitrosodiphenylamine
Concen: 42.932 ng/ul
RT: 15.751 min Scan# 2153
Delta R.T. 0.000 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27

Tgt Ion:169 Resp: 1334622
Ion Ratio Lower Upper
169 100
168 64.9 52.4 78.6
167 34.9 27.5 41.3









#68
4-Bromophenyl-phenylether
Concen: 37.559 ng/ul
RT: 16.434 min Scan# 2269
Delta R.T. 0.000 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:248 Resp: 406460
Ion Ratio Lower Upper
248 100
250 101.2 77.8 116.8
141 93.6 62.6 93.8



#69
Hexachlorobenzene
Concen: 36.839 ng/ul
RT: 16.551 min Scan# 2289
Delta R.T. 0.000 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27

Tgt Ion:284 Resp: 454384
Ion Ratio Lower Upper
284 100
142 44.2 27.8 41.8#
249 33.2 25.7 38.5





#70

Atrazine

Concen: 38.445 ng/ul

RT: 16.704 min Scan# 2315

Delta R.T. 0.000 min

Lab File: BN019540.D

Acq: 27 Apr 2022 11:27

Instrument :

BNA_N

ClientSampleId :

SSTD040213





Tgt Ion:200 Resp: 446880

Ion Ratio Lower Upper

200 100

173 29.3 20.5 30.7

215 47.0 38.9 58.3



#71

Pentachlorophenol

Concen: 39.205 ng/ul

RT: 16.886 min Scan# 2346

Delta R.T. 0.000 min

Lab File: BN019540.D

Acq: 27 Apr 2022 11:27




Tgt Ion:266 Resp: 293689


Ion Ratio Lower Upper

266 100

264 64.5 50.2 75.2

268 63.5 49.3 73.9









#72
Phenanthrene
Concen: 41.553 ng/ul
RT: 17.281 min Scan# 2413
Delta R.T. 0.000 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:178 Resp: 2367717
Ion Ratio Lower Upper
178 100
179 14.6 12.2 18.4
176 18.9 15.2 22.8



#74
Anthracene
Concen: 41.995 ng/ul
RT: 17.375 min Scan# 2429
Delta R.T. 0.006 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27

Tgt Ion:178 Resp: 2397859
Ion Ratio Lower Upper
178 100
179 15.1 13.0 19.4
176 18.7 15.5 23.3









#75
1,2,3,4-Tetrachlorobenzene
Concen: 40.215 ng/uL
RT: 13.304 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:216 Resp: 657066
Ion Ratio Lower Upper
216 100
214 74.4 63.2 94.8
218 47.7 37.9 56.9



#76
Pentachlorobenzene
Concen: 38.660 ng/uL
RT: 14.822 min Scan# 1995
Delta R.T. 0.006 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27

Tgt Ion:250 Resp: 576218
Ion Ratio Lower Upper
250 100
248 60.0 53.0 79.4
252 59.4 53.4 80.2









#77
 Carbazole
 Concen: 44.310 ng/ul
 RT: 17.639 min Scan# 2473
 Delta R.T. 0.006 min
 Lab File: BN019540.D
 Acq: 27 Apr 2022 11:27




Instrument :
 BNA_N
 ClientSampleId :
 SSTD040213


Tgt Ion:167 Resp: 2195509
 Ion Ratio Lower Upper
 167 100
 166 20.6 16.1 24.1
 139 13.1 10.2 15.2



#78
 Di-n-butylphthalate
 Concen: 42.828 ng/ul
 RT: 18.216 min Scan# 2572
 Delta R.T. 0.000 min
 Lab File: BN019540.D
 Acq: 27 Apr 2022 11:27

Tgt Ion:149 Resp: 2654095
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.1 10.7
 104 5.1 4.1 6.1





#80

Fluoranthene

Concen: 37.622 ng/ul

RT: 19.292 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN019540.D

Acq: 27 Apr 2022 11:27

Instrument :

BNA_N

ClientSampleId :

SSTD040213





Tgt Ion:202 Resp: 2616618

Ion Ratio Lower Upper

202 100

101 13.4 8.9 13.3#

100 10.2 6.7 10.1#



#82

Pyrene

Concen: 37.328 ng/ul

RT: 19.657 min Scan# 2817

Delta R.T. 0.000 min

Lab File: BN019540.D

Acq: 27 Apr 2022 11:27




Tgt Ion:202 Resp: 2631135


Ion Ratio Lower Upper

202 100

101 15.0 9.7 14.5#

100 12.6 8.0 12.0#









#83
Butylbenzylphthalate
Concen: 41.724 ng/ul
RT: 20.569 min Scan# 2971
Delta R.T. 0.006 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:149 Resp: 1172343
Ion Ratio Lower Upper
149 100
91 70.4 58.8 88.2
206 19.9 18.1 27.1



#84
3,3'-Dichlorobenzidine
Concen: 45.448 ng/ul
RT: 21.339 min Scan# 3103
Delta R.T. 0.000 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27

Tgt Ion:252 Resp: 859642
Ion Ratio Lower Upper
252 100
254 64.4 50.8 76.2
126 14.1 8.2 12.4#









#85
Benzo(a)anthracene
Concen: 39.379 ng/ul
RT: 21.404 min Scan# 3114
Delta R.T. 0.000 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:228 Resp: 2426299
Ion Ratio Lower Upper
228 100
229 19.3 16.3 24.5
226 26.4 23.2 34.8



#86
Bis(2-ethylhexyl)phthalate
Concen: 42.767 ng/ul
RT: 21.351 min Scan# 3105
Delta R.T. 0.000 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27

Tgt Ion:149 Resp: 1725606
Ion Ratio Lower Upper
149 100
167 26.6 22.5 33.7
279 3.3 3.5 5.3#









#87
Chrysene
Concen: 39.654 ng/ul
RT: 21.463 min Scan# 3123
Delta R.T. 0.006 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:228 Resp: 2363737
Ion Ratio Lower Upper
228 100
226 30.4 24.2 36.2
229 19.0 16.2 24.2



#89
Di-n-octyl phthalate
Concen: 36.275 ng/ul
RT: 22.269 min Scan# 3261
Delta R.T. 0.000 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27

Tgt Ion:149 Resp: 2999663









#90
Benzo(b)fluoranthene
Concen: 38.946 ng/u1
RT: 23.063 min Scan# 31
Delta R.T. 0.000 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:252 Resp: 2646624
Ion Ratio Lower Upper
252 100
253 20.9 16.5 24.7
125 11.9 7.8 11.8#



#91
Benzo(k)fluoranthene
Concen: 37.276 ng/u1
RT: 23.116 min Scan# 3405
Delta R.T. 0.006 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27

Tgt Ion:252 Resp: 2353267
Ion Ratio Lower Upper
252 100
253 21.4 18.2 27.4
125 12.3 8.2 12.4









#93
Benzo(a)pyrene
Concen: 39.709 ng/ul
RT: 23.680 min Scan# 31
Delta R.T. 0.012 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:252 Resp: 2486958
Ion Ratio Lower Upper
252 100
253 21.5 18.7 28.1
125 13.5 9.2 13.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 44.996 ng/ul
RT: 26.209 min Scan# 3931
Delta R.T. 0.006 min
Lab File: BN019540.D
Acq: 27 Apr 2022 11:27

Tgt Ion:276 Resp: 2909319
Ion Ratio Lower Upper
276 100
138 29.2 19.5 29.3
277 24.8 21.4 32.0





#95

Dibenzo(a,h)anthracene

Concen: 44.348 ng/ul

RT: 26.233 min Scan# 3933

Delta R.T. 0.012 min

Lab File: BN019540.D

Acq: 27 Apr 2022 11:27

Instrument :

BNA_N

ClientSampleId :

SSTD040213





Tgt Ion:278 Resp: 2472251

Ion Ratio Lower Upper

278 100

139 19.2 13.8 20.8

279 23.5 18.3 27.5



#96

Benzo(g,h,i)perylene

Concen: 47.306 ng/ul

RT: 26.957 min Scan# 4058

Delta R.T. 0.012 min

Lab File: BN019540.D

Acq: 27 Apr 2022 11:27

Tgt Ion:276 Resp: 2592646

Ion Ratio Lower Upper

276 100

138 26.5 17.0 25.6#

277 23.0 17.8 26.6

