

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042722\
 Data File : BN019542.D
 Acq On : 27 Apr 2022 12:39
 Operator : CG/JU
 Sample : SSTD16015
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD160215

Quant Time: Apr 27 13:28:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN042722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 27 12:47:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.881	152	163687	20.000	ng/u1	0.00
20) Naphthalene-d8	10.675	136	663544	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.504	164	389488	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.239	188	747411	20.000	ng/u1	0.00
79) Chrysene-d12	21.427	240	603819	20.000	ng/u1	# 0.00
88) Perylene-d12	23.780	264	687866	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.340	96	276977	70.545	ng/uL	0.00
4) Pyridine-d5	3.752	84	1757056	168.331	ng/u1	0.00
7) Phenol-d5	7.052	99	2269637	161.742	ng/u1	0.01
9) Bis-(2-Chloroethyl)eth...	7.216	67	1265222	145.146	ng/u1	0.00
11) 2-Chlorophenol-d4	7.416	132	1820430	170.148	ng/u1	0.00
15) 4-Methylphenol-d8	8.599	113	1764437	156.786	ng/u1	0.02
21) Nitrobenzene-d5	9.040	128	873867	172.292	ng/u1	0.01
24) 2-Nitrophenol-d4	9.758	143	955068	169.527	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.299	165	1574211	153.733	ng/u1	0.01
31) 4-Chloroaniline-d4	10.810	131	2311731	163.780	ng/u1	0.01
46) Dimethylphthalate-d6	13.922	166	4177328	145.370	ng/u1	0.01
49) Acenaphthylene-d8	14.198	160	5413131	152.637	ng/u1	0.00
54) 4-Nitrophenol-d4	14.710	143	912635	173.011	ng/u1	0.03
60) Fluorene-d10	15.498	176	3414625	142.031	ng/u1	0.01
65) 4,6-Dinitro-2-methylph...	15.616	200	715437	155.965	ng/u1	0.02
73) Anthracene-d10	17.345	188	5017489	144.857	ng/u1	0.01
81) Pyrene-d10	19.633	212	5377047	142.837	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.639	264	5415937	154.375	ng/u1	0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.376	88	272827	69.263	ng/uL	95
5) Pyridine	3.770	79	1839746	167.849	ng/u1	97
6) Benzaldehyde	7.022	77	1030471	137.073	ng/u1	91
8) Phenol	7.081	94	2247966	159.234	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.311	93	1710545	151.503	ng/u1	99
12) 2-Chlorophenol	7.452	128	1830363	167.160	ng/u1	97
13) 2-Methylphenol	8.322	108	1688861	159.467	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.416	45	2407810	145.980	ng/u1	99
16) Acetophenone	8.711	105	2490726	148.354	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.716	70	1212649	136.121	ng/u1	98
18) 4-Methylphenol	8.669	108	1793081	155.826	ng/u1	99
19) Hexachloroethane	8.963	117	718568	157.463	ng/u1	87
22) Nitrobenzene	9.081	77	1899326	151.824	ng/u1	95
23) Isophorone	9.628	82	3517849	148.061	ng/u1	99
25) 2-Nitrophenol	9.793	139	993133	169.738	ng/u1	98
26) 2,4-Dimethylphenol	9.858	107	1898678	153.261	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.093	93	2147037	146.757	ng/u1	98
29) 2,4-Dichlorophenol	10.328	162	1540016	152.684	ng/u1	97
30) Naphthalene	10.728	128	5285156	156.895	ng/u1	99
32) 4-Chloroaniline	10.834	127	2275734	161.659	ng/u1	100
33) Hexachlorobutadiene	11.022	225	844168	126.101	ng/u1	99
34) Caprolactam	11.640	113	270827	80.780	ng/u1	95
35) 4-Chloro-3-methylphenol	11.963	107	1756484	154.562	ng/u1	99
36) 2-Methylnaphthalene	12.334	142	3434835	148.490	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042722\
 Data File : BN019542.D
 Acq On : 27 Apr 2022 12:39
 Operator : CG/JU
 Sample : SSTD16015
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD160215

Quant Time: Apr 27 13:28:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN042722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 27 12:47:24 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.557	142	3444459	147.141	ng/ul	95
39) 1,2,4,5-Tetrachloroben...	12.704	216	1525303	132.264	ng/ul	96
40) Hexachlorocyclopentadiene	12.693	237	1017210	153.191	ng/ul	94
41) 2,4,6-Trichlorophenol	12.940	196	1126415	154.669	ng/ul	93
42) 2,4,5-Trichlorophenol	13.016	196	1207255	149.841	ng/ul	99
43) 1,1'-Biphenyl	13.346	154	4244568	147.367	ng/ul	98
44) 2-Chloronaphthalene	13.387	162	3354341	150.979	ng/ul	98
45) 2-Nitroaniline	13.587	65	1125700	166.878	ng/ul	93
47) Dimethylphthalate	13.975	163	4061198	144.650	ng/ul	99
48) 2,6-Dinitrotoluene	14.087	165	907071	163.602	ng/ul	93
50) Acenaphthylene	14.228	152	5306106	150.083	ng/ul	99
51) 3-Nitroaniline	14.404	138	633383	110.875	ng/ul	100
52) Acenaphthene	14.569	153	3446420	151.013	ng/ul	98
53) 2,4-Dinitrophenol	14.610	184	547777	154.228	ng/ul	99
55) 4-Nitrophenol	14.722	109	688253	155.026	ng/ul	95
56) Dibenzofuran	14.904	168	4745578	143.963	ng/ul	96
57) 2,4-Dinitrotoluene	14.869	165	1282008	158.685	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.128	232	947780	145.902	ng/ul	93
59) Diethylphthalate	15.340	149	4090360	144.717	ng/ul	99
61) Fluorene	15.557	166	3363440	130.770	ng/ul	96
62) 4-Chlorophenyl-phenyle...	15.545	204	1555104	119.132	ng/ul	91
63) 4-Nitroaniline	15.569	138	461295	84.205	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.634	198	694774	154.204	ng/ul	97
67) N-Nitrosodiphenylamine	15.757	169	3254852	152.138	ng/ul	98
68) 4-Bromophenyl-phenylether	16.440	248	1035000	138.969	ng/ul	92
69) Hexachlorobenzene	16.557	284	1146924	135.115	ng/ul	94
70) Atrazine	16.716	200	1165403	145.684	ng/ul	95
71) Pentachlorophenol	16.892	266	827502	160.510	ng/ul	95
72) Phenanthrene	17.287	178	5758415	146.845	ng/ul	98
74) Anthracene	17.381	178	5701495	145.093	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.310	216	1623155	144.352	ng/ul	94
76) Pentachlorobenzene	14.828	250	1431840	139.590	ng/ul	97
77) Carbazole	17.645	167	5520121	161.881	ng/ul	96
78) Di-n-butylphthalate	18.222	149	6833544	160.230	ng/ul	99
80) Fluoranthene	19.298	202	6339270	145.408	ng/ul#	94
82) Pyrene	19.663	202	6086068	137.746	ng/ul#	90
83) Butylbenzylphthalate	20.569	149	3162660	179.571	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.345	252	1657686	139.815	ng/ul#	98
85) Benzo(a)anthracene	21.410	228	5843213	151.296	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.357	149	4131052	163.336	ng/ul	99
87) Chrysene	21.469	228	5302828	141.923	ng/ul	98
89) Di-n-octyl phthalate	22.269	149	7714300	138.741	ng/ul	100
90) Benzo(b)fluoranthene	23.080	252	6540775	143.143	ng/ul	98
91) Benzo(k)fluoranthene	23.127	252	5846746	137.735	ng/ul	96
93) Benzo(a)pyrene	23.692	252	6200660	147.244	ng/ul	96
94) Indeno(1,2,3-cd)pyrene	26.233	276	7338947	168.806	ng/ul	93
95) Dibenzo(a,h)anthracene	26.257	278	5971415	159.306	ng/ul	98
96) Benzo(g,h,i)perylene	26.986	276	6504506	176.509	ng/ul#	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042722\
 Data File : BN019542.D
 Acq On : 27 Apr 2022 12:39
 Operator : CG/JU
 Sample : SSTD16015
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD160215

Quant Time: Apr 27 13:28:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN042722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 27 12:47:24 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 69.263 ng/uL
RT: 3.376 min Scan# 49
Delta R.T. -0.006 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion: 88 Resp: 272827
Ion Ratio Lower Upper
88 100
43 29.9 22.4 33.6
58 80.8 61.0 91.4



#5
Pyridine
Concen: 167.849 ng/uL
RT: 3.770 min Scan# 116
Delta R.T. -0.006 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39

Tgt Ion: 79 Resp: 1839746
Ion Ratio Lower Upper
79 100
52 84.5 65.0 97.4
51 38.3 31.6 47.4









#6
Benzaldehyde
Concen: 137.073 ng/ul
RT: 7.022 min Scan# 60
Delta R.T. 0.006 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion: 77 Resp: 1030471
Ion Ratio Lower Upper
77 100
105 100.4 74.7 112.1
106 98.7 70.2 105.4



#8
Phenol
Concen: 159.234 ng/ul
RT: 7.081 min Scan# 679
Delta R.T. 0.018 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39

Tgt Ion: 94 Resp: 2247966
Ion Ratio Lower Upper
94 100
65 28.3 23.6 35.4
66 38.9 31.2 46.8









#10
Bis(2-Chloroethyl)ether
Concen: 151.503 ng/ul
RT: 7.311 min Scan# 716
Delta R.T. 0.012 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion: 93 Resp: 1710545
Ion Ratio Lower Upper
93 100
63 77.7 63.4 95.0
95 32.7 26.2 39.2



#12
2-Chlorophenol
Concen: 167.160 ng/ul
RT: 7.452 min Scan# 742
Delta R.T. 0.012 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39

Tgt Ion:128 Resp: 1830363
Ion Ratio Lower Upper
128 100
64 45.8 38.6 57.8
130 32.2 25.0 37.4





#13

2-Methylphenol

Concen: 159.467 ng/ul

RT: 8.322 min Scan# 890

Delta R.T. 0.006 min

Lab File: BN019542.D

Acq: 27 Apr 2022 12:39

Instrument :

BNA_N

ClientSampleId :





SSTD160215

Tgt Ion:108 Resp: 1688861

Ion Ratio Lower Upper

108 100

107 89.3 71.5 107.3



#14

2,2'-oxybis(1-Chloropropane)

Concen: 145.980 ng/ul

RT: 8.416 min Scan# 906

Delta R.T. 0.000 min

Lab File: BN019542.D

Acq: 27 Apr 2022 12:39




Tgt Ion: 45 Resp: 2407810


Ion Ratio Lower Upper

45 100

77 13.3 10.9 16.3

79 10.2 8.6 13.0









#16
Acetophenone
Concen: 148.354 ng/ul
RT: 8.711 min Scan# 91
Delta R.T. 0.012 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:105 Resp: 2490726
Ion Ratio Lower Upper
105 100
77 76.3 63.5 95.3
51 27.8 23.5 35.3



#17
N-Nitroso-di-n-propylamine
Concen: 136.121 ng/ul
RT: 8.716 min Scan# 957
Delta R.T. 0.030 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39

Tgt Ion: 70 Resp: 1212649
Ion Ratio Lower Upper
70 100
42 59.8 47.8 71.6
101 10.3 7.8 11.6
130 22.5 15.8 23.8









#18
4-Methylphenol
Concen: 155.826 ng/ul
RT: 8.669 min Scan# 944
Delta R.T. 0.030 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:108 Resp: 1793081
Ion Ratio Lower Upper
108 100
107 110.7 88.0 132.0



#19
Hexachloroethane
Concen: 157.463 ng/ul
RT: 8.963 min Scan# 999
Delta R.T. 0.000 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39

Tgt Ion:117 Resp: 718568
Ion Ratio Lower Upper
117 100
201 89.0 83.2 124.8
199 54.8 50.2 75.4









#22
Nitrobenzene
Concen: 151.824 ng/ul
RT: 9.081 min Scan# 1017
Delta R.T. 0.012 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion: 77 Resp: 1899326
Ion Ratio Lower Upper
77 100
123 47.4 34.5 51.7
65 14.3 11.1 16.7



#23
Isophorone
Concen: 148.061 ng/ul
RT: 9.628 min Scan# 1112
Delta R.T. 0.024 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39

Tgt Ion: 82 Resp: 3517849
Ion Ratio Lower Upper
82 100
95 7.5 5.8 8.8
138 18.0 13.8 20.8





#25

2-Nitrophenol

Concen: 169.738 ng/ul

RT: 9.793 min Scan# 1138

Delta R.T. 0.012 min

Lab File: BN019542.D

Acq: 27 Apr 2022 12:39

Instrument :

BNA_N

ClientSampleId :

SSTD160215





Tgt Ion:139 Resp: 993133

Ion Ratio Lower Upper

139 100

65 52.2 43.7 65.5

109 31.9 25.8 38.6



#26

2,4-Dimethylphenol

Concen: 153.261 ng/ul

RT: 9.858 min Scan# 1151

Delta R.T. 0.012 min

Lab File: BN019542.D

Acq: 27 Apr 2022 12:39




Tgt Ion:107 Resp: 1898678


Ion Ratio Lower Upper

107 100

121 49.1 38.3 57.5

122 84.3 66.3 99.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 146.757 ng/ul
 RT: 10.093 min Scan# 1190
 Delta R.T. 0.006 min
 Lab File: BN019542.D
 Acq: 27 Apr 2022 12:39




Instrument :
 BNA_N
 ClientSampleId :
 SSTD160215


Tgt Ion: 93 Resp: 2147037
 Ion Ratio Lower Upper
 93 100
 95 32.3 25.4 38.0
 123 11.6 8.3 12.5



#29
 2,4-Dichlorophenol
 Concen: 152.684 ng/ul
 RT: 10.328 min Scan# 1231
 Delta R.T. 0.012 min
 Lab File: BN019542.D
 Acq: 27 Apr 2022 12:39

Tgt Ion:162 Resp: 1540016
 Ion Ratio Lower Upper
 162 100
 164 62.9 52.9 79.3
 98 38.2 29.8 44.8









#30
Naphthalene
Concen: 156.895 ng/ul
RT: 10.728 min Scan# 1298
Delta R.T. 0.006 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:128 Resp: 5285156
Ion Ratio Lower Upper
128 100
129 11.0 8.7 13.1
127 13.0 10.6 16.0



#32
4-Chloroaniline
Concen: 161.659 ng/ul
RT: 10.834 min Scan# 1317
Delta R.T. 0.012 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39

Tgt Ion:127 Resp: 2275734
Ion Ratio Lower Upper
127 100
129 32.4 26.1 39.1









#33
Hexachlorobutadiene
Concen: 126.101 ng/ul
RT: 11.022 min Scan# 11
Delta R.T. 0.006 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:225 Resp: 844168
Ion Ratio Lower Upper
225 100
223 65.1 51.8 77.6
227 64.3 50.6 76.0



#34
Caprolactam
Concen: 80.780 ng/ul
RT: 11.640 min Scan# 1454
Delta R.T. 0.053 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39

Tgt Ion:113 Resp: 270827
Ion Ratio Lower Upper
113 100
55 170.4 145.5 218.3
56 131.2 106.5 159.7









#35
 4-Chloro-3-methylphenol
 Concen: 154.562 ng/ul
 RT: 11.963 min Scan# 1509
 Delta R.T. 0.018 min
 Lab File: BN019542.D
 Acq: 27 Apr 2022 12:39




Instrument :
 BNA_N
 ClientSampleId :
 SSTD160215


Tgt Ion:107 Resp: 1756484
 Ion Ratio Lower Upper
 107 100
 144 25.5 20.6 31.0
 142 77.2 62.2 93.4



#36
 2-Methylnaphthalene
 Concen: 148.490 ng/ul
 RT: 12.334 min Scan# 1572
 Delta R.T. 0.006 min
 Lab File: BN019542.D
 Acq: 27 Apr 2022 12:39

Tgt Ion:142 Resp: 3434835
 Ion Ratio Lower Upper
 142 100
 141 87.6 70.6 105.8









#37
1-Methylnaphthalene
Concen: 147.141 ng/ul
RT: 12.557 min Scan# 1610
Delta R.T. 0.012 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:142 Resp: 3444459
Ion Ratio Lower Upper
142 100
141 93.1 70.9 106.3
116 4.0 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 132.264 ng/ul
RT: 12.704 min Scan# 1635
Delta R.T. 0.006 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39

Tgt Ion:216 Resp: 1525303
Ion Ratio Lower Upper
216 100
214 75.3 63.4 95.0
179 23.6 17.8 26.8
108 21.3 15.1 22.7









#40
Hexachlorocyclopentadiene
Concen: 153.191 ng/ul
RT: 12.693 min Scan# 1633
Delta R.T. 0.006 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:237 Resp: 1017210
Ion Ratio Lower Upper
237 100
235 56.5 48.8 73.2
272 13.3 9.3 13.9



#41
2,4,6-Trichlorophenol
Concen: 154.669 ng/ul
RT: 12.940 min Scan# 1675
Delta R.T. 0.006 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39

Tgt Ion:196 Resp: 1126415
Ion Ratio Lower Upper
196 100
198 91.6 79.8 119.8
200 29.3 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 149.841 ng/ul
RT: 13.016 min Scan# 1
Delta R.T. 0.018 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:196 Resp: 1207255
Ion Ratio Lower Upper
196 100
198 96.1 77.6 116.4
200 30.5 24.4 36.6



#43
1,1'-Biphenyl
Concen: 147.367 ng/ul
RT: 13.346 min Scan# 1744
Delta R.T. 0.006 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39

Tgt Ion:154 Resp: 4244568
Ion Ratio Lower Upper
154 100
153 41.1 31.5 47.3
76 13.1 10.5 15.7









#44
2-Chloronaphthalene
Concen: 150.979 ng/ul
RT: 13.387 min Scan# 1751
Delta R.T. 0.012 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:162 Resp: 3354341
Ion Ratio Lower Upper
162 100
164 32.3 25.8 38.6
127 41.7 31.4 47.2



#45
2-Nitroaniline
Concen: 166.878 ng/ul
RT: 13.587 min Scan# 1785
Delta R.T. 0.018 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39

Tgt Ion: 65 Resp: 1125700
Ion Ratio Lower Upper
65 100
92 68.8 54.1 81.1
138 107.4 77.0 115.4









#47
Dimethylphthalate
Concen: 144.650 ng/ul
RT: 13.975 min Scan# 1851
Delta R.T. 0.018 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:163 Resp: 4061198
Ion Ratio Lower Upper
163 100
194 4.0 3.4 5.0
164 9.9 7.7 11.5



#48
2,6-Dinitrotoluene
Concen: 163.602 ng/ul
RT: 14.087 min Scan# 1870
Delta R.T. 0.018 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39

Tgt Ion:165 Resp: 907071
Ion Ratio Lower Upper
165 100
89 56.9 50.6 75.8
121 20.6 17.9 26.9









#50
Acenaphthylene
Concen: 150.083 ng/ul
RT: 14.228 min Scan# 1893
Delta R.T. 0.006 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:152 Resp: 5306106
Ion Ratio Lower Upper
152 100
151 20.4 16.0 24.0
153 13.3 10.7 16.1



#51
3-Nitroaniline
Concen: 110.875 ng/ul
RT: 14.404 min Scan# 1924
Delta R.T. 0.012 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39

Tgt Ion:138 Resp: 633383
Ion Ratio Lower Upper
138 100
108 11.0 8.6 12.8
92 116.1 92.9 139.3









#52
Acenaphthene
Concen: 151.013 ng/ul
RT: 14.569 min Scan# 1951
Delta R.T. 0.006 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:153 Resp: 3446420
Ion Ratio Lower Upper
153 100
152 48.5 38.4 57.6
154 89.7 69.7 104.5



#53
2,4-Dinitrophenol
Concen: 154.228 ng/ul
RT: 14.610 min Scan# 1959
Delta R.T. 0.012 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39

Tgt Ion:184 Resp: 547777
Ion Ratio Lower Upper
184 100
63 69.2 55.9 83.9
154 66.0 53.1 79.7









#55
4-Nitrophenol
Concen: 155.026 ng/ul
RT: 14.722 min Scan# 1973
Delta R.T. 0.030 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:109 Resp: 688253
Ion Ratio Lower Upper
109 100
139 134.1 100.1 150.1
65 132.6 104.6 157.0



#56
Dibenzofuran
Concen: 143.963 ng/ul
RT: 14.904 min Scan# 2009
Delta R.T. 0.006 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39

Tgt Ion:168 Resp: 4745578
Ion Ratio Lower Upper
168 100
139 41.0 31.1 46.7









#57
2,4-Dinitrotoluene
Concen: 158.685 ng/ul
RT: 14.869 min Scan# 2000
Delta R.T. 0.018 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:165 Resp: 1282008
Ion Ratio Lower Upper
165 100
63 45.2 36.3 54.5
182 2.9 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 145.902 ng/ul
RT: 15.128 min Scan# 2047
Delta R.T. 0.006 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39

Tgt Ion:232 Resp: 947780
Ion Ratio Lower Upper
232 100
131 54.2 38.5 57.7
130 2.8 2.2 3.2
166 35.4 25.8 38.8









#59
Diethylphthalate
Concen: 144.717 ng/ul
RT: 15.340 min Scan# 2083
Delta R.T. 0.018 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:149 Resp: 4090360
Ion Ratio Lower Upper
149 100
177 21.6 17.8 26.8
150 12.9 10.5 15.7



#61
Fluorene
Concen: 130.770 ng/ul
RT: 15.557 min Scan# 2120
Delta R.T. 0.012 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39

Tgt Ion:166 Resp: 3363440
Ion Ratio Lower Upper
166 100
165 101.3 77.7 116.5
167 13.9 11.1 16.7









#62
4-Chlorophenyl-phenylether
Concen: 119.132 ng/ul
RT: 15.545 min Scan# 2117
Delta R.T. 0.006 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:204 Resp: 1555104
Ion Ratio Lower Upper
204 100
206 32.1 26.7 40.1
141 71.9 50.2 75.2



#63
4-Nitroaniline
Concen: 84.205 ng/ul
RT: 15.569 min Scan# 2122
Delta R.T. 0.018 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39

Tgt Ion:138 Resp: 461295
Ion Ratio Lower Upper
138 100
92 51.3 41.0 61.6
108 77.8 65.3 97.9









#66
 4,6-Dinitro-2-methylphenol
 Concen: 154.204 ng/ul
 RT: 15.634 min Scan# 2129
 Delta R.T. 0.024 min
 Lab File: BN019542.D
 Acq: 27 Apr 2022 12:39




Instrument :
 BNA_N
 ClientSampleId :
 SSTD160215


Tgt Ion:198 Resp: 694774
 Ion Ratio Lower Upper
 198 100
 182 3.6 3.0 4.4
 77 26.8 20.1 30.1



#67
 N-Nitrosodiphenylamine
 Concen: 152.138 ng/ul
 RT: 15.757 min Scan# 2154
 Delta R.T. 0.006 min
 Lab File: BN019542.D
 Acq: 27 Apr 2022 12:39

Tgt Ion:169 Resp: 3254852
 Ion Ratio Lower Upper
 169 100
 168 66.5 52.4 78.6
 167 35.8 27.5 41.3









#68
4-Bromophenyl-phenylether
Concen: 138.969 ng/ul
RT: 16.440 min Scan# 21
Delta R.T. 0.006 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:248 Resp: 1035000
Ion Ratio Lower Upper
248 100
250 99.2 77.8 116.8
141 91.1 62.6 93.8



#69
Hexachlorobenzene
Concen: 135.115 ng/ul
RT: 16.557 min Scan# 2290
Delta R.T. 0.006 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39

Tgt Ion:284 Resp: 1146924
Ion Ratio Lower Upper
284 100
142 41.5 27.8 41.8
249 32.5 25.7 38.5





#70

Atrazine

Concen: 145.684 ng/ul

RT: 16.716 min Scan# 21

Delta R.T. 0.012 min

Lab File: BN019542.D

Acq: 27 Apr 2022 12:39

Instrument :

BNA_N

ClientSampleId :

SSTD160215





Tgt Ion:200 Resp: 1165403

Ion Ratio Lower Upper

200 100

173 29.2 20.5 30.7

215 46.4 38.9 58.3



#71

Pentachlorophenol

Concen: 160.510 ng/ul

RT: 16.892 min Scan# 2347

Delta R.T. 0.006 min

Lab File: BN019542.D

Acq: 27 Apr 2022 12:39




Tgt Ion:266 Resp: 827502


Ion Ratio Lower Upper

266 100

264 65.7 50.2 75.2

268 66.5 49.3 73.9





#72

Phenanthrene

Concen: 146.845 ng/u1

RT: 17.287 min Scan# 2413

Delta R.T. 0.006 min

Lab File: BN019542.D

Acq: 27 Apr 2022 12:39

Instrument :

BNA_N

ClientSampleId :

SSTD160215





Tgt Ion:178 Resp: 5758415

Ion Ratio Lower Upper

178 100

179 16.1 12.2 18.4

176 20.0 15.2 22.8



#74

Anthracene

Concen: 145.093 ng/u1

RT: 17.381 min Scan# 2430

Delta R.T. 0.012 min

Lab File: BN019542.D

Acq: 27 Apr 2022 12:39




Tgt Ion:178 Resp: 5701495


Ion Ratio Lower Upper

178 100

179 15.2 13.0 19.4

176 18.9 15.5 23.3









#75
1,2,3,4-Tetrachlorobenzene
Concen: 144.352 ng/uL
RT: 13.310 min Scan# 11
Delta R.T. 0.006 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:216 Resp: 1623155
Ion Ratio Lower Upper
216 100
214 83.5 63.2 94.8
218 52.3 37.9 56.9



#76
Pentachlorobenzene
Concen: 139.590 ng/uL
RT: 14.828 min Scan# 1996
Delta R.T. 0.012 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39

Tgt Ion:250 Resp: 1431840
Ion Ratio Lower Upper
250 100
248 62.6 53.0 79.4
252 66.1 53.4 80.2









#77
 Carbazole
 Concen: 161.881 ng/ul
 RT: 17.645 min Scan# 2475
 Delta R.T. 0.012 min
 Lab File: BN019542.D
 Acq: 27 Apr 2022 12:39




Instrument :
 BNA_N
 ClientSampleId :
 SSTD160215


Tgt Ion:167 Resp: 5520121
 Ion Ratio Lower Upper
 167 100
 166 21.8 16.1 24.1
 139 14.6 10.2 15.2



#78
 Di-n-butylphthalate
 Concen: 160.230 ng/ul
 RT: 18.222 min Scan# 2573
 Delta R.T. 0.006 min
 Lab File: BN019542.D
 Acq: 27 Apr 2022 12:39

Tgt Ion:149 Resp: 6833544
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.1 10.7
 104 5.3 4.1 6.1









#80
Fluoranthene
Concen: 145.408 ng/u1
RT: 19.298 min Scan# 21
Delta R.T. 0.006 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:202 Resp: 6339270
Ion Ratio Lower Upper
202 100
101 13.6 8.9 13.3#
100 10.5 6.7 10.1#



#82
Pyrene
Concen: 137.746 ng/u1
RT: 19.663 min Scan# 2818
Delta R.T. 0.006 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39

Tgt Ion:202 Resp: 6086068
Ion Ratio Lower Upper
202 100
101 16.3 9.7 14.5#
100 13.3 8.0 12.0#









#83
 Butylbenzylphthalate
 Concen: 179.571 ng/ul
 RT: 20.569 min Scan# 2972
 Delta R.T. 0.006 min
 Lab File: BN019542.D
 Acq: 27 Apr 2022 12:39




Instrument :
 BNA_N
 ClientSampleId :
 SSTD160215


Tgt Ion:149 Resp: 3162660
 Ion Ratio Lower Upper
 149 100
 91 70.6 58.8 88.2
 206 19.4 18.1 27.1



#84
 3,3'-Dichlorobenzidine
 Concen: 139.815 ng/ul
 RT: 21.345 min Scan# 3104
 Delta R.T. 0.006 min
 Lab File: BN019542.D
 Acq: 27 Apr 2022 12:39

Tgt Ion:252 Resp: 1657686
 Ion Ratio Lower Upper
 252 100
 254 64.5 50.8 76.2
 126 13.2 8.2 12.4#









#85
Benzo(a)anthracene
Concen: 151.296 ng/ul
RT: 21.410 min Scan# 3115
Delta R.T. 0.006 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:228 Resp: 5843213
Ion Ratio Lower Upper
228 100
229 19.1 16.3 24.5
226 27.6 23.2 34.8



#86
Bis(2-ethylhexyl)phthalate
Concen: 163.336 ng/ul
RT: 21.357 min Scan# 3106
Delta R.T. 0.006 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39

Tgt Ion:149 Resp: 4131052
Ion Ratio Lower Upper
149 100
167 27.3 22.5 33.7
279 4.1 3.5 5.3









#87
Chrysene
Concen: 141.923 ng/ul
RT: 21.469 min Scan# 3125
Delta R.T. 0.012 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:228 Resp: 5302828
Ion Ratio Lower Upper
228 100
226 31.6 24.2 36.2
229 20.3 16.2 24.2



#89
Di-n-octyl phthalate
Concen: 138.741 ng/ul
RT: 22.269 min Scan# 3261
Delta R.T. 0.000 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39

Tgt Ion:149 Resp: 7714300









#90
Benzo(b)fluoranthene
Concen: 143.143 ng/ul
RT: 23.080 min Scan# 31
Delta R.T. 0.018 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:252 Resp: 6540775
Ion Ratio Lower Upper
252 100
253 21.1 16.5 24.7
125 11.5 7.8 11.8



#91
Benzo(k)fluoranthene
Concen: 137.735 ng/ul
RT: 23.127 min Scan# 3407
Delta R.T. 0.018 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39

Tgt Ion:252 Resp: 5846746
Ion Ratio Lower Upper
252 100
253 20.7 18.2 27.4
125 11.6 8.2 12.4









#93
Benzo(a)pyrene
Concen: 147.244 ng/ul
RT: 23.692 min Scan# 31
Delta R.T. 0.024 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:252 Resp: 6200660
Ion Ratio Lower Upper
252 100
253 21.3 18.7 28.1
125 13.3 9.2 13.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 168.806 ng/ul
RT: 26.233 min Scan# 3935
Delta R.T. 0.030 min
Lab File: BN019542.D
Acq: 27 Apr 2022 12:39

Tgt Ion:276 Resp: 7338947
Ion Ratio Lower Upper
276 100
138 29.1 19.5 29.3
277 24.3 21.4 32.0





#95

Dibenzo(a,h)anthracene

Concen: 159.306 ng/ul

RT: 26.257 min Scan# 3933

Delta R.T. 0.036 min

Lab File: BN019542.D

Acq: 27 Apr 2022 12:39

Instrument :

BNA_N

ClientSampleId :

SSTD160215





Tgt Ion:278 Resp: 5971415

Ion Ratio Lower Upper

278 100

139 18.9 13.8 20.8

279 23.4 18.3 27.5



#96

Benzo(g,h,i)perylene

Concen: 176.509 ng/ul

RT: 26.986 min Scan# 4063

Delta R.T. 0.041 min

Lab File: BN019542.D

Acq: 27 Apr 2022 12:39

Tgt Ion:276 Resp: 6504506

Ion Ratio Lower Upper

276 100

138 26.6 17.0 25.6#

277 22.5 17.8 26.6

