

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042722\
 Data File : BN019543.D
 Acq On : 27 Apr 2022 13:52
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SICV216

Quant Time: Apr 27 19:04:51 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN042722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 27 13:41:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.869	152	168601	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.663	136	673581	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.498	164	374584	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.233	188	688491	20.000	ng/u1	0.00
79) Chrysene-d12	21.421	240	591907	20.000	ng/u1	0.00
88) Perylene-d12	23.780	264	596962	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.340	96	37415	8.375	ng/uL	0.00
4) Pyridine-d5	3.746	84	229142	20.254	ng/u1	0.00
7) Phenol-d5	7.028	99	298357	20.420	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	7.199	67	173175	20.434	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.399	132	242705	20.942	ng/u1	-0.01
15) 4-Methylphenol-d8	8.569	113	226954	20.277	ng/u1	-0.01
21) Nitrobenzene-d5	9.022	128	107424	21.873	ng/u1	0.00
24) 2-Nitrophenol-d4	9.746	143	106377	21.979	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.281	165	206594	21.324	ng/u1	0.00
31) 4-Chloroaniline-d4	10.793	131	300516	20.433	ng/u1	0.00
46) Dimethylphthalate-d6	13.910	166	547537	21.282	ng/u1	0.00
49) Acenaphthylene-d8	14.187	160	742724	21.351	ng/u1	0.00
54) 4-Nitrophenol-d4	14.675	143	98242	19.634	ng/u1	0.00
60) Fluorene-d10	15.486	176	473696	20.857	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.598	200	64009	18.211	ng/u1	0.00
73) Anthracene-d10	17.333	188	684343	21.106	ng/u1	0.00
81) Pyrene-d10	19.627	212	725882	21.330	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.627	264	669369	21.935	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.369	88	37175	8.159	ng/uL	94
5) Pyridine	3.764	79	245048	20.555	ng/u1	98
6) Benzaldehyde	7.005	77	161398	21.553	ng/u1	95
8) Phenol	7.058	94	304248	20.694	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.293	93	234828	20.585	ng/u1	98
12) 2-Chlorophenol	7.434	128	248982	21.155	ng/u1	98
13) 2-Methylphenol	8.305	108	219081	20.125	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.405	45	336106	20.500	ng/u1	99
16) Acetophenone	8.687	105	353523	21.048	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.675	70	166500	21.299	ng/u1	97
18) 4-Methylphenol	8.634	108	242100	20.771	ng/u1	100
19) Hexachloroethane	8.957	117	97190	20.370	ng/u1	89
22) Nitrobenzene	9.063	77	251766	21.743	ng/u1	95
23) Isophorone	9.593	82	440948	21.352	ng/u1	99
25) 2-Nitrophenol	9.775	139	120718	22.088	ng/u1	96
26) 2,4-Dimethylphenol	9.840	107	260967	21.435	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.075	93	289532	21.321	ng/u1	99
29) 2,4-Dichlorophenol	10.304	162	207382	21.350	ng/u1	98
30) Naphthalene	10.716	128	751041	20.819	ng/u1	99
32) 4-Chloroaniline	10.816	127	309583	20.943	ng/u1	99
33) Hexachlorobutadiene	11.016	225	122968	21.101	ng/u1	98
34) Caprolactam	11.575	113	53981	19.650	ng/u1	89
35) 4-Chloro-3-methylphenol	11.940	107	217150	21.121	ng/u1	98
36) 2-Methylnaphthalene	12.322	142	491617	21.175	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042722\
 Data File : BN019543.D
 Acq On : 27 Apr 2022 13:52
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SICV216


Quant Time: Apr 27 19:04:51 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN042722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 27 13:41:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.545	142	493102	21.215	ng/ul	96
39) 1,2,4,5-Tetrachloroben...	12.693	216	224044	21.556	ng/ul	97
40) Hexachlorocyclopentadiene	12.681	237	132370	20.698	ng/ul	98
41) 2,4,6-Trichlorophenol	12.928	196	137436	21.197	ng/ul	99
42) 2,4,5-Trichlorophenol	12.998	196	148917	21.153	ng/ul	98
43) 1,1'-Biphenyl	13.334	154	624654	21.092	ng/ul	99
44) 2-Chloronaphthalene	13.375	162	474961	20.990	ng/ul	97
45) 2-Nitroaniline	13.569	65	117760	21.687	ng/ul	93
47) Dimethylphthalate	13.951	163	543378	21.267	ng/ul	99
48) 2,6-Dinitrotoluene	14.063	165	96089	21.801	ng/ul	96
50) Acenaphthylene	14.216	152	760035	21.266	ng/ul	99
51) 3-Nitroaniline	14.387	138	75317	16.671	ng/ul#	95
52) Acenaphthene	14.557	153	488456	20.883	ng/ul	97
53) 2,4-Dinitrophenol	14.592	184	40768	16.675	ng/ul	97
55) 4-Nitrophenol	14.687	109	78630	19.776	ng/ul	95
56) Dibenzofuran	14.892	168	674478	20.643	ng/ul	98
57) 2,4-Dinitrotoluene	14.845	165	141018	22.002	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.116	232	110366	21.493	ng/ul	92
59) Diethylphthalate	15.322	149	518052	20.615	ng/ul	99
61) Fluorene	15.539	166	537017	21.172	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.539	204	248408	21.660	ng/ul	97
63) 4-Nitroaniline	15.545	138	52807	13.638	ng/ul	90
66) 4,6-Dinitro-2-methylph...	15.610	198	69749	19.511	ng/ul#	98
67) N-Nitrosodiphenylamine	15.751	169	456285	21.599	ng/ul	99
68) 4-Bromophenyl-phenylether	16.434	248	132389	20.468	ng/ul	94
69) Hexachlorobenzene	16.545	284	147135	20.224	ng/ul#	91
70) Atrazine	16.698	200	137730	20.438	ng/ul	97
71) Pentachlorophenol	16.886	266	84961	19.244	ng/ul	98
72) Phenanthrene	17.281	178	818676	21.278	ng/ul	99
74) Anthracene	17.369	178	816292	21.238	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.298	216	230276	21.697	ng/ul	98
76) Pentachlorobenzene	14.816	250	198322	21.450	ng/ul	96
77) Carbazole	17.633	167	739700	21.170	ng/ul	98
78) Di-n-butylphthalate	18.216	149	766485	20.507	ng/ul	99
80) Fluoranthene	19.292	202	874224	21.793	ng/ul	96
82) Pyrene	19.657	202	899642	21.531	ng/ul#	94
83) Butylbenzylphthalate	20.569	149	329611	21.326	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.339	252	238314	20.048	ng/ul#	96
85) Benzo(a)anthracene	21.410	228	817144	21.555	ng/ul	94
86) Bis(2-ethylhexyl)phtha...	21.357	149	502159	21.174	ng/ul	99
87) Chrysene	21.457	228	791957	21.267	ng/ul	98
89) Di-n-octyl phthalate	22.269	149	805573	19.712	ng/ul	100
90) Benzo(b)fluoranthene	23.063	252	809443	20.960	ng/ul#	96
91) Benzo(k)fluoranthene	23.110	252	784930	22.408	ng/ul	97
93) Benzo(a)pyrene	23.674	252	823142	22.445	ng/ul	95
94) Indeno(1,2,3-cd)pyrene	26.209	276	913742	21.742	ng/ul#	91
95) Dibenzo(a,h)anthracene	26.221	278	796462	22.163	ng/ul	97
96) Benzo(g,h,i)perylene	26.951	276	804924	21.775	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 27 19:04:51 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN042722.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Apr 27 13:41:44 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.159 ng/uL
RT: 3.369 min Scan# 41
Delta R.T. -0.011 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SICV216


Tgt Ion: 88 Resp: 37175
Ion Ratio Lower Upper
88 100
43 31.7 22.4 33.6
58 80.6 61.0 91.4



#5
Pyridine
Concen: 20.555 ng/uL
RT: 3.764 min Scan# 115
Delta R.T. -0.012 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52

Tgt Ion: 79 Resp: 245048
Ion Ratio Lower Upper
79 100
52 82.9 65.0 97.4
51 39.1 31.6 47.4









#6
Benzaldehyde
Concen: 21.553 ng/ul
RT: 7.005 min Scan# 666
Delta R.T. -0.012 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SICV216


Tgt Ion: 77 Resp: 161398
Ion Ratio Lower Upper
77 100
105 95.8 74.7 112.1
106 94.4 70.2 105.4



#8
Phenol
Concen: 20.694 ng/ul
RT: 7.058 min Scan# 675
Delta R.T. -0.006 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52

Tgt Ion: 94 Resp: 304248
Ion Ratio Lower Upper
94 100
65 28.2 23.6 35.4
66 37.5 31.2 46.8









#10
Bis(2-Chloroethyl)ether
Concen: 20.585 ng/ul
RT: 7.293 min Scan# 716
Delta R.T. -0.006 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SICV216


Tgt Ion: 93 Resp: 234828
Ion Ratio Lower Upper
93 100
63 77.1 63.4 95.0
95 32.5 26.2 39.2



#12
2-Chlorophenol
Concen: 21.155 ng/ul
RT: 7.434 min Scan# 739
Delta R.T. -0.006 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52

Tgt Ion:128 Resp: 248982
Ion Ratio Lower Upper
128 100
64 47.7 38.6 57.8
130 33.3 25.0 37.4





#13

2-Methylphenol

Concen: 20.125 ng/ul

RT: 8.305 min Scan# 889

Delta R.T. -0.012 min

Lab File: BN019543.D

Acq: 27 Apr 2022 13:52

Instrument :

BNA_N

ClientSampleId :





SICV216

Tgt Ion:108 Resp: 219081

Ion Ratio Lower Upper

108 100

107 88.8 71.5 107.3



#14

2,2'-oxybis(1-Chloropropane)

Concen: 20.500 ng/ul

RT: 8.405 min Scan# 904

Delta R.T. -0.012 min

Lab File: BN019543.D

Acq: 27 Apr 2022 13:52




Tgt Ion: 45 Resp: 336106


Ion Ratio Lower Upper

45 100

77 13.7 10.9 16.3

79 10.0 8.6 13.0









#16
Acetophenone
Concen: 21.048 ng/ul
RT: 8.687 min Scan# 91
Delta R.T. -0.012 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SICV216


Tgt Ion:105 Resp: 353523
Ion Ratio Lower Upper
105 100
77 76.4 63.5 95.3
51 28.1 23.5 35.3



#17
N-Nitroso-di-n-propylamine
Concen: 21.299 ng/ul
RT: 8.675 min Scan# 950
Delta R.T. -0.012 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52

Tgt Ion: 70 Resp: 166500
Ion Ratio Lower Upper
70 100
42 61.9 47.8 71.6
101 9.5 7.8 11.6
130 21.1 15.8 23.8









#18
4-Methylphenol
Concen: 20.771 ng/ul
RT: 8.634 min Scan# 944
Delta R.T. -0.006 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SICV216


Tgt Ion:108 Resp: 242100
Ion Ratio Lower Upper
108 100
107 109.6 88.0 132.0



#19
Hexachloroethane
Concen: 20.370 ng/ul
RT: 8.957 min Scan# 998
Delta R.T. -0.006 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52

Tgt Ion:117 Resp: 97190
Ion Ratio Lower Upper
117 100
201 91.8 83.2 124.8
199 54.7 50.2 75.4





#22

Nitrobenzene

Concen: 21.743 ng/ul

RT: 9.063 min Scan# 1017

Delta R.T. -0.006 min

Lab File: BN019543.D

Acq: 27 Apr 2022 13:52

Instrument :

BNA_N

ClientSampleId :

SICV216





Tgt Ion: 77 Resp: 251766

Ion Ratio Lower Upper

77 100

123 46.9 34.5 51.7

65 14.0 11.1 16.7



#23

Isophorone

Concen: 21.352 ng/ul

RT: 9.593 min Scan# 1106

Delta R.T. -0.012 min

Lab File: BN019543.D

Acq: 27 Apr 2022 13:52




Tgt Ion: 82 Resp: 440948


Ion Ratio Lower Upper

82 100

95 7.4 5.8 8.8

138 18.1 13.8 20.8





#25

2-Nitrophenol

Concen: 22.088 ng/ul

RT: 9.775 min Scan# 1137

Delta R.T. -0.006 min

Lab File: BN019543.D

Acq: 27 Apr 2022 13:52

Instrument :

BNA_N

ClientSampleId :

SICV216





Tgt Ion:139 Resp: 120718

Ion Ratio Lower Upper

139 100

65 50.6 43.7 65.5

109 31.9 25.8 38.6



#26

2,4-Dimethylphenol

Concen: 21.435 ng/ul

RT: 9.840 min Scan# 1148

Delta R.T. -0.006 min

Lab File: BN019543.D

Acq: 27 Apr 2022 13:52




Tgt Ion:107 Resp: 260967


Ion Ratio Lower Upper

107 100

121 50.1 38.3 57.5

122 86.4 66.3 99.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 21.321 ng/ul
 RT: 10.075 min Scan# 1190
 Delta R.T. -0.012 min
 Lab File: BN019543.D
 Acq: 27 Apr 2022 13:52




Instrument :
 BNA_N
 ClientSampleId :
 SICV216


Tgt Ion: 93 Resp: 289532
 Ion Ratio Lower Upper
 93 100
 95 31.9 25.4 38.0
 123 11.7 8.3 12.5



#29
 2,4-Dichlorophenol
 Concen: 21.350 ng/ul
 RT: 10.304 min Scan# 1227
 Delta R.T. -0.012 min
 Lab File: BN019543.D
 Acq: 27 Apr 2022 13:52

Tgt Ion:162 Resp: 207382
 Ion Ratio Lower Upper
 162 100
 164 63.8 52.9 79.3
 98 37.5 29.8 44.8









#30
Naphthalene
Concen: 20.819 ng/ul
RT: 10.716 min Scan# 11
Delta R.T. -0.006 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SICV216


Tgt Ion:128 Resp: 751041
Ion Ratio Lower Upper
128 100
129 11.1 8.7 13.1
127 13.8 10.6 16.0



#32
4-Chloroaniline
Concen: 20.943 ng/ul
RT: 10.816 min Scan# 1314
Delta R.T. -0.006 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52

Tgt Ion:127 Resp: 309583
Ion Ratio Lower Upper
127 100
129 32.2 26.1 39.1









#33
Hexachlorobutadiene
Concen: 21.101 ng/ul
RT: 11.016 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SICV216


Tgt Ion:225 Resp: 122968
Ion Ratio Lower Upper
225 100
223 61.9 51.8 77.6
227 62.4 50.6 76.0



#34
Caprolactam
Concen: 19.650 ng/ul
RT: 11.575 min Scan# 1443
Delta R.T. -0.012 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52

Tgt Ion:113 Resp: 53981
Ion Ratio Lower Upper
113 100
55 165.3 145.5 218.3
56 121.2 106.5 159.7









#35
4-Chloro-3-methylphenol
Concen: 21.121 ng/ul
RT: 11.940 min Scan# 1506
Delta R.T. -0.006 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SICV216


Tgt Ion:107 Resp: 217150
Ion Ratio Lower Upper
107 100
144 24.1 20.6 31.0
142 79.4 62.2 93.4



#36
2-Methylnaphthalene
Concen: 21.175 ng/ul
RT: 12.322 min Scan# 1570
Delta R.T. -0.006 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52

Tgt Ion:142 Resp: 491617
Ion Ratio Lower Upper
142 100
141 87.4 70.6 105.8





#37

1-Methylnaphthalene

Concen: 21.215 ng/ul

RT: 12.545 min Scan# 1608

Delta R.T. 0.000 min

Lab File: BN019543.D

Acq: 27 Apr 2022 13:52

Instrument :

BNA_N

ClientSampleId :

SICV216





Tgt Ion:142 Resp: 493102

Ion Ratio Lower Upper

142 100

141 92.4 70.9 106.3

116 3.9 3.1 4.7



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.556 ng/ul

RT: 12.693 min Scan# 1633

Delta R.T. -0.006 min

Lab File: BN019543.D

Acq: 27 Apr 2022 13:52

Tgt Ion:216 Resp: 224044




Ion Ratio Lower Upper


216 100

214 80.1 63.4 95.0

179 24.1 17.8 26.8

108 22.3 15.1 22.7









#40
Hexachlorocyclopentadiene
Concen: 20.698 ng/ul
RT: 12.681 min Scan# 1631
Delta R.T. -0.006 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SICV216


Tgt Ion:237 Resp: 132370
Ion Ratio Lower Upper
237 100
235 60.0 48.8 73.2
272 12.6 9.3 13.9



#41
2,4,6-Trichlorophenol
Concen: 21.197 ng/ul
RT: 12.928 min Scan# 1673
Delta R.T. -0.006 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52

Tgt Ion:196 Resp: 137436
Ion Ratio Lower Upper
196 100
198 99.1 79.8 119.8
200 32.8 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 21.153 ng/ul
RT: 12.998 min Scan# 1685
Delta R.T. 0.000 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SICV216


Tgt Ion:196 Resp: 148917
Ion Ratio Lower Upper
196 100
198 94.2 77.6 116.4
200 30.6 24.4 36.6



#43
1,1'-Biphenyl
Concen: 21.092 ng/ul
RT: 13.334 min Scan# 1742
Delta R.T. -0.006 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52

Tgt Ion:154 Resp: 624654
Ion Ratio Lower Upper
154 100
153 39.8 31.5 47.3
76 12.4 10.5 15.7









#44
2-Chloronaphthalene
Concen: 20.990 ng/ul
RT: 13.375 min Scan# 1749
Delta R.T. 0.000 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SICV216


Tgt Ion:162 Resp: 474961
Ion Ratio Lower Upper
162 100
164 33.8 25.8 38.6
127 40.9 31.4 47.2



#45
2-Nitroaniline
Concen: 21.687 ng/ul
RT: 13.569 min Scan# 1782
Delta R.T. 0.000 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52

Tgt Ion: 65 Resp: 117760
Ion Ratio Lower Upper
65 100
92 67.7 54.1 81.1
138 107.4 77.0 115.4





#47

Dimethylphthalate

Concen: 21.267 ng/ul

RT: 13.951 min Scan# 1847

Delta R.T. -0.006 min

Lab File: BN019543.D

Acq: 27 Apr 2022 13:52

Instrument :

BNA_N

ClientSampleId :

SICV216





Tgt Ion:163 Resp: 543378

Ion Ratio Lower Upper

163 100

194 3.7 3.4 5.0

164 9.9 7.7 11.5



#48

2,6-Dinitrotoluene

Concen: 21.801 ng/ul

RT: 14.063 min Scan# 1866

Delta R.T. -0.006 min

Lab File: BN019543.D

Acq: 27 Apr 2022 13:52




Tgt Ion:165 Resp: 96089


Ion Ratio Lower Upper

165 100

89 60.1 50.6 75.8

121 24.1 17.9 26.9









#50
Acenaphthylene
Concen: 21.266 ng/ul
RT: 14.216 min Scan# 1892
Delta R.T. -0.006 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SICV216


Tgt Ion:152 Resp: 760035
Ion Ratio Lower Upper
152 100
151 19.4 16.0 24.0
153 12.8 10.7 16.1



#51
3-Nitroaniline
Concen: 16.671 ng/ul
RT: 14.387 min Scan# 1921
Delta R.T. -0.006 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52

Tgt Ion:138 Resp: 75317
Ion Ratio Lower Upper
138 100
108 13.2 8.6 12.8#
92 121.9 92.9 139.3









#52
Acenaphthene
Concen: 20.883 ng/ul
RT: 14.557 min Scan# 1950
Delta R.T. -0.006 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SICV216


Tgt Ion:153 Resp: 488456
Ion Ratio Lower Upper
153 100
152 46.5 38.4 57.6
154 89.7 69.7 104.5



#53
2,4-Dinitrophenol
Concen: 16.675 ng/ul
RT: 14.592 min Scan# 1956
Delta R.T. -0.006 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52

Tgt Ion:184 Resp: 40768
Ion Ratio Lower Upper
184 100
63 71.9 55.9 83.9
154 69.9 53.1 79.7









#55
4-Nitrophenol
Concen: 19.776 ng/ul
RT: 14.687 min Scan# 1973
Delta R.T. -0.006 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SICV216


Tgt Ion:109 Resp: 78630
Ion Ratio Lower Upper
109 100
139 134.5 100.1 150.1
65 132.0 104.6 157.0



#56
Dibenzofuran
Concen: 20.643 ng/ul
RT: 14.892 min Scan# 2007
Delta R.T. -0.006 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52

Tgt Ion:168 Resp: 674478
Ion Ratio Lower Upper
168 100
139 40.0 31.1 46.7









#57
2,4-Dinitrotoluene
Concen: 22.002 ng/ul
RT: 14.845 min Scan# 1999
Delta R.T. -0.006 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SICV216


Tgt Ion:165 Resp: 141018
Ion Ratio Lower Upper
165 100
63 45.3 36.3 54.5
182 2.7 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.493 ng/ul
RT: 15.116 min Scan# 2045
Delta R.T. -0.006 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52

Tgt Ion:232 Resp: 110366
Ion Ratio Lower Upper
232 100
131 54.6 38.5 57.7
130 2.9 2.2 3.2
166 35.4 25.8 38.8









#59
Diethylphthalate
Concen: 20.615 ng/ul
RT: 15.322 min Scan# 2080
Delta R.T. 0.000 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SICV216


Tgt Ion:149 Resp: 518052
Ion Ratio Lower Upper
149 100
177 22.1 17.8 26.8
150 12.7 10.5 15.7



#61
Fluorene
Concen: 21.172 ng/ul
RT: 15.539 min Scan# 2117
Delta R.T. -0.006 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52

Tgt Ion:166 Resp: 537017
Ion Ratio Lower Upper
166 100
165 98.7 77.7 116.5
167 13.8 11.1 16.7









#62
4-Chlorophenyl-phenylether
Concen: 21.660 ng/ul
RT: 15.539 min Scan# 2117
Delta R.T. 0.000 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SICV216


Tgt Ion:204 Resp: 248408
Ion Ratio Lower Upper
204 100
206 34.1 26.7 40.1
141 66.0 50.2 75.2



#63
4-Nitroaniline
Concen: 13.638 ng/ul
RT: 15.545 min Scan# 2118
Delta R.T. -0.006 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52

Tgt Ion:138 Resp: 52807
Ion Ratio Lower Upper
138 100
92 50.1 41.0 61.6
108 67.9 65.3 97.9









#66
 4,6-Dinitro-2-methylphenol
 Concen: 19.511 ng/ul
 RT: 15.610 min Scan# 2129
 Delta R.T. 0.000 min
 Lab File: BN019543.D
 Acq: 27 Apr 2022 13:52




Instrument :
 BNA_N
 ClientSampleId :
 SICV216


Tgt Ion:198 Resp: 69749
 Ion Ratio Lower Upper
 198 100
 182 5.1 3.0 4.4#
 77 25.9 20.1 30.1



#67
 N-Nitrosodiphenylamine
 Concen: 21.599 ng/ul
 RT: 15.751 min Scan# 2153
 Delta R.T. 0.000 min
 Lab File: BN019543.D
 Acq: 27 Apr 2022 13:52

Tgt Ion:169 Resp: 456285
 Ion Ratio Lower Upper
 169 100
 168 64.3 52.4 78.6
 167 34.1 27.5 41.3









#68
4-Bromophenyl-phenylether
Concen: 20.468 ng/ul
RT: 16.434 min Scan# 2269
Delta R.T. 0.000 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SICV216


Tgt Ion:248 Resp: 132389
Ion Ratio Lower Upper
248 100
250 95.4 77.8 116.8
141 87.8 62.6 93.8



#69
Hexachlorobenzene
Concen: 20.224 ng/ul
RT: 16.545 min Scan# 2288
Delta R.T. -0.006 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52

Tgt Ion:284 Resp: 147135
Ion Ratio Lower Upper
284 100
142 43.5 27.8 41.8#
249 30.6 25.7 38.5









#70
Atrazine
Concen: 20.438 ng/ul
RT: 16.698 min Scan# 2314
Delta R.T. -0.006 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SICV216


Tgt Ion:200 Resp: 137730
Ion Ratio Lower Upper
200 100
173 29.7 20.5 30.7
215 49.1 38.9 58.3



#71
Pentachlorophenol
Concen: 19.244 ng/ul
RT: 16.886 min Scan# 2346
Delta R.T. 0.000 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52

Tgt Ion:266 Resp: 84961
Ion Ratio Lower Upper
266 100
264 60.5 50.2 75.2
268 61.9 49.3 73.9









#72
 Phenanthrene
 Concen: 21.278 ng/ul
 RT: 17.281 min Scan# 2413
 Delta R.T. 0.000 min
 Lab File: BN019543.D
 Acq: 27 Apr 2022 13:52

Instrument :
 BNA_N
 ClientSampleId :
 SICV216

Tgt Ion:178 Resp: 818676




Ion	Ratio	Lower	Upper
178	100		
179	14.4	12.2	18.4
176	18.8	15.2	22.8




#74
 Anthracene
 Concen: 21.238 ng/ul
 RT: 17.369 min Scan# 2428
 Delta R.T. 0.000 min
 Lab File: BN019543.D
 Acq: 27 Apr 2022 13:52

Tgt Ion:178 Resp: 816292

Ion	Ratio	Lower	Upper
178	100		
179	15.7	13.0	19.4
176	18.8	15.5	23.3





#75

1,2,3,4-Tetrachlorobenzene

Concen: 21.697 ng/uL

RT: 13.298 min Scan# 1736

Delta R.T. -0.006 min

Lab File: BN019543.D

Acq: 27 Apr 2022 13:52

Instrument :

BNA_N

ClientSampleId :

SICV216





Tgt Ion:216 Resp: 230276

Ion Ratio Lower Upper

216 100

214 81.9 63.2 94.8

218 48.1 37.9 56.9



#76

Pentachlorobenzene

Concen: 21.450 ng/uL

RT: 14.816 min Scan# 1994

Delta R.T. 0.000 min

Lab File: BN019543.D

Acq: 27 Apr 2022 13:52




Tgt Ion:250 Resp: 198322


Ion Ratio Lower Upper

250 100

248 63.7 53.0 79.4

252 63.6 53.4 80.2









#77
 Carbazole
 Concen: 21.170 ng/ul
 RT: 17.633 min Scan# 2473
 Delta R.T. 0.000 min
 Lab File: BN019543.D
 Acq: 27 Apr 2022 13:52




Instrument :
 BNA_N
 ClientSampleId :
 SICV216


Tgt Ion:167 Resp: 739700
 Ion Ratio Lower Upper
 167 100
 166 20.8 16.1 24.1
 139 13.5 10.2 15.2



#78
 Di-n-butylphthalate
 Concen: 20.507 ng/ul
 RT: 18.216 min Scan# 2572
 Delta R.T. 0.000 min
 Lab File: BN019543.D
 Acq: 27 Apr 2022 13:52

Tgt Ion:149 Resp: 766485
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.1 10.7
 104 5.2 4.1 6.1









#80
Fluoranthene
Concen: 21.793 ng/u1
RT: 19.292 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SICV216


Tgt Ion:202 Resp: 874224
Ion Ratio Lower Upper
202 100
101 12.3 8.9 13.3
100 10.1 6.7 10.1



#82
Pyrene
Concen: 21.531 ng/u1
RT: 19.657 min Scan# 2817
Delta R.T. 0.000 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52

Tgt Ion:202 Resp: 899642
Ion Ratio Lower Upper
202 100
101 14.6 9.7 14.5#
100 12.1 8.0 12.0#









#83
Butylbenzylphthalate
Concen: 21.326 ng/ul
RT: 20.569 min Scan# 2971
Delta R.T. 0.006 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SICV216


Tgt Ion:149 Resp: 329611
Ion Ratio Lower Upper
149 100
91 69.5 58.8 88.2
206 19.1 18.1 27.1



#84
3,3'-Dichlorobenzidine
Concen: 20.048 ng/ul
RT: 21.339 min Scan# 3103
Delta R.T. 0.000 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52

Tgt Ion:252 Resp: 238314
Ion Ratio Lower Upper
252 100
254 65.6 50.8 76.2
126 13.8 8.2 12.4#









#85
Benzo(a)anthracene
Concen: 21.555 ng/ul
RT: 21.410 min Scan# 3115
Delta R.T. 0.006 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SICV216


Tgt Ion:228 Resp: 817144
Ion Ratio Lower Upper
228 100
229 18.5 16.3 24.5
226 25.1 23.2 34.8



#86
Bis(2-ethylhexyl)phthalate
Concen: 21.174 ng/ul
RT: 21.357 min Scan# 3106
Delta R.T. 0.006 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52

Tgt Ion:149 Resp: 502159
Ion Ratio Lower Upper
149 100
167 27.5 22.5 33.7
279 3.9 3.5 5.3





#87

Chrysene

Concen: 21.267 ng/ul

RT: 21.457 min Scan# 3123

Delta R.T. 0.000 min

Lab File: BN019543.D

Acq: 27 Apr 2022 13:52

Instrument :

BNA_N

ClientSampleId :

SICV216





Tgt Ion:228 Resp: 791957

Ion Ratio Lower Upper

228 100

226 28.8 24.2 36.2

229 19.9 16.2 24.2



#89

Di-n-octyl phthalate

Concen: 19.712 ng/ul




RT: 22.269 min Scan# 3261


Delta R.T. 0.000 min

Lab File: BN019543.D

Acq: 27 Apr 2022 13:52

Tgt Ion:149 Resp: 805573









#90
Benzo(b)fluoranthene
Concen: 20.960 ng/ul
RT: 23.063 min Scan# 3396
Delta R.T. 0.000 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SICV216


Tgt Ion:252 Resp: 809443
Ion Ratio Lower Upper
252 100
253 21.5 16.5 24.7
125 12.6 7.8 11.8#



#91
Benzo(k)fluoranthene
Concen: 22.408 ng/ul
RT: 23.110 min Scan# 3404
Delta R.T. 0.000 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52

Tgt Ion:252 Resp: 784930
Ion Ratio Lower Upper
252 100
253 21.5 18.2 27.4
125 11.8 8.2 12.4









#93
Benzo(a)pyrene
Concen: 22.445 ng/ul
RT: 23.674 min Scan# 31
Delta R.T. 0.006 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52




Instrument :
BNA_N
ClientSampleId :
SICV216


Tgt Ion:252 Resp: 823142
Ion Ratio Lower Upper
252 100
253 20.7 18.7 28.1
125 13.5 9.2 13.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 21.742 ng/ul
RT: 26.209 min Scan# 3931
Delta R.T. 0.006 min
Lab File: BN019543.D
Acq: 27 Apr 2022 13:52

Tgt Ion:276 Resp: 913742
Ion Ratio Lower Upper
276 100
138 32.2 19.5 29.3#
277 25.6 21.4 32.0





#95

Dibenzo(a,h)anthracene

Concen: 22.163 ng/ul

RT: 26.221 min Scan# 3933

Delta R.T. 0.000 min

Lab File: BN019543.D

Acq: 27 Apr 2022 13:52

Instrument :

BNA_N

ClientSampleId :

SICV216





Tgt Ion:278 Resp: 796462

Ion Ratio Lower Upper

278 100

139 19.3 13.8 20.8

279 22.3 18.3 27.5



#96

Benzo(g,h,i)perylene

Concen: 21.775 ng/ul

RT: 26.951 min Scan# 4057

Delta R.T. 0.006 min

Lab File: BN019543.D

Acq: 27 Apr 2022 13:52

Tgt Ion:276 Resp: 804924

Ion Ratio Lower Upper

276 100

138 25.3 17.0 25.6

277 23.3 17.8 26.6

