

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN042723\  
 Data File : BN025168.D  
 Acq On : 27 Apr 2023 09:25  
 Operator : CG/JU  
 Sample : PB152345BL  
 Misc :  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 SBLK345

Quant Time: Apr 28 01:19:29 2023  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\SFAM-EPA-SIM-BN042023.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Apr 27 15:27:43 2023  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc  | Units | Dev(Min) |
|-----------------------------|--------|------|----------|-------|-------|----------|
| -----                       |        |      |          |       |       |          |
| Internal Standards          |        |      |          |       |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 7.917  | 152  | 6607     | 0.400 | ng/ul | 0.00     |
| 4) Naphthalene-d8           | 10.726 | 136  | 21352    | 0.400 | ng/ul | 0.00     |
| 9) Acenaphthene-d10         | 14.561 | 164  | 13056    | 0.400 | ng/ul | 0.00     |
| 13) Phenanthrene-d10        | 17.309 | 188  | 27684    | 0.400 | ng/ul | 0.00     |
| 17) Chrysene-d12            | 21.502 | 240  | 20101    | 0.400 | ng/ul | 0.00     |
| 23) Perylene-d12            | 23.926 | 264  | 20458    | 0.400 | ng/ul | #-0.02   |
| System Monitoring Compounds |        |      |          |       |       |          |
| 3) 1,4-Dioxane-d8           | 3.293  | 96   | 40347    | 5.898 | ng/ul | 0.00     |
| 6) 2-Methylnaphthalene-d10  | 12.321 | 152  | 8806     | 0.321 | ng/ul | 0.00     |
| 18) Fluoranthene-d10        | 19.335 | 212  | 19249    | 0.363 | ng/ul | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN042723\  
Data File : BN025168.D  
Acq On : 27 Apr 2023 09:25  
Operator : CG/JU  
Sample : PB152345BL  
Misc :  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
SBLK345

Quant Time: Apr 28 01:19:29 2023  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\SFAM-EPA-SIM-BN042023.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Apr 27 15:27:43 2023  
Response via : Initial Calibration

