

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042724\
 Data File : BN031283.D
 Acq On : 27 Apr 2024 23:19
 Operator : MA/JU
 Sample : PB160538BL
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK538

Quant Time: Apr 28 21:17:36 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 26 22:57:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.622	152	218283	20.000	ng/u1	0.00
20) Naphthalene-d8	10.399	136	1049448	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.269	164	771034	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.016	188	1702178	20.000	ng/u1	0.00
79) Chrysene-d12	21.251	240	1562752	20.000	ng/u1	0.00
88) Perylene-d12	24.151	264	1624292	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.134	96	22872	4.895	ng/uL	0.00
4) Pyridine-d5	3.540	84	329749	24.507	ng/u1	0.00
7) Phenol-d5	6.799	99	439960	23.893	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.963	67	276248	25.786	ng/u1	0.00
11) 2-Chlorophenol-d4	7.152	132	372532	25.984	ng/u1	0.00
15) 4-Methylphenol-d8	8.328	113	380519	24.572	ng/u1	0.00
21) Nitrobenzene-d5	8.775	128	197479	26.190	ng/u1	0.00
24) 2-Nitrophenol-d4	9.493	143	217981	27.811	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.022	165	374477	24.228	ng/u1	0.00
31) 4-Chloroaniline-d4	10.546	131	588045	24.346	ng/u1	0.00
46) Dimethylphthalate-d6	13.687	166	1374103	25.610	ng/u1	0.00
49) Acenaphthylene-d8	13.957	160	1507617	24.796	ng/u1	0.00
54) 4-Nitrophenol-d4	14.487	143	244046	21.792	ng/u1	0.00
60) Fluorene-d10	15.263	176	1150275	24.627	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.392	200	193069	22.482	ng/u1	0.00
73) Anthracene-d10	17.116	188	1851949	25.101	ng/u1	0.00
81) Pyrene-d10	19.422	212	2162162	28.156	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.951	264	1983933	25.757	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042724\
Data File : BN031283.D
Acq On : 27 Apr 2024 23:19
Operator : MA/JU
Sample : PB160538BL
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK538

Quant Time: Apr 28 21:17:36 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Apr 26 22:57:25 2024
Response via : Initial Calibration

