

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042724\
 Data File : BN031292.D
 Acq On : 28 Apr 2024 05:56
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020388

Quant Time: Apr 28 21:19:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 26 22:57:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.616	152	253989	20.000	ng/u1	0.00
20) Naphthalene-d8	10.399	136	1235034	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.269	164	866546	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.016	188	1905778	20.000	ng/u1	0.00
79) Chrysene-d12	21.257	240	1762853	20.000	ng/u1	0.00
88) Perylene-d12	24.145	264	1828979	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.134	96	37239	6.850	ng/uL	0.00
4) Pyridine-d5	3.540	84	285729	18.250	ng/u1	0.00
7) Phenol-d5	6.793	99	370943	17.313	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	6.963	67	223484	17.928	ng/u1	0.00
11) 2-Chlorophenol-d4	7.152	132	296852	17.794	ng/u1	0.00
15) 4-Methylphenol-d8	8.328	113	314090	17.431	ng/u1	0.00
21) Nitrobenzene-d5	8.775	128	158844	17.901	ng/u1	0.00
24) 2-Nitrophenol-d4	9.487	143	175318	19.007	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.022	165	319366	17.557	ng/u1	0.00
31) 4-Chloroaniline-d4	10.546	131	476688	16.770	ng/u1	0.00
46) Dimethylphthalate-d6	13.687	166	1047991	17.379	ng/u1	0.00
49) Acenaphthylene-d8	13.957	160	1185231	17.345	ng/u1	0.00
54) 4-Nitrophenol-d4	14.487	143	209283	16.628	ng/u1	0.00
60) Fluorene-d10	15.263	176	887923	16.915	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.392	200	173404	18.035	ng/u1	0.00
73) Anthracene-d10	17.116	188	1410534	17.076	ng/u1	0.00
81) Pyrene-d10	19.422	212	1678892	19.381	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.945	264	1506490	17.370	ng/u1	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.170	88	39121	6.790	ng/uL	93
5) Pyridine	3.558	79	281139	17.795	ng/u1	97
6) Benzaldehyde	6.769	77	225264	21.412	ng/u1	99
8) Phenol	6.822	94	384053	17.386	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.058	93	315061	18.077	ng/u1	98
12) 2-Chlorophenol	7.181	128	311640	17.972	ng/u1	99
13) 2-Methylphenol	8.063	108	298754	17.452	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.140	45	423721	18.647	ng/u1	100
16) Acetophenone	8.440	105	504074	18.056	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.428	70	255854	18.386	ng/u1	99
18) 4-Methylphenol	8.393	108	335455	17.494	ng/u1	96
19) Hexachloroethane	8.681	117	128218	17.819	ng/u1	99
22) Nitrobenzene	8.816	77	370641	17.697	ng/u1	99
23) Isophorone	9.340	82	752447	17.605	ng/u1	98
25) 2-Nitrophenol	9.522	139	189572	18.576	ng/u1	94
26) 2,4-Dimethylphenol	9.587	107	356930	17.150	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.822	93	460836	17.751	ng/u1	98
29) 2,4-Dichlorophenol	10.052	162	319505	17.586	ng/u1	99
30) Naphthalene	10.446	128	1085449	16.958	ng/u1	99
32) 4-Chloroaniline	10.569	127	466684	16.897	ng/u1	97
33) Hexachlorobutadiene	10.722	225	188042	16.978	ng/u1	98
34) Caprolactam	11.351	113	78452	10.960	ng/u1	95
35) 4-Chloro-3-methylphenol	11.698	107	375144	17.809	ng/u1	98
36) 2-Methylnaphthalene	12.063	142	776709	17.208	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042724\
 Data File : BN031292.D
 Acq On : 28 Apr 2024 05:56
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 28 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SSTD020388


Quant Time: Apr 28 21:19:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 26 22:57:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.287	142	774763	16.963	ng/ul	94
39) 1,2,4,5-Tetrachloroben...	12.440	216	384460	16.955	ng/ul	99
40) Hexachlorocyclopentadiene	12.410	237	204558	16.423	ng/ul	98
41) 2,4,6-Trichlorophenol	12.687	196	273892	18.147	ng/ul	97
42) 2,4,5-Trichlorophenol	12.757	196	297786	17.796	ng/ul	98
43) 1,1'-Biphenyl	13.093	154	1037031	17.401	ng/ul	98
44) 2-Chloronaphthalene	13.134	162	812040	17.051	ng/ul	98
45) 2-Nitroaniline	13.351	65	244515	18.598	ng/ul	99
47) Dimethylphthalate	13.734	163	1064254	17.108	ng/ul	100
48) 2,6-Dinitrotoluene	13.857	165	219522	18.799	ng/ul	98
50) Acenaphthylene	13.987	152	1383489	17.333	ng/ul	99
51) 3-Nitroaniline	14.187	138	248240	18.517	ng/ul#	91
52) Acenaphthene	14.334	153	889725	16.815	ng/ul	96
53) 2,4-Dinitrophenol	14.392	184	108968	16.501	ng/ul	97
55) 4-Nitrophenol	14.498	109	166971	16.347	ng/ul	97
56) Dibenzofuran	14.669	168	1240907	16.962	ng/ul	98
57) 2,4-Dinitrotoluene	14.645	165	327487	18.457	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	14.898	232	256243	17.719	ng/ul	93
59) Diethylphthalate	15.104	149	1104669	17.267	ng/ul	100
61) Fluorene	15.322	166	1004597	16.686	ng/ul	96
62) 4-Chlorophenyl-phenyle...	15.316	204	483692	16.867	ng/ul	99
63) 4-Nitroaniline	15.351	138	245573	16.990	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.410	198	187860	18.196	ng/ul#	95
67) N-Nitrosodiphenylamine	15.534	169	871189	17.055	ng/ul	98
68) 4-Bromophenyl-phenylether	16.216	248	310075	17.204	ng/ul	87
69) Hexachlorobenzene	16.316	284	366456	17.637	ng/ul	98
70) Atrazine	16.492	200	311733	17.383	ng/ul	95
71) Pentachlorophenol	16.669	266	225810	16.727	ng/ul	97
72) Phenanthrene	17.063	178	1694436	16.810	ng/ul	99
74) Anthracene	17.151	178	1725476	16.915	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.051	216	391644	17.520	ng/ul	99
76) Pentachlorobenzene	14.581	250	391940	16.979	ng/ul	96
77) Carbazole	17.428	167	1578598	16.941	ng/ul	98
78) Di-n-butylphthalate	17.992	149	2003701	17.642	ng/ul	99
80) Fluoranthene	19.080	202	2040703	19.905	ng/ul	99
82) Pyrene	19.451	202	2067804	19.196	ng/ul	98
83) Butylbenzylphthalate	20.363	149	936413	19.701	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.169	252	622648	16.951	ng/ul	97
85) Benzo(a)anthracene	21.239	228	2014985	16.832	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.163	149	1320249	18.499	ng/ul	99
87) Chrysene	21.298	228	1871618	16.471	ng/ul	98
89) Di-n-octyl phthalate	22.263	149	2337665	22.174	ng/ul	100
90) Benzo(b)fluoranthene	23.233	252	1928283	18.738	ng/ul	100
91) Benzo(k)fluoranthene	23.292	252	1949238	18.833	ng/ul	99
93) Benzo(a)pyrene	24.010	252	1771641	17.257	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	27.386	276	1895376	14.878	ng/ul	99
95) Dibenzo(a,h)anthracene	27.445	278	1573344	15.047	ng/ul	96
96) Benzo(g,h,i)perylene	28.403	276	1466790	14.298	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 28 21:19:23 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Apr 26 22:57:25 2024
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 6.790 ng/uL
RT: 3.170 min Scan# 31
Delta R.T. -0.006 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56




Instrument :
BNA_N
ClientSampleId :
SSTD020388


Tgt Ion: 88 Resp: 39121
Ion Ratio Lower Upper
88 100
43 30.0 22.2 33.2
58 68.9 50.2 75.2



#5
Pyridine
Concen: 17.795 ng/uL
RT: 3.558 min Scan# 97
Delta R.T. -0.006 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56

Tgt Ion: 79 Resp: 281139
Ion Ratio Lower Upper
79 100
52 76.2 59.9 89.9
51 33.5 29.0 43.6









#6
Benzaldehyde
Concen: 21.412 ng/ul
RT: 6.769 min Scan# 644
Delta R.T. -0.006 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56




Instrument :
BNA_N
ClientSampleId :
SSTD020388


Tgt Ion: 77 Resp: 225264
Ion Ratio Lower Upper
77 100
105 98.8 80.0 120.0
106 94.8 75.0 112.6



#8
Phenol
Concen: 17.386 ng/ul
RT: 6.822 min Scan# 652
Delta R.T. -0.006 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56

Tgt Ion: 94 Resp: 384053
Ion Ratio Lower Upper
94 100
65 25.6 22.0 33.0
66 36.3 29.8 44.6









#10
Bis(2-Chloroethyl)ether
Concen: 18.077 ng/ul
RT: 7.058 min Scan# 692
Delta R.T. 0.000 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56




Instrument :
BNA_N
ClientSampleId :
SSTD020388


Tgt Ion: 93 Resp: 315061
Ion Ratio Lower Upper
93 100
63 71.8 57.1 85.7
95 33.3 24.9 37.3



#12
2-Chlorophenol
Concen: 17.972 ng/ul
RT: 7.181 min Scan# 713
Delta R.T. -0.006 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56

Tgt Ion:128 Resp: 311640
Ion Ratio Lower Upper
128 100
64 43.7 34.9 52.3
130 32.4 27.0 40.4





#13

2-Methylphenol

Concen: 17.452 ng/ul

RT: 8.063 min Scan# 8

Delta R.T. -0.006 min

Lab File: BN031292.D

Acq: 28 Apr 2024 05:56

Instrument :

BNA_N

ClientSampleId :





SSTD020388

Tgt Ion:108 Resp: 298754

Ion Ratio Lower Upper

108 100

107 89.0 70.1 105.1



#14

2,2'-oxybis(1-Chloropropane)

Concen: 18.647 ng/ul

RT: 8.140 min Scan# 876

Delta R.T. -0.006 min

Lab File: BN031292.D

Acq: 28 Apr 2024 05:56




Tgt Ion: 45 Resp: 423721


Ion Ratio Lower Upper

45 100

77 15.8 12.5 18.7

79 11.8 9.6 14.4









#16
Acetophenone
Concen: 18.056 ng/ul
RT: 8.440 min Scan# 91
Delta R.T. -0.006 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56




Instrument :
BNA_N
ClientSampleId :
SSTD020388


Tgt Ion:105 Resp: 504074
Ion Ratio Lower Upper
105 100
77 77.4 60.2 90.4
51 25.4 20.6 31.0



#17
N-Nitroso-di-n-propylamine
Concen: 18.386 ng/ul
RT: 8.428 min Scan# 925
Delta R.T. -0.006 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56

Tgt Ion: 70 Resp: 255854
Ion Ratio Lower Upper
70 100
42 55.2 45.0 67.6
101 9.8 8.1 12.1
130 24.0 19.2 28.8









#18
4-Methylphenol
Concen: 17.494 ng/ul
RT: 8.393 min Scan# 919
Delta R.T. -0.006 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56




Instrument :
BNA_N
ClientSampleId :
SSTD020388


Tgt Ion:108 Resp: 335455
Ion Ratio Lower Upper
108 100
107 108.3 89.8 134.8



#19
Hexachloroethane
Concen: 17.819 ng/ul
RT: 8.681 min Scan# 968
Delta R.T. -0.006 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56

Tgt Ion:117 Resp: 128218
Ion Ratio Lower Upper
117 100
201 97.5 77.4 116.0
199 60.3 49.6 74.4









#22
Nitrobenzene
Concen: 17.697 ng/ul
RT: 8.816 min Scan# 991
Delta R.T. -0.006 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56




Instrument :
BNA_N
ClientSampleId :
SSTD020388


Tgt Ion: 77 Resp: 370641
Ion Ratio Lower Upper
77 100
123 48.2 38.2 57.2
65 13.5 10.7 16.1



#23
Isophorone
Concen: 17.605 ng/ul
RT: 9.340 min Scan# 1080
Delta R.T. -0.006 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56

Tgt Ion: 82 Resp: 752447
Ion Ratio Lower Upper
82 100
95 7.3 5.9 8.9
138 19.3 14.7 22.1









#25
2-Nitrophenol
Concen: 18.576 ng/ul
RT: 9.522 min Scan# 1111
Delta R.T. -0.006 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56




Instrument :
BNA_N
ClientSampleId :
SSTD020388


Tgt Ion:139 Resp: 189572
Ion Ratio Lower Upper
139 100
65 50.2 35.2 52.8
109 32.6 25.4 38.0



#26
2,4-Dimethylphenol
Concen: 17.150 ng/ul
RT: 9.587 min Scan# 1122
Delta R.T. 0.000 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56

Tgt Ion:107 Resp: 356930
Ion Ratio Lower Upper
107 100
121 50.7 40.3 60.5
122 86.7 69.8 104.8





#27

Bis(2-Chloroethoxy)methane

Concen: 17.751 ng/ul

RT: 9.822 min Scan# 1162

Delta R.T. -0.006 min

Lab File: BN031292.D

Acq: 28 Apr 2024 05:56

Instrument :

BNA_N

ClientSampleId :

SSTD020388





Tgt Ion: 93 Resp: 460836

Ion Ratio Lower Upper

93 100

95 31.1 25.8 38.6

123 11.4 9.8 14.6



#29

2,4-Dichlorophenol

Concen: 17.586 ng/ul

RT: 10.052 min Scan# 1201

Delta R.T. -0.006 min

Lab File: BN031292.D

Acq: 28 Apr 2024 05:56




Tgt Ion:162 Resp: 319505


Ion Ratio Lower Upper

162 100

164 64.0 51.0 76.6

98 34.9 28.6 42.8





#30

Naphthalene

Concen: 16.958 ng/ul

RT: 10.446 min Scan# 1268

Delta R.T. -0.006 min

Lab File: BN031292.D

Acq: 28 Apr 2024 05:56

Instrument :

BNA_N

ClientSampleId :

SSTD020388





Tgt Ion:128 Resp: 1085449

Ion Ratio Lower Upper

128 100

129 10.9 9.0 13.4

127 13.4 10.5 15.7



#32

4-Chloroaniline

Concen: 16.897 ng/ul

RT: 10.569 min Scan# 1289

Delta R.T. 0.000 min

Lab File: BN031292.D




Acq: 28 Apr 2024 05:56


Tgt Ion:127 Resp: 466684

Ion Ratio Lower Upper

127 100

129 33.3 25.4 38.2





#33

Hexachlorobutadiene

Concen: 16.978 ng/ul

RT: 10.722 min Scan# 1315

Delta R.T. -0.006 min

Lab File: BN031292.D

Acq: 28 Apr 2024 05:56

Instrument :

BNA_N

ClientSampleId :

SSTD020388





Tgt Ion:225 Resp: 188042

Ion Ratio Lower Upper

225 100

223 63.0 49.8 74.6

227 63.6 48.8 73.2



#34

Caprolactam

Concen: 10.960 ng/ul

RT: 11.351 min Scan# 1422

Delta R.T. 0.000 min

Lab File: BN031292.D

Acq: 28 Apr 2024 05:56




Tgt Ion:113 Resp: 78452


Ion Ratio Lower Upper

113 100

55 148.5 112.4 168.6

56 113.1 88.0 132.0









#35
4-Chloro-3-methylphenol
Concen: 17.809 ng/ul
RT: 11.698 min Scan# 1481
Delta R.T. -0.006 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56




Instrument :
BNA_N
ClientSampleId :
SSTD020388


Tgt Ion:107 Resp: 375144
Ion Ratio Lower Upper
107 100
144 26.6 20.6 31.0
142 82.6 67.4 101.2



#36
2-Methylnaphthalene
Concen: 17.208 ng/ul
RT: 12.063 min Scan# 1543
Delta R.T. -0.006 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56

Tgt Ion:142 Resp: 776709
Ion Ratio Lower Upper
142 100
141 89.2 69.6 104.4





#37

1-Methylnaphthalene

Concen: 16.963 ng/ul

RT: 12.287 min Scan# 1581

Delta R.T. -0.006 min

Lab File: BN031292.D

Acq: 28 Apr 2024 05:56

Instrument :

BNA_N

ClientSampleId :

SSTD020388





Tgt Ion:142 Resp: 774763

Ion Ratio Lower Upper

142 100

141 93.4 70.2 105.4

116 3.7 3.1 4.7



#39

1,2,4,5-Tetrachlorobenzene

Concen: 16.955 ng/ul

RT: 12.440 min Scan# 1607

Delta R.T. 0.000 min

Lab File: BN031292.D

Acq: 28 Apr 2024 05:56

Tgt Ion:216 Resp: 384460




Ion Ratio Lower Upper


216 100

214 77.8 63.1 94.7

179 21.7 17.0 25.6

108 17.8 14.2 21.4









#40
Hexachlorocyclopentadiene
Concen: 16.423 ng/ul
RT: 12.410 min Scan# 1602
Delta R.T. -0.006 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56




Instrument :
BNA_N
ClientSampleId :
SSTD020388


Tgt Ion:237 Resp: 204558
Ion Ratio Lower Upper
237 100
235 61.1 47.6 71.4
272 12.4 10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 18.147 ng/ul
RT: 12.687 min Scan# 1649
Delta R.T. -0.006 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56

Tgt Ion:196 Resp: 273892
Ion Ratio Lower Upper
196 100
198 91.1 69.9 104.9
200 29.5 23.4 35.2





#42

2,4,5-Trichlorophenol

Concen: 17.796 ng/ul

RT: 12.757 min Scan# 1

Delta R.T. -0.006 min

Lab File: BN031292.D

Acq: 28 Apr 2024 05:56

Instrument :

BNA_N

ClientSampleId :

SSTD020388





Tgt Ion:196 Resp: 297786

Ion Ratio Lower Upper

196 100

198 96.9 75.9 113.9

200 30.1 24.7 37.1



#43

1,1'-Biphenyl

Concen: 17.401 ng/ul

RT: 13.093 min Scan# 1718

Delta R.T. -0.006 min

Lab File: BN031292.D

Acq: 28 Apr 2024 05:56




Tgt Ion:154 Resp: 1037031


Ion Ratio Lower Upper

154 100

153 41.6 31.8 47.8

76 12.9 10.6 15.8









#44
 2-Chloronaphthalene
 Concen: 17.051 ng/ul
 RT: 13.134 min Scan# 1725
 Delta R.T. -0.006 min
 Lab File: BN031292.D
 Acq: 28 Apr 2024 05:56




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020388


Tgt Ion:	162	Resp:	812040
Ion Ratio	Lower	Upper	
162	100		
164	31.7	25.8	38.8
127	40.1	33.7	50.5



#45
 2-Nitroaniline
 Concen: 18.598 ng/ul
 RT: 13.351 min Scan# 1762
 Delta R.T. -0.006 min
 Lab File: BN031292.D
 Acq: 28 Apr 2024 05:56

Tgt Ion:	65	Resp:	244515
Ion Ratio	Lower	Upper	
65	100		
92	70.0	57.4	86.2
138	113.9	91.6	137.4









#47
Dimethylphthalate
Concen: 17.108 ng/ul
RT: 13.734 min Scan# 1827
Delta R.T. 0.000 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56




Instrument :
BNA_N
ClientSampleId :
SSTD020388


Tgt Ion:163 Resp: 1064254
Ion Ratio Lower Upper
163 100
194 3.9 3.0 4.6
164 9.9 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 18.799 ng/ul
RT: 13.857 min Scan# 1848
Delta R.T. 0.000 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56

Tgt Ion:165 Resp: 219522
Ion Ratio Lower Upper
165 100
89 54.2 44.7 67.1
121 20.4 17.0 25.4









#50
 Acenaphthylene
 Concen: 17.333 ng/ul
 RT: 13.987 min Scan# 1870
 Delta R.T. -0.006 min
 Lab File: BN031292.D
 Acq: 28 Apr 2024 05:56

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020388

Tgt Ion:152 Resp: 1383489




Ion	Ratio	Lower	Upper
152	100		
151	19.4	15.8	23.8
153	12.9	10.5	15.7




#51
 3-Nitroaniline
 Concen: 18.517 ng/ul
 RT: 14.187 min Scan# 1904
 Delta R.T. -0.006 min
 Lab File: BN031292.D
 Acq: 28 Apr 2024 05:56

Tgt Ion:138 Resp: 248240

Ion	Ratio	Lower	Upper
138	100		
108	10.9	11.1	16.7#
92	107.6	93.9	140.9









#52
 Acenaphthene
 Concen: 16.815 ng/ul
 RT: 14.334 min Scan# 1929
 Delta R.T. 0.000 min
 Lab File: BN031292.D
 Acq: 28 Apr 2024 05:56

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020388

Tgt Ion:153 Resp: 889725




Ion	Ratio	Lower	Upper
153	100		
152	46.4	36.3	54.5
154	85.6	72.4	108.6




#53
 2,4-Dinitrophenol
 Concen: 16.501 ng/ul
 RT: 14.392 min Scan# 1939
 Delta R.T. -0.006 min
 Lab File: BN031292.D
 Acq: 28 Apr 2024 05:56

Tgt Ion:184 Resp: 108968

Ion	Ratio	Lower	Upper
184	100		
63	67.6	50.4	75.6
154	65.8	53.0	79.6









#55
4-Nitrophenol
Concen: 16.347 ng/ul
RT: 14.498 min Scan# 1957
Delta R.T. -0.006 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56




Instrument :
BNA_N
ClientSampleId :
SSTD020388


Tgt Ion:109 Resp: 166971
Ion Ratio Lower Upper
109 100
139 133.9 105.6 158.4
65 128.6 98.2 147.2



#56
Dibenzofuran
Concen: 16.962 ng/ul
RT: 14.669 min Scan# 1986
Delta R.T. -0.006 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56

Tgt Ion:168 Resp: 1240907
Ion Ratio Lower Upper
168 100
139 37.3 31.0 46.6









#57
2,4-Dinitrotoluene
Concen: 18.457 ng/ul
RT: 14.645 min Scan# 1982
Delta R.T. -0.006 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56




Instrument :
BNA_N
ClientSampleId :
SSTD020388


Tgt Ion:165 Resp: 327487
Ion Ratio Lower Upper
165 100
63 40.9 33.9 50.9
182 2.7 2.6 4.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 17.719 ng/ul
RT: 14.898 min Scan# 2025
Delta R.T. 0.000 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56

Tgt Ion:232 Resp: 256243
Ion Ratio Lower Upper
232 100
131 45.4 41.2 61.8
130 2.5 2.2 3.2
166 31.9 27.3 40.9









#59
 Diethylphthalate
 Concen: 17.267 ng/ul
 RT: 15.104 min Scan# 2060
 Delta R.T. -0.006 min
 Lab File: BN031292.D
 Acq: 28 Apr 2024 05:56

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020388

Tgt Ion:149 Resp: 1104669




Ion	Ratio	Lower	Upper
149	100		
177	21.3	16.8	25.2
150	12.7	10.1	15.1




#61
 Fluorene
 Concen: 16.686 ng/ul
 RT: 15.322 min Scan# 2097
 Delta R.T. 0.000 min
 Lab File: BN031292.D
 Acq: 28 Apr 2024 05:56

Tgt Ion:166 Resp: 1004597

Ion	Ratio	Lower	Upper
166	100		
165	97.1	74.6	112.0
167	12.9	10.6	16.0









#62
4-Chlorophenyl-phenylether
Concen: 16.867 ng/ul
RT: 15.316 min Scan# 2096
Delta R.T. -0.006 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56




Instrument :
BNA_N
ClientSampleId :
SSTD020388


Tgt Ion:204 Resp: 483692
Ion Ratio Lower Upper
204 100
206 35.1 26.4 39.6
141 63.4 50.7 76.1



#63
4-Nitroaniline
Concen: 16.990 ng/ul
RT: 15.351 min Scan# 2102
Delta R.T. -0.006 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56

Tgt Ion:138 Resp: 245573
Ion Ratio Lower Upper
138 100
92 51.4 38.4 57.6
108 80.8 63.8 95.8









#66
4,6-Dinitro-2-methylphenol
Concen: 18.196 ng/ul
RT: 15.410 min Scan# 2112
Delta R.T. 0.000 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56




Instrument :
BNA_N
ClientSampleId :
SSTD020388


Tgt Ion:198 Resp: 187860
Ion Ratio Lower Upper
198 100
182 3.6 4.0 6.0#
77 25.2 22.2 33.4



#67
N-Nitrosodiphenylamine
Concen: 17.055 ng/ul
RT: 15.534 min Scan# 2133
Delta R.T. -0.006 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56

Tgt Ion:169 Resp: 871189
Ion Ratio Lower Upper
169 100
168 67.4 52.6 79.0
167 35.3 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 17.204 ng/ul
RT: 16.216 min Scan# 2149
Delta R.T. 0.000 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56




Instrument :
BNA_N
ClientSampleId :
SSTD020388


Tgt Ion:248 Resp: 310075
Ion Ratio Lower Upper
248 100
250 90.6 78.2 117.2
141 71.5 71.4 107.0



#69
Hexachlorobenzene
Concen: 17.637 ng/ul
RT: 16.316 min Scan# 2266
Delta R.T. -0.006 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56

Tgt Ion:284 Resp: 366456
Ion Ratio Lower Upper
284 100
142 39.2 30.0 45.0
249 32.1 25.5 38.3









#70
 Atrazine
 Concen: 17.383 ng/ul
 RT: 16.492 min Scan# 21
 Delta R.T. -0.006 min
 Lab File: BN031292.D
 Acq: 28 Apr 2024 05:56

Instrument :
 BNA_N
 ClientSampleId :
 SST020388

Tgt Ion:200 Resp: 311733




Ion	Ratio	Lower	Upper
200	100		
173	28.1	20.6	31.0
215	50.2	37.6	56.4




#71
 Pentachlorophenol
 Concen: 16.727 ng/ul
 RT: 16.669 min Scan# 2326
 Delta R.T. -0.006 min
 Lab File: BN031292.D
 Acq: 28 Apr 2024 05:56

Tgt Ion:266 Resp: 225810

Ion	Ratio	Lower	Upper
266	100		
264	64.0	49.1	73.7
268	64.0	49.2	73.8









#72
Phenanthrene
Concen: 16.810 ng/ul
RT: 17.063 min Scan# 2393
Delta R.T. 0.000 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56




Instrument :
BNA_N
ClientSampleId :
SSTD020388


Tgt Ion:178 Resp: 1694436
Ion Ratio Lower Upper
178 100
179 15.0 12.4 18.6
176 18.0 14.6 21.8



#74
Anthracene
Concen: 16.915 ng/ul
RT: 17.151 min Scan# 2408
Delta R.T. -0.006 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56

Tgt Ion:178 Resp: 1725476
Ion Ratio Lower Upper
178 100
179 14.8 12.2 18.4
176 18.3 14.4 21.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 17.520 ng/uL
RT: 13.051 min Scan# 11
Delta R.T. -0.006 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56




Instrument :
BNA_N
ClientSampleId :
SSTD020388


Tgt Ion:216 Resp: 391644
Ion Ratio Lower Upper
216 100
214 78.2 63.0 94.6
218 47.0 38.3 57.5



#76
Pentachlorobenzene
Concen: 16.979 ng/uL
RT: 14.581 min Scan# 1971
Delta R.T. -0.006 min
Lab File: BN031292.D
Acq: 28 Apr 2024 05:56

Tgt Ion:250 Resp: 391940
Ion Ratio Lower Upper
250 100
248 66.5 49.3 73.9
252 64.6 52.9 79.3





#77

Carbazole

Concen: 16.941 ng/ul

RT: 17.428 min Scan# 2455

Delta R.T. -0.006 min

Lab File: BN031292.D

Acq: 28 Apr 2024 05:56

Instrument :

BNA_N

ClientSampleId :

SSTD020388





Tgt Ion:167 Resp: 1578598

Ion Ratio Lower Upper

167 100

166 21.9 16.2 24.2

139 12.9 10.6 15.8



#78

Di-n-butylphthalate

Concen: 17.642 ng/ul

RT: 17.992 min Scan# 2551

Delta R.T. -0.006 min

Lab File: BN031292.D

Acq: 28 Apr 2024 05:56




Tgt Ion:149 Resp: 2003701


Ion Ratio Lower Upper

149 100

150 9.5 7.1 10.7

104 5.2 3.9 5.9









#80
 Fluoranthene
 Concen: 19.905 ng/u1
 RT: 19.080 min Scan# 2737
 Delta R.T. -0.006 min
 Lab File: BN031292.D
 Acq: 28 Apr 2024 05:56




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020388


Tgt Ion:202 Resp: 2040703
 Ion Ratio Lower Upper
 202 100
 101 11.9 9.0 13.6
 100 8.8 6.9 10.3



#82
 Pyrene
 Concen: 19.196 ng/u1
 RT: 19.451 min Scan# 2799
 Delta R.T. 0.000 min
 Lab File: BN031292.D
 Acq: 28 Apr 2024 05:56

Tgt Ion:202 Resp: 2067804
 Ion Ratio Lower Upper
 202 100
 101 13.0 11.1 16.7
 100 10.6 9.3 13.9





#83

Butylbenzylphthalate

Concen: 19.701 ng/ul

RT: 20.363 min Scan# 2954

Delta R.T. 0.000 min

Lab File: BN031292.D

Acq: 28 Apr 2024 05:56

Instrument :

BNA_N

ClientSampleId :

SSTD020388





Tgt Ion:149 Resp: 936413

Ion Ratio Lower Upper

149 100

91 68.4 57.3 85.9

206 20.0 16.2 24.4



#84

3,3'-Dichlorobenzidine

Concen: 16.951 ng/ul

RT: 21.169 min Scan# 3091

Delta R.T. -0.006 min

Lab File: BN031292.D

Acq: 28 Apr 2024 05:56




Tgt Ion:252 Resp: 622648


Ion Ratio Lower Upper

252 100

254 65.7 54.5 81.7

126 13.1 9.8 14.6





#85

Benzo(a)anthracene

Concen: 16.832 ng/ul

RT: 21.239 min Scan# 3103

Delta R.T. 0.000 min

Lab File: BN031292.D

Acq: 28 Apr 2024 05:56

Instrument :

BNA_N

ClientSampleId :

SSTD020388





Tgt Ion:228 Resp: 2014985

Ion Ratio Lower Upper

228 100

229 19.7 16.7 25.1

226 28.2 21.2 31.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 18.499 ng/ul

RT: 21.163 min Scan# 3090

Delta R.T. -0.006 min

Lab File: BN031292.D

Acq: 28 Apr 2024 05:56




Tgt Ion:149 Resp: 1320249


Ion Ratio Lower Upper

149 100

167 27.4 22.3 33.5

279 3.8 3.0 4.6





#87

Chrysene

Concen: 16.471 ng/ul

RT: 21.298 min Scan# 3114

Delta R.T. 0.000 min

Lab File: BN031292.D

Acq: 28 Apr 2024 05:56

Instrument :

BNA_N

ClientSampleId :

SSTD020388





Tgt Ion:228 Resp: 1871618

Ion Ratio Lower Upper

228 100

226 30.5 22.6 34.0

229 19.8 15.8 23.8



#89

Di-n-octyl phthalate

Concen: 22.174 ng/ul




RT: 22.263 min Scan# 3277


Delta R.T. 0.000 min

Lab File: BN031292.D

Acq: 28 Apr 2024 05:56

Tgt Ion:149 Resp: 2337665





#90

Benzo(b)fluoranthene

Concen: 18.738 ng/ul

RT: 23.233 min Scan# 3443

Delta R.T. 0.000 min

Lab File: BN031292.D

Acq: 28 Apr 2024 05:56

Instrument :

BNA_N

ClientSampleId :

SSTD020388





Tgt Ion:252 Resp: 1928283

Ion Ratio Lower Upper

252 100

253 22.0 17.6 26.4

125 11.0 8.6 12.8



#91

Benzo(k)fluoranthene

Concen: 18.833 ng/ul

RT: 23.292 min Scan# 3452

Delta R.T. -0.006 min

Lab File: BN031292.D

Acq: 28 Apr 2024 05:56




Tgt Ion:252 Resp: 1949238


Ion Ratio Lower Upper

252 100

253 21.7 17.4 26.0

125 10.8 9.4 14.2





#93

Benzo(a)pyrene

Concen: 17.257 ng/ul

RT: 24.010 min Scan# 3

Delta R.T. -0.012 min

Lab File: BN031292.D

Acq: 28 Apr 2024 05:56

Instrument :

BNA_N

ClientSampleId :

SSTD020388





Tgt Ion:252 Resp: 1771641

Ion Ratio Lower Upper

252 100

253 23.6 17.7 26.5

125 12.6 9.9 14.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 14.878 ng/ul

RT: 27.386 min Scan# 4148

Delta R.T. -0.012 min

Lab File: BN031292.D

Acq: 28 Apr 2024 05:56




Tgt Ion:276 Resp: 1895376


Ion Ratio Lower Upper

276 100

138 22.3 17.4 26.2

277 22.5 18.3 27.5





#95

Dibenzo(a,h)anthracene

Concen: 15.047 ng/ul

RT: 27.445 min Scan# 4158

Delta R.T. -0.012 min

Lab File: BN031292.D

Acq: 28 Apr 2024 05:56

Instrument :

BNA_N

ClientSampleId :

SSTD020388





Tgt Ion:278 Resp: 1573344

Ion Ratio Lower Upper

278 100

139 17.6 13.3 19.9

279 23.7 17.0 25.6



#96

Benzo(g,h,i)perylene

Concen: 14.298 ng/ul

RT: 28.403 min Scan# 4321

Delta R.T. -0.006 min

Lab File: BN031292.D

Acq: 28 Apr 2024 05:56

Tgt Ion:276 Resp: 1466790

Ion Ratio Lower Upper

276 100

138 23.2 19.0 28.6

277 22.5 19.3 28.9

