

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042724\
 Data File : BN031288.D
 Acq On : 28 Apr 2024 02:38
 Operator : MA/JU
 Sample : P2203-23
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 E1D03

Quant Time: Apr 28 21:18:21 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 26 22:57:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.616	152	194297	20.000	ng/ul	0.00
20) Naphthalene-d8	10.399	136	945334	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.269	164	706621	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.016	188	1601272	20.000	ng/ul	0.00
79) Chrysene-d12	21.251	240	1465962	20.000	ng/ul	0.00
88) Perylene-d12	24.145	264	1496044	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.134	96	12744	3.064	ng/uL	0.00
4) Pyridine-d5	3.552	84	58684	4.900	ng/ul	0.00
7) Phenol-d5	6.799	99	100566	6.136	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	6.963	67	287714	30.171	ng/ul	0.00
11) 2-Chlorophenol-d4	7.152	132	297362	23.301	ng/ul	0.00
15) 4-Methylphenol-d8	8.328	113	204730	14.853	ng/ul	0.00
21) Nitrobenzene-d5	8.775	128	207190	30.504	ng/ul	0.00
24) 2-Nitrophenol-d4	9.493	143	214997	30.452	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.022	165	376156	27.017	ng/ul	0.00
31) 4-Chloroaniline-d4	10.546	131	242316	11.137	ng/ul	0.00
46) Dimethylphthalate-d6	13.687	166	1543914	31.398	ng/ul	0.00
49) Acenaphthylene-d8	13.957	160	1690951	30.347	ng/ul	0.00
54) 4-Nitrophenol-d4	14.487	143	49985	4.870	ng/ul	0.00
60) Fluorene-d10	15.263	176	1333817	31.160	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.392	200	232466	28.775	ng/ul	0.00
73) Anthracene-d10	17.116	188	2281149	32.867	ng/ul	0.00
81) Pyrene-d10	19.422	212	2646962	36.745	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.951	264	2446572	34.486	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042724\
Data File : BN031288.D
Acq On : 28 Apr 2024 02:38
Operator : MA/JU
Sample : P2203-23
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
E1D03

Quant Time: Apr 28 21:18:21 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Apr 26 22:57:25 2024
Response via : Initial Calibration

