

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042821\
 Data File : BN014467.D
 Acq On : 28 Apr 2021 10:32
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 28 12:30:35 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 27 19:23:29 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	99	0.00
2	1,4-Dioxane	0.400	0.406	-1.5	99	0.00
3	n-Nitrosodimethylamine	0.400	0.327	18.3	92	0.02
4 S	2-Fluorophenol	0.400	0.433	-8.2	107	0.00
5 S	Phenol-d6	0.400	0.423	-5.7	109	0.00
6	bis(2-Chloroethyl)ether	0.400	0.430	-7.5	105	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	100	0.00
8 S	Nitrobenzene-d5	0.400	0.445	-11.2	116	-0.01
9	Naphthalene	0.400	0.431	-7.7	109	0.00
10	Hexachlorobutadiene	0.400	0.443	-10.7	108	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.423	-5.7	108	0.00
12	2-Methylnaphthalene	0.400	0.422	-5.5	107	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	97	0.00
14 S	2,4,6-Tribromophenol	0.400	0.365	8.8	108	0.00
15 S	2-Fluorobiphenyl	0.400	0.435	-8.7	109	0.00
16	Acenaphthylene	0.400	0.422	-5.5	110	0.00
17	Acenaphthene	0.400	0.415	-3.7	107	-0.01
18	Fluorene	0.400	0.412	-3.0	107	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	98	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.399	0.3	102	0.00
21	4-Bromophenyl-phenylether	0.400	0.428	-7.0	107	0.00
22	Hexachlorobenzene	0.400	0.474	-18.5	115	0.00
23	Atrazine	0.400	0.353	11.8	98	0.00
24	Pentachlorophenol	0.400	0.448	-12.0	96	-0.01
25	Phenanthrene	0.400	0.430	-7.5	107	0.00
26	Anthracene	0.400	0.402	-0.5	105	0.00
27 SURR	Fluoranthene-d10	0.400	0.404	-1.0	103	0.00
28	Fluoranthene	0.400	0.405	-1.3	103	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	89	0.00
30	Pyrene	0.400	0.427	-6.7	99	0.00
31 S	Terphenyl-d14	0.400	0.443	-10.7	102	0.00
32	Benzo(a)anthracene	0.400	0.397	0.8	89	0.00
33	Chrysene	0.400	0.454	-13.5	100	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.327	18.3	76	0.00
35 I	Perylene-d12	0.400	0.400	0.0	82	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.409	-2.2	83	0.00
37	Benzo(b)fluoranthene	0.400	0.437	-9.2	91	0.00
38	Benzo(k)fluoranthene	0.400	0.432	-8.0	90	0.00
39 C	Benzo(a)pyrene	0.400	0.407	-1.7	83	0.00
40	Dibenzo(a,h)anthracene	0.400	0.413	-3.2	84	0.00
41	Benzo(g,h,i)perylene	0.400	0.446	-11.5	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4