

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042821\
 Data File : BN014457.D
 Acq On : 27 Apr 2021 15:15
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: Apr 27 18:53:16 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 27 18:52:25 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.676	152	7364	0.40	ng	0.00
7) Naphthalene-d8	10.442	136	27268	0.40	ng	0.00
13) Acenaphthene-d10	14.308	164	14540	0.40	ng	0.00
19) Phenanthrene-d10	17.055	188	31515	0.40	ng	0.00
29) Chrysene-d12	21.250	240	30599	0.40	ng	0.00
35) Perylene-d12	23.469	264	26505	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.276	112	1482	0.10	ng	0.00
5) Phenol-d6	6.847	99	1705	0.09	ng	0.00
8) Nitrobenzene-d5	8.819	82	1701	0.07	ng	0.00
11) 2-Methylnaphthalene-d10	12.040	152	4058	0.10	ng	0.00
14) 2,4,6-Tribromophenol	15.805	330	411	0.09	ng	0.01
15) 2-Fluorobiphenyl	12.925	172	5385	0.09	ng	0.00
27) Fluoranthene-d10	19.089	212	8387	0.10	ng	0.00
31) Terphenyl-d14	19.695	244	6291	0.10	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.239	88	1030	0.11	ng	# 84
6) bis(2-Chloroethyl)ether	7.104	93	1831	0.10	ng	99
9) Naphthalene	10.492	128	6773	0.10	ng	97
10) Hexachlorobutadiene	10.784	225	1318	0.09	ng	# 99
12) 2-Methylnaphthalene	12.115	142	4353	0.10	ng	97
16) Acenaphthylene	14.016	152	5090	0.08	ng	99
17) Acenaphthene	14.372	154	3866	0.09	ng	99
18) Fluorene	15.361	166	4960	0.10	ng	99
20) 4,6-Dinitro-2-methylph...	15.450	198	185	0.05	ng	# 1
21) 4-Bromophenyl-phenylether	16.252	248	1581	0.09	ng	96
22) Hexachlorobenzene	16.362	284	2019	0.08	ng	97
23) Atrazine	16.520	200	1002	0.10	ng	# 94
25) Phenanthrene	17.092	178	8847	0.10	ng	99
26) Anthracene	17.177	178	6266	0.09	ng	99
28) Fluoranthene	19.119	202	9510	0.10	ng	100
30) Pyrene	19.481	202	9890	0.11	ng	100
32) Benzo(a)anthracene	21.239	228	7991	0.10	ng	99
33) Chrysene	21.282	228	9279	0.10	ng	99
34) Bis(2-ethylhexyl)phtha...	21.176	149	4331	0.15	ng	99
36) Indeno(1,2,3-cd)pyrene	25.697	276	8959	0.08	ng	98
37) Benzo(b)fluoranthene	22.808	252	9015	0.09	ng	# 76
38) Benzo(k)fluoranthene	22.849	252	9215	0.09	ng	# 82
39) Benzo(a)pyrene	23.376	252	8051	0.10	ng	# 58
40) Dibenzo(a,h)anthracene	25.714	278	7393	0.08	ng	# 89
41) Benzo(g,h,i)perylene	26.375	276	8632	0.09	ng	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042821\
Data File : BN014457.D
Acq On : 27 Apr 2021 15:15
Operator : CG/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Quant Time: Apr 27 18:53:16 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 27 18:52:25 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.676 min Scan# 608
 Delta R.T. 0.000 min
 Lab File: BN014457.D
 Acq: 27 Apr 2021 15:15

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Tgt Ion: 152 Resp: 7364




Ion	Ratio	Lower	Upper
152	100		
150	152.3	121.2	181.8
115	57.4	46.0	69.0




#2
 1,4-Dioxane
 Concen: 0.11 ng
 RT: 3.239 min Scan# 57
 Delta R.T. 0.008 min
 Lab File: BN014457.D
 Acq: 27 Apr 2021 15:15

Tgt Ion: 88 Resp: 1030

Ion	Ratio	Lower	Upper
88	100		
58	52.0	53.1	79.7#
43	19.2	20.3	30.5#









#4
2-Fluorophenol
Concen: 0.10 ng
RT: 5.276 min Scan# 310
Delta R.T. -0.000 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:112 Resp: 1482
Ion Ratio Lower Upper
112 100
64 49.1 41.4 62.2
63 28.8 21.8 32.8



#5
Phenol-d6
Concen: 0.09 ng
RT: 6.847 min Scan# 505
Delta R.T. 0.008 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15

Tgt Ion: 99 Resp: 1705
Ion Ratio Lower Upper
99 100
42 20.3 13.8 20.6
71 32.9 25.5 38.3









#6
bis(2-Chloroethyl)ether
Concen: 0.10 ng
RT: 7.104 min Scan# 537
Delta R.T. -0.000 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion: 93 Resp: 1831
Ion Ratio Lower Upper
93 100
63 70.1 56.7 85.1
95 32.1 26.3 39.5



#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.442 min Scan# 855
Delta R.T. -0.000 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15

Tgt Ion: 136 Resp: 27268
Ion Ratio Lower Upper
136 100
137 11.4 9.0 13.4
54 7.0 5.7 8.5
68 6.7 5.2 7.8









#8
Nitrobenzene-d5
Concen: 0.07 ng
RT: 8.819 min Scan# 727
Delta R.T. -0.000 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion: 82 Resp: 1701
Ion Ratio Lower Upper
82 100
128 48.6 38.5 57.7
54 46.3 32.1 48.1



#9
Naphthalene
Concen: 0.10 ng
RT: 10.492 min Scan# 859
Delta R.T. -0.000 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15

Tgt Ion: 128 Resp: 6773
Ion Ratio Lower Upper
128 100
129 12.7 9.0 13.4
127 14.2 10.5 15.7









#10
Hexachlorobutadiene
Concen: 0.09 ng
RT: 10.784 min Scan# 882
Delta R.T. -0.000 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:225 Resp: 1318
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.8 51.5 77.3



#11
2-Methylnaphthalene-d10
Concen: 0.10 ng
RT: 12.040 min Scan# 1059
Delta R.T. -0.000 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15

Tgt Ion:152 Resp: 4058
Ion Ratio Lower Upper
152 100
151 19.8 16.2 24.4









#12
2-Methylnaphthalene
Concen: 0.10 ng
RT: 12.115 min Scan# 1076
Delta R.T. -0.000 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:	142	Resp:	4353
Ion	Ratio	Lower	Upper
142	100		
141	88.7	69.6	104.4
115	37.8	27.0	40.6



#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.308 min Scan# 1306
Delta R.T. -0.000 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15

Tgt Ion:	164	Resp:	14540
Ion	Ratio	Lower	Upper
164	100		
162	98.6	78.3	117.5
160	46.2	36.6	54.8









#14
2,4,6-Tribromophenol
Concen: 0.09 ng
RT: 15.805 min Scan# 1424
Delta R.T. 0.013 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:330 Resp: 411
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 39.9 29.3 43.9



#15
2-Fluorobiphenyl
Concen: 0.09 ng
RT: 12.925 min Scan# 1197
Delta R.T. -0.000 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15

Tgt Ion:172 Resp: 5385
Ion Ratio Lower Upper
172 100
171 35.1 27.0 40.4
170 23.9 18.2 27.4









#16
Acenaphthylene
Concen: 0.08 ng
RT: 14.016 min Scan# 1283
Delta R.T. -0.000 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	152	Resp:	5090
Ion	Ratio	Lower	Upper
152	100		
151	19.4	15.4	23.2
153	13.5	10.4	15.6



#17
Acenaphthene
Concen: 0.09 ng
RT: 14.372 min Scan# 1311
Delta R.T. -0.000 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15

Tgt Ion:	154	Resp:	3866
Ion	Ratio	Lower	Upper
154	100		
153	113.8	90.5	135.7
152	58.7	47.2	70.8









#18
Fluorene
Concen: 0.10 ng
RT: 15.361 min Scan# 1389
Delta R.T. -0.000 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:166 Resp: 4960
Ion Ratio Lower Upper
166 100
165 96.2 78.1 117.1
167 13.6 11.0 16.6



#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.055 min Scan# 1526
Delta R.T. -0.000 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15

Tgt Ion:188 Resp: 31515
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 6.7 5.4 8.2









#20
4,6-Dinitro-2-methylphenol
Concen: 0.05 ng
RT: 15.450 min Scan# 1396
Delta R.T. 0.013 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:198 Resp: 185
Ion Ratio Lower Upper
198 100
51 467.3 104.5 156.7#
105 197.2 62.2 93.2#



#21
4-Bromophenyl-phenylether
Concen: 0.09 ng
RT: 16.252 min Scan# 1460
Delta R.T. -0.000 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15

Tgt Ion:248 Resp: 1581
Ion Ratio Lower Upper
248 100
250 103.1 81.8 122.6
141 49.7 34.2 51.2









#22
Hexachlorobenzene
Concen: 0.08 ng
RT: 16.362 min Scan# 1469
Delta R.T. -0.000 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	284	Resp:	2019
Ion	Ratio	Lower	Upper
284	100		
142	35.3	27.2	40.8
249	31.4	23.4	35.2



#23
Atrazine
Concen: 0.10 ng
RT: 16.520 min Scan# 1482
Delta R.T. -0.000 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15

Tgt Ion:	200	Resp:	1002
Ion	Ratio	Lower	Upper
200	100		
173	32.0	20.8	31.2#
215	51.3	39.3	58.9









#25
Phenanthrene
Concen: 0.10 ng
RT: 17.092 min Scan# 1529
Delta R.T. -0.000 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	178	Resp:	8847
Ion	Ratio	Lower	Upper
178	100		
176	19.0	15.1	22.7
179	15.7	12.2	18.4



#26
Anthracene
Concen: 0.09 ng
RT: 17.177 min Scan# 1536
Delta R.T. -0.000 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15

Tgt Ion:	178	Resp:	6266
Ion	Ratio	Lower	Upper
178	100		
176	18.9	14.6	22.0
179	15.0	12.0	18.0









#27
Fluoranthene-d10
Concen: 0.10 ng
RT: 19.089 min Scan# 1821
Delta R.T. -0.000 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:212 Resp: 8387
Ion Ratio Lower Upper
212 100
106 12.7 10.2 15.4
104 7.0 5.6 8.4



#28
Fluoranthene
Concen: 0.10 ng
RT: 19.119 min Scan# 1827
Delta R.T. -0.000 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15

Tgt Ion:202 Resp: 9510
Ion Ratio Lower Upper
202 100
101 10.5 8.5 12.7
203 17.2 13.7 20.5









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.250 min Scan# 2123
Delta R.T. -0.000 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:240 Resp: 30599
Ion Ratio Lower Upper
240 100
120 10.8 9.0 13.6
236 26.5 21.4 32.0



#30
Pyrene
Concen: 0.11 ng
RT: 19.481 min Scan# 1900
Delta R.T. -0.005 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15

Tgt Ion:202 Resp: 9890
Ion Ratio Lower Upper
202 100
200 20.5 16.4 24.6
203 17.4 13.9 20.9









#31
 Terphenyl-d14
 Concen: 0.10 ng
 RT: 19.695 min Scan# 1943
 Delta R.T. -0.000 min
 Lab File: BN014457.D
 Acq: 27 Apr 2021 15:15




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:244	Resp:	6291
Ion Ratio	Lower	Upper
244	100	
212	8.3	6.2 9.2
122	12.6	9.8 14.8



#32
 Benzo(a)anthracene
 Concen: 0.10 ng
 RT: 21.239 min Scan# 2122
 Delta R.T. -0.000 min
 Lab File: BN014457.D
 Acq: 27 Apr 2021 15:15

Tgt Ion:228	Resp:	7991
Ion Ratio	Lower	Upper
228	100	
226	26.4	20.8 31.2
229	21.1	16.1 24.1









#33
Chrysene
Concen: 0.10 ng
RT: 21.282 min Scan# 2126
Delta R.T. -0.000 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Tgt Ion:228 Resp: 9279




Ion	Ratio	Lower	Upper
228	100		
226	31.0	24.8	37.2
229	19.8	15.3	22.9




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.15 ng
RT: 21.176 min Scan# 2116
Delta R.T. -0.000 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15

Tgt Ion:149 Resp: 4331


Ion	Ratio	Lower	Upper
149	100		
167	24.5	19.4	29.2
279	4.3	2.9	4.3








#35
Perylene-d12
Concen: 0.40 ng
RT: 23.469 min Scan# 2542
Delta R.T. -0.003 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1







Tgt Ion:264 Resp: 26505
Ion Ratio Lower Upper
264 100
260 25.1 19.8 29.6
265 41.4 31.6 47.4



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.08 ng
RT: 25.697 min Scan# 3193
Delta R.T. -0.000 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15

Tgt Ion:276 Resp: 8959
Ion Ratio Lower Upper
276 100
138 21.2 18.5 27.7
277 24.8 19.9 29.9









#37
Benzo(b)fluoranthene
Concen: 0.09 ng
RT: 22.808 min Scan# 2349
Delta R.T. -0.000 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:252 Resp: 9015
Ion Ratio Lower Upper
252 100
253 32.5 19.3 28.9#
125 30.7 12.8 19.2#



#38
Benzo(k)fluoranthene
Concen: 0.09 ng
RT: 22.849 min Scan# 2361
Delta R.T. -0.000 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15


Tgt Ion:252 Resp: 9215
Ion Ratio Lower Upper
252 100
253 32.3 19.3 28.9#
125 20.9 10.0 15.0#








#39
Benzo(a)pyrene
Concen: 0.10 ng
RT: 23.376 min Scan# 2515
Delta R.T. -0.000 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1







Tgt Ion:252 Resp: 8051
Ion Ratio Lower Upper
252 100
253 39.4 21.0 31.6#
125 55.9 20.8 31.2#



#40
Dibenzo(a,h)anthracene
Concen: 0.08 ng
RT: 25.714 min Scan# 3198
Delta R.T. 0.003 min
Lab File: BN014457.D
Acq: 27 Apr 2021 15:15

Tgt Ion:278 Resp: 7393
Ion Ratio Lower Upper
278 100
139 20.5 14.3 21.5
279 32.8 20.3 30.5#





#41
 Benzo(g,h,i)perylene
 Concen: 0.09 ng
 RT: 26.375 min Scan# 3391
 Delta R.T. -0.000 min
 Lab File: BN014457.D
 Acq: 27 Apr 2021 15:15

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Tgt Ion:	276	Resp:	8632
Ion	Ratio	Lower	Upper
276	100		
277	27.0	19.5	29.3
138	25.8	18.3	27.5

