

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042821\
 Data File : BN014459.D
 Acq On : 27 Apr 2021 16:26
 Operator : CG/JU
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4

Quant Time: Apr 27 18:56:00 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 27 18:52:25 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.676	152	7728	0.40	ng	0.00
7) Naphthalene-d8	10.442	136	27772	0.40	ng	0.00
13) Acenaphthene-d10	14.308	164	14683	0.40	ng	0.00
19) Phenanthrene-d10	17.055	188	29726	0.40	ng	0.00
29) Chrysene-d12	21.250	240	31356	0.40	ng	0.00
35) Perylene-d12	23.472	264	27575	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.276	112	5809	0.37	ng	0.00
5) Phenol-d6	6.839	99	6876	0.35	ng	0.00
8) Nitrobenzene-d5	8.819	82	7121	0.31	ng	0.00
11) 2-Methylnaphthalene-d10	12.040	152	16456	0.39	ng	0.00
14) 2,4,6-Tribromophenol	15.793	330	1589	0.34	ng	0.00
15) 2-Fluorobiphenyl	12.925	172	21485	0.37	ng	0.00
27) Fluoranthene-d10	19.089	212	31410	0.40	ng	0.00
31) Terphenyl-d14	19.695	244	24165	0.36	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.231	88	3472	0.37	ng	100
3) n-Nitrosodimethylamine	3.593	42	1466	0.30	ng	# 100
6) bis(2-Chloroethyl)ether	7.104	93	7445	0.39	ng	100
9) Naphthalene	10.492	128	27301	0.40	ng	100
10) Hexachlorobutadiene	10.784	225	5448	0.36	ng	# 100
12) 2-Methylnaphthalene	12.115	142	17778	0.39	ng	100
16) Acenaphthylene	14.016	152	21498	0.34	ng	100
17) Acenaphthene	14.372	154	15281	0.37	ng	100
18) Fluorene	15.361	166	18995	0.37	ng	100
20) 4,6-Dinitro-2-methylph...	15.437	198	876	0.26	ng	100
21) 4-Bromophenyl-phenylether	16.252	248	6110	0.36	ng	100
22) Hexachlorobenzene	16.362	284	8105	0.35	ng	100
23) Atrazine	16.520	200	3472	0.36	ng	100
24) Pentachlorophenol	16.715	266	1191	0.28	ng	100
25) Phenanthrene	17.092	178	32590	0.39	ng	100
26) Anthracene	17.177	178	24435	0.37	ng	100
28) Fluoranthene	19.119	202	35328	0.40	ng	100
30) Pyrene	19.486	202	35686	0.38	ng	100
32) Benzo(a)anthracene	21.239	228	34260	0.41	ng	100
33) Chrysene	21.282	228	38185	0.39	ng	100
34) Bis(2-ethylhexyl)phtha...	21.176	149	14045	0.48	ng	100
36) Indeno(1,2,3-cd)pyrene	25.697	276	42754	0.37	ng	100
37) Benzo(b)fluoranthene	22.808	252	38492	0.38	ng	100
38) Benzo(k)fluoranthene	22.849	252	39699	0.38	ng	100
39) Benzo(a)pyrene	23.376	252	35449	0.41	ng	100
40) Dibenzo(a,h)anthracene	25.711	278	35448	0.37	ng	100
41) Benzo(g,h,i)perylene	26.375	276	37976	0.38	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042821\
Data File : BN014459.D
Acq On : 27 Apr 2021 16:26
Operator : CG/JU
Sample : SSTDICCC0.4
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Quant Time: Apr 27 18:56:00 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 27 18:52:25 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.676 min Scan# 608
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:152 Resp: 7728
Ion Ratio Lower Upper
152 100
150 151.5 121.2 181.8
115 57.5 46.0 69.0



#2
1,4-Dioxane
Concen: 0.37 ng
RT: 3.231 min Scan# 56
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26

Tgt Ion: 88 Resp: 3472
Ion Ratio Lower Upper
88 100
58 66.4 53.1 79.7
43 25.4 20.3 30.5









#3
n-Nitrosodimethylamine
Concen: 0.30 ng
RT: 3.593 min Scan# 101
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion: 42 Resp: 1466
Ion Ratio Lower Upper
42 100
74 155.9 124.7 187.1
44 0.0 0.0 0.0



#4
2-Fluorophenol
Concen: 0.37 ng
RT: 5.276 min Scan# 310
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26

Tgt Ion: 112 Resp: 5809
Ion Ratio Lower Upper
112 100
64 51.8 41.4 62.2
63 27.3 21.8 32.8









#5
Phenol-d6
Concen: 0.35 ng
RT: 6.839 min Scan# 504
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion: 99 Resp: 6876
Ion Ratio Lower Upper
99 100
42 17.2 13.8 20.6
71 31.9 25.5 38.3



#6
bis(2-Chloroethyl)ether
Concen: 0.39 ng
RT: 7.104 min Scan# 537
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26

Tgt Ion: 93 Resp: 7445
Ion Ratio Lower Upper
93 100
63 70.9 56.7 85.1
95 32.9 26.3 39.5









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.442 min Scan# 855
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	136	Resp:	27772
Ion Ratio	Lower	Upper	
136	100		
137	11.2	9.0	13.4
54	7.1	5.7	8.5
68	6.5	5.2	7.8



#8
Nitrobenzene-d5
Concen: 0.31 ng
RT: 8.819 min Scan# 727
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26

Tgt Ion:	82	Resp:	7121
Ion Ratio	Lower	Upper	
82	100		
128	48.1	38.5	57.7
54	40.1	32.1	48.1









#9
Naphthalene
Concen: 0.40 ng
RT: 10.492 min Scan# 859
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	128	Resp:	27301
Ion	Ratio	Lower	Upper
128	100		
129	11.2	9.0	13.4
127	13.1	10.5	15.7



#10
Hexachlorobutadiene
Concen: 0.36 ng
RT: 10.784 min Scan# 882
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26

Tgt Ion:	225	Resp:	5448
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.4	51.5	77.3









#11
2-Methylnaphthalene-d10
Concen: 0.39 ng
RT: 12.040 min Scan# 1059
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:152 Resp: 16456
Ion Ratio Lower Upper
152 100
151 20.3 16.2 24.4



#12
2-Methylnaphthalene
Concen: 0.39 ng
RT: 12.115 min Scan# 1076
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26

Tgt Ion:142 Resp: 17778
Ion Ratio Lower Upper
142 100
141 87.0 69.6 104.4
115 33.8 27.0 40.6









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.308 min Scan# 1306
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:164 Resp: 14683
Ion Ratio Lower Upper
164 100
162 97.9 78.3 117.5
160 45.7 36.6 54.8



#14
2,4,6-Tribromophenol
Concen: 0.34 ng
RT: 15.793 min Scan# 1423
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26

Tgt Ion:330 Resp: 1589
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 36.6 29.3 43.9









#15
2-Fluorobiphenyl
Concen: 0.37 ng
RT: 12.925 min Scan# 1197
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	172	Resp:	21485
Ion	Ratio	Lower	Upper
172	100		
171	33.7	27.0	40.4
170	22.8	18.2	27.4



#16
Acenaphthylene
Concen: 0.34 ng
RT: 14.016 min Scan# 1283
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26

Tgt Ion:	152	Resp:	21498
Ion	Ratio	Lower	Upper
152	100		
151	19.3	15.4	23.2
153	13.0	10.4	15.6









#17
Acenaphthene
Concen: 0.37 ng
RT: 14.372 min Scan# 1311
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:154 Resp: 15281
Ion Ratio Lower Upper
154 100
153 113.1 90.5 135.7
152 59.0 47.2 70.8



#18
Fluorene
Concen: 0.37 ng
RT: 15.361 min Scan# 1389
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26

Tgt Ion:166 Resp: 18995
Ion Ratio Lower Upper
166 100
165 97.6 78.1 117.1
167 13.8 11.0 16.6









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.055 min Scan# 1526
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Tgt Ion:188 Resp: 29726




Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	6.8	5.4	8.2




#20
4,6-Dinitro-2-methylphenol
Concen: 0.26 ng
RT: 15.437 min Scan# 1395
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26

Tgt Ion:198 Resp: 876

Ion	Ratio	Lower	Upper
198	100		
51	130.6	104.5	156.7
105	77.7	62.2	93.2









#21
4-Bromophenyl-phenylether
Concen: 0.36 ng
RT: 16.252 min Scan# 1460
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	248	Resp:	6110
Ion	Ratio	Lower	Upper
248	100		
250	102.2	81.8	122.6
141	42.7	34.2	51.2



#22
Hexachlorobenzene
Concen: 0.35 ng
RT: 16.362 min Scan# 1469
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26

Tgt Ion:	284	Resp:	8105
Ion	Ratio	Lower	Upper
284	100		
142	34.0	27.2	40.8
249	29.3	23.4	35.2









#23
Atrazine
Concen: 0.36 ng
RT: 16.520 min Scan# 1482
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	200	Resp:	3472
Ion Ratio	Lower	Upper	
200	100		
173	26.0	20.8	31.2
215	49.1	39.3	58.9



#24
Pentachlorophenol
Concen: 0.28 ng
RT: 16.715 min Scan# 1498
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26

Tgt Ion:	266	Resp:	1191
Ion Ratio	Lower	Upper	
266	100		
264	62.4	49.9	74.9
268	64.5	51.6	77.4









#25
Phenanthrene
Concen: 0.39 ng
RT: 17.092 min Scan# 1529
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	178	Resp:	32590
Ion	Ratio	Lower	Upper
178	100		
176	18.9	15.1	22.7
179	15.3	12.2	18.4



#26
Anthracene
Concen: 0.37 ng
RT: 17.177 min Scan# 1536
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26

Tgt Ion:	178	Resp:	24435
Ion	Ratio	Lower	Upper
178	100		
176	18.3	14.6	22.0
179	15.0	12.0	18.0









#27
Fluoranthene-d10
Concen: 0.40 ng
RT: 19.089 min Scan# 1821
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:212 Resp: 31410
Ion Ratio Lower Upper
212 100
106 12.8 10.2 15.4
104 7.0 5.6 8.4



#28
Fluoranthene
Concen: 0.40 ng
RT: 19.119 min Scan# 1827
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26

Tgt Ion:202 Resp: 35328
Ion Ratio Lower Upper
202 100
101 10.6 8.5 12.7
203 17.1 13.7 20.5









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.250 min Scan# 2123
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:240 Resp: 31356
Ion Ratio Lower Upper
240 100
120 11.3 9.0 13.6
236 26.7 21.4 32.0



#30
Pyrene
Concen: 0.38 ng
RT: 19.486 min Scan# 1901
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26

Tgt Ion:202 Resp: 35686
Ion Ratio Lower Upper
202 100
200 20.5 16.4 24.6
203 17.4 13.9 20.9









#31
 Terphenyl-d14
 Concen: 0.36 ng
 RT: 19.695 min Scan# 1943
 Delta R.T. 0.000 min
 Lab File: BN014459.D
 Acq: 27 Apr 2021 16:26




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.4


Tgt Ion:244 Resp: 24165
 Ion Ratio Lower Upper
 244 100
 212 7.7 6.2 9.2
 122 12.3 9.8 14.8



#32
 Benzo(a)anthracene
 Concen: 0.41 ng
 RT: 21.239 min Scan# 2122
 Delta R.T. 0.000 min
 Lab File: BN014459.D
 Acq: 27 Apr 2021 16:26

Tgt Ion:228 Resp: 34260
 Ion Ratio Lower Upper
 228 100
 226 26.0 20.8 31.2
 229 20.1 16.1 24.1









#33
Chrysene
Concen: 0.39 ng
RT: 21.282 min Scan# 2126
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	228	Resp:	38185
Ion Ratio	Lower	Upper	
228	100		
226	31.0	24.8	37.2
229	19.1	15.3	22.9



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.48 ng
RT: 21.176 min Scan# 2116
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26

Tgt Ion:	149	Resp:	14045
Ion Ratio	Lower	Upper	
149	100		
167	24.3	19.4	29.2
279	3.6	2.9	4.3









#35
Perylene-d12
Concen: 0.40 ng
RT: 23.472 min Scan# 2543
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:264 Resp: 27575
Ion Ratio Lower Upper
264 100
260 24.7 19.8 29.6
265 39.5 31.6 47.4



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.37 ng
RT: 25.697 min Scan# 3193
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26

Tgt Ion:276 Resp: 42754
Ion Ratio Lower Upper
276 100
138 23.1 18.5 27.7
277 24.9 19.9 29.9





#37

Benzo(b)fluoranthene

Concen: 0.38 ng

RT: 22.808 min Scan# 2349

Delta R.T. 0.000 min

Lab File: BN014459.D

Acq: 27 Apr 2021 16:26





Tgt Ion:252 Resp: 38492

Ion Ratio Lower Upper

252 100

253 24.1 19.3 28.9

125 16.0 12.8 19.2



#38

Benzo(k)fluoranthene

Concen: 0.38 ng

RT: 22.849 min Scan# 2361

Delta R.T. 0.000 min

Lab File: BN014459.D

Acq: 27 Apr 2021 16:26




Tgt Ion:252 Resp: 39699

Ion Ratio Lower Upper

252 100

253 24.1 19.3 28.9

125 12.5 10.0 15.0




Instrument :

BNA_N


ClientSampleId :

SSTDICCC0.4





#39
Benzo(a)pyrene
Concen: 0.41 ng
RT: 23.376 min Scan# 2515
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4







Tgt Ion:252 Resp: 35449
Ion Ratio Lower Upper
252 100
253 26.3 21.0 31.6
125 26.0 20.8 31.2



#40
Dibenzo(a,h)anthracene
Concen: 0.37 ng
RT: 25.711 min Scan# 3197
Delta R.T. 0.000 min
Lab File: BN014459.D
Acq: 27 Apr 2021 16:26

Tgt Ion:278 Resp: 35448
Ion Ratio Lower Upper
278 100
139 17.9 14.3 21.5
279 25.4 20.3 30.5





#41
 Benzo(g,h,i)perylene
 Concen: 0.38 ng
 RT: 26.375 min Scan# 3391
 Delta R.T. 0.000 min
 Lab File: BN014459.D
 Acq: 27 Apr 2021 16:26

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4

Tgt Ion:	276	Resp:	37976
Ion	Ratio	Lower	Upper
276	100		
277	24.4	19.5	29.3
138	22.9	18.3	27.5

