

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042821\
 Data File : BN014460.D
 Acq On : 27 Apr 2021 17:02
 Operator : CG/JU
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Quant Time: Apr 27 18:56:57 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 27 18:52:25 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.668	152	7334	0.40	ng	0.00
7) Naphthalene-d8	10.441	136	26898	0.40	ng	0.00
13) Acenaphthene-d10	14.308	164	14055	0.40	ng	0.00
19) Phenanthrene-d10	17.055	188	30080	0.40	ng	0.00
29) Chrysene-d12	21.247	240	32007	0.40	ng	0.00
35) Perylene-d12	23.469	264	24556	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.276	112	10813	0.72	ng	0.00
5) Phenol-d6	6.839	99	13207	0.71	ng	0.00
8) Nitrobenzene-d5	8.807	82	15229	0.68	ng	-0.01
11) 2-Methylnaphthalene-d10	12.040	152	33328	0.82	ng	0.00
14) 2,4,6-Tribromophenol	15.793	330	3524	0.78	ng	0.00
15) 2-Fluorobiphenyl	12.925	172	46354	0.83	ng	0.00
27) Fluoranthene-d10	19.091	212	66482	0.83	ng	0.00
31) Terphenyl-d14	19.697	244	53436	0.79	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.231	88	6371	0.71	ng	97
3) n-Nitrosodimethylamine	3.577	42	2946	0.63	ng	# 100
6) bis(2-Chloroethyl)ether	7.104	93	14440	0.80	ng	99
9) Naphthalene	10.492	128	55069	0.82	ng	100
10) Hexachlorobutadiene	10.784	225	10574	0.72	ng	# 100
12) 2-Methylnaphthalene	12.111	142	36209	0.83	ng	99
16) Acenaphthylene	14.016	152	44840	0.75	ng	100
17) Acenaphthene	14.371	154	31535	0.80	ng	99
18) Fluorene	15.361	166	39896	0.82	ng	100
20) 4,6-Dinitro-2-methylph...	15.437	198	2081	0.62	ng	# 52
21) 4-Bromophenyl-phenylether	16.252	248	12789	0.75	ng	97
22) Hexachlorobenzene	16.361	284	16997	0.73	ng	99
23) Atrazine	16.519	200	7485	0.76	ng	99
24) Pentachlorophenol	16.702	266	2781	0.65	ng	99
25) Phenanthrene	17.091	178	67453	0.80	ng	100
26) Anthracene	17.177	178	52135	0.79	ng	100
28) Fluoranthene	19.121	202	75538	0.84	ng	100
30) Pyrene	19.483	202	73670	0.77	ng	100
32) Benzo(a)anthracene	21.237	228	70744	0.84	ng	99
33) Chrysene	21.290	228	79090	0.79	ng	97
34) Bis(2-ethylhexyl)phtha...	21.173	149	26672	0.89	ng	98
36) Indeno(1,2,3-cd)pyrene	25.694	276	77207	0.75	ng	98
37) Benzo(b)fluoranthene	22.805	252	71662	0.79	ng	# 94
38) Benzo(k)fluoranthene	22.850	252	76915	0.84	ng	97
39) Benzo(a)pyrene	23.377	252	63368	0.83	ng	# 87
40) Dibenzo(a,h)anthracene	25.708	278	64977	0.76	ng	97
41) Benzo(g,h,i)perylene	26.372	276	69297	0.77	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042821\
Data File : BN014460.D
Acq On : 27 Apr 2021 17:02
Operator : CG/JU
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Quant Time: Apr 27 18:56:57 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 27 18:52:25 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.668 min Scan# 607
Delta R.T. -0.008 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02




Instrument :
BNA_N
ClientSampled :
SSTDIC0.8


Tgt Ion:152 Resp: 7334
Ion Ratio Lower Upper
152 100
150 157.8 121.2 181.8
115 60.5 46.0 69.0



#2
1,4-Dioxane
Concen: 0.71 ng
RT: 3.231 min Scan# 56
Delta R.T. -0.000 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02

Tgt Ion: 88 Resp: 6371
Ion Ratio Lower Upper
88 100
58 68.8 53.1 79.7
43 26.0 20.3 30.5









#3
n-Nitrosodimethylamine
Concen: 0.63 ng
RT: 3.577 min Scan# 99
Delta R.T. -0.016 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion: 42 Resp: 2946
Ion Ratio Lower Upper
42 100
74 155.7 124.7 187.1
44 20.0 0.0 0.0#



#4
2-Fluorophenol
Concen: 0.72 ng
RT: 5.276 min Scan# 310
Delta R.T. -0.000 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02

Tgt Ion: 112 Resp: 10813
Ion Ratio Lower Upper
112 100
64 52.0 41.4 62.2
63 26.8 21.8 32.8









#5
Phenol-d6
Concen: 0.71 ng
RT: 6.839 min Scan# 504
Delta R.T. -0.000 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02

Instrument :
BNA_N
ClientSampled :
SSTDICC0.8

Tgt Ion: 99 Resp: 13207




Ion	Ratio	Lower	Upper
99	100		
42	16.9	13.8	20.6
71	32.3	25.5	38.3




#6
bis(2-Chloroethyl)ether
Concen: 0.80 ng
RT: 7.104 min Scan# 537
Delta R.T. -0.000 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02

Tgt Ion: 93 Resp: 14440

Ion	Ratio	Lower	Upper
93	100		
63	71.5	56.7	85.1
95	33.6	26.3	39.5









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.441 min Scan# 855
Delta R.T. -0.000 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	136	Resp:	26898
Ion	Ratio	Lower	Upper
136	100		
137	11.2	9.0	13.4
54	6.7	5.7	8.5
68	6.3	5.2	7.8



#8
Nitrobenzene-d5
Concen: 0.68 ng
RT: 8.807 min Scan# 726
Delta R.T. -0.013 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02

Tgt Ion:	82	Resp:	15229
Ion	Ratio	Lower	Upper
82	100		
128	33.5	38.5	57.7#
54	48.9	32.1	48.1#









#9
Naphthalene
Concen: 0.82 ng
RT: 10.492 min Scan# 859
Delta R.T. -0.000 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:128 Resp: 55069
Ion Ratio Lower Upper
128 100
129 11.1 9.0 13.4
127 13.0 10.5 15.7



#10
Hexachlorobutadiene
Concen: 0.72 ng
RT: 10.784 min Scan# 882
Delta R.T. -0.000 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02

Tgt Ion:225 Resp: 10574
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.2 51.5 77.3









#11
2-Methylnaphthalene-d10
Concen: 0.82 ng
RT: 12.040 min Scan# 1059
Delta R.T. -0.000 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:152 Resp: 33328
Ion Ratio Lower Upper
152 100
151 20.3 16.2 24.4



#12
2-Methylnaphthalene
Concen: 0.83 ng
RT: 12.111 min Scan# 1075
Delta R.T. -0.005 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02

Tgt Ion:142 Resp: 36209
Ion Ratio Lower Upper
142 100
141 88.4 69.6 104.4
115 34.3 27.0 40.6









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.308 min Scan# 1306
Delta R.T. -0.000 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:164 Resp: 14055
Ion Ratio Lower Upper
164 100
162 99.3 78.3 117.5
160 46.7 36.6 54.8



#14
2,4,6-Tribromophenol
Concen: 0.78 ng
RT: 15.793 min Scan# 1423
Delta R.T. -0.000 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02

Tgt Ion:330 Resp: 3524
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 34.8 29.3 43.9









#15
2-Fluorobiphenyl
Concen: 0.83 ng
RT: 12.925 min Scan# 1197
Delta R.T. -0.000 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	172	Resp:	46354
Ion	Ratio	Lower	Upper
172	100		
171	33.9	27.0	40.4
170	22.8	18.2	27.4



#16
Acenaphthylene
Concen: 0.75 ng
RT: 14.016 min Scan# 1283
Delta R.T. -0.000 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02

Tgt Ion:	152	Resp:	44840
Ion	Ratio	Lower	Upper
152	100		
151	19.2	15.4	23.2
153	13.2	10.4	15.6









#17
Acenaphthene
Concen: 0.80 ng
RT: 14.371 min Scan# 1311
Delta R.T. -0.000 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:154 Resp: 31535
Ion Ratio Lower Upper
154 100
153 112.3 90.5 135.7
152 58.3 47.2 70.8



#18
Fluorene
Concen: 0.82 ng
RT: 15.361 min Scan# 1389
Delta R.T. -0.000 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02

Tgt Ion:166 Resp: 39896
Ion Ratio Lower Upper
166 100
165 97.7 78.1 117.1
167 13.5 11.0 16.6









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.055 min Scan# 1526
Delta R.T. -0.001 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02




Instrument :
BNA_N
ClientSampled :
SSTDIC0.8


Tgt Ion:188 Resp: 30080
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 6.7 5.4 8.2



#20
4,6-Dinitro-2-methylphenol
Concen: 0.62 ng
RT: 15.437 min Scan# 1395
Delta R.T. -0.000 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02

Tgt Ion:198 Resp: 2081
Ion Ratio Lower Upper
198 100
51 66.1 104.5 156.7#
105 47.1 62.2 93.2#









#21
4-Bromophenyl-phenylether
Concen: 0.75 ng
RT: 16.252 min Scan# 1460
Delta R.T. -0.001 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	248	Resp:	12789
Ion	Ratio	Lower	Upper
248	100		
250	103.3	81.8	122.6
141	38.4	34.2	51.2



#22
Hexachlorobenzene
Concen: 0.73 ng
RT: 16.361 min Scan# 1469
Delta R.T. -0.001 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02

Tgt Ion:	284	Resp:	16997
Ion	Ratio	Lower	Upper
284	100		
142	33.3	27.2	40.8
249	29.1	23.4	35.2









#23
Atrazine
Concen: 0.76 ng
RT: 16.519 min Scan# 1482
Delta R.T. -0.001 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	200	Resp:	7485
Ion Ratio	Lower	Upper	
200	100		
173	24.7	20.8	31.2
215	48.9	39.3	58.9



#24
Pentachlorophenol
Concen: 0.65 ng
RT: 16.702 min Scan# 1497
Delta R.T. -0.013 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02

Tgt Ion:	266	Resp:	2781
Ion Ratio	Lower	Upper	
266	100		
264	62.7	49.9	74.9
268	63.7	51.6	77.4









#25
Phenanthrene
Concen: 0.80 ng
RT: 17.091 min Scan# 1529
Delta R.T. -0.001 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:178 Resp: 67453
Ion Ratio Lower Upper
178 100
176 18.7 15.1 22.7
179 15.3 12.2 18.4



#26
Anthracene
Concen: 0.79 ng
RT: 17.177 min Scan# 1536
Delta R.T. -0.001 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02

Tgt Ion:178 Resp: 52135
Ion Ratio Lower Upper
178 100
176 18.2 14.6 22.0
179 15.3 12.0 18.0









#27
Fluoranthene-d10
Concen: 0.83 ng
RT: 19.091 min Scan# 1823
Delta R.T. 0.002 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:212 Resp: 66482
Ion Ratio Lower Upper
212 100
106 12.7 10.2 15.4
104 6.9 5.6 8.4



#28
Fluoranthene
Concen: 0.84 ng
RT: 19.121 min Scan# 1829
Delta R.T. 0.002 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02

Tgt Ion:202 Resp: 75538
Ion Ratio Lower Upper
202 100
101 10.7 8.5 12.7
203 17.2 13.7 20.5









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.247 min Scan# 2124
Delta R.T. -0.003 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:240 Resp: 32007
Ion Ratio Lower Upper
240 100
120 13.3 9.0 13.6
236 27.0 21.4 32.0



#30
Pyrene
Concen: 0.77 ng
RT: 19.483 min Scan# 1902
Delta R.T. -0.003 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02

Tgt Ion:202 Resp: 73670
Ion Ratio Lower Upper
202 100
200 20.6 16.4 24.6
203 17.4 13.9 20.9









#31
Terphenyl-d14
Concen: 0.79 ng
RT: 19.697 min Scan# 1945
Delta R.T. 0.002 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:244 Resp: 53436
Ion Ratio Lower Upper
244 100
212 7.3 6.2 9.2
122 10.9 9.8 14.8



#32
Benzo(a)anthracene
Concen: 0.84 ng
RT: 21.237 min Scan# 2123
Delta R.T. -0.003 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02

Tgt Ion:228 Resp: 70744
Ion Ratio Lower Upper
228 100
226 26.3 20.8 31.2
229 19.7 16.1 24.1









#33
Chrysene
Concen: 0.79 ng
RT: 21.290 min Scan# 2128
Delta R.T. 0.008 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:228	Resp:	79090
Ion Ratio	Lower	Upper
228	100	
226	28.8	24.8 37.2
229	19.8	15.3 22.9



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.89 ng
RT: 21.173 min Scan# 2117
Delta R.T. -0.003 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02


Tgt Ion:149	Resp:	26672
Ion Ratio	Lower	Upper
149	100	
167	25.2	19.4 29.2
279	3.5	2.9 4.3








#35
Perylene-d12
Concen: 0.40 ng
RT: 23.469 min Scan# 2544
Delta R.T. -0.003 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8







Tgt Ion:264 Resp: 24556
Ion Ratio Lower Upper
264 100
260 24.8 19.8 29.6
265 41.7 31.6 47.4



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.75 ng
RT: 25.694 min Scan# 3194
Delta R.T. -0.003 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02

Tgt Ion:276 Resp: 77207
Ion Ratio Lower Upper
276 100
138 25.2 18.5 27.7
277 24.7 19.9 29.9









#37
Benzo(b)fluoranthene
Concen: 0.79 ng
RT: 22.805 min Scan# 2350
Delta R.T. -0.003 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:252 Resp: 71662
Ion Ratio Lower Upper
252 100
253 22.7 19.3 28.9
125 11.9 12.8 19.2#



#38
Benzo(k)fluoranthene
Concen: 0.84 ng
RT: 22.850 min Scan# 2363
Delta R.T. 0.000 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02

Tgt Ion:252 Resp: 76915
Ion Ratio Lower Upper
252 100
253 23.0 19.3 28.9
125 11.0 10.0 15.0









#39
Benzo(a)pyrene
Concen: 0.83 ng
RT: 23.377 min Scan# 2517
Delta R.T. 0.001 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:252 Resp: 63368
Ion Ratio Lower Upper
252 100
253 23.8 21.0 31.6
125 15.4 20.8 31.2#



#40
Dibenzo(a,h)anthracene
Concen: 0.76 ng
RT: 25.708 min Scan# 3198
Delta R.T. -0.003 min
Lab File: BN014460.D
Acq: 27 Apr 2021 17:02

Tgt Ion:278 Resp: 64977
Ion Ratio Lower Upper
278 100
139 17.1 14.3 21.5
279 23.9 20.3 30.5





#41
 Benzo(g,h,i)perylene
 Concen: 0.77 ng
 RT: 26.372 min Scan# 3392
 Delta R.T. -0.003 min
 Lab File: BN014460.D
 Acq: 27 Apr 2021 17:02

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Tgt Ion:	276	Resp:	69297
Ion	Ratio	Lower	Upper
276	100		
277	24.1	19.5	29.3
138	21.9	18.3	27.5

