

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042821\
 Data File : BN014467.D
 Acq On : 28 Apr 2021 10:32
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Manual Integrations
 APPROVED

mohammad
 4/29/2021 5:44:00 PM

Quant Time: Apr 28 12:30:35 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 27 19:23:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.668	152	7630	0.40	ng	0.00
7) Naphthalene-d8	10.442	136	27822	0.40	ng	0.00
13) Acenaphthene-d10	14.308	164	14273	0.40	ng	0.00
19) Phenanthrene-d10	17.043	188	29139	0.40	ng	#-0.01
29) Chrysene-d12	21.247	240	27939	0.40	ng	0.00
35) Perylene-d12	23.466	264	22521	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.276	112	6213	0.43	ng	0.00
5) Phenol-d6	6.839	99	7501	0.42	ng	0.00
8) Nitrobenzene-d5	8.807	82	8260	0.44	ng	-0.01
11) 2-Methylnaphthalene-d10	12.040	152	17763	0.42	ng	0.00
14) 2,4,6-Tribromophenol	15.793	330	1718	0.37	ng	0.00
15) 2-Fluorobiphenyl	12.925	172	23395	0.44	ng	0.00
27) Fluoranthene-d10	19.086	212	32221	0.40	ng	0.00
31) Terphenyl-d14	19.692	244	24758	0.44	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.231	88	3453	0.41	ng	95
3) n-Nitrosodimethylamine	3.609	42	1348	0.33	ng	# 95
6) bis(2-Chloroethyl)ether	7.104	93	7841	0.43	ng	99
9) Naphthalene	10.492	128	29659	0.43	ng	100
10) Hexachlorobutadiene	10.784	225	5898	0.44	ng	# 100
12) 2-Methylnaphthalene	12.111	142	19094	0.42	ng	98
16) Acenaphthylene	14.016	152	23640	0.42	ng	100
17) Acenaphthene	14.359	154	16302	0.41	ng	98
18) Fluorene	15.349	166	20295	0.41	ng	100
20) 4,6-Dinitro-2-methylph...	15.437	198	897	0.40	ng	# 90
21) 4-Bromophenyl-phenylether	16.252	248	6525	0.43	ng	97
22) Hexachlorobenzene	16.361	284	9319	0.47	ng	100
23) Atrazine	16.519	200	3401	0.35	ng	98
24) Pentachlorophenol	16.702	266	1146	0.45	ng	99
25) Phenanthrene	17.091	178	34793	0.43	ng	100
26) Anthracene	17.177	178	25650	0.40	ng	100
28) Fluoranthene	19.116	202	36265	0.40	ng	100
30) Pyrene	19.483	202	35223	0.43	ng	100
32) Benzo(a)anthracene	21.237	228	30580	0.40	ng	99
33) Chrysene	21.279	228	38078	0.45	ng	100
34) Bis(2-ethylhexyl)phtha...	21.173	149	10681m	0.33	ng	
36) Indeno(1,2,3-cd)pyrene	25.691	276	35496	0.41	ng	98
37) Benzo(b)fluoranthene	22.802	252	34994	0.44	ng	# 96
38) Benzo(k)fluoranthene	22.846	252	35836	0.43	ng	99
39) Benzo(a)pyrene	23.370	252	29394	0.41	ng	# 89
40) Dibenzo(a,h)anthracene	25.708	278	29782	0.41	ng	99
41) Benzo(g,h,i)perylene	26.369	276	34122	0.45	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042821\
Data File : BN014467.D
Acq On : 28 Apr 2021 10:32
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_N
Client Sampled :
SSTDCCC0.4

Manual Integrations
APPROVED


mohammad
4/29/2021 5:44:00 PM

Quant Time: Apr 28 12:30:35 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 27 19:23:29 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.668 min Scan# 607
Delta R.T. -0.008 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32





Tgt Ion: 152 Resp: 7630
Ion Ratio Lower Upper
152 100
150 157.2 121.2 181.8
115 60.1 46.0 69.0



#2
1,4-Dioxane
Concen: 0.41 ng
RT: 3.231 min Scan# 56
Delta R.T. -0.000 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32




Tgt Ion: 88 Resp: 3453
Ion Ratio Lower Upper
88 100
58 62.9 53.1 79.7
43 23.0 20.3 30.5



Instrument :
BNA_N
ClientSampled :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad
4/29/2021 5:44:00 PM







#3
n-Nitrosodimethylamine
Concen: 0.33 ng
RT: 3.609 min Scan# 103
Delta R.T. 0.016 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32

Tgt Ion: 42 Resp: 1348
Ion Ratio Lower Upper
42 100
74 149.8 124.7 187.1
44 0.0 0.0 0.0

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
4/29/2021 5:44:00 PM



#4
2-Fluorophenol
Concen: 0.43 ng
RT: 5.276 min Scan# 310
Delta R.T. -0.000 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32

Tgt Ion: 112 Resp: 6213
Ion Ratio Lower Upper
112 100
64 52.3 41.4 62.2
63 27.7 21.8 32.8









#5
Phenol-d6
Concen: 0.42 ng
RT: 6.839 min Scan# 504
Delta R.T. -0.000 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32

Tgt Ion:	99	Resp:	7501
Ion	Ratio	Lower	Upper
99	100		
42	17.9	13.8	20.6
71	32.3	25.5	38.3

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
4/29/2021 5:44:00 PM



#6
bis(2-Chloroethyl)ether
Concen: 0.43 ng
RT: 7.104 min Scan# 537
Delta R.T. -0.000 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32


Tgt Ion:	93	Resp:	7841
Ion	Ratio	Lower	Upper
93	100		
63	71.7	56.7	85.1
95	34.1	26.3	39.5





#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.442 min Scan# 855
Delta R.T. -0.000 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion:	136	Resp:	27822
Ion	Ratio	Lower	Upper
136	100		
137	11.1	9.0	13.4
54	6.3	5.7	8.5
68	5.7	5.2	7.8

Manual Integrations
APPROVED



mohammad
4/29/2021 5:44:00 PM




#8
Nitrobenzene-d5
Concen: 0.44 ng
RT: 8.807 min Scan# 726
Delta R.T. -0.013 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32



Tgt Ion:	82	Resp:	8260
Ion	Ratio	Lower	Upper
82	100		
128	34.2	38.5	57.7#
54	50.0	32.1	48.1#









#9
Naphthalene
Concen: 0.43 ng
RT: 10.492 min Scan# 859
Delta R.T. -0.000 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32

Tgt Ion:	128	Resp:	29659
Ion	Ratio	Lower	Upper
128	100		
129	11.4	9.0	13.4
127	13.1	10.5	15.7

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
4/29/2021 5:44:00 PM



#10
Hexachlorobutadiene
Concen: 0.44 ng
RT: 10.784 min Scan# 882
Delta R.T. -0.000 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32

Tgt Ion:	225	Resp:	5898
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.3	51.5	77.3









#11
2-Methylnaphthalene-d10
Concen: 0.42 ng
RT: 12.040 min Scan# 1059
Delta R.T. -0.000 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32

Tgt Ion:152 Resp: 17763
Ion Ratio Lower Upper
152 100
151 20.3 16.2 24.4

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
4/29/2021 5:44:00 PM



#12
2-Methylnaphthalene
Concen: 0.42 ng
RT: 12.111 min Scan# 1075
Delta R.T. -0.005 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32


Tgt Ion:142 Resp: 19094
Ion Ratio Lower Upper
142 100
141 88.6 69.6 104.4
115 35.5 27.0 40.6





#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.308 min Scan# 1306
Delta R.T. -0.000 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:164 Resp: 14273
Ion Ratio Lower Upper
164 100
162 97.8 78.3 117.5
160 46.8 36.6 54.8




Manual Integrations
APPROVED


mohammad
4/29/2021 5:44:00 PM



#14
2,4,6-Tribromophenol
Concen: 0.37 ng
RT: 15.793 min Scan# 1423
Delta R.T. -0.000 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32

Tgt Ion:330 Resp: 1718
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 35.0 29.3 43.9









#15
2-Fluorobiphenyl
Concen: 0.44 ng
RT: 12.925 min Scan# 1197
Delta R.T. -0.000 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32

Tgt Ion:	172	Resp:	23395
Ion Ratio	Lower	Upper	
172	100		
171	34.0	27.0	40.4
170	22.7	18.2	27.4

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
4/29/2021 5:44:00 PM



#16
Acenaphthylene
Concen: 0.42 ng
RT: 14.016 min Scan# 1283
Delta R.T. -0.000 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32


Tgt Ion:	152	Resp:	23640
Ion Ratio	Lower	Upper	
152	100		
151	19.4	15.4	23.2
153	13.0	10.4	15.6





#17
Acenaphthene
Concen: 0.41 ng
RT: 14.359 min Scan# 1310
Delta R.T. -0.013 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion:154 Resp: 16302
Ion Ratio Lower Upper
154 100
153 114.2 90.5 135.7
152 61.5 47.2 70.8

Manual Integrations
APPROVED



mohammad
4/29/2021 5:44:00 PM




#18
Fluorene
Concen: 0.41 ng
RT: 15.349 min Scan# 1388
Delta R.T. -0.013 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32



Tgt Ion:166 Resp: 20295
Ion Ratio Lower Upper
166 100
165 98.0 78.1 117.1
167 13.5 11.0 16.6









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.043 min Scan# 1525
Delta R.T. -0.013 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	12.9	5.4	8.2#




Manual Integrations
APPROVED


mohammad
4/29/2021 5:44:00 PM




#20
4,6-Dinitro-2-methylphenol
Concen: 0.40 ng
RT: 15.437 min Scan# 1395
Delta R.T. 0.000 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32

Tgt Ion	Ratio	Lower	Upper
198	100		
51	131.3	104.5	156.7
105	56.5	62.2	93.2#









#21
4-Bromophenyl-phenylether
Concen: 0.43 ng
RT: 16.252 min Scan# 1460
Delta R.T. -0.001 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32





Tgt Ion:248 Resp: 6525
Ion Ratio Lower Upper
248 100
250 103.2 81.8 122.6
141 37.1 34.2 51.2



#22
Hexachlorobenzene
Concen: 0.47 ng
RT: 16.361 min Scan# 1469
Delta R.T. -0.001 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32




Tgt Ion:284 Resp: 9319
Ion Ratio Lower Upper
284 100
142 34.0 27.2 40.8
249 29.2 23.4 35.2




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED




mohammad
4/29/2021 5:44:00 PM




#23
Atrazine
Concen: 0.35 ng
RT: 16.519 min Scan# 1482
Delta R.T. -0.001 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32





Tgt Ion:200 Resp: 3401
Ion Ratio Lower Upper
200 100
173 24.6 20.8 31.2
215 49.8 39.3 58.9



#24
Pentachlorophenol
Concen: 0.45 ng
RT: 16.702 min Scan# 1497
Delta R.T. -0.013 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32




Tgt Ion:266 Resp: 1146
Ion Ratio Lower Upper
266 100
264 63.8 49.9 74.9
268 64.6 51.6 77.4



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad
4/29/2021 5:44:00 PM







#25
Phenanthrene
Concen: 0.43 ng
RT: 17.091 min Scan# 1529
Delta R.T. -0.001 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32

Tgt Ion:178	Resp:	34793
Ion Ratio	Lower	Upper
178	100	
176	18.9	15.1 22.7
179	15.3	12.2 18.4

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
4/29/2021 5:44:00 PM



#26
Anthracene
Concen: 0.40 ng
RT: 17.177 min Scan# 1536
Delta R.T. -0.001 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32


Tgt Ion:178	Resp:	25650
Ion Ratio	Lower	Upper
178	100	
176	18.3	14.6 22.0
179	15.3	12.0 18.0





#27
Fluoranthene-d10
Concen: 0.40 ng
RT: 19.086 min Scan# 1822
Delta R.T. -0.003 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion:212 Resp: 32221
Ion Ratio Lower Upper
212 100
106 12.5 10.2 15.4
104 7.0 5.6 8.4

Manual Integrations
APPROVED



mohammad
4/29/2021 5:44:00 PM




#28
Fluoranthene
Concen: 0.40 ng
RT: 19.116 min Scan# 1828
Delta R.T. -0.003 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32




Tgt Ion:202 Resp: 36265
Ion Ratio Lower Upper
202 100
101 10.8 8.5 12.7
203 17.2 13.7 20.5









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.247 min Scan# 2124
Delta R.T. -0.003 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32





Tgt Ion:240 Resp: 27939
Ion Ratio Lower Upper
240 100
120 9.5 9.0 13.6
236 26.4 21.4 32.0



#30
Pyrene
Concen: 0.43 ng
RT: 19.483 min Scan# 1902
Delta R.T. -0.003 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32




Tgt Ion:202 Resp: 35223
Ion Ratio Lower Upper
202 100
200 20.5 16.4 24.6
203 17.4 13.9 20.9




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED




mohammad
4/29/2021 5:44:00 PM




#31
Terphenyl-d14
Concen: 0.44 ng
RT: 19.692 min Scan# 1944
Delta R.T. -0.003 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32





Tgt Ion:244 Resp: 24758
Ion Ratio Lower Upper
244 100
212 7.7 6.2 9.2
122 12.2 9.8 14.8



#32
Benzo(a)anthracene
Concen: 0.40 ng
RT: 21.237 min Scan# 2123
Delta R.T. -0.003 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32




Tgt Ion:228 Resp: 30580
Ion Ratio Lower Upper
228 100
226 26.3 20.8 31.2
229 19.8 16.1 24.1



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad
4/29/2021 5:44:00 PM







#33
Chrysene
Concen: 0.45 ng
RT: 21.279 min Scan# 2127
Delta R.T. -0.003 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32

Tgt Ion:	228	Resp:	38078
Ion Ratio	Lower	Upper	
228	100		
226	30.8	24.8	37.2
229	19.2	15.3	22.9

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
4/29/2021 5:44:00 PM




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.33 ng m
RT: 21.173 min Scan# 2117
Delta R.T. -0.003 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32

Tgt Ion:	149	Resp:	10681
Ion Ratio	Lower	Upper	
149	100		
167	23.8	19.4	29.2
279	3.4	2.9	4.3







#35
Perylene-d12
Concen: 0.40 ng
RT: 23.466 min Scan# 2543
Delta R.T. -0.006 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32



Tgt Ion: 264 Resp: 22521


Ion	Ratio	Lower	Upper
264	100		
260	24.6	19.8	29.6
265	41.9	31.6	47.4




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad
4/29/2021 5:44:00 PM






#36
Indeno(1,2,3-cd)pyrene
Concen: 0.41 ng
RT: 25.691 min Scan# 3193
Delta R.T. -0.006 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32



Tgt Ion: 276 Resp: 35496

Ion	Ratio	Lower	Upper
276	100		
138	24.9	18.5	27.7
277	24.8	19.9	29.9









#37
Benzo(b)fluoranthene
Concen: 0.44 ng
RT: 22.802 min Scan# 2349
Delta R.T. -0.006 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.9	19.3	28.9
125	12.5	12.8	19.2#

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
4/29/2021 5:44:00 PM




#38
Benzo(k)fluoranthene
Concen: 0.43 ng
RT: 22.846 min Scan# 2362
Delta R.T. -0.003 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.7	19.3	28.9
125	11.8	10.0	15.0









#39
Benzo(a)pyrene
Concen: 0.41 ng
RT: 23.370 min Scan# 2515
Delta R.T. -0.006 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32





Tgt Ion:252 Resp: 29394
Ion Ratio Lower Upper
252 100
253 25.2 21.0 31.6
125 15.6 20.8 31.2#



#40
Dibenzo(a,h)anthracene
Concen: 0.41 ng
RT: 25.708 min Scan# 3198
Delta R.T. -0.003 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32




Tgt Ion:278 Resp: 29782
Ion Ratio Lower Upper
278 100
139 17.2 14.3 21.5
279 25.2 20.3 30.5



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad
4/29/2021 5:44:00 PM



#41
Benzo(g,h,i)perylene
Concen: 0.45 ng
RT: 26.369 min Scan# 3391
Delta R.T. -0.006 min
Lab File: BN014467.D
Acq: 28 Apr 2021 10:32

Tgt Ion:	276	Resp:	34122
Ion	Ratio	Lower	Upper
276	100		
277	24.4	19.5	29.3
138	22.4	18.3	27.5



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad
4/29/2021 5:44:00 PM