

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042821\
 Data File : BN014469.D
 Acq On : 28 Apr 2021 12:28
 Operator : CG/JU
 Sample : M1458-01
 Misc : LOD-MDL-SOIL-0.1PPM
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 LOD-MDL-SOIL-01-QT1-2021

Manual Integrations
 APPROVED


mohammad
 4/29/2021 5:44:03 PM

Quant Time: Apr 28 13:14:15 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 28 13:14:06 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.668	152	5099	0.40	ng	0.00
7) Naphthalene-d8	10.441	136	18122	0.40	ng	0.00
13) Acenaphthene-d10	14.308	164	9107	0.40	ng	0.00
19) Phenanthrene-d10	17.055	188	18622	0.40	ng	0.00
29) Chrysene-d12	21.250	240	16306	0.40	ng	0.00
35) Perylene-d12	23.469	264	14231	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.268	112	4790	0.50	ng	0.00
5) Phenol-d6	6.839	99	5420	0.46	ng	0.00
8) Nitrobenzene-d5	8.807	82	6021m	0.50	ng	-0.01
11) 2-Methylnaphthalene-d10	12.040	152	10798	0.39	ng	0.00
14) 2,4,6-Tribromophenol	15.805	330	1066	0.35	ng	0.01
15) 2-Fluorobiphenyl	12.925	172	17693	0.52	ng	0.00
27) Fluoranthene-d10	19.089	212	18648	0.37	ng	0.00
31) Terphenyl-d14	19.695	244	17749	0.54	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.239	88	721	0.13	ng	# 83
3) n-Nitrosodimethylamine	3.593	42	106	0.04	ng	# 1
6) bis(2-Chloroethyl)ether	7.104	93	1495	0.12	ng	99
9) Naphthalene	10.492	128	5391	0.12	ng	96
10) Hexachlorobutadiene	10.784	225	1110	0.13	ng	# 99
12) 2-Methylnaphthalene	12.111	142	3380	0.11	ng	95
16) Acenaphthylene	14.016	152	4187	0.12	ng	98
17) Acenaphthene	14.371	154	2777	0.11	ng	97
18) Fluorene	15.361	166	3365	0.11	ng	100
20) 4,6-Dinitro-2-methylph...	15.450	198	136m	0.08	ng	
21) 4-Bromophenyl-phenylether	16.252	248	1098	0.11	ng	96
22) Hexachlorobenzene	16.362	284	1670	0.13	ng	98
23) Atrazine	16.520	200	537	0.09	ng	# 91
24) Pentachlorophenol	16.715	266	142	0.06	ng	93
25) Phenanthrene	17.092	178	5968	0.12	ng	99
26) Anthracene	17.177	178	4230	0.10	ng	99
28) Fluoranthene	19.119	202	5926	0.10	ng	100
30) Pyrene	19.481	202	5669	0.12	ng	99
32) Benzo(a)anthracene	21.229	228	4737	0.11	ng	97
33) Chrysene	21.282	228	5758	0.12	ng	98
34) Bis(2-ethylhexyl)phtha...	21.176	149	2004	0.11	ng	100
36) Indeno(1,2,3-cd)pyrene	25.697	276	5773	0.11	ng	99
37) Benzo(b)fluoranthene	22.805	252	5744	0.11	ng	# 79
38) Benzo(k)fluoranthene	22.849	252	5906	0.11	ng	# 78
39) Benzo(a)pyrene	23.373	252	5282	0.12	ng	# 70
40) Dibenzo(a,h)anthracene	25.711	278	4623	0.10	ng	# 85
41) Benzo(g,h,i)perylene	26.375	276	5611	0.12	ng	93




(#) = qualifier out of range (m) = manual integration (+) = signals summed




#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.668 min Scan# 607
Delta R.T. -0.008 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28





Tgt Ion: 152 Resp: 5099
Ion Ratio Lower Upper
152 100
150 156.3 121.2 181.8
115 61.4 46.0 69.0



#2
1,4-Dioxane
Concen: 0.13 ng
RT: 3.239 min Scan# 57
Delta R.T. 0.008 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28




Tgt Ion: 88 Resp: 721
Ion Ratio Lower Upper
88 100
58 51.3 53.1 79.7#
43 18.4 20.3 30.5#



Instrument :
BNA_N
ClientSampleId :
LOD-MDL-SOIL-01-QT1-2021

Manual Integrations
APPROVED

mohammad
4/29/2021 5:44:03 PM







#3
n-Nitrosodimethylamine
Concen: 0.04 ng
RT: 3.593 min Scan# 101
Delta R.T. -0.000 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28

Tgt Ion: 42 Resp: 106
Ion Ratio Lower Upper
42 100
74 0.0 124.7 187.1#
44 0.0 0.0 0.0

Instrument :
BNA_N
ClientSampleId :
LOD-MDL-SOIL-01-QT1-2021




Manual Integrations
APPROVED


mohammad
4/29/2021 5:44:03 PM



#4
2-Fluorophenol
Concen: 0.50 ng
RT: 5.268 min Scan# 309
Delta R.T. -0.008 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28

Tgt Ion: 112 Resp: 4790
Ion Ratio Lower Upper
112 100
64 52.1 41.4 62.2
63 27.7 21.8 32.8









#5
Phenol-d6
Concen: 0.46 ng
RT: 6.839 min Scan# 504
Delta R.T. -0.000 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28

Tgt Ion:	99	Resp:	5420
Ion	Ratio	Lower	Upper
99	100		
42	18.5	13.8	20.6
71	33.0	25.5	38.3

Instrument :
BNA_N
ClientSampleId :
LOD-MDL-SOIL-01-QT1-2021




Manual Integrations
APPROVED


mohammad
4/29/2021 5:44:03 PM



#6
bis(2-Chloroethyl)ether
Concen: 0.12 ng
RT: 7.104 min Scan# 537
Delta R.T. -0.000 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28


Tgt Ion:	93	Resp:	1495
Ion	Ratio	Lower	Upper
93	100		
63	71.7	56.7	85.1
95	34.2	26.3	39.5





#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.441 min Scan# 855
Delta R.T. -0.000 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28



Instrument :
BNA_N
ClientSampleId :
LOD-MDL-SOIL-01-QT1-2021




Tgt Ion:	136	Resp:	18122
Ion	Ratio	Lower	Upper
136	100		
137	11.6	9.0	13.4
54	6.8	5.7	8.5
68	6.6	5.2	7.8

Manual Integrations
APPROVED



mohammad
4/29/2021 5:44:03 PM




#8
Nitrobenzene-d5
Concen: 0.50 ng m
RT: 8.807 min Scan# 726
Delta R.T. -0.013 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28



Tgt Ion:	82	Resp:	6021
Ion	Ratio	Lower	Upper
82	100		
128	33.8	38.5	57.7#
54	50.6	32.1	48.1#









#9
Naphthalene
Concen: 0.12 ng
RT: 10.492 min Scan# 859
Delta R.T. -0.000 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28

Tgt Ion:	128	Resp:	5391
Ion	Ratio	Lower	Upper
128	100		
129	13.1	9.0	13.4
127	14.2	10.5	15.7

Instrument :
BNA_N
ClientSampleId :
LOD-MDL-SOIL-01-QT1-2021




Manual Integrations
APPROVED


mohammad
4/29/2021 5:44:03 PM



#10
Hexachlorobutadiene
Concen: 0.13 ng
RT: 10.784 min Scan# 882
Delta R.T. -0.000 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28

Tgt Ion:	225	Resp:	1110
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	65.0	51.5	77.3









#11
2-Methylnaphthalene-d10
Concen: 0.39 ng
RT: 12.040 min Scan# 1059
Delta R.T. -0.000 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28

Tgt Ion:152 Resp: 10798
Ion Ratio Lower Upper
152 100
151 19.0 16.2 24.4

Instrument :
BNA_N
ClientSampleId :
LOD-MDL-SOIL-01-QT1-2021




Manual Integrations
APPROVED


mohammad
4/29/2021 5:44:03 PM



#12
2-Methylnaphthalene
Concen: 0.11 ng
RT: 12.111 min Scan# 1075
Delta R.T. -0.005 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28


Tgt Ion:142 Resp: 3380
Ion Ratio Lower Upper
142 100
141 89.8 69.6 104.4
115 39.1 27.0 40.6





#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.308 min Scan# 1306
Delta R.T. -0.000 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28




Instrument :
BNA_N
ClientSampleId :
LOD-MDL-SOIL-01-QT1-2021




Tgt Ion:164 Resp: 9107
Ion Ratio Lower Upper
164 100
162 97.0 78.3 117.5
160 49.8 36.6 54.8

Manual Integrations
APPROVED



mohammad
4/29/2021 5:44:03 PM




#14
2,4,6-Tribromophenol
Concen: 0.35 ng
RT: 15.805 min Scan# 1424
Delta R.T. 0.013 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28



Tgt Ion:330 Resp: 1066
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 35.6 29.3 43.9









#15
2-Fluorobiphenyl
Concen: 0.52 ng
RT: 12.925 min Scan# 1197
Delta R.T. -0.000 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28

Tgt Ion:	172	Resp:	17693
Ion Ratio	Lower	Upper	
172	100		
171	34.3	27.0	40.4
170	22.8	18.2	27.4

Instrument :
BNA_N
ClientSampleId :
LOD-MDL-SOIL-01-QT1-2021




Manual Integrations
APPROVED


mohammad
4/29/2021 5:44:03 PM



#16
Acenaphthylene
Concen: 0.12 ng
RT: 14.016 min Scan# 1283
Delta R.T. -0.000 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28


Tgt Ion:	152	Resp:	4187
Ion Ratio	Lower	Upper	
152	100		
151	20.8	15.4	23.2
153	13.3	10.4	15.6





#17
Acenaphthene
Concen: 0.11 ng
RT: 14.371 min Scan# 1311
Delta R.T. -0.000 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28




Instrument :
BNA_N
ClientSampleId :
LOD-MDL-SOIL-01-QT1-2021




Tgt Ion:154 Resp: 2777
Ion Ratio Lower Upper
154 100
153 114.8 90.5 135.7
152 64.2 47.2 70.8

Manual Integrations
APPROVED



mohammad
4/29/2021 5:44:03 PM




#18
Fluorene
Concen: 0.11 ng
RT: 15.361 min Scan# 1389
Delta R.T. -0.000 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28




Tgt Ion:166 Resp: 3365
Ion Ratio Lower Upper
166 100
165 97.8 78.1 117.1
167 13.8 11.0 16.6





#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.055 min Scan# 1526
Delta R.T. -0.000 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28



Instrument :
BNA_N
ClientSampleId :
LOD-MDL-SOIL-01-QT1-2021




Tgt Ion:188 Resp: 18622
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.0 5.4 8.2

Manual Integrations
APPROVED



mohammad
4/29/2021 5:44:03 PM




#20
4,6-Dinitro-2-methylphenol
Concen: 0.08 ng m
RT: 15.450 min Scan# 1396
Delta R.T. 0.013 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28




Tgt Ion:198 Resp: 136
Ion Ratio Lower Upper
198 100
51 625.6 104.5 156.7#
105 156.4 62.2 93.2#






#21
4-Bromophenyl-phenylether
Concen: 0.11 ng
RT: 16.252 min Scan# 1460
Delta R.T. -0.000 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28




Tgt Ion:248 Resp: 1098
Ion Ratio Lower Upper
248 100
250 104.0 81.8 122.6
141 48.9 34.2 51.2




Instrument :
BNA_N
ClientSampleId :
LOD-MDL-SOIL-01-QT1-2021

Manual Integrations
APPROVED



mohammad
4/29/2021 5:44:03 PM




#22
Hexachlorobenzene
Concen: 0.13 ng
RT: 16.362 min Scan# 1469
Delta R.T. -0.000 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28




Tgt Ion:284 Resp: 1670
Ion Ratio Lower Upper
284 100
142 34.4 27.2 40.8
249 31.4 23.4 35.2









#23
Atrazine
Concen: 0.09 ng
RT: 16.520 min Scan# 1482
Delta R.T. -0.000 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28





Tgt Ion:200 Resp: 537
Ion Ratio Lower Upper
200 100
173 34.8 20.8 31.2#
215 52.6 39.3 58.9



#24
Pentachlorophenol
Concen: 0.06 ng
RT: 16.715 min Scan# 1498
Delta R.T. -0.000 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28




Tgt Ion:266 Resp: 142
Ion Ratio Lower Upper
266 100
264 54.2 49.9 74.9
268 61.3 51.6 77.4



Instrument :
BNA_N
ClientSampleId :
LOD-MDL-SOIL-01-QT1-2021

Manual Integrations
APPROVED

mohammad
4/29/2021 5:44:03 PM







#25
Phenanthrene
Concen: 0.12 ng
RT: 17.092 min Scan# 1529
Delta R.T. -0.000 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28

Tgt Ion:	178	Resp:	5968
Ion	Ratio	Lower	Upper
178	100		
176	19.1	15.1	22.7
179	15.8	12.2	18.4

Instrument :
BNA_N
ClientSampleId :
LOD-MDL-SOIL-01-QT1-2021




Manual Integrations
APPROVED


mohammad
4/29/2021 5:44:03 PM



#26
Anthracene
Concen: 0.10 ng
RT: 17.177 min Scan# 1536
Delta R.T. -0.000 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28


Tgt Ion:	178	Resp:	4230
Ion	Ratio	Lower	Upper
178	100		
176	19.1	14.6	22.0
179	15.2	12.0	18.0





#27
Fluoranthene-d10
Concen: 0.37 ng
RT: 19.089 min Scan# 1821
Delta R.T. -0.000 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28



Instrument :
BNA_N
ClientSampleId :
LOD-MDL-SOIL-01-QT1-2021




Tgt Ion:212 Resp: 18648
Ion Ratio Lower Upper
212 100
106 12.9 10.2 15.4
104 7.1 5.6 8.4

Manual Integrations
APPROVED



mohammad
4/29/2021 5:44:03 PM




#28
Fluoranthene
Concen: 0.10 ng
RT: 19.119 min Scan# 1827
Delta R.T. -0.000 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28



Tgt Ion:202 Resp: 5926
Ion Ratio Lower Upper
202 100
101 10.9 8.5 12.7
203 17.1 13.7 20.5









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.250 min Scan# 2123
Delta R.T. -0.000 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28

Tgt Ion:240 Resp: 16306
Ion Ratio Lower Upper
240 100
120 10.5 9.0 13.6
236 26.5 21.4 32.0

Instrument :
BNA_N
ClientSampleId :
LOD-MDL-SOIL-01-QT1-2021




Manual Integrations
APPROVED


mohammad
4/29/2021 5:44:03 PM



#30
Pyrene
Concen: 0.12 ng
RT: 19.481 min Scan# 1900
Delta R.T. -0.005 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28

Tgt Ion:202 Resp: 5669
Ion Ratio Lower Upper
202 100
200 20.4 16.4 24.6
203 17.9 13.9 20.9









#31
Terphenyl-d14
Concen: 0.54 ng
RT: 19.695 min Scan# 1943
Delta R.T. -0.000 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28

Tgt Ion:244 Resp: 17749
Ion Ratio Lower Upper
244 100
212 7.7 6.2 9.2
122 12.0 9.8 14.8

Instrument :
BNA_N
ClientSampleId :
LOD-MDL-SOIL-01-QT1-2021




Manual Integrations
APPROVED


mohammad
4/29/2021 5:44:03 PM




#32
Benzo(a)anthracene
Concen: 0.11 ng
RT: 21.229 min Scan# 2121
Delta R.T. -0.011 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28

Tgt Ion:228 Resp: 4737
Ion Ratio Lower Upper
228 100
226 29.1 20.8 31.2
229 20.1 16.1 24.1









#33
Chrysene
Concen: 0.12 ng
RT: 21.282 min Scan# 2126
Delta R.T. -0.000 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28





Tgt Ion:228 Resp: 5758
Ion Ratio Lower Upper
228 100
226 31.5 24.8 37.2
229 20.4 15.3 22.9



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.11 ng
RT: 21.176 min Scan# 2116
Delta R.T. -0.000 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28




Tgt Ion:149 Resp: 2004
Ion Ratio Lower Upper
149 100
167 24.5 19.4 29.2
279 3.6 2.9 4.3




Instrument :
BNA_N
ClientSampleId :
LOD-MDL-SOIL-01-QT1-2021

Manual Integrations
APPROVED



mohammad
4/29/2021 5:44:03 PM



#35
Perylene-d12
Concen: 0.40 ng
RT: 23.469 min Scan# 2542
Delta R.T. -0.004 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28




Tgt Ion:264 Resp: 14231
Ion Ratio Lower Upper
264 100
260 24.7 19.8 29.6
265 52.4 31.6 47.4#




Instrument :
BNA_N
ClientSampleId :
LOD-MDL-SOIL-01-QT1-2021

Manual Integrations
APPROVED



mohammad
4/29/2021 5:44:03 PM




#36
Indeno(1,2,3-cd)pyrene
Concen: 0.11 ng
RT: 25.697 min Scan# 3193
Delta R.T. -0.000 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28



Tgt Ion:276 Resp: 5773
Ion Ratio Lower Upper
276 100
138 22.3 18.5 27.7
277 25.3 19.9 29.9









#37
Benzo(b)fluoranthene
Concen: 0.11 ng
RT: 22.805 min Scan# 2348
Delta R.T. -0.004 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28

Tgt Ion:252 Resp: 5744
Ion Ratio Lower Upper
252 100
253 33.4 19.3 28.9#
125 26.2 12.8 19.2#

Instrument :
BNA_N
ClientSampleId :
LOD-MDL-SOIL-01-QT1-2021




Manual Integrations
APPROVED


mohammad
4/29/2021 5:44:03 PM



#38
Benzo(k)fluoranthene
Concen: 0.11 ng
RT: 22.849 min Scan# 2361
Delta R.T. -0.000 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28

Tgt Ion:252 Resp: 5906
Ion Ratio Lower Upper
252 100
253 34.4 19.3 28.9#
125 22.6 10.0 15.0#









#39
Benzo(a)pyrene
Concen: 0.12 ng
RT: 23.373 min Scan# 2514
Delta R.T. -0.004 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28

Tgt Ion:	252	Resp:	5282
Ion Ratio	Lower	Upper	
252	100		
253	40.5	21.0	31.6#
125	42.9	20.8	31.2#

Instrument :
BNA_N
ClientSampleId :
LOD-MDL-SOIL-01-QT1-2021




Manual Integrations
APPROVED


mohammad
4/29/2021 5:44:03 PM



#40
Dibenzo(a,h)anthracene
Concen: 0.10 ng
RT: 25.711 min Scan# 3197
Delta R.T. -0.000 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28


Tgt Ion:	278	Resp:	4623
Ion Ratio	Lower	Upper	
278	100		
139	22.0	14.3	21.5#
279	34.7	20.3	30.5#





#41
Benzo(g,h,i)perylene
Concen: 0.12 ng
RT: 26.375 min Scan# 3391
Delta R.T. -0.000 min
Lab File: BN014469.D
Acq: 28 Apr 2021 12:28

Tgt Ion:	276	Resp:	5611
Ion Ratio	Lower	Upper	
276	100		
277	28.6	19.5	29.3
138	25.1	18.3	27.5



Instrument :
BNA_N
ClientSampleId :
LOD-MDL-SOIL-01-QT1-2021

Manual Integrations
APPROVED

mohammad
4/29/2021 5:44:03 PM