

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042823\
 Data File : BN025185.D
 Acq On : 28 Apr 2023 11:20
 Operator : CG/JU
 Sample : 02520-03
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GWOW-06B-40-45-042623-FD

Quant Time: Apr 28 13:32:51 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN040523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 28 11:24:46 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.918	152	6795	0.400	ng	0.00
7) Naphthalene-d8	10.728	136	22439	0.400	ng	0.01
13) Acenaphthene-d10	14.555	164	14045	0.400	ng	0.00
19) Phenanthrene-d10	17.298	188	31148	0.400	ng	0.00
29) Chrysene-d12	21.504	240	23540	0.400	ng	0.00
35) Perylene-d12	23.937	264	20123	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.470	112	1356	0.086	ng	0.00
5) Phenol-d6	7.080	99	976	0.049	ng	0.00
8) Nitrobenzene-d5	9.084	82	5545	0.307	ng	0.01
11) 2-Methylnaphthalene-d10	12.313	152	10195	0.343	ng	0.00
14) 2,4,6-Tribromophenol	16.057	330	1686	0.267	ng	0.01
15) 2-Fluorobiphenyl	13.176	172	15942	0.297	ng	0.00
27) Fluoranthene-d10	19.338	212	28797	0.408	ng	0.00
31) Terphenyl-d14	19.928	244	26025	0.488	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.325	88	87705	11.746	ng	98
9) Naphthalene	10.771	128	1625	0.028	ng	# 82
25) Phenanthrene	17.348	178	2450	0.027	ng	# 92
30) Pyrene	19.730	202	2973	0.036	ng	# 80
32) Benzo(a)anthracene	21.486	228	1608	0.020	ng	# 70
34) Bis(2-ethylhexyl)phtha...	21.405	149	8446	0.198	ng	# 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042823\
Data File : BN025185.D
Acq On : 28 Apr 2023 11:20
Operator : CG/JU
Sample : 02520-03
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GWOW-06B-40-45-042623-FD

Quant Time: Apr 28 13:32:51 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN040523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 28 11:24:46 2023
Response via : Initial Calibration

