


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042823\
Data File : BN025181.D
Acq On : 28 Apr 2023 08:57
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN040523.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Fri Apr 28 15:09:05 2023






AutoFind: Scans 2474, 2475, 2476; Background Corrected with Scan 2466

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	51.3	327708	PASS
68	69	0.00	2	1.6	4444	PASS
69	198	0.00	100	42.8	273454	PASS
70	69	0.00	2	0.4	1204	PASS
127	198	10	80	46.0	293525	PASS
197	198	0.00	2	0.4	2298	PASS
198	198	100	100	100.0	638528	PASS
199	198	5	9	6.5	41546	PASS
275	198	10	60	26.8	170989	PASS
365	198	1	100	3.8	24085	PASS
441	198	0.01	100	14.5	92896	PASS
442	442	50	100	100.0	604181	PASS
443	442	15	24	17.3	104373	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042823\
 Data File : BN025181.D
 Acq On : 28 Apr 2023 08:57
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Quant Time: May 01 06:00:34 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 01 06:00:29 2023
 Response via : Initial Calibration



TIC: BN025181.D\data.ms

(70) Pentachlorophenol (C)

16.957min (0.000) 23335.58 ng




response 813105

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	62.93
264.00	61.60	61.03
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042823\
 Data File : BN025181.D
 Acq On : 28 Apr 2023 08:57
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Quant Time: May 01 06:00:34 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 01 06:00:29 2023
 Response via : Initial Calibration



TIC: BN025181.D\data.ms

(77) Benzidine

19.557min (0.000) 709994.38 ng

response 3983223

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.58
183.00	13.20	11.98
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
4/28/2023	BNA_N	<u>BN025181.D</u>
Compound Name	Response	Retention Time
DDT	1556670	20.786
DDD	531253	20.351
DDE	8477	19.839
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
539730	2096400	25.75

Instrument :
BNA_N
ClientSampleId :
DFTPP