



4,6-Dinitro-2-methylphenol



Response = $5.366e-002 * \text{Amt} - 3.324e-002$
 Coef of Det (r^2) = 0.991194 Curve Fit: Linear
 Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN042921.M
 Calibration Table Last Updated: Thu Apr 29 15:11:14 2021

Pentachlorophenol

Response Ratio



$$\text{Response} = 7.320\text{e-}002 * \text{Amt} - 5.421\text{e-}002$$


Coef of Det (r^2) = 0.993718 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN042921.M

Calibration Table Last Updated: Thu Apr 29 15:11:14 2021

n-Nitrosodimethylamine

Response Ratio



$$\text{Response} = 2.632\text{e-}001 * \text{Amt} - 1.099\text{e-}001$$

Coef of Det (r^2) = 0.999276 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN042921.M

Calibration Table Last Updated: Thu Apr 29 15:11:14 2021