

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042921\
 Data File : BN014483.D
 Acq On : 29 Apr 2021 10:18
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Apr 29 15:14:40 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 29 15:11:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.668	152	7629	0.40	ng	0.00
7) Naphthalene-d8	10.441	136	27536	0.40	ng	0.00
13) Acenaphthene-d10	14.295	164	13512	0.40	ng	0.00
19) Phenanthrene-d10	17.043	188	27194	0.40	ng	0.00
29) Chrysene-d12	21.247	240	23086	0.40	ng	# 0.00
35) Perylene-d12	23.466	264	19158	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.276	112	5464	0.40	ng	0.00
5) Phenol-d6	6.838	99	6566	0.39	ng	0.00
8) Nitrobenzene-d5	8.807	82	5334	0.38	ng	0.00
11) 2-Methylnaphthalene-d10	12.035	152	17098	0.30	ng	0.00
14) 2,4,6-Tribromophenol	15.792	330	1329	0.39	ng	0.00
15) 2-Fluorobiphenyl	12.925	172	22720	0.43	ng	0.00
27) Fluoranthene-d10	19.086	212	28842	0.30	ng	0.00
31) Terphenyl-d14	19.692	244	22289	0.42	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.231	88	3641	0.43	ng	97
3) n-Nitrosodimethylamine	3.609	42	1318	0.43	ng	# 87
6) bis(2-Chloroethyl)ether	7.104	93	7645	0.41	ng	99
9) Naphthalene	10.492	128	29213	0.40	ng	100
10) Hexachlorobutadiene	10.784	225	5693	0.41	ng	# 99
12) 2-Methylnaphthalene	12.111	142	18462	0.39	ng	99
16) Acenaphthylene	14.016	152	19092	0.38	ng	99
17) Acenaphthene	14.359	154	15494	0.40	ng	99
18) Fluorene	15.348	166	18602	0.40	ng	100
20) 4,6-Dinitro-2-methylph...	15.437	198	688	0.44	ng	91
21) 4-Bromophenyl-phenylether	16.252	248	5978	0.41	ng	99
22) Hexachlorobenzene	16.361	284	8551	0.43	ng	100
23) Atrazine	16.519	200	2626	0.38	ng	99
24) Pentachlorophenol	16.702	266	948	0.49	ng	98
25) Phenanthrene	17.091	178	32258	0.41	ng	100
26) Anthracene	17.176	178	22709	0.38	ng	100
28) Fluoranthene	19.116	202	32524	0.38	ng	100
30) Pyrene	19.478	202	31573	0.41	ng	100
32) Benzo(a)anthracene	21.236	228	23721	0.38	ng	98
33) Chrysene	21.279	228	32672	0.42	ng	99
34) Bis(2-ethylhexyl)phtha...	21.173	149	6853	0.41	ng	99
36) Indeno(1,2,3-cd)pyrene	25.687	276	29903	0.39	ng	100
37) Benzo(b)fluoranthene	22.802	252	29339	0.40	ng	99
38) Benzo(k)fluoranthene	22.843	252	33086	0.41	ng	100
39) Benzo(a)pyrene	23.370	252	24891	0.39	ng	99
40) Dibenzo(a,h)anthracene	25.701	278	24921	0.38	ng	98
41) Benzo(g,h,i)perylene	26.365	276	29483	0.41	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042921\
Data File : BN014483.D
Acq On : 29 Apr 2021 10:18
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Apr 29 15:14:40 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 29 15:11:14 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.668 min Scan# 607
 Delta R.T. -0.000 min
 Lab File: BN014483.D
 Acq: 29 Apr 2021 10:18




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:152 Resp: 7629
 Ion Ratio Lower Upper
 152 100
 150 155.8 126.6 189.8
 115 59.3 47.7 71.5



#2
 1,4-Dioxane
 Concen: 0.43 ng
 RT: 3.231 min Scan# 56
 Delta R.T. -0.000 min
 Lab File: BN014483.D
 Acq: 29 Apr 2021 10:18

Tgt Ion: 88 Resp: 3641
 Ion Ratio Lower Upper
 88 100
 58 67.7 55.4 83.0
 43 24.1 21.6 32.4









#3
n-Nitrosodimethylamine
Concen: 0.43 ng
RT: 3.609 min Scan# 103
Delta R.T. 0.008 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 42 Resp: 1318
Ion Ratio Lower Upper
42 100
74 134.1 120.3 180.5
44 11.6 6.7 10.1#



#4
2-Fluorophenol
Concen: 0.40 ng
RT: 5.276 min Scan# 310
Delta R.T. -0.000 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18

Tgt Ion: 112 Resp: 5464
Ion Ratio Lower Upper
112 100
64 52.9 42.4 63.6
63 28.3 22.2 33.4









#5
Phenol-d6
Concen: 0.39 ng
RT: 6.838 min Scan# 504
Delta R.T. -0.000 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 99 Resp: 6566
Ion Ratio Lower Upper
99 100
42 18.0 13.8 20.8
71 32.4 26.0 39.0



#6
bis(2-Chloroethyl)ether
Concen: 0.41 ng
RT: 7.104 min Scan# 537
Delta R.T. -0.000 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18

Tgt Ion: 93 Resp: 7645
Ion Ratio Lower Upper
93 100
63 72.0 58.3 87.5
95 33.2 27.0 40.6









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.441 min Scan# 855
Delta R.T. -0.000 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	27536
Ion	Ratio	Lower	Upper
136	100		
137	11.0	9.0	13.6
54	5.7	4.8	7.2
68	5.3	4.6	6.8



#8
Nitrobenzene-d5
Concen: 0.38 ng
RT: 8.807 min Scan# 726
Delta R.T. -0.000 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18

Tgt Ion:	82	Resp:	5334
Ion	Ratio	Lower	Upper
82	100		
128	32.5	26.0	39.0
54	49.2	39.5	59.3









#9
Naphthalene
Concen: 0.40 ng
RT: 10.492 min Scan# 859
Delta R.T. -0.000 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	128	Resp:	29213
Ion	Ratio	Lower	Upper
128	100		
129	11.1	9.0	13.6
127	12.8	10.2	15.4



#10
Hexachlorobutadiene
Concen: 0.41 ng
RT: 10.784 min Scan# 882
Delta R.T. -0.000 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18

Tgt Ion:	225	Resp:	5693
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.7	51.8	77.6









#11
2-Methylnaphthalene-d10
Concen: 0.30 ng
RT: 12.035 min Scan# 1058
Delta R.T. -0.000 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 17098
Ion Ratio Lower Upper
152 100
151 20.3 16.3 24.5



#12
2-Methylnaphthalene
Concen: 0.39 ng
RT: 12.111 min Scan# 1075
Delta R.T. -0.000 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18

Tgt Ion:142 Resp: 18462
Ion Ratio Lower Upper
142 100
141 88.7 70.5 105.7
115 34.9 27.5 41.3









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.295 min Scan# 1305
Delta R.T. -0.000 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:164 Resp: 13512
Ion Ratio Lower Upper
164 100
162 103.8 82.4 123.6
160 51.5 40.5 60.7



#14
2,4,6-Tribromophenol
Concen: 0.39 ng
RT: 15.792 min Scan# 1423
Delta R.T. -0.000 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18

Tgt Ion:330 Resp: 1329
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 35.9 28.5 42.7









#15
2-Fluorobiphenyl
Concen: 0.43 ng
RT: 12.925 min Scan# 1197
Delta R.T. -0.000 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:172	Resp:	22720
Ion Ratio	Lower	Upper
172	100	
171	33.2	26.6 40.0
170	22.4	18.2 27.2



#16
Acenaphthylene
Concen: 0.38 ng
RT: 14.016 min Scan# 1283
Delta R.T. -0.000 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18

Tgt Ion:152	Resp:	19092
Ion Ratio	Lower	Upper
152	100	
151	19.3	15.6 23.4
153	12.9	10.7 16.1









#17
Acenaphthene
Concen: 0.40 ng
RT: 14.359 min Scan# 1310
Delta R.T. -0.000 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 15494
Ion Ratio Lower Upper
154 100
153 112.3 90.3 135.5
152 59.9 47.3 70.9



#18
Fluorene
Concen: 0.40 ng
RT: 15.348 min Scan# 1388
Delta R.T. -0.000 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18

Tgt Ion:166 Resp: 18602
Ion Ratio Lower Upper
166 100
165 97.3 78.2 117.2
167 13.5 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.043 min Scan# 1525
Delta R.T. -0.001 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:188 Resp: 27194
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.2 8.8 13.2



#20
4,6-Dinitro-2-methylphenol
Concen: 0.44 ng
RT: 15.437 min Scan# 1395
Delta R.T. 0.000 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18

Tgt Ion:198 Resp: 688
Ion Ratio Lower Upper
198 100
51 133.1 116.0 174.0
105 50.7 45.3 67.9









#21
4-Bromophenyl-phenylether
Concen: 0.41 ng
RT: 16.252 min Scan# 1460
Delta R.T. -0.001 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	248	Resp:	5978
Ion	Ratio	Lower	Upper
248	100		
250	105.2	84.6	126.8
141	35.1	26.7	40.1



#22
Hexachlorobenzene
Concen: 0.43 ng
RT: 16.361 min Scan# 1469
Delta R.T. -0.001 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18

Tgt Ion:	284	Resp:	8551
Ion	Ratio	Lower	Upper
284	100		
142	33.6	27.0	40.6
249	29.6	23.6	35.4









#23
Atrazine
Concen: 0.38 ng
RT: 16.519 min Scan# 1482
Delta R.T. -0.001 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:200 Resp: 2626
Ion Ratio Lower Upper
200 100
173 24.6 20.0 30.0
215 49.7 39.0 58.6



#24
Pentachlorophenol
Concen: 0.49 ng
RT: 16.702 min Scan# 1497
Delta R.T. -0.000 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18

Tgt Ion:266 Resp: 948
Ion Ratio Lower Upper
266 100
264 63.8 49.9 74.9
268 63.8 49.9 74.9









#25
Phenanthrene
Concen: 0.41 ng
RT: 17.091 min Scan# 1529
Delta R.T. -0.001 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178 Resp: 32258
Ion Ratio Lower Upper
178 100
176 18.9 15.1 22.7
179 15.4 12.2 18.4



#26
Anthracene
Concen: 0.38 ng
RT: 17.176 min Scan# 1536
Delta R.T. -0.001 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18

Tgt Ion:178 Resp: 22709
Ion Ratio Lower Upper
178 100
176 18.5 14.6 22.0
179 15.0 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.30 ng
RT: 19.086 min Scan# 1822
Delta R.T. 0.002 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:212 Resp: 28842
Ion Ratio Lower Upper
212 100
106 12.6 10.2 15.4
104 7.1 5.6 8.4



#28
Fluoranthene
Concen: 0.38 ng
RT: 19.116 min Scan# 1828
Delta R.T. 0.002 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18

Tgt Ion:202 Resp: 32524
Ion Ratio Lower Upper
202 100
101 10.8 8.7 13.1
203 17.1 13.6 20.4









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.247 min Scan# 2124
Delta R.T. -0.003 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:240 Resp: 23086
Ion Ratio Lower Upper
240 100
120 9.6 6.3 9.5#
236 26.5 20.9 31.3



#30
Pyrene
Concen: 0.41 ng
RT: 19.478 min Scan# 1901
Delta R.T. -0.003 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18

Tgt Ion:202 Resp: 31573
Ion Ratio Lower Upper
202 100
200 20.3 16.4 24.6
203 17.4 13.9 20.9









#31
Terphenyl-d14
Concen: 0.42 ng
RT: 19.692 min Scan# 1944
Delta R.T. 0.002 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:244 Resp: 22289
Ion Ratio Lower Upper
244 100
212 7.8 6.3 9.5
122 11.8 10.6 16.0



#32
Benzo(a)anthracene
Concen: 0.38 ng
RT: 21.236 min Scan# 2123
Delta R.T. 0.008 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18

Tgt Ion:228 Resp: 23721
Ion Ratio Lower Upper
228 100
226 26.1 21.8 32.8
229 20.2 15.8 23.6









#33
Chrysene
Concen: 0.42 ng
RT: 21.279 min Scan# 2127
Delta R.T. -0.003 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:228 Resp: 32672
Ion Ratio Lower Upper
228 100
226 31.1 23.9 35.9
229 19.4 15.5 23.3



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.41 ng
RT: 21.173 min Scan# 2117
Delta R.T. -0.003 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18

Tgt Ion:149 Resp: 6853
Ion Ratio Lower Upper
149 100
167 26.5 21.4 32.2
279 3.8 3.3 4.9









#35
Perylene-d12
Concen: 0.40 ng
RT: 23.466 min Scan# 2543
Delta R.T. 0.001 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:264	Resp:	19158
Ion Ratio	Lower	Upper
264	100	
260	24.6	19.6 29.4
265	43.9	35.5 53.3



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.39 ng
RT: 25.687 min Scan# 3192
Delta R.T. 0.001 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18

Tgt Ion:276	Resp:	29903
Ion Ratio	Lower	Upper
276	100	
138	24.9	20.1 30.1
277	24.8	19.8 29.8





#37

Benzo(b)fluoranthene

Concen: 0.40 ng

RT: 22.802 min Scan# 2349





Delta R.T. 0.001 min

Lab File: BN014483.D

Acq: 29 Apr 2021 10:18

Tgt Ion:	252	Resp:	29339
Ion	Ratio	Lower	Upper
252	100		
253	24.0	19.7	29.5
125	12.7	10.0	15.0

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



#38

Benzo(k)fluoranthene

Concen: 0.41 ng




RT: 22.843 min Scan# 2361


Delta R.T. -0.003 min

Lab File: BN014483.D

Acq: 29 Apr 2021 10:18

Tgt Ion:	252	Resp:	33086
Ion	Ratio	Lower	Upper
252	100		
253	24.1	19.4	29.2
125	12.3	10.0	15.0









#39
Benzo(a)pyrene
Concen: 0.39 ng
RT: 23.370 min Scan# 2515
Delta R.T. 0.001 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252	Resp:	24891
Ion Ratio	Lower	Upper
252	100	
253	25.8	20.8 31.2
125	16.6	12.6 19.0



#40
Dibenzo(a,h)anthracene
Concen: 0.38 ng
RT: 25.701 min Scan# 3196
Delta R.T. 0.001 min
Lab File: BN014483.D
Acq: 29 Apr 2021 10:18

Tgt Ion:278	Resp:	24921
Ion Ratio	Lower	Upper
278	100	
139	17.1	14.2 21.4
279	25.0	21.0 31.4





#41
 Benzo(g,h,i)perylene
 Concen: 0.41 ng
 RT: 26.365 min Scan# 3390
 Delta R.T. 0.001 min
 Lab File: BN014483.D
 Acq: 29 Apr 2021 10:18

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Tgt Ion:	276	Resp:	29483
Ion	Ratio	Lower	Upper
276	100		
277	24.6	19.4	29.2
138	22.4	17.5	26.3

