

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042921\
 Data File : BN014488.D
 Acq On : 29 Apr 2021 13:16
 Operator : CG/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Apr 29 13:47:50 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 29 12:40:58 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.668	152	7471	0.40	ng	0.00
7) Naphthalene-d8	10.442	136	26159	0.40	ng	0.00
13) Acenaphthene-d10	14.295	164	13677	0.40	ng	0.00
19) Phenanthrene-d10	17.043	188	27031	0.40	ng	0.00
29) Chrysene-d12	21.250	240	26484	0.40	ng	0.00
35) Perylene-d12	23.465	264	20674	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.276	112	21840	1.55	ng	0.00
5) Phenol-d6	6.839	99	26839	1.54	ng	0.00
8) Nitrobenzene-d5	8.807	82	22790	1.31	ng	0.00
11) 2-Methylnaphthalene-d10	12.035	152	88717	2.25	ng	0.00
14) 2,4,6-Tribromophenol	15.793	330	6248	1.39	ng	0.00
15) 2-Fluorobiphenyl	12.925	172	88000	1.71	ng	0.00
27) Fluoranthene-d10	19.084	212	163140	2.20	ng	0.00
31) Terphenyl-d14	19.690	244	98977	1.87	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.231	88	13715	1.65	ng	97
3) n-Nitrosodimethylamine	3.569	42	6852	1.70	ng	# 95
6) bis(2-Chloroethyl)ether	7.104	93	30097	1.69	ng	99
9) Naphthalene	10.492	128	113932	1.76	ng	99
10) Hexachlorobutadiene	10.784	225	21533	1.72	ng	# 99
12) 2-Methylnaphthalene	12.111	142	74742	1.76	ng	99
16) Acenaphthylene	14.016	152	84606	1.58	ng	99
17) Acenaphthene	14.359	154	64556	1.71	ng	98
18) Fluorene	15.349	166	80386	1.70	ng	100
20) 4,6-Dinitro-2-methylph...	15.437	198	4098	1.73	ng	# 25
21) 4-Bromophenyl-phenylether	16.252	248	24767	1.75	ng	99
22) Hexachlorobenzene	16.362	284	32503	1.78	ng	100
23) Atrazine	16.520	200	12072	1.35	ng	97
24) Pentachlorophenol	16.702	266	5485	1.47	ng	98
25) Phenanthrene	17.092	178	132305	1.76	ng	100
26) Anthracene	17.177	178	101765	1.72	ng	100
28) Fluoranthene	19.114	202	149588	1.80	ng	99
30) Pyrene	19.481	202	140639	1.80	ng	100
32) Benzo(a)anthracene	21.229	228	118307	1.62	ng	99
33) Chrysene	21.282	228	147553	1.86	ng	100
34) Bis(2-ethylhexyl)phtha...	21.176	149	28836	0.93	ng	99
36) Indeno(1,2,3-cd)pyrene	25.687	276	149687	1.88	ng	100
37) Benzo(b)fluoranthene	22.801	252	134582	1.83	ng	95
38) Benzo(k)fluoranthene	22.846	252	152105	2.00	ng	95
39) Benzo(a)pyrene	23.369	252	116517	1.76	ng	# 93
40) Dibenzo(a,h)anthracene	25.704	278	127329	1.92	ng	96
41) Benzo(g,h,i)perylene	26.368	276	134152	1.91	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042921\
Data File : BN014488.D
Acq On : 29 Apr 2021 13:16
Operator : CG/JU
Sample : SSTDICC1.6
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Quant Time: Apr 29 13:47:50 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 29 12:40:58 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.668 min Scan# 607
 Delta R.T. 0.000 min
 Lab File: BN014488.D
 Acq: 29 Apr 2021 13:16

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Tgt Ion:152 Resp: 7471




Ion	Ratio	Lower	Upper
152	100		
150	157.4	126.6	189.8
115	59.1	47.7	71.5




#2
 1,4-Dioxane
 Concen: 1.65 ng
 RT: 3.231 min Scan# 56
 Delta R.T. 0.000 min
 Lab File: BN014488.D
 Acq: 29 Apr 2021 13:16

Tgt Ion: 88 Resp: 13715

Ion	Ratio	Lower	Upper
88	100		
58	66.7	55.4	83.0
43	24.9	21.6	32.4









#3
n-Nitrosodimethylamine
Concen: 1.70 ng
RT: 3.569 min Scan# 98
Delta R.T. -0.032 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 42 Resp: 6852
Ion Ratio Lower Upper
42 100
74 144.2 120.3 180.5
44 4.8 6.7 10.1#



#4
2-Fluorophenol
Concen: 1.55 ng
RT: 5.276 min Scan# 310
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16

Tgt Ion: 112 Resp: 21840
Ion Ratio Lower Upper
112 100
64 52.6 42.4 63.6
63 27.5 22.2 33.4









#5
Phenol-d6
Concen: 1.54 ng
RT: 6.839 min Scan# 504
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 99 Resp: 26839
Ion Ratio Lower Upper
99 100
42 17.2 13.8 20.8
71 32.1 26.0 39.0



#6
bis(2-Chloroethyl)ether
Concen: 1.69 ng
RT: 7.104 min Scan# 537
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16

Tgt Ion: 93 Resp: 30097
Ion Ratio Lower Upper
93 100
63 71.9 58.3 87.5
95 33.4 27.0 40.6









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.442 min Scan# 855
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	136	Resp:	26159
Ion Ratio	Lower	Upper	
136	100		
137	11.0	9.0	13.6
54	5.7	4.8	7.2
68	5.6	4.6	6.8



#8
Nitrobenzene-d5
Concen: 1.31 ng
RT: 8.807 min Scan# 726
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16

Tgt Ion:	82	Resp:	22790
Ion Ratio	Lower	Upper	
82	100		
128	33.8	26.0	39.0
54	47.5	39.5	59.3









#9
Naphthalene
Concen: 1.76 ng
RT: 10.492 min Scan# 859
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:128 Resp: 113932
Ion Ratio Lower Upper
128 100
129 11.0 9.0 13.6
127 12.6 10.2 15.4



#10
Hexachlorobutadiene
Concen: 1.72 ng
RT: 10.784 min Scan# 882
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16

Tgt Ion:225 Resp: 21533
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.0 51.8 77.6









#11
2-Methylnaphthalene-d10
Concen: 2.25 ng
RT: 12.035 min Scan# 1058
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 88717
Ion Ratio Lower Upper
152 100
151 20.3 16.3 24.5



#12
2-Methylnaphthalene
Concen: 1.76 ng
RT: 12.111 min Scan# 1075
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16

Tgt Ion:142 Resp: 74742
Ion Ratio Lower Upper
142 100
141 87.6 70.5 105.7
115 33.7 27.5 41.3









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.295 min Scan# 1305
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:164 Resp: 13677
Ion Ratio Lower Upper
164 100
162 102.9 82.4 123.6
160 49.7 40.5 60.7



#14
2,4,6-Tribromophenol
Concen: 1.39 ng
RT: 15.793 min Scan# 1423
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16

Tgt Ion:330 Resp: 6248
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 35.4 28.5 42.7









#15
2-Fluorobiphenyl
Concen: 1.71 ng
RT: 12.925 min Scan# 1197
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:172 Resp: 88000
Ion Ratio Lower Upper
172 100
171 33.3 26.6 40.0
170 21.8 18.2 27.2



#16
Acenaphthylene
Concen: 1.58 ng
RT: 14.016 min Scan# 1283
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16

Tgt Ion:152 Resp: 84606
Ion Ratio Lower Upper
152 100
151 19.2 15.6 23.4
153 13.0 10.7 16.1









#17
Acenaphthene
Concen: 1.71 ng
RT: 14.359 min Scan# 1310
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:154 Resp: 64556
Ion Ratio Lower Upper
154 100
153 110.9 90.3 135.5
152 56.8 47.3 70.9



#18
Fluorene
Concen: 1.70 ng
RT: 15.349 min Scan# 1388
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16

Tgt Ion:166 Resp: 80386
Ion Ratio Lower Upper
166 100
165 97.2 78.2 117.2
167 13.4 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.043 min Scan# 1525
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:188	Resp:	27031
Ion Ratio	Lower	Upper
188	100	
94	0.0	0.0
80	11.3	8.8



#20
4,6-Dinitro-2-methylphenol
Concen: 1.73 ng
RT: 15.437 min Scan# 1395
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16

Tgt Ion:198	Resp:	4098
Ion Ratio	Lower	Upper
198	100	
51	36.0	116.0
105	26.1	45.3









#21
4-Bromophenyl-phenylether
Concen: 1.75 ng
RT: 16.252 min Scan# 1460
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:248 Resp: 24767
Ion Ratio Lower Upper
248 100
250 105.5 84.6 126.8
141 30.7 26.7 40.1



#22
Hexachlorobenzene
Concen: 1.78 ng
RT: 16.362 min Scan# 1469
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16

Tgt Ion:284 Resp: 32503
Ion Ratio Lower Upper
284 100
142 34.1 27.0 40.6
249 29.7 23.6 35.4









#23
Atrazine
Concen: 1.35 ng
RT: 16.520 min Scan# 1482
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:200 Resp: 12072
Ion Ratio Lower Upper
200 100
173 21.8 20.0 30.0
215 50.0 39.0 58.6



#24
Pentachlorophenol
Concen: 1.47 ng
RT: 16.702 min Scan# 1497
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16

Tgt Ion:266 Resp: 5485
Ion Ratio Lower Upper
266 100
264 63.2 49.9 74.9
268 64.0 49.9 74.9









#25
Phenanthrene
Concen: 1.76 ng
RT: 17.092 min Scan# 1529
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:178 Resp: 132305
Ion Ratio Lower Upper
178 100
176 18.8 15.1 22.7
179 15.2 12.2 18.4



#26
Anthracene
Concen: 1.72 ng
RT: 17.177 min Scan# 1536
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16

Tgt Ion:178 Resp: 101765
Ion Ratio Lower Upper
178 100
176 18.3 14.6 22.0
179 15.4 12.2 18.2









#27
Fluoranthene-d10
Concen: 2.20 ng
RT: 19.084 min Scan# 1820
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:212 Resp: 163140
Ion Ratio Lower Upper
212 100
106 12.7 10.2 15.4
104 6.9 5.6 8.4



#28
Fluoranthene
Concen: 1.80 ng
RT: 19.114 min Scan# 1826
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16

Tgt Ion:202 Resp: 149588
Ion Ratio Lower Upper
202 100
101 10.8 8.7 13.1
203 17.3 13.6 20.4









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.250 min Scan# 2123
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	240	Resp:	26484
Ion Ratio	Lower	Upper	
240	100		
120	7.4	6.3	9.5
236	25.5	20.9	31.3



#30
Pyrene
Concen: 1.80 ng
RT: 19.481 min Scan# 1900
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16

Tgt Ion:	202	Resp:	140639
Ion Ratio	Lower	Upper	
202	100		
200	20.5	16.4	24.6
203	17.4	13.9	20.9









#31
Terphenyl-d14
Concen: 1.87 ng
RT: 19.690 min Scan# 1942
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:244 Resp: 98977
Ion Ratio Lower Upper
244 100
212 7.5 6.3 9.5
122 12.4 10.6 16.0



#32
Benzo(a)anthracene
Concen: 1.62 ng
RT: 21.229 min Scan# 2121
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16

Tgt Ion:228 Resp: 118307
Ion Ratio Lower Upper
228 100
226 27.5 21.8 32.8
229 19.0 15.8 23.6









#33
Chrysene
Concen: 1.86 ng
RT: 21.282 min Scan# 2126
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:228 Resp: 147553
Ion Ratio Lower Upper
228 100
226 29.7 23.9 35.9
229 19.4 15.5 23.3



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.93 ng
RT: 21.176 min Scan# 2116
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16

Tgt Ion:149 Resp: 28836
Ion Ratio Lower Upper
149 100
167 26.4 21.4 32.2
279 3.6 3.3 4.9









#35
Perylene-d12
Concen: 0.40 ng
RT: 23.465 min Scan# 2541
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Tgt Ion:264 Resp: 20674




Ion	Ratio	Lower	Upper
264	100		
260	24.9	19.6	29.4
265	42.7	35.5	53.3




#36
Indeno(1,2,3-cd)pyrene
Concen: 1.88 ng
RT: 25.687 min Scan# 3190
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16

Tgt Ion:276 Resp: 149687

Ion	Ratio	Lower	Upper
276	100		
138	25.0	20.1	30.1
277	24.8	19.8	29.8









#37
Benzo(b)fluoranthene
Concen: 1.83 ng
RT: 22.801 min Scan# 2347
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:252 Resp: 134582
Ion Ratio Lower Upper
252 100
253 22.3 19.7 29.5
125 10.5 10.0 15.0



#38
Benzo(k)fluoranthene
Concen: 2.00 ng
RT: 22.846 min Scan# 2360
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16

Tgt Ion:252 Resp: 152105
Ion Ratio Lower Upper
252 100
253 22.2 19.4 29.2
125 10.1 10.0 15.0









#39
Benzo(a)pyrene
Concen: 1.76 ng
RT: 23.369 min Scan# 2513
Delta R.T. 0.000 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:252 Resp: 116517
Ion Ratio Lower Upper
252 100
253 22.6 20.8 31.2
125 12.4 12.6 19.0#



#40
Dibenzo(a,h)anthracene
Concen: 1.92 ng
RT: 25.704 min Scan# 3195
Delta R.T. 0.003 min
Lab File: BN014488.D
Acq: 29 Apr 2021 13:16

Tgt Ion:278 Resp: 127329
Ion Ratio Lower Upper
278 100
139 16.5 14.2 21.4
279 23.5 21.0 31.4





#41
 Benzo(g,h,i)perylene
 Concen: 1.91 ng
 RT: 26.368 min Scan# 3389
 Delta R.T. 0.003 min
 Lab File: BN014488.D
 Acq: 29 Apr 2021 13:16

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Tgt Ion:276 Resp: 134152
 Ion Ratio Lower Upper
 276 100
 277 23.7 19.4 29.2
 138 21.6 17.5 26.3

