

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042922\
 Data File : BN019587.D
 Acq On : 29 Apr 2022 12:18
 Operator : CG/JU
 Sample : N2469-02DL 5X
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BFFF9DL

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 05/04/2022
 Supervised By :Yogesh Patel 05/05/2022

Quant Time: Apr 30 03:18:05 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN042922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 28 13:24:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.880	152	5209	0.400	ng/ul	0.00
4) Naphthalene-d8	10.665	136	14646	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.496	164	7537	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.232	188	14596	0.400	ng/ul	0.00
17) Chrysene-d12	21.417	240	12942	0.400	ng/ul	0.00
23) Perylene-d12	23.767	264	12192	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.348	96	3766	0.611	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.255	152	973	0.044	ng/ul	0.00
18) Fluoranthene-d10	19.260	212	1993	0.049	ng/ul	0.00
Target Compounds						
					Qvalue	
10) Acenaphthylene	14.213	152	3979	0.107	ng/ul	96
12) Fluorene	15.541	166	679	0.021	ng/ul#	87
15) Phenanthrene	17.275	178	16925	0.315	ng/ul	99
16) Anthracene	17.363	178	4370	0.100	ng/ul	97
19) Fluoranthene	19.288	202	57767	0.975	ng/ul	98
20) Pyrene	19.651	202	60806	1.028	ng/ul	100
21) Benzo(a)anthracene	21.399	228	24629	0.509	ng/ul	96
22) Chrysene	21.452	228	40143	0.704	ng/ul	98
24) Benzo(b)fluoranthene	23.054	252	66276m	1.071	ng/ul	
25) Benzo(k)fluoranthene	23.098	252	16894m	0.283	ng/ul	
26) Benzo(a)pyrene	23.662	252	33468	0.635	ng/ul#	79
27) Indeno(1,2,3-cd)pyrene	26.184	276	30456	0.463	ng/ul#	98
28) Dibenzo(a,h)anthracene	26.191	278	6727	0.122	ng/ul#	74
29) Benzo(g,h,i)perylene	26.926	276	27107	0.451	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042922\
Data File : BN019587.D
Acq On : 29 Apr 2022 12:18
Operator : CG/JU
Sample : N2469-02DL 5X
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BFFF9DL

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022

Quant Time: Apr 30 03:18:05 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN042922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 28 13:24:26 2022
Response via : Initial Calibration

