


Data Path : Z:\HPCHEM1\BNA_N\Data\BN043018\
 Data File : BN000992.D
 Acq On : 30 Apr 2018 14:24
 Operator : JU/SJ
 Sample : DFTPP54
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP54

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN042418MA.M
 Title : SVOA CALIBRATION
 Last Update : Tue May 01 01:47:00 2018



AutoFind: Scans 2459, 2460, 2461; Background Corrected with Scan 2454

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.4	58189	PASS
68	69	0.00	2	1.7	1225	PASS
69	198	0.00	100	44.6	71386	PASS
70	69	0.00	2	0.5	363	PASS
127	198	10	80	51.4	82269	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	160018	PASS
199	198	5	9	7.0	11229	PASS
275	198	10	60	25.2	40344	PASS
365	198	1	100	3.0	4821	PASS
441	443	0.01	100	79.7	20284	PASS
442	198	50	100	80.9	129517	PASS
443	442	15	24	19.6	25447	PASS

m/z	Abundance
38.10	537.0
39.10	3730.0
40.00	523.0
44.00	1438.0
49.10	405.0
50.00	13608.0
51.10	51424.0
52.00	2799.0
55.00	336.0
56.00	1898.0
57.00	3840.0
61.00	641.0
62.00	788.0
63.00	2466.0
63.90	338.0
65.00	1325.0
68.00	1096.0
69.00	62984.0
70.00	322.0
73.10	571.0
74.00	6165.0
75.00	10219.0
76.10	3466.0
77.10	74944.0
78.00	5425.0
79.00	4582.0
80.00	3531.0
81.00	5092.0
82.00	1175.0
83.00	1229.0
85.00	974.0
86.00	1309.0
87.00	588.0
91.00	1318.0
92.00	1178.0
93.00	7804.0
94.00	572.0
96.00	428.0
98.00	5803.0
99.00	5491.0
100.00	458.0
101.00	2776.0
103.00	898.0
104.00	1697.0
105.00	1618.0
106.10	597.0
107.00	20040.0
108.00	3284.0
109.00	774.0
110.00	37928.0
111.00	6186.0
112.00	768.0
116.00	1166.0
117.00	16520.0
118.00	1324.0
120.00	310.0
122.00	1322.0
123.00	1931.0
124.00	1113.0
125.00	1027.0
127.00	67792.0
128.00	5434.0
129.00	25864.0
130.00	2490.0
131.10	458.0
133.90	805.0
135.00	2071.0
136.00	869.0
137.00	1255.0
137.90	307.0
140.00	398.0
141.00	3465.0
142.00	1084.0
143.00	737.0
145.90	534.0
147.00	1740.0
148.00	3941.0
149.00	1070.0
151.00	508.0
151.90	309.0
153.00	1084.0
154.00	757.0
155.00	1745.0
156.00	2692.0
157.10	508.0
157.90	570.0
159.00	524.0
160.00	1021.0
161.00	1722.0
162.00	457.0
165.00	1276.0

Instrument :
BNA_N
ClientSampleId :
DFTPP54

166.00	912.0
167.00	6635.0
168.00	4728.0
169.10	640.0
171.90	655.0
173.00	844.0
174.00	1348.0
175.00	2595.0
176.10	770.0
177.00	1296.0
178.00	426.0
179.00	4738.0
180.00	3484.0
181.00	1464.0
183.90	307.0
185.10	2454.0
186.10	17208.0
187.00	4999.0
188.10	526.0
189.00	989.0
191.00	529.0
192.00	1560.0
193.00	1522.0
194.00	359.0
196.00	3557.0
198.00	123192.0
199.00	9114.0
200.00	739.0
201.30	633.0
203.00	935.0
204.00	4029.0
205.10	7074.0
206.10	31008.0
207.00	5005.0
208.00	1168.0
209.00	443.0
210.20	533.0
211.10	1222.0
215.00	354.0
216.10	581.0
217.00	7793.0
218.00	1058.0
221.00	4968.0
221.80	1045.0
223.00	1892.0
224.10	16784.0
225.10	4285.0
226.10	527.0
227.00	6821.0
228.00	1037.0
229.00	1687.0
231.00	489.0
234.00	328.0
235.00	547.0
235.90	421.0
237.00	723.0
239.00	347.0
241.00	364.0
242.00	717.0
243.00	953.0
244.10	12442.0
245.10	1514.0
246.00	2568.0
247.00	514.0
249.00	469.0
255.00	62784.0
256.10	9449.0
257.00	753.0
258.00	3644.0
259.00	606.0
265.00	1368.0
273.00	1948.0
274.10	5064.0
275.10	28048.0
276.10	4039.0
277.00	2473.0
278.00	531.0
285.00	462.0
293.00	452.0
296.00	7097.0
297.00	987.0
303.00	882.0
314.00	447.0
315.00	913.0
316.10	448.0
323.10	2374.0
324.10	399.0
326.90	420.0
334.00	1485.0
335.00	456.0
346.00	490.0
352.10	615.0
353.00	427.0
354.10	775.0
365.00	2895.0

Instrument :
BNA_N
ClientSampleId :
DFTPP54

366.00	433.0
372.10	1185.0
373.00	308.0
382.90	345.0
402.00	517.0
403.00	702.0
421.00	626.0
422.00	573.0
423.00	3864.0
424.10	1340.0
441.10	10790.0
442.10	70024.0
443.10	14671.0
444.10	1433.0

Instrument :
BNA_N
ClientSampleId :
DFTPP54

m/z	Abundance
38.00	692.0
39.10	4718.0
40.00	454.0
44.00	1308.0
49.10	528.0
50.00	17024.0
51.10	68504.0
52.00	3627.0
55.00	427.0
56.00	1853.0
57.00	4676.0
61.00	991.0
62.00	1071.0
63.10	2836.0
64.00	429.0
65.10	1608.0
68.10	1377.0
69.00	83040.0
70.00	392.0
73.00	697.0
74.00	8424.0
75.00	13291.0
76.10	4687.0
77.10	97712.0
78.10	7072.0
79.00	6214.0
80.00	4547.0
81.00	6742.0
82.00	1636.0
83.00	1690.0
85.00	1302.0
86.00	2004.0
87.00	842.0
88.00	409.0
91.00	1429.0
92.00	1702.0
93.00	10204.0
94.00	651.0
96.00	563.0
98.00	8114.0
99.00	6796.0
100.00	504.0
101.00	3685.0
103.10	1088.0
104.00	2357.0
105.00	2486.0
106.00	683.0
107.00	28104.0
108.00	4516.0
109.00	720.0
110.00	52592.0
111.00	8065.0
112.00	966.0
113.00	410.0
116.00	1472.0
117.00	22536.0
118.00	1870.0
119.90	335.0
122.00	1715.0
123.00	2774.0
124.00	1255.0
125.00	1137.0
127.00	95864.0
128.00	7515.0
129.00	36480.0
130.00	2939.0
131.00	697.0
132.00	345.0
134.00	1056.0
135.00	2880.0
136.00	1182.0
137.10	1773.0
137.90	344.0
140.00	486.0
141.00	4638.0
142.00	1525.0
142.90	995.0
146.00	843.0
147.00	2437.0
148.00	5636.0
149.00	1248.0
150.10	348.0
151.00	575.0
151.80	357.0
153.00	1469.0
154.10	1089.0
155.00	2754.0
156.00	4177.0
157.10	751.0
157.90	730.0
159.00	657.0

Instrument :
BNA_N
ClientSampleId :
DFTPP54

160.00	1483.0
161.00	2192.0
162.00	603.0
165.00	1683.0
166.00	1463.0
167.00	9336.0
168.00	5831.0
169.00	832.0
170.00	362.0
170.90	497.0
172.00	837.0
173.00	1093.0
174.00	2095.0
175.00	3603.0
176.00	1183.0
177.00	1639.0
178.00	636.0
179.00	6747.0
180.00	5177.0
181.00	2314.0
182.00	393.0
184.00	570.0
185.00	3449.0
186.00	26248.0
187.00	7471.0
188.00	860.0
189.00	1756.0
191.00	817.0
192.00	2266.0
193.00	2663.0
194.00	519.0
195.10	384.0
196.10	5578.0
198.00	186560.0
199.00	13267.0
200.00	1045.0
201.40	762.0
203.00	1168.0
204.00	6773.0
205.10	11331.0
206.10	49064.0
207.10	7012.0
208.10	1617.0
209.00	653.0
210.10	763.0
211.10	1785.0
215.00	528.0
216.10	1038.0
217.00	12610.0
218.00	1705.0
221.00	7034.0
221.70	1731.0
223.00	2683.0
224.10	26352.0
225.10	6888.0
226.10	842.0
227.00	10218.0
228.00	1554.0
229.00	2333.0
230.00	304.0
231.10	925.0
234.00	613.0
235.00	755.0
236.00	535.0
237.00	1005.0
239.00	328.0
239.90	381.0
241.10	592.0
242.00	1381.0
243.10	1518.0
244.10	20224.0
245.10	2843.0
246.00	3775.0
247.00	818.0
249.00	821.0
253.00	482.0
255.00	103840.0
256.10	14905.0
257.10	1135.0
258.00	5597.0
259.00	995.0
265.10	2506.0
266.00	333.0
273.00	2980.0
274.10	8503.0
275.10	46288.0
276.10	6708.0
277.00	3870.0
278.00	807.0
283.00	428.0
284.10	357.0
285.10	588.0
293.00	911.0
294.00	341.0
295.20	340.0

Instrument :

BNA_N

ClientSampleId :

DFTPP54

296.00	12942.0
297.00	1724.0
303.10	1626.0
304.00	424.0
314.10	732.0
315.00	1468.0
316.10	724.0
321.00	386.0
323.10	4049.0
324.10	725.0
327.00	873.0
328.00	389.0
333.00	376.0
334.10	2769.0
335.10	898.0
341.10	487.0
346.00	948.0
352.00	1062.0
353.00	899.0
354.10	1470.0
365.00	5565.0
366.00	869.0
371.10	387.0
372.00	2263.0
373.10	642.0
383.00	570.0
402.00	984.0
403.10	1358.0
404.10	364.0
421.00	1137.0
422.10	1199.0
423.10	8072.0
424.10	1946.0
441.10	23088.0
442.10	145920.0
443.10	29096.0
444.10	2772.0

Instrument :
BNA_N
ClientSampleId :
DFTPP54

m/z	Abundance
38.00	744.0
39.00	3870.0
40.00	472.0
44.00	1352.0
49.00	384.0
50.00	13821.0
51.10	54640.0
52.00	2676.0
55.00	412.0
56.00	1770.0
57.00	3769.0
61.00	764.0
62.00	735.0
63.00	2409.0
64.00	320.0
65.00	1283.0
68.10	1203.0
69.00	68136.0
70.00	376.0
73.00	560.0
74.00	7219.0
75.00	11374.0
76.10	3770.0
77.10	81656.0
78.10	5460.0
79.00	5366.0
80.00	4033.0
81.00	5687.0
82.00	1459.0
83.00	1188.0
85.00	1241.0
86.00	1352.0
87.00	724.0
88.00	301.0
91.00	1350.0
92.00	1247.0
93.00	8751.0
94.00	693.0
96.00	399.0
98.00	6577.0
99.00	5654.0
99.90	477.0
101.00	3100.0
103.00	1044.0
104.00	1817.0
105.00	1791.0
106.10	665.0
107.00	24112.0
108.00	3881.0
109.10	675.0
110.00	44872.0
111.00	6426.0
112.00	827.0
116.00	1264.0
117.00	19040.0
118.00	1519.0
119.00	310.0
120.00	327.0
122.00	1344.0
123.00	2264.0
124.00	1125.0
125.00	937.0
127.00	83152.0
128.00	5636.0
129.00	30816.0
130.00	2558.0
130.90	383.0
132.00	394.0
134.00	867.0
135.00	2378.0
136.00	1054.0
137.00	1470.0
140.00	324.0
141.00	4131.0
142.00	1368.0
143.00	853.0
145.90	679.0
147.00	2152.0
148.00	4927.0
149.00	902.0
150.00	344.0
151.00	484.0
152.90	1301.0
154.00	964.0
155.00	2286.0
156.10	3601.0
157.10	686.0
157.90	682.0
159.00	688.0
160.00	1362.0
161.00	1731.0

Instrument :
BNA_N
ClientSampleId :
DFTPP54

162.00	592.0
165.00	1590.0
166.10	1242.0
167.00	8477.0
168.00	4793.0
169.00	809.0
170.90	353.0
172.00	615.0
172.90	856.0
174.00	1678.0
175.00	3296.0
176.00	765.0
177.00	1564.0
178.00	610.0
179.00	5937.0
180.00	4363.0
181.00	1831.0
183.90	516.0
185.00	2934.0
186.10	23904.0
187.00	7005.0
188.10	690.0
189.00	1430.0
191.10	773.0
192.00	2000.0
193.00	2321.0
194.00	469.0
196.10	4673.0
198.00	170304.0
199.00	11306.0
200.00	963.0
201.40	727.0
203.00	1173.0
204.10	5951.0
205.10	10147.0
206.10	43304.0
207.00	6105.0
208.00	1564.0
209.10	475.0
209.90	520.0
211.00	1710.0
215.00	469.0
216.10	1009.0
217.00	11682.0
218.00	1699.0
221.10	6761.0
221.80	1494.0
223.00	2750.0
224.10	24112.0
225.10	6212.0
226.00	789.0
227.00	10358.0
228.00	1352.0
229.00	2153.0
231.00	835.0
234.00	560.0
235.00	670.0
235.90	460.0
237.00	996.0
239.00	490.0
240.00	333.0
241.00	566.0
242.00	1244.0
243.10	1428.0
244.10	18784.0
245.10	2718.0
246.10	3572.0
247.00	686.0
249.00	811.0
253.00	500.0
255.00	99776.0
256.00	14186.0
257.00	1217.0
258.00	5921.0
259.10	805.0
265.00	2271.0
266.10	312.0
272.20	343.0
273.00	2656.0
274.10	8803.0
275.10	46696.0
276.10	6062.0
277.10	3902.0
278.00	650.0
283.00	379.0
284.00	306.0
285.10	658.0
293.00	1028.0
294.10	309.0
296.00	12889.0
297.00	1815.0
302.10	318.0
303.10	1518.0
304.10	398.0
314.00	579.0

Instrument :
BNA_N
ClientSampleId :
DFTPP54

315.00	1324.0
316.10	757.0
321.00	421.0
323.10	4542.0
324.10	798.0
327.00	770.0
328.00	481.0
332.10	416.0
333.00	490.0
334.10	2773.0
335.00	832.0
341.00	424.0
346.00	1222.0
352.00	1365.0
353.10	848.0
354.10	1665.0
355.10	386.0
365.00	6004.0
366.10	847.0
370.90	355.0
372.00	2502.0
373.10	648.0
383.00	641.0
402.00	1048.0
403.10	1594.0
404.10	538.0
421.00	1324.0
422.10	1430.0
423.10	9619.0
424.10	2477.0
440.10	381.0
441.10	26976.0
442.10	172608.0
443.10	32576.0
444.10	2943.0

Instrument :
BNA_N
ClientSampleId :
DFTPP54

m/z	Abundance
39.90	334.0
44.00	1366.0
207.00	602.0

Instrument :
BNA_N
ClientSampleId :
DFTPP54