

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN043021\
 Data File : BN014505.D
 Acq On : 30 Apr 2021 16:30
 Operator : CG/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: Apr 30 17:06:39 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN043021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 30 15:34:39 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.628	152	9392	0.40	ng	0.00
7) Naphthalene-d8	10.404	136	31146	0.40	ng	0.00
13) Acenaphthene-d10	14.270	164	16478	0.40	ng	0.00
19) Phenanthrene-d10	17.019	188	31158	0.40	ng	0.00
29) Chrysene-d12	21.208	240	30686	0.40	ng	#-0.01
35) Perylene-d12	23.414	264	23053	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.228	112	62154	3.72	ng	0.00
5) Phenol-d6	6.798	99	78012	3.80	ng	0.00
8) Nitrobenzene-d5	8.769	82	65209	4.07	ng	0.00
11) 2-Methylnaphthalene-d10	11.995	152	217224	3.42	ng	0.00
14) 2,4,6-Tribromophenol	15.767	330	21189	5.04	ng	0.00
15) 2-Fluorobiphenyl	12.887	172	211136	3.27	ng	0.00
27) Fluoranthene-d10	19.054	212	390623	3.58	ng	0.00
31) Terphenyl-d14	19.660	244	227859	3.23	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.174	88	33915	3.25	ng	89
3) n-Nitrosodimethylamine	3.496	42	19605	4.18	ng	# 57
6) bis(2-Chloroethyl)ether	7.056	93	73837	3.23	ng	98
9) Naphthalene	10.454	128	271963	3.31	ng	99
10) Hexachlorobutadiene	10.746	225	52354	3.33	ng	# 100
12) 2-Methylnaphthalene	12.071	142	178949	3.38	ng	100
16) Acenaphthylene	13.978	152	233220	3.81	ng	100
17) Acenaphthene	14.334	154	158791	3.39	ng	97
18) Fluorene	15.323	166	195126	3.41	ng	100
20) 4,6-Dinitro-2-methylph...	15.399	198	13128	5.15	ng	# 1
21) 4-Bromophenyl-phenylether	16.216	248	60930	3.63	ng	96
22) Hexachlorobenzene	16.325	284	73806	3.23	ng	100
23) Atrazine	16.483	200	36154	4.61	ng	98
24) Pentachlorophenol	16.666	266	19449	5.20	ng	# 41
25) Phenanthrene	17.055	178	307549	3.39	ng	100
26) Anthracene	17.141	178	257608	3.75	ng	100
28) Fluoranthene	19.084	202	344104	3.53	ng	99
30) Pyrene	19.447	202	334978	3.29	ng	100
32) Benzo(a)anthracene	21.197	228	298330	3.63	ng	99
33) Chrysene	21.250	228	337034	3.27	ng	99
34) Bis(2-ethylhexyl)phtha...	21.144	149	95184	4.26	ng	97
36) Indeno(1,2,3-cd)pyrene	25.611	276	323497	3.46	ng	99
37) Benzo(b)fluoranthene	22.753	252	306819	3.47	ng	# 94
38) Benzo(k)fluoranthene	22.798	252	330028	3.41	ng	95
39) Benzo(a)pyrene	23.315	252	267182	3.52	ng	# 88
40) Dibenzo(a,h)anthracene	25.629	278	271563	3.46	ng	96
41) Benzo(g,h,i)perylene	26.282	276	297296	3.41	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN043021\
Data File : BN014505.D
Acq On : 30 Apr 2021 16:30
Operator : CG/JU
Sample : SSTDIC3.2
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Quant Time: Apr 30 17:06:39 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN043021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 30 15:34:39 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.628 min Scan# 602
 Delta R.T. -0.000 min
 Lab File: BN014505.D
 Acq: 30 Apr 2021 16:30




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:	152	Resp:	9392
Ion	Ratio	Lower	Upper
152	100		
150	155.5	123.8	185.8
115	57.5	47.0	70.4



1,4-Dioxane
 Concen: 3.25 ng
 RT: 3.174 min Scan# 49
 Delta R.T. -0.000 min
 Lab File: BN014505.D
 Acq: 30 Apr 2021 16:30

Tgt Ion:	88	Resp:	33915
Ion	Ratio	Lower	Upper
88	100		
58	64.5	44.4	66.6
43	23.8	15.9	23.9









#3
n-Nitrosodimethylamine
Concen: 4.18 ng
RT: 3.496 min Scan# 89
Delta R.T. -0.056 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	42	Resp:	19605
Ion	Ratio	Lower	Upper
42	100		
74	148.3	175.1	262.7#
44	5.4	7.5	11.3#



#4
2-Fluorophenol
Concen: 3.72 ng
RT: 5.228 min Scan# 304
Delta R.T. -0.000 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30

Tgt Ion:	112	Resp:	62154
Ion	Ratio	Lower	Upper
112	100		
64	51.6	41.8	62.6
63	27.0	22.0	33.0









Phenol-d6
 Concen: 3.80 ng
 RT: 6.798 min Scan# 499
 Delta R.T. -0.000 min
 Lab File: BN014505.D
 Acq: 30 Apr 2021 16:30




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Tgt Ion:	99	Resp:	78012
Ion	Ratio	Lower	Upper
99	100		
42	16.4	13.2	19.8
71	31.8	26.2	39.2



bis(2-Chloroethyl)ether
 Concen: 3.23 ng
 RT: 7.056 min Scan# 531
 Delta R.T. -0.008 min
 Lab File: BN014505.D
 Acq: 30 Apr 2021 16:30


Tgt Ion:	93	Resp:	73837
Ion	Ratio	Lower	Upper
93	100		
63	70.2	58.0	87.0
95	33.2	27.0	40.6






Abundance Scan 852 (10.404 min): BN014502.D\data.ms (-8457-)

Naphthalene-d8
Concen: 0.40 ng
RT: 10.404 min Scan# 852
Delta R.T. -0.000 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2






Tgt Ion:	136	Resp:	31146
Ion Ratio	Lower	Upper	
136	100		
137	11.0	8.9	13.3
54	4.7	3.8	5.8
68	4.5	3.8	5.6




Abundance Scan 723 (8.769 min): BN014502.D\data.ms (-7207-)

Nitrobenzene-d5
Concen: 4.07 ng
RT: 8.769 min Scan# 723
Delta R.T. -0.000 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30

Tgt Ion:	82	Resp:	65209
Ion Ratio	Lower	Upper	
82	100		
128	40.0	29.5	44.3
54	42.0	36.9	55.3









Naphthalene
Concen: 3.31 ng
RT: 10.454 min Scan# 856
Delta R.T. -0.000 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:	128	Resp:	271963
Ion	Ratio	Lower	Upper
128	100		
129	11.0	9.0	13.4
127	12.4	10.4	15.6




Hexachlorobutadiene
Concen: 3.33 ng
RT: 10.746 min Scan# 879
Delta R.T. -0.000 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30

Tgt Ion:	225	Resp:	52354
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.0	50.9	76.3



Abundance Scan 1050 (12.000 min): BN014502.D\data.ms (-1034) (-)





2-Methylnaphthalene-d10
Concen: 3.42 ng
RT: 11.995 min Scan# 1049
Delta R.T. -0.005 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30


Tgt Ion:152 Resp: 217224
Ion Ratio Lower Upper
152 100
151 20.1 16.2 24.2

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Abundance Scan 1049 (11.995 min): BN014505.D\data.ms



Abundance Scan 1049 (11.995 min): BN014505.D\data.ms (-1005) (-)





Abundance Scan 1067 (12.075 min): BN014502.D\data.ms (-1036) (-)




2-Methylnaphthalene
Concen: 3.38 ng
RT: 12.071 min Scan# 1066
Delta R.T. -0.005 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30

Tgt Ion:142 Resp: 178949
Ion Ratio Lower Upper
142 100
141 87.6 70.2 105.4
115 33.3 26.7 40.1

Abundance Scan 1066 (12.071 min): BN014505.D\data.ms




Abundance Scan 1066 (12.071 min): BN014505.D\data.ms (-1045) (-)






Abundance Scan 1303 (14.270 min): BN014502.D\data.ms (-1295) (-)

Acenaphthene-d10
Concen: 0.40 ng
RT: 14.270 min Scan# 1303
Delta R.T. -0.000 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2




Tgt Ion:	164	Resp:	16478
Ion Ratio	Lower	Upper	
164	100		
162	97.7	78.6	118.0
160	44.8	35.8	53.8






Abundance Scan 1421 (15.767 min): BN014502.D\data.ms (-1414) (-)

2,4,6-Tribromophenol
Concen: 5.04 ng
RT: 15.767 min Scan# 1421
Delta R.T. -0.000 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30




Tgt Ion:	330	Resp:	21189
Ion Ratio	Lower	Upper	
330	100		
332	0.0	0.0	0.0
141	33.3	27.0	40.4





Abundance Scan 1194 (12.887 min): BN014502.D\data.ms (-1196) (-)

2-Fluorobiphenyl
Concen: 3.27 ng
RT: 12.887 min Scan# 1194
Delta R.T. -0.000 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30


Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2



Tgt Ion:	172	Resp:	211136
Ion	Ratio	Lower	Upper
172	100		
171	33.3	27.1	40.7
170	21.9	18.3	27.5





Abundance Scan 1194 (12.887 min): BN014505.D\data.ms (-1186) (-)




Abundance Scan 1280 (13.978 min): BN014502.D\data.ms (-1276) (-)

Acenaphthylene
Concen: 3.81 ng
RT: 13.978 min Scan# 1280
Delta R.T. -0.000 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30

Tgt Ion:	152	Resp:	233220
Ion	Ratio	Lower	Upper
152	100		
151	19.3	15.6	23.4
153	13.0	10.5	15.7




Abundance Scan 1280 (13.978 min): BN014505.D\data.ms (-1264) (-)






Abundance Scan 1308 (14.334 min): BN014502.D\data.ms (-1307) (-)

Acenaphthene
Concen: 3.39 ng
RT: 14.334 min Scan# 1308
Delta R.T. -0.000 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2




Tgt Ion:	154	Resp:	158791
Ion	Ratio	Lower	Upper
154	100		
153	110.5	90.2	135.4
152	54.9	46.2	69.2






Abundance Scan 1386 (15.323 min): BN014502.D\data.ms (-1385) (-)

Fluorene
Concen: 3.41 ng
RT: 15.323 min Scan# 1386
Delta R.T. -0.000 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30




Tgt Ion:	166	Resp:	195126
Ion	Ratio	Lower	Upper
166	100		
165	97.5	78.4	117.6
167	13.6	10.7	16.1





Abundance Scan 1523 (17.019 min): BN014502.D\data.ms (-1519) (-)

Phenanthrene-d10
Concen: 0.40 ng
RT: 17.019 min Scan# 1523
Delta R.T. -0.000 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30


Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2



Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	5.8	5.2	7.8






Abundance Scan 1523 (17.019 min): BN014505.D\data.ms (-1507) (-)




Abundance Scan 1394 (15.425 min): BN014502.D\data.ms (-1329) (-)

4,6-Dinitro-2-methylphenol
Concen: 5.15 ng
RT: 15.399 min Scan# 1392
Delta R.T. -0.025 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30


Tgt Ion	Ratio	Lower	Upper
198	100		
51	35.5	865.3	1297.9#
105	35.1	93.6	140.4#



Abundance Scan 1392 (15.399 min): BN014505.D\data.ms (-1386) (-)





Abundance Scan 1457 (16.216 min): BN014502.D\data.ms (-1423) (-)




4-Bromophenyl-phenylether
Concen: 3.63 ng
RT: 16.216 min Scan# 1457
Delta R.T. -0.000 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	248	Resp:	60930
Ion	Ratio	Lower	Upper
248	100		
250	101.2	80.5	120.7
141	38.2	37.2	55.8



Abundance Scan 1457 (16.216 min): BN014505.D\data.ms (-1449) (-)





Abundance Scan 1466 (16.325 min): BN014502.D\data.ms (-1422) (-)




Hexachlorobenzene
Concen: 3.23 ng
RT: 16.325 min Scan# 1466
Delta R.T. -0.000 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30

Tgt Ion:	284	Resp:	73806
Ion	Ratio	Lower	Upper
284	100		
142	33.2	26.6	39.8
249	29.6	23.6	35.4




Abundance Scan 1466 (16.325 min): BN014505.D\data.ms (-1458) (-)






Abundance Scan 1479 (16.483 min): BN014502.D\data.ms (-1423) (-)

Atrazine
Concen: 4.61 ng
RT: 16.483 min Scan# 1479
Delta R.T. -0.000 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2




Tgt Ion	Ratio	Lower	Upper
200	100		
173	26.3	23.0	34.6
215	48.1	38.6	57.8






Abundance Scan 1497 (16.703 min): BN014502.D\data.ms (-1424) (-)


Pentachlorophenol
Concen: 5.20 ng
RT: 16.666 min Scan# 1494
Delta R.T. -0.037 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30



Tgt Ion	Ratio	Lower	Upper
266	100		
264	63.0	24.2	36.4#
268	63.8	25.2	37.8#






Abundance Scan 1526 (17.055 min): BN014502.D\data.ms (-1525) (-)




Phenanthrene
Concen: 3.39 ng
RT: 17.055 min Scan# 1526
Delta R.T. -0.000 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30

Instrument :
BNA_N
ClientSampled :
SSTDICC3.2

Tgt Ion:	178	Resp:	307549
Ion Ratio	Lower	Upper	
178	100		
176	18.8	15.2	22.8
179	15.3	12.4	18.6






Abundance Scan 1534 (17.153 min): BN014502.D\data.ms (-1526) (-)



Anthracene
Concen: 3.75 ng
RT: 17.141 min Scan# 1533
Delta R.T. -0.012 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30


Tgt Ion:	178	Resp:	257608
Ion Ratio	Lower	Upper	
178	100		
176	18.4	14.8	22.2
179	15.3	12.2	18.2






Abundance Scan 1814 (19.054 min): BN014502.D\data.ms (-1827) (-)

Fluoranthene-d10
Concen: 3.58 ng
RT: 19.054 min Scan# 1814
Delta R.T. -0.000 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2







Tgt Ion:	212	Resp:	390623
Ion Ratio	Lower	Upper	
212	100		
106	12.4	10.1	15.1
104	6.8	5.4	8.2




Abundance Scan 1820 (19.084 min): BN014502.D\data.ms (-1823) (-)

Fluoranthene
Concen: 3.53 ng
RT: 19.084 min Scan# 1820
Delta R.T. -0.000 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30

Tgt Ion:	202	Resp:	344104
Ion Ratio	Lower	Upper	
202	100		
101	10.5	8.4	12.6
203	17.4	13.5	20.3






Abundance Scan 2120 (21.218 min): BN014502.D\data.ms (-2120) (-)




Chrysene-d12
Concen: 0.40 ng
RT: 21.208 min Scan# 2119
Delta R.T. -0.011 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:240 Resp: 30686
Ion Ratio Lower Upper
240 100
120 13.4 5.8 8.8#
236 27.0 20.6 30.8






Abundance Scan 1893 (19.447 min): BN014502.D\data.ms (-1893) (-)




Pyrene
Concen: 3.29 ng
RT: 19.447 min Scan# 1893
Delta R.T. -0.000 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30

Tgt Ion:202 Resp: 334978
Ion Ratio Lower Upper
202 100
200 20.5 16.3 24.5
203 17.6 14.0 21.0






Abundance Scan 1936 (19.660 min): BN014502.D\data.ms (-1936) (-)




Terphenyl-d14
Concen: 3.23 ng
RT: 19.660 min Scan# 1936
Delta R.T. -0.000 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:	244	Resp:	227859
Ion	Ratio	Lower	Upper
244	100		
212	7.2	6.0	9.0
122	10.8	9.3	13.9






Abundance Scan 2118 (21.197 min): BN014502.D\data.ms (-2118) (-)




Benzo(a)anthracene
Concen: 3.63 ng
RT: 21.197 min Scan# 2118
Delta R.T. -0.000 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30

Tgt Ion:	228	Resp:	298330
Ion	Ratio	Lower	Upper
228	100		
226	26.2	21.7	32.5
229	19.7	15.7	23.5






Abundance Scan 2123 (21.250 min): BN014502.D\data.ms (-2123) (-)




Chrysene
Concen: 3.27 ng
RT: 21.250 min Scan# 2123
Delta R.T. -0.000 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:	228	Resp:	337034
Ion	Ratio	Lower	Upper
228	100		
226	29.1	23.8	35.6
229	19.8	15.6	23.4






Abundance Scan 2113 (21.144 min): BN014502.D\data.ms (-2113) (-)



Bis(2-ethylhexyl)phthalate
Concen: 4.26 ng
RT: 21.144 min Scan# 2113
Delta R.T. -0.000 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30


Tgt Ion:	149	Resp:	95184
Ion	Ratio	Lower	Upper
149	100		
167	27.4	20.5	30.7
279	3.8	2.9	4.3






Abundance Scan 2527 (23.417 min): BN014502.D\data.ms (-2505) (-)

Perylene-d12
Concen: 0.40 ng
RT: 23.414 min Scan# 2526
Delta R.T. -0.003 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2




Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.8	19.5	29.3
265	42.6	35.4	53.0






Abundance Scan 3170 (25.618 min): BN014502.D\data.ms (-3148) (-)


Indeno(1,2,3-cd)pyrene
Concen: 3.46 ng
RT: 25.611 min Scan# 3168
Delta R.T. -0.007 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30



Tgt Ion	Ratio	Lower	Upper
276	100		
138	24.6	18.8	28.2
277	24.8	19.8	29.8





Abundance Scan 2334 (22.757 min): BN014502.D\data.ms (-2275) (-)




Benzo(b)fluoranthene
Concen: 3.47 ng
RT: 22.753 min Scan# 2333
Delta R.T. -0.003 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30

Tgt Ion:	252	Resp:	306819
Ion Ratio	Lower	Upper	
252	100		
253	21.8	19.6	29.4
125	9.8	10.3	15.5#


Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2



Abundance Scan 2333 (22.753 min): BN014505.D\data.ms (-2276) (-)





Abundance Scan 2347 (22.801 min): BN014502.D\data.ms (-2346) (-)




Benzo(k)fluoranthene
Concen: 3.41 ng
RT: 22.798 min Scan# 2346
Delta R.T. -0.003 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30

Tgt Ion:	252	Resp:	330028
Ion Ratio	Lower	Upper	
252	100		
253	22.0	19.7	29.5
125	9.7	9.7	14.5




Abundance Scan 2346 (22.798 min): BN014505.D\data.ms (-2259) (-)





Abundance Scan 2499 (23.322 min): BN014502.D\data.ms (-2441) (-)

Benzo(a)pyrene
Concen: 3.52 ng
RT: 23.315 min Scan# 2497
Delta R.T. -0.007 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30


Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2



Tgt Ion:	252	Resp:	267182
Ion Ratio	Lower	Upper	
252	100		
253	22.1	21.8	32.6
125	11.4	14.9	22.3#





Abundance Scan 2497 (23.315 min): BN014505.D\data.ms (-2441) (-)




Abundance Scan 3175 (25.635 min): BN014502.D\data.ms (-3152) (-)

Dibenzo(a,h)anthracene
Concen: 3.46 ng
RT: 25.629 min Scan# 3173
Delta R.T. -0.007 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30


Tgt Ion:	278	Resp:	271563
Ion Ratio	Lower	Upper	
278	100		
139	16.2	13.9	20.9
279	23.6	20.7	31.1



Abundance Scan 3173 (25.629 min): BN014505.D\data.ms (-3152) (-)




Abundance Scan 3366 (26.289 min): BN014502.D\data.ms (-3341)



Benzo(g,h,i)perylene
Concen: 3.41 ng
RT: 26.282 min Scan# 3364
Delta R.T. -0.007 min
Lab File: BN014505.D
Acq: 30 Apr 2021 16:30

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Abundance Scan 3364 (26.282 min): BN014505.D\data.ms



Tgt Ion:	276	Resp:	297296
Ion	Ratio	Lower	Upper
276	100		
277	23.7	19.6	29.4
138	21.1	17.5	26.3

Abundance Scan 3364 (26.282 min): BN014505.D\data.ms (-3308)

