

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN043021\
 Data File : BN014504.D
 Acq On : 30 Apr 2021 15:55
 Operator : CG/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Apr 30 16:24:04 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN043021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 30 15:34:39 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.628	152	9234	0.40	ng	0.00
7) Naphthalene-d8	10.404	136	32085	0.40	ng	0.00
13) Acenaphthene-d10	14.270	164	16882	0.40	ng	0.00
19) Phenanthrene-d10	17.019	188	31714	0.40	ng	0.00
29) Chrysene-d12	21.218	240	30006	0.40	ng	0.00
35) Perylene-d12	23.417	264	23314	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.228	112	29524	1.80	ng	0.00
5) Phenol-d6	6.798	99	35696	1.77	ng	0.00
8) Nitrobenzene-d5	8.769	82	30821	1.87	ng	0.00
11) 2-Methylnaphthalene-d10	12.000	152	109711	1.68	ng	0.00
14) 2,4,6-Tribromophenol	15.767	330	9374	2.18	ng	0.00
15) 2-Fluorobiphenyl	12.887	172	107612	1.63	ng	0.00
27) Fluoranthene-d10	19.054	212	192131	1.73	ng	0.00
31) Terphenyl-d14	19.660	244	112930	1.63	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.174	88	16989	1.65	ng	# 87
3) n-Nitrosodimethylamine	3.513	42	8578	1.86	ng	# 60
6) bis(2-Chloroethyl)ether	7.064	93	37101	1.65	ng	98
9) Naphthalene	10.454	128	138724	1.64	ng	99
10) Hexachlorobutadiene	10.746	225	26859	1.66	ng	# 99
12) 2-Methylnaphthalene	12.071	142	91239	1.67	ng	100
16) Acenaphthylene	13.978	152	107994	1.72	ng	100
17) Acenaphthene	14.334	154	78786	1.64	ng	99
18) Fluorene	15.323	166	97035	1.65	ng	100
20) 4,6-Dinitro-2-methylph...	15.399	198	5132	1.98	ng	# 1
21) 4-Bromophenyl-phenylether	16.216	248	30077	1.76	ng	98
22) Hexachlorobenzene	16.325	284	37876	1.63	ng	100
23) Atrazine	16.484	200	16466	2.06	ng	97
24) Pentachlorophenol	16.678	266	8071	2.12	ng	# 41
25) Phenanthrene	17.055	178	155604	1.69	ng	99
26) Anthracene	17.141	178	122434	1.75	ng	100
28) Fluoranthene	19.084	202	172068	1.74	ng	99
30) Pyrene	19.447	202	164017	1.65	ng	100
32) Benzo(a)anthracene	21.197	228	143733	1.79	ng	100
33) Chrysene	21.250	228	160338	1.59	ng	99
34) Bis(2-ethylhexyl)phtha...	21.144	149	43326	1.98	ng	97
36) Indeno(1,2,3-cd)pyrene	25.615	276	155814	1.65	ng	99
37) Benzo(b)fluoranthene	22.757	252	144724	1.62	ng	# 94
38) Benzo(k)fluoranthene	22.798	252	169414	1.73	ng	95
39) Benzo(a)pyrene	23.318	252	128050	1.67	ng	# 90
40) Dibenzo(a,h)anthracene	25.629	278	129952	1.64	ng	97
41) Benzo(g,h,i)perylene	26.286	276	145898	1.66	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN043021\
Data File : BN014504.D
Acq On : 30 Apr 2021 15:55
Operator : CG/JU
Sample : SSTDIC1.6
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDIC1.6

Quant Time: Apr 30 16:24:04 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN043021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 30 15:34:39 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.628 min Scan# 602
 Delta R.T. -0.000 min
 Lab File: BN014504.D
 Acq: 30 Apr 2021 15:55




Instrument :
 BNA_N
 ClientSampled :
 SSTDICC1.6


Tgt Ion:152 Resp: 9234
 Ion Ratio Lower Upper
 152 100
 150 156.0 123.8 185.8
 115 58.6 47.0 70.4



#2
 1,4-Dioxane
 Concen: 1.65 ng
 RT: 3.174 min Scan# 49
 Delta R.T. 0.000 min
 Lab File: BN014504.D
 Acq: 30 Apr 2021 15:55

Tgt Ion: 88 Resp: 16989
 Ion Ratio Lower Upper
 88 100
 58 66.1 44.4 66.6
 43 24.5 15.9 23.9#









#3
n-Nitrosodimethylamine
Concen: 1.86 ng
RT: 3.513 min Scan# 91
Delta R.T. -0.040 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 42 Resp: 8578
Ion Ratio Lower Upper
42 100
74 152.4 175.1 262.7#
44 3.0 7.5 11.3#



#4
2-Fluorophenol
Concen: 1.80 ng
RT: 5.228 min Scan# 304
Delta R.T. 0.000 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55

Tgt Ion: 112 Resp: 29524
Ion Ratio Lower Upper
112 100
64 51.8 41.8 62.6
63 27.1 22.0 33.0









#5
Phenol-d6
Concen: 1.77 ng
RT: 6.798 min Scan# 499
Delta R.T. 0.000 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 99 Resp: 35696
Ion Ratio Lower Upper
99 100
42 16.5 13.2 19.8
71 32.0 26.2 39.2



#6
bis(2-Chloroethyl)ether
Concen: 1.65 ng
RT: 7.064 min Scan# 532
Delta R.T. 0.000 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55

Tgt Ion: 93 Resp: 37101
Ion Ratio Lower Upper
93 100
63 70.1 58.0 87.0
95 32.9 27.0 40.6









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.404 min Scan# 852
Delta R.T. 0.000 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	136	Resp:	32085
Ion Ratio	Lower	Upper	
136	100		
137	11.1	8.9	13.3
54	4.7	3.8	5.8
68	4.5	3.8	5.6



#8
Nitrobenzene-d5
Concen: 1.87 ng
RT: 8.769 min Scan# 723
Delta R.T. 0.000 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55

Tgt Ion:	82	Resp:	30821
Ion Ratio	Lower	Upper	
82	100		
128	38.6	29.5	44.3
54	43.4	36.9	55.3









#9
Naphthalene
Concen: 1.64 ng
RT: 10.454 min Scan# 856
Delta R.T. 0.000 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:128 Resp: 138724
Ion Ratio Lower Upper
128 100
129 10.9 9.0 13.4
127 12.5 10.4 15.6



#10
Hexachlorobutadiene
Concen: 1.66 ng
RT: 10.746 min Scan# 879
Delta R.T. 0.000 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55

Tgt Ion:225 Resp: 26859
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.2 50.9 76.3









#11
2-Methylnaphthalene-d10
Concen: 1.68 ng
RT: 12.000 min Scan# 1050
Delta R.T. 0.000 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 109711
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.2



#12
2-Methylnaphthalene
Concen: 1.67 ng
RT: 12.071 min Scan# 1066
Delta R.T. -0.004 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55

Tgt Ion:142 Resp: 91239
Ion Ratio Lower Upper
142 100
141 87.8 70.2 105.4
115 33.7 26.7 40.1









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.270 min Scan# 1303
Delta R.T. 0.000 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:164 Resp: 16882
Ion Ratio Lower Upper
164 100
162 97.6 78.6 118.0
160 44.5 35.8 53.8



#14
2,4,6-Tribromophenol
Concen: 2.18 ng
RT: 15.767 min Scan# 1421
Delta R.T. 0.000 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55

Tgt Ion:330 Resp: 9374
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 33.3 27.0 40.4









#15
2-Fluorobiphenyl
Concen: 1.63 ng
RT: 12.887 min Scan# 1194
Delta R.T. 0.000 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:172 Resp: 107612
Ion Ratio Lower Upper
172 100
171 33.8 27.1 40.7
170 22.5 18.3 27.5



#16
Acenaphthylene
Concen: 1.72 ng
RT: 13.978 min Scan# 1280
Delta R.T. 0.000 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55

Tgt Ion:152 Resp: 107994
Ion Ratio Lower Upper
152 100
151 19.3 15.6 23.4
153 13.0 10.5 15.7









#17
Acenaphthene
Concen: 1.64 ng
RT: 14.334 min Scan# 1308
Delta R.T. 0.000 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:154 Resp: 78786
Ion Ratio Lower Upper
154 100
153 111.7 90.2 135.4
152 55.9 46.2 69.2



#18
Fluorene
Concen: 1.65 ng
RT: 15.323 min Scan# 1386
Delta R.T. 0.000 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55

Tgt Ion:166 Resp: 97035
Ion Ratio Lower Upper
166 100
165 97.5 78.4 117.6
167 13.5 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.019 min Scan# 1523
Delta R.T. 0.000 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	188	Resp:	31714
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	6.4	5.2	7.8



#20
4,6-Dinitro-2-methylphenol
Concen: 1.98 ng
RT: 15.399 min Scan# 1392
Delta R.T. -0.025 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55

Tgt Ion:	198	Resp:	5132
Ion Ratio	Lower	Upper	
198	100		
51	71.6	865.3	1297.9#
105	47.0	93.6	140.4#









#21
4-Bromophenyl-phenylether
Concen: 1.76 ng
RT: 16.216 min Scan# 1457
Delta R.T. 0.000 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55




Instrument :
BNA_N
ClientSampled :
SSTDICC1.6


Tgt Ion:248 Resp: 30077
Ion Ratio Lower Upper
248 100
250 100.2 80.5 120.7
141 41.9 37.2 55.8



#22
Hexachlorobenzene
Concen: 1.63 ng
RT: 16.325 min Scan# 1466
Delta R.T. 0.000 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55

Tgt Ion:284 Resp: 37876
Ion Ratio Lower Upper
284 100
142 33.2 26.6 39.8
249 29.6 23.6 35.4









#23
Atrazine
Concen: 2.06 ng
RT: 16.484 min Scan# 1479
Delta R.T. 0.000 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:200 Resp: 16466
Ion Ratio Lower Upper
200 100
173 25.3 23.0 34.6
215 48.4 38.6 57.8



#24
Pentachlorophenol
Concen: 2.12 ng
RT: 16.678 min Scan# 1495
Delta R.T. -0.025 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55

Tgt Ion:266 Resp: 8071
Ion Ratio Lower Upper
266 100
264 62.6 24.2 36.4#
268 64.1 25.2 37.8#









#25
Phenanthrene
Concen: 1.69 ng
RT: 17.055 min Scan# 1526
Delta R.T. 0.000 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:178 Resp: 155604
Ion Ratio Lower Upper
178 100
176 18.8 15.2 22.8
179 15.3 12.4 18.6



#26
Anthracene
Concen: 1.75 ng
RT: 17.141 min Scan# 1533
Delta R.T. -0.012 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55

Tgt Ion:178 Resp: 122434
Ion Ratio Lower Upper
178 100
176 18.3 14.8 22.2
179 15.2 12.2 18.2









#27
Fluoranthene-d10
Concen: 1.73 ng
RT: 19.054 min Scan# 1814
Delta R.T. 0.000 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:212 Resp: 192131
Ion Ratio Lower Upper
212 100
106 12.4 10.1 15.1
104 6.8 5.4 8.2



#28
Fluoranthene
Concen: 1.74 ng
RT: 19.084 min Scan# 1820
Delta R.T. 0.000 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55

Tgt Ion:202 Resp: 172068
Ion Ratio Lower Upper
202 100
101 10.6 8.4 12.6
203 17.3 13.5 20.3









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.218 min Scan# 2120
Delta R.T. 0.000 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:240 Resp: 30006
Ion Ratio Lower Upper
240 100
120 6.7 5.8 8.8
236 25.6 20.6 30.8



#30
Pyrene
Concen: 1.65 ng
RT: 19.447 min Scan# 1893
Delta R.T. 0.000 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55

Tgt Ion:202 Resp: 164017
Ion Ratio Lower Upper
202 100
200 20.5 16.3 24.5
203 17.4 14.0 21.0









#31
 Terphenyl-d14
 Concen: 1.63 ng
 RT: 19.660 min Scan# 1936
 Delta R.T. 0.000 min
 Lab File: BN014504.D
 Acq: 30 Apr 2021 15:55




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:244 Resp: 112930
 Ion Ratio Lower Upper
 244 100
 212 7.3 6.0 9.0
 122 11.2 9.3 13.9



#32
 Benzo(a)anthracene
 Concen: 1.79 ng
 RT: 21.197 min Scan# 2118
 Delta R.T. 0.000 min
 Lab File: BN014504.D
 Acq: 30 Apr 2021 15:55

Tgt Ion:228 Resp: 143733
 Ion Ratio Lower Upper
 228 100
 226 26.7 21.7 32.5
 229 19.6 15.7 23.5









#33
Chrysene
Concen: 1.59 ng
RT: 21.250 min Scan# 2123
Delta R.T. 0.000 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:228 Resp: 160338
Ion Ratio Lower Upper
228 100
226 29.3 23.8 35.6
229 19.7 15.6 23.4



#34
Bis(2-ethylhexyl)phthalate
Concen: 1.98 ng
RT: 21.144 min Scan# 2113
Delta R.T. 0.000 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55

Tgt Ion:149 Resp: 43326
Ion Ratio Lower Upper
149 100
167 27.5 20.5 30.7
279 3.7 2.9 4.3









#35
Perylene-d12
Concen: 0.40 ng
RT: 23.417 min Scan# 2527
Delta R.T. 0.000 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:264	Resp:	23314
Ion Ratio	Lower	Upper
264	100	
260	24.8	19.5 29.3
265	43.8	35.4 53.0



#36
Indeno(1,2,3-cd)pyrene
Concen: 1.65 ng
RT: 25.615 min Scan# 3169
Delta R.T. -0.003 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55

Tgt Ion:276	Resp:	155814
Ion Ratio	Lower	Upper
276	100	
138	24.4	18.8 28.2
277	24.8	19.8 29.8









#37
Benzo(b)fluoranthene
Concen: 1.62 ng
RT: 22.757 min Scan# 2334
Delta R.T. 0.000 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:252 Resp: 144724
Ion Ratio Lower Upper
252 100
253 22.1 19.6 29.4
125 10.2 10.3 15.5#



#38
Benzo(k)fluoranthene
Concen: 1.73 ng
RT: 22.798 min Scan# 2346
Delta R.T. -0.003 min
Lab File: BN014504.D
Acq: 30 Apr 2021 15:55

Tgt Ion:252 Resp: 169414
Ion Ratio Lower Upper
252 100
253 22.2 19.7 29.5
125 10.0 9.7 14.5





#39

Benzo(a)pyrene

Concen: 1.67 ng

RT: 23.318 min Scan# 2498

Delta R.T. -0.003 min

Lab File: BN014504.D

Acq: 30 Apr 2021 15:55





Tgt Ion:252 Resp: 128050

Ion Ratio Lower Upper

252 100

253 22.9 21.8 32.6

125 12.6 14.9 22.3#



#40

Dibenzo(a,h)anthracene

Concen: 1.64 ng

RT: 25.629 min Scan# 3173

Delta R.T. -0.007 min

Lab File: BN014504.D

Acq: 30 Apr 2021 15:55




Tgt Ion:278 Resp: 129952


Ion Ratio Lower Upper

278 100

139 16.6 13.9 20.9


279 24.0 20.7 31.1





#41
 Benzo(g,h,i)perylene
 Concen: 1.66 ng
 RT: 26.286 min Scan# 3365
 Delta R.T. -0.003 min
 Lab File: BN014504.D
 Acq: 30 Apr 2021 15:55

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6



Tgt Ion: 276 Resp: 145898
 Ion Ratio Lower Upper
 276 100
 277 23.6 19.6 29.4
 138 21.2 17.5 26.3

