

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN043021\
 Data File : BN014507.D
 Acq On : 30 Apr 2021 17:42
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN043021

Quant Time: Apr 30 18:13:58 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN043021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 30 17:46:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.628	152	9452	0.40	ng	0.00
7) Naphthalene-d8	10.404	136	34279	0.40	ng	0.00
13) Acenaphthene-d10	14.270	164	18051	0.40	ng	0.00
19) Phenanthrene-d10	17.018	188	35908	0.40	ng	0.00
29) Chrysene-d12	21.215	240	30882	0.40	ng	0.00
35) Perylene-d12	23.415	264	26017	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.228	112	6630	0.35	ng	0.00
5) Phenol-d6	6.798	99	7654	0.35	ng	0.00
8) Nitrobenzene-d5	8.769	82	6799	0.35	ng	0.00
11) 2-Methylnaphthalene-d10	12.000	152	25636	0.36	ng	0.00
14) 2,4,6-Tribromophenol	15.767	330	1834	0.33	ng	0.00
15) 2-Fluorobiphenyl	12.887	172	25351	0.37	ng	0.00
27) Fluoranthene-d10	19.052	212	43624	0.34	ng	0.00
31) Terphenyl-d14	19.662	244	26669	0.37	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.174	88	3986	0.35	ng	88
3) n-Nitrosodimethylamine	3.561	42	1410	0.38	ng	# 60
6) bis(2-Chloroethyl)ether	7.064	93	8087	0.36	ng	99
9) Naphthalene	10.454	128	32816	0.36	ng	100
10) Hexachlorobutadiene	10.746	225	6560	0.37	ng	# 100
12) 2-Methylnaphthalene	12.075	142	21338	0.36	ng	99
16) Acenaphthylene	13.978	152	23245	0.34	ng	99
17) Acenaphthene	14.334	154	17972	0.35	ng	100
18) Fluorene	15.323	166	22135	0.35	ng	100
20) 4,6-Dinitro-2-methylph...	15.412	198	592	0.36	ng	# 1
21) 4-Bromophenyl-phenylether	16.215	248	6882	0.35	ng	98
22) Hexachlorobenzene	16.325	284	9160	0.36	ng	100
23) Atrazine	16.483	200	3622	0.32	ng	99
24) Pentachlorophenol	16.690	266	1161	0.45	ng	# 33
25) Phenanthrene	17.055	178	36809	0.36	ng	100
26) Anthracene	17.140	178	26597	0.33	ng	99
28) Fluoranthene	19.081	202	38258	0.34	ng	100
30) Pyrene	19.449	202	37669	0.36	ng	100
32) Benzo(a)anthracene	21.194	228	29447	0.34	ng	100
33) Chrysene	21.247	228	36956	0.37	ng	100
34) Bis(2-ethylhexyl)phtha...	21.141	149	10129	0.37	ng	# 98
36) Indeno(1,2,3-cd)pyrene	25.612	276	33075	0.34	ng	99
37) Benzo(b)fluoranthene	22.757	252	32701	0.34	ng	100
38) Benzo(k)fluoranthene	22.799	252	38885	0.35	ng	100
39) Benzo(a)pyrene	23.319	252	30075	0.35	ng	97
40) Dibenzo(a,h)anthracene	25.633	278	27082	0.34	ng	100
41) Benzo(g,h,i)perylene	26.283	276	33756	0.35	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN043021\
Data File : BN014507.D
Acq On : 30 Apr 2021 17:42
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN043021

Quant Time: Apr 30 18:13:58 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN043021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 30 17:46:00 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.628 min Scan# 602
 Delta R.T. -0.000 min
 Lab File: BN014507.D
 Acq: 30 Apr 2021 17:42




Instrument :
 BNA_N
 ClientSampled :
 ICVBN043021


Tgt Ion:152 Resp: 9452
 Ion Ratio Lower Upper
 152 100
 150 158.2 123.8 185.8
 115 58.1 47.0 70.4



#2
 1,4-Dioxane
 Concen: 0.35 ng
 RT: 3.174 min Scan# 49
 Delta R.T. 0.000 min
 Lab File: BN014507.D
 Acq: 30 Apr 2021 17:42

Tgt Ion: 88 Resp: 3986
 Ion Ratio Lower Upper
 88 100
 58 64.9 44.4 66.6
 43 23.6 15.9 23.9









#3
n-Nitrosodimethylamine
Concen: 0.38 ng
RT: 3.561 min Scan# 97
Delta R.T. 0.008 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42




Instrument :
BNA_N
ClientSampleId :
ICVBN043021


Tgt Ion: 42 Resp: 1410
Ion Ratio Lower Upper
42 100
74 152.5 175.1 262.7#
44 5.5 7.5 11.3#



#4
2-Fluorophenol
Concen: 0.35 ng
RT: 5.228 min Scan# 304
Delta R.T. -0.000 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42

Tgt Ion: 112 Resp: 6630
Ion Ratio Lower Upper
112 100
64 52.1 41.8 62.6
63 27.6 22.0 33.0









#5
Phenol-d6
Concen: 0.35 ng
RT: 6.798 min Scan# 499
Delta R.T. -0.000 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42




Instrument :
BNA_N
ClientSampled :
ICVBN043021


Tgt Ion: 99 Resp: 7654
Ion Ratio Lower Upper
99 100
42 16.6 13.2 19.8
71 32.0 26.2 39.2



#6
bis(2-Chloroethyl)ether
Concen: 0.36 ng
RT: 7.064 min Scan# 532
Delta R.T. 0.000 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42

Tgt Ion: 93 Resp: 8087
Ion Ratio Lower Upper
93 100
63 71.9 58.0 87.0
95 34.6 27.0 40.6









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.404 min Scan# 852
Delta R.T. 0.000 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42




Instrument :
BNA_N
ClientSampleId :
ICVBN043021


Tgt Ion:	136	Resp:	34279
Ion Ratio	Lower	Upper	
136	100		
137	11.0	8.9	13.3
54	4.9	3.8	5.8
68	4.7	3.8	5.6



#8
Nitrobenzene-d5
Concen: 0.35 ng
RT: 8.769 min Scan# 723
Delta R.T. 0.000 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42

Tgt Ion:	82	Resp:	6799
Ion Ratio	Lower	Upper	
82	100		
128	36.2	29.5	44.3
54	46.5	36.9	55.3









#9
Naphthalene
Concen: 0.36 ng
RT: 10.454 min Scan# 856
Delta R.T. 0.000 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42




Instrument :
BNA_N
ClientSampleId :
ICVBN043021


Tgt Ion:128 Resp: 32816
Ion Ratio Lower Upper
128 100
129 11.3 9.0 13.4
127 12.8 10.4 15.6



#10
Hexachlorobutadiene
Concen: 0.37 ng
RT: 10.746 min Scan# 879
Delta R.T. 0.000 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42

Tgt Ion:225 Resp: 6560
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.9 50.9 76.3









#11
2-Methylnaphthalene-d10
Concen: 0.36 ng
RT: 12.000 min Scan# 1050
Delta R.T. -0.000 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42




Instrument :
BNA_N
ClientSampleId :
ICVBN043021


Tgt Ion:152 Resp: 25636
Ion Ratio Lower Upper
152 100
151 20.3 16.2 24.2



#12
2-Methylnaphthalene
Concen: 0.36 ng
RT: 12.075 min Scan# 1067
Delta R.T. 0.000 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42

Tgt Ion:142 Resp: 21338
Ion Ratio Lower Upper
142 100
141 86.9 70.2 105.4
115 33.5 26.7 40.1









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.270 min Scan# 1303
Delta R.T. 0.000 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42




Instrument :
BNA_N
ClientSampleId :
ICVBN043021


Tgt Ion:164 Resp: 18051
Ion Ratio Lower Upper
164 100
162 97.4 78.6 118.0
160 44.6 35.8 53.8



#14
2,4,6-Tribromophenol
Concen: 0.33 ng
RT: 15.767 min Scan# 1421
Delta R.T. 0.000 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42

Tgt Ion:330 Resp: 1834
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 32.8 27.0 40.4









#15
2-Fluorobiphenyl
Concen: 0.37 ng
RT: 12.887 min Scan# 1194
Delta R.T. -0.000 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42




Instrument :
BNA_N
ClientSampleId :
ICVBN043021


Tgt Ion:	172	Resp:	25351
Ion Ratio	Lower	Upper	
172	100		
171	33.9	27.1	40.7
170	22.5	18.3	27.5



#16
Acenaphthylene
Concen: 0.34 ng
RT: 13.978 min Scan# 1280
Delta R.T. 0.000 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42

Tgt Ion:	152	Resp:	23245
Ion Ratio	Lower	Upper	
152	100		
151	19.2	15.6	23.4
153	13.0	10.5	15.7









#17
Acenaphthene
Concen: 0.35 ng
RT: 14.334 min Scan# 1308
Delta R.T. -0.000 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42




Instrument :
BNA_N
ClientSampleId :
ICVBN043021


Tgt Ion:154 Resp: 17972
Ion Ratio Lower Upper
154 100
153 113.2 90.2 135.4
152 57.6 46.2 69.2



#18
Fluorene
Concen: 0.35 ng
RT: 15.323 min Scan# 1386
Delta R.T. 0.000 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42

Tgt Ion:166 Resp: 22135
Ion Ratio Lower Upper
166 100
165 98.1 78.4 117.6
167 13.6 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.018 min Scan# 1523
Delta R.T. -0.001 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42




Instrument :
BNA_N
ClientSampleId :
ICVBN043021


Tgt Ion:188 Resp: 35908
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 6.4 5.2 7.8



#20
4,6-Dinitro-2-methylphenol
Concen: 0.36 ng
RT: 15.412 min Scan# 1393
Delta R.T. -0.013 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42

Tgt Ion:198 Resp: 592
Ion Ratio Lower Upper
198 100
51 317.5 865.3 1297.9#
105 99.5 93.6 140.4









#21
4-Bromophenyl-phenylether
Concen: 0.35 ng
RT: 16.215 min Scan# 1457
Delta R.T. -0.001 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42




Instrument :
BNA_N
ClientSampleId :
ICVBN043021


Tgt Ion:248 Resp: 6882
Ion Ratio Lower Upper
248 100
250 98.7 80.5 120.7
141 48.4 37.2 55.8



#22
Hexachlorobenzene
Concen: 0.36 ng
RT: 16.325 min Scan# 1466
Delta R.T. -0.001 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42

Tgt Ion:284 Resp: 9160
Ion Ratio Lower Upper
284 100
142 33.0 26.6 39.8
249 29.5 23.6 35.4









#23
Atrazine
Concen: 0.32 ng
RT: 16.483 min Scan# 1479
Delta R.T. -0.001 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42




Instrument :
BNA_N
ClientSampleId :
ICVBN043021


Tgt Ion:200 Resp: 3622
Ion Ratio Lower Upper
200 100
173 28.5 23.0 34.6
215 47.8 38.6 57.8



#24
Pentachlorophenol
Concen: 0.45 ng
RT: 16.690 min Scan# 1496
Delta R.T. -0.013 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42

Tgt Ion:266 Resp: 1161
Ion Ratio Lower Upper
266 100
264 62.4 24.2 36.4#
268 72.8 25.2 37.8#









#25
Phenanthrene
Concen: 0.36 ng
RT: 17.055 min Scan# 1526
Delta R.T. -0.001 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42




Instrument :
BNA_N
ClientSampleId :
ICVBN043021


Tgt Ion:178	Resp:	36809
Ion Ratio	Lower	Upper
178	100	
176	18.9	15.2 22.8
179	15.3	12.4 18.6



#26
Anthracene
Concen: 0.33 ng
RT: 17.140 min Scan# 1533
Delta R.T. -0.013 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42

Tgt Ion:178	Resp:	26597
Ion Ratio	Lower	Upper
178	100	
176	18.3	14.8 22.2
179	15.4	12.2 18.2









#27
Fluoranthene-d10
Concen: 0.34 ng
RT: 19.052 min Scan# 1815
Delta R.T. -0.003 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42




Instrument :
BNA_N
ClientSampleId :
ICVBN043021


Tgt Ion:212 Resp: 43624
Ion Ratio Lower Upper
212 100
106 12.4 10.1 15.1
104 6.8 5.4 8.2



#28
Fluoranthene
Concen: 0.34 ng
RT: 19.081 min Scan# 1821
Delta R.T. -0.003 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42

Tgt Ion:202 Resp: 38258
Ion Ratio Lower Upper
202 100
101 10.3 8.4 12.6
203 17.0 13.5 20.3









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.215 min Scan# 2121
Delta R.T. -0.003 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42




Instrument :
BNA_N
ClientSampleId :
ICVBN043021


Tgt Ion:240 Resp: 30882
Ion Ratio Lower Upper
240 100
120 8.3 5.8 8.8
236 25.8 20.6 30.8



#30
Pyrene
Concen: 0.36 ng
RT: 19.449 min Scan# 1895
Delta R.T. 0.002 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42

Tgt Ion:202 Resp: 37669
Ion Ratio Lower Upper
202 100
200 20.3 16.3 24.5
203 17.4 14.0 21.0









#31
Terphenyl-d14
Concen: 0.37 ng
RT: 19.662 min Scan# 1938
Delta R.T. 0.002 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42




Instrument :
BNA_N
ClientSampleId :
ICVBN043021


Tgt Ion:244 Resp: 26669
Ion Ratio Lower Upper
244 100
212 7.6 6.0 9.0
122 11.2 9.3 13.9



#32
Benzo(a)anthracene
Concen: 0.34 ng
RT: 21.194 min Scan# 2119
Delta R.T. -0.003 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42

Tgt Ion:228 Resp: 29447
Ion Ratio Lower Upper
228 100
226 27.3 21.7 32.5
229 19.6 15.7 23.5





#33

Chrysene

Concen: 0.37 ng

RT: 21.247 min Scan# 2124

Delta R.T. -0.003 min

Lab File: BN014507.D

Acq: 30 Apr 2021 17:42





Tgt Ion:228	Resp:	36956
Ion Ratio	Lower	Upper
228	100	
226	29.6	23.8 35.6
229	19.4	15.6 23.4

Instrument :

BNA_N

ClientSampleId :

ICVBN043021



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.37 ng




RT: 21.141 min Scan# 2114


Delta R.T. -0.003 min

Lab File: BN014507.D

Acq: 30 Apr 2021 17:42

Tgt Ion:149	Resp:	10129
Ion Ratio	Lower	Upper
149	100	
167	26.4	20.5 30.7
279	4.5	2.9 4.3









#35
Perylene-d12
Concen: 0.40 ng
RT: 23.415 min Scan# 2528
Delta R.T. -0.003 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42




Instrument :
BNA_N
ClientSampleId :
ICVBN043021


Tgt Ion:264	Resp:	26017
Ion Ratio	Lower	Upper
264	100	
260	24.6	19.5 29.3
265	42.0	35.4 53.0



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.34 ng
RT: 25.612 min Scan# 3170
Delta R.T. -0.006 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42

Tgt Ion:276	Resp:	33075
Ion Ratio	Lower	Upper
276	100	
138	22.1	18.8 28.2
277	24.8	19.8 29.8









#37
Benzo(b)fluoranthene
Concen: 0.34 ng
RT: 22.757 min Scan# 2336
Delta R.T. 0.001 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42




Instrument :
BNA_N
ClientSampled :
ICVBN043021


Tgt Ion:252 Resp: 32701
Ion Ratio Lower Upper
252 100
253 24.4 19.6 29.4
125 13.3 10.3 15.5



#38
Benzo(k)fluoranthene
Concen: 0.35 ng
RT: 22.799 min Scan# 2348
Delta R.T. -0.003 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42

Tgt Ion:252 Resp: 38885
Ion Ratio Lower Upper
252 100
253 24.6 19.7 29.5
125 11.9 9.7 14.5









#39
Benzo(a)pyrene
Concen: 0.35 ng
RT: 23.319 min Scan# 2500
Delta R.T. -0.003 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42




Instrument :
BNA_N
ClientSampleId :
ICVBN043021


Tgt Ion:252	Resp:	30075
Ion Ratio	Lower	Upper
252	100	
253	27.2	21.8 32.6
125	21.6	14.9 22.3



#40
Dibenzo(a,h)anthracene
Concen: 0.34 ng
RT: 25.633 min Scan# 3176
Delta R.T. -0.003 min
Lab File: BN014507.D
Acq: 30 Apr 2021 17:42

Tgt Ion:278	Resp:	27082
Ion Ratio	Lower	Upper
278	100	
139	17.4	13.9 20.9
279	26.3	20.7 31.1





#41
 Benzo(g,h,i)perylene
 Concen: 0.35 ng
 RT: 26.283 min Scan# 3366
 Delta R.T. -0.006 min
 Lab File: BN014507.D
 Acq: 30 Apr 2021 17:42

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN043021

Tgt Ion:	276	Resp:	33756
Ion Ratio	Lower	Upper	
276	100		
277	24.7	19.6	29.4
138	21.5	17.5	26.3

