

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN043022\
 Data File : BN019609.D
 Acq On : 30 Apr 2022 02:31
 Operator : CG/JU
 Sample : PB144473BS
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS473

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 05/02/2022
 Supervised By :Yogesh Patel 05/05/2022

Quant Time: Apr 30 03:09:48 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN042722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 27 13:41:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.869	152	141786	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.663	136	572165	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.493	164	320060	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.234	188	646330	20.000	ng/u1	0.00
79) Chrysene-d12	21.410	240	599890	20.000	ng/u1	-0.01
88) Perylene-d12	23.769	264	612386	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.340	96	26061	6.936	ng/uL	0.00
4) Pyridine-d5	3.746	84	324724	34.131	ng/u1	0.00
7) Phenol-d5	7.034	99	418447	34.056	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.199	67	255520	35.853	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.405	132	357684	36.700	ng/u1	0.00
15) 4-Methylphenol-d8	8.575	113	334988	35.589	ng/u1	0.00
21) Nitrobenzene-d5	9.022	128	170473	40.864	ng/u1	0.00
24) 2-Nitrophenol-d4	9.746	143	174892	42.540	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.281	165	308186	37.448	ng/u1	0.00
31) 4-Chloroaniline-d4	10.793	131	411706	32.955	ng/u1	0.00
46) Dimethylphthalate-d6	13.904	166	863398	39.275	ng/u1	0.00
49) Acenaphthylene-d8	14.187	160	1051760	35.385	ng/u1	0.00
54) 4-Nitrophenol-d4	14.675	143	174218m	40.749	ng/u1	0.00
60) Fluorene-d10	15.481	176	731098	37.674	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.593	200	126303	38.278	ng/u1	0.00
73) Anthracene-d10	17.328	188	1084141	35.618	ng/u1	0.00
81) Pyrene-d10	19.622	212	1232509	35.736	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.616	264	1183065	37.792	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.370	88	50861	13.274	ng/uL#	92
5) Pyridine	3.764	79	349425	34.854	ng/u1	97
6) Benzaldehyde	7.011	77	281741	44.739	ng/u1	95
8) Phenol	7.064	94	441696	35.724	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.293	93	334988	34.918	ng/u1	99
12) 2-Chlorophenol	7.434	128	367311	37.111	ng/u1	98
13) 2-Methylphenol	8.311	108	321750	35.146	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.399	45	503053	36.486	ng/u1	96
16) Acetophenone	8.687	105	503549	35.649	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.681	70	247947	37.717	ng/u1	93
18) 4-Methylphenol	8.634	108	352213	35.934	ng/u1	97
19) Hexachloroethane	8.952	117	148971	37.128	ng/u1	85
22) Nitrobenzene	9.063	77	375756	38.203	ng/u1	94
23) Isophorone	9.593	82	673754	38.408	ng/u1	98
25) 2-Nitrophenol	9.775	139	189035	40.719	ng/u1	96
26) 2,4-Dimethylphenol	9.840	107	358166	34.633	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.075	93	422836	36.656	ng/u1	98
29) 2,4-Dichlorophenol	10.310	162	309920	37.562	ng/u1	100
30) Naphthalene	10.710	128	1076982	35.146	ng/u1	99
32) 4-Chloroaniline	10.816	127	411479	32.770	ng/u1	100
33) Hexachlorobutadiene	11.010	225	182361	36.840	ng/u1	97
34) Caprolactam	11.569	113	100803m	43.199	ng/u1	
35) 4-Chloro-3-methylphenol	11.940	107	341153	39.064	ng/u1	99
36) 2-Methylnaphthalene	12.322	142	708063	35.904	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN043022\
 Data File : BN019609.D
 Acq On : 30 Apr 2022 02:31
 Operator : CG/JU
 Sample : PB144473BS
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :

BNA_N

ClientSampleId :

SLCS473

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 05/02/2022

Supervised By :Yogesh Patel 05/05/2022

Quant Time: Apr 30 03:09:48 2022

Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN042722.M


Quant Title : SVOA CALIBRATION

QLast Update : Wed Apr 27 13:41:44 2022

Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.540	142	723424	36.641	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.693	216	330219	37.184	ng/ul	98
40) Hexachlorocyclopentadiene	12.681	237	170897	31.274	ng/ul	99
41) 2,4,6-Trichlorophenol	12.928	196	221096	39.909	ng/ul	96
42) 2,4,5-Trichlorophenol	12.993	196	241332	40.121	ng/ul	97
43) 1,1'-Biphenyl	13.328	154	913999	36.119	ng/ul	98
44) 2-Chloronaphthalene	13.369	162	700388	36.226	ng/ul	96
45) 2-Nitroaniline	13.563	65	210132	45.292	ng/ul	92
47) Dimethylphthalate	13.951	163	839241	38.443	ng/ul	99
48) 2,6-Dinitrotoluene	14.063	165	170756	45.341	ng/ul	96
50) Acenaphthylene	14.216	152	1133393	37.115	ng/ul	98
51) 3-Nitroaniline	14.387	138	188545	48.844	ng/ul	98
52) Acenaphthene	14.557	153	728466	36.450	ng/ul	99
53) 2,4-Dinitrophenol	14.593	184	85008	40.693	ng/ul	96
55) 4-Nitrophenol	14.693	109	145994	42.973	ng/ul	97
56) Dibenzofuran	14.893	168	1016916	36.425	ng/ul	98
57) 2,4-Dinitrotoluene	14.845	165	251393	45.906	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.116	232	182478	41.591	ng/ul	91
59) Diethylphthalate	15.316	149	846636	39.429	ng/ul	97
61) Fluorene	15.540	166	786623	36.296	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.534	204	362445	36.988	ng/ul	93
63) 4-Nitroaniline	15.545	138	199129	60.190	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.610	198	126912	37.816	ng/ul	96
67) N-Nitrosodiphenylamine	15.745	169	697168	35.155	ng/ul	98
68) 4-Bromophenyl-phenylether	16.428	248	211967	34.910	ng/ul	92
69) Hexachlorobenzene	16.545	284	237962	34.843	ng/ul#	88
70) Atrazine	16.692	200	240339	37.990	ng/ul	97
71) Pentachlorophenol	16.881	266	157759	38.063	ng/ul	96
72) Phenanthrene	17.275	178	1265008	35.023	ng/ul	100
74) Anthracene	17.363	178	1286422	35.654	ng/ul	96
75) 1,2,3,4-Tetrachloroben...	13.293	216	331294	33.251	ng/uL	97
76) Pentachlorobenzene	14.816	250	298586	34.401	ng/uL	95
77) Carbazole	17.628	167	1218478	37.148	ng/ul	97
78) Di-n-butylphthalate	18.204	149	1428692	40.717	ng/ul	99
80) Fluoranthene	19.286	202	1452337	35.723	ng/ul	97
82) Pyrene	19.651	202	1495385	35.312	ng/ul	97
83) Butylbenzylphthalate	20.557	149	656961	41.939	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.328	252	438631	36.408	ng/ul#	95
85) Benzo(a)anthracene	21.398	228	1409921	36.696	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.339	149	1014515	42.209	ng/ul#	97
87) Chrysene	21.451	228	1339534	35.493	ng/ul	98
89) Di-n-octyl phthalate	22.251	149	1741452	41.539	ng/ul	100
90) Benzo(b)fluoranthene	23.051	252	1518103	38.320	ng/ul	97
91) Benzo(k)fluoranthene	23.104	252	1326583	36.916	ng/ul	97
93) Benzo(a)pyrene	23.663	252	1438896	38.247	ng/ul	95
94) Indeno(1,2,3-cd)pyrene	26.192	276	1622570	37.635	ng/ul#	92
95) Dibenzo(a,h)anthracene	26.204	278	1379618	37.424	ng/ul	98
96) Benzo(g,h,i)perylene	26.939	276	1399389	36.902	ng/ul	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 13.274 ng/uL
RT: 3.370 min Scan# 41
Delta R.T. -0.011 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Instrument :
BNA_N
ClientSampleId :
SLCS473



Tgt Ion: 88 Resp: 5086
Ion Ratio Lower Upper
88 100
43 34.8 22.4 33.6
58 81.2 61.0 91.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#5
Pyridine
Concen: 34.854 ng/uL
RT: 3.764 min Scan# 115
Delta R.T. -0.012 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31


Tgt Ion: 79 Resp: 349425
Ion Ratio Lower Upper
79 100
52 84.4 65.0 97.4
51 41.3 31.6 47.4





#6
Benzaldehyde
Concen: 44.739 ng/ul
RT: 7.011 min Scan# 667
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Instrument :
BNA_N
ClientSampleId :
SLCS473




Tgt Ion: 77 Resp: 28174
Ion Ratio Lower Upper
77 100
105 97.5 74.7 112.1
106 94.1 70.2 105.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#8
Phenol
Concen: 35.724 ng/ul
RT: 7.064 min Scan# 676
Delta R.T. 0.000 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31



Tgt Ion: 94 Resp: 441696
Ion Ratio Lower Upper
94 100
65 29.8 23.6 35.4
66 38.9 31.2 46.8






#10
Bis(2-Chloroethyl)ether
Concen: 34.918 ng/ul
RT: 7.293 min Scan# 715
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31

Instrument :

BNA_N




ClientSampleId :
SLCS473




Tgt Ion: 93 Resp: 334988
Ion Ratio Lower Upper
93 100
63 80.1 63.4 95.0
95 31.6 26.2 39.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#12
2-Chlorophenol
Concen: 37.111 ng/ul
RT: 7.434 min Scan# 739
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Tgt Ion:128 Resp: 367311
Ion Ratio Lower Upper
128 100
64 47.0 38.6 57.8
130 32.0 25.0 37.4





#13
2-Methylphenol
Concen: 35.146 ng/ul
RT: 8.311 min Scan# 888
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Instrument :
BNA_N
ClientSampleId :
SLCS473



Tgt Ion:108 Resp: 321750
Ion Ratio Lower Upper
108 100
107 87.7 71.5 107.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#14
2,2'-oxybis(1-Chloropropane)
Concen: 36.486 ng/ul
RT: 8.399 min Scan# 903
Delta R.T. -0.017 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31


Tgt Ion: 45 Resp: 503053
Ion Ratio Lower Upper
45 100
77 12.0 10.9 16.3
79 9.2 8.6 13.0





#16
Acetophenone
Concen: 35.649 ng/ul
RT: 8.687 min Scan# 91
Delta R.T. -0.012 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Instrument :
BNA_N
ClientSampleId :
SLCS473




Tgt Ion: 105 Resp: 503549
Ion Ratio Lower Upper
105 100
77 75.9 63.5 95.3
51 29.1 23.5 35.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#17
N-Nitroso-di-n-propylamine
Concen: 37.717 ng/ul
RT: 8.681 min Scan# 951
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Tgt Ion: 70 Resp: 247947
Ion Ratio Lower Upper
70 100
42 66.3 47.8 71.6
101 10.0 7.8 11.6
130 22.6 15.8 23.8





#18
4-Methylphenol
Concen: 35.934 ng/ul
RT: 8.634 min Scan# 943
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Instrument :
BNA_N
ClientSampleId :
SLCS473




Tgt Ion:108 Resp: 35221
Ion Ratio Lower Upper
108 100
107 112.7 88.0 132.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#19
Hexachloroethane
Concen: 37.128 ng/ul
RT: 8.952 min Scan# 997
Delta R.T. -0.012 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Tgt Ion:117 Resp: 148971
Ion Ratio Lower Upper
117 100
201 87.7 83.2 124.8
199 52.4 50.2 75.4





#22
Nitrobenzene
Concen: 38.203 ng/u1
RT: 9.063 min Scan# 1016
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Instrument :
BNA_N
ClientSampleId :
SLCS473




Tgt Ion: 77 Resp: 375750
Ion Ratio Lower Upper
77 100
123 48.0 34.5 51.7
65 15.0 11.1 16.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#23
Isophorone
Concen: 38.408 ng/u1
RT: 9.593 min Scan# 1106
Delta R.T. -0.012 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31



Tgt Ion: 82 Resp: 673754
Ion Ratio Lower Upper
82 100
95 7.4 5.8 8.8
138 18.5 13.8 20.8





#25
2-Nitrophenol
Concen: 40.719 ng/ul
RT: 9.775 min Scan# 1137
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31

Instrument :

BNA_N

ClientSampleId :





SLCS473

Tgt Ion:139 Resp: 189035
Ion Ratio Lower Upper
139 100
65 50.3 43.7 65.5
109 32.6 25.8 38.6

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#26
2,4-Dimethylphenol
Concen: 34.633 ng/ul
RT: 9.840 min Scan# 1148
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31

Tgt Ion:107 Resp: 358166
Ion Ratio Lower Upper
107 100
121 48.9 38.3 57.5
122 85.0 66.3 99.5









#27
Bis(2-Chloroethoxy)methane
Concen: 36.656 ng/ul
RT: 10.075 min Scan# 1188
Delta R.T. -0.012 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31

Instrument :
BNA_N
ClientSampleId :
SLCS473

Tgt Ion: 93 Resp: 422830
Ion Ratio Lower Upper
93 100
95 32.6 25.4 38.0
123 12.1 8.3 12.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#29
2,4-Dichlorophenol
Concen: 37.562 ng/ul
RT: 10.310 min Scan# 1228
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31


Tgt Ion:162 Resp: 309920
Ion Ratio Lower Upper
162 100
164 66.4 52.9 79.3
98 37.6 29.8 44.8





#30
Naphthalene
Concen: 35.146 ng/ul
RT: 10.710 min Scan# 1296
Delta R.T. -0.012 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Instrument :
BNA_N
ClientSampleId :
SLCS473



Tgt Ion:128 Resp: 107698
Ion Ratio Lower Upper
128 100
129 11.3 8.7 13.1
127 13.4 10.6 16.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#32
4-Chloroaniline
Concen: 32.770 ng/ul
RT: 10.816 min Scan# 1314
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31


Tgt Ion:127 Resp: 411479
Ion Ratio Lower Upper
127 100
129 32.6 26.1 39.1





#33
Hexachlorobutadiene
Concen: 36.840 ng/ul
RT: 11.010 min Scan# 11
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Instrument :
BNA_N
ClientSampleId :
SLCS473




Tgt Ion:225 Resp: 18236
Ion Ratio Lower Upper
225 100
223 63.8 51.8 77.6
227 66.5 50.6 76.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#34
Caprolactam
Concen: 43.199 ng/ul m
RT: 11.569 min Scan# 1442
Delta R.T. -0.017 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31



Tgt Ion:113 Resp: 100803
Ion Ratio Lower Upper
113 100
55 180.0 145.5 218.3
56 136.6 106.5 159.7






#35
4-Chloro-3-methylphenol
Concen: 39.064 ng/ul
RT: 11.940 min Scan# 1505
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31

Instrument :

BNA_N

ClientSampleId :

SLCS473






Tgt Ion:107 Resp: 34115
Ion Ratio Lower Upper
107 100
144 24.6 20.6 31.0
142 77.2 62.2 93.4


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#36
2-Methylnaphthalene
Concen: 35.904 ng/ul
RT: 12.322 min Scan# 1570
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31



Tgt Ion:142 Resp: 708063
Ion Ratio Lower Upper
142 100
141 89.0 70.6 105.8





#37
1-Methylnaphthalene
Concen: 36.641 ng/ul
RT: 12.540 min Scan# 1607
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31

Instrument :

BNA_N

ClientSampleId :





SLCS473

Tgt Ion:142 Resp: 723424
Ion Ratio Lower Upper
142 100
141 90.8 70.9 106.3
116 3.9 3.1 4.7

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#39
1,2,4,5-Tetrachlorobenzene
Concen: 37.184 ng/ul
RT: 12.693 min Scan# 1633
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31


Tgt Ion:216 Resp: 330219
Ion Ratio Lower Upper
216 100
214 77.3 63.4 95.0
179 22.6 17.8 26.8
108 20.0 15.1 22.7





#40
Hexachlorocyclopentadiene
Concen: 31.274 ng/ul
RT: 12.681 min Scan# 1631
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Instrument :
BNA_N
ClientSampleId :
SLCS473




Tgt Ion:237 Resp: 170897
Ion Ratio Lower Upper
237 100
235 61.8 48.8 73.2
272 12.8 9.3 13.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#41
2,4,6-Trichlorophenol
Concen: 39.909 ng/ul
RT: 12.928 min Scan# 1673
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31



Tgt Ion:196 Resp: 221096
Ion Ratio Lower Upper
196 100
198 95.4 79.8 119.8
200 32.5 25.0 37.6






#42
2,4,5-Trichlorophenol
Concen: 40.121 ng/ul
RT: 12.993 min Scan# 10
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31

Instrument :

BNA_N

ClientSampleId :

SLCS473






Tgt Ion:196 Resp: 241332
Ion Ratio Lower Upper
196 100
198 94.3 77.6 116.4
200 29.4 24.4 36.6


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#43
1,1'-Biphenyl
Concen: 36.119 ng/ul
RT: 13.328 min Scan# 1741
Delta R.T. -0.012 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Tgt Ion:154 Resp: 913999
Ion Ratio Lower Upper
154 100
153 40.5 31.5 47.3
76 12.8 10.5 15.7





#44
2-Chloronaphthalene
Concen: 36.226 ng/ul
RT: 13.369 min Scan# 1748
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Instrument :
BNA_N
ClientSampleId :
SLCS473




Tgt Ion:162 Resp: 700388
Ion Ratio Lower Upper
162 100
164 32.7 25.8 38.6
127 43.1 31.4 47.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#45
2-Nitroaniline
Concen: 45.292 ng/ul
RT: 13.563 min Scan# 1781
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Tgt Ion: 65 Resp: 210132
Ion Ratio Lower Upper
65 100
92 68.4 54.1 81.1
138 108.1 77.0 115.4





#47
Dimethylphthalate
Concen: 38.443 ng/ul
RT: 13.951 min Scan# 1847
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Instrument :
BNA_N
ClientSampleId :
SLCS473




Tgt Ion:163 Resp: 83924
Ion Ratio Lower Upper
163 100
194 3.8 3.4 5.0
164 10.2 7.7 11.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#48
2,6-Dinitrotoluene
Concen: 45.341 ng/ul
RT: 14.063 min Scan# 1866
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Tgt Ion:165 Resp: 170756
Ion Ratio Lower Upper
165 100
89 60.0 50.6 75.8
121 21.1 17.9 26.9





#50
Acenaphthylene
Concen: 37.115 ng/ul
RT: 14.216 min Scan# 1892
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Instrument : BNA_N
ClientSampleId : SLCS473



Tgt Ion:152 Resp: 113339
Ion Ratio Lower Upper
152 100
151 19.2 16.0 24.0
153 12.5 10.7 16.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#51
3-Nitroaniline
Concen: 48.844 ng/ul
RT: 14.387 min Scan# 1921
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31


Tgt Ion:138 Resp: 188545
Ion Ratio Lower Upper
138 100
108 11.5 8.6 12.8
92 117.8 92.9 139.3





#52
Acenaphthene
Concen: 36.450 ng/ul
RT: 14.557 min Scan# 1950
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Instrument :
BNA_N
ClientSampleId :
SLCS473




Tgt Ion:153 Resp: 728460
Ion Ratio Lower Upper
153 100
152 47.0 38.4 57.6
154 87.9 69.7 104.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#53
2,4-Dinitrophenol
Concen: 40.693 ng/ul
RT: 14.593 min Scan# 1956
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Tgt Ion:184 Resp: 85008
Ion Ratio Lower Upper
184 100
63 73.0 55.9 83.9
154 70.5 53.1 79.7





#55
4-Nitrophenol
Concen: 42.973 ng/ul
RT: 14.693 min Scan# 1973
Delta R.T. 0.000 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Instrument :
BNA_N
ClientSampleId :
SLCS473



Tgt Ion:109 Resp: 145994
Ion Ratio Lower Upper
109 100
139 128.2 100.1 150.1
65 127.5 104.6 157.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#56
Dibenzofuran
Concen: 36.425 ng/ul
RT: 14.893 min Scan# 2007
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31

Tgt Ion:168 Resp: 1016916
Ion Ratio Lower Upper
168 100
139 40.1 31.1 46.7









#57
2,4-Dinitrotoluene
Concen: 45.906 ng/ul
RT: 14.845 min Scan# 1999
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31

Instrument :
BNA_N
ClientSampleId :
SLCS473

Tgt Ion:165 Resp: 25139
Ion Ratio Lower Upper
165 100
63 45.1 36.3 54.5
182 2.9 2.3 3.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#58
2,3,4,6-Tetrachlorophenol
Concen: 41.591 ng/ul
RT: 15.116 min Scan# 2045
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31

Tgt Ion:232 Resp: 182478
Ion Ratio Lower Upper
232 100
131 55.5 38.5 57.7
130 2.8 2.2 3.2
166 36.9 25.8 38.8





#59
Diethylphthalate
Concen: 39.429 ng/ul
RT: 15.316 min Scan# 2079
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31

Instrument :

BNA_N

ClientSampleId :





SLCS473

Tgt Ion:149 Resp: 846630
Ion Ratio Lower Upper
149 100
177 20.6 17.8 26.8
150 12.2 10.5 15.7

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#61
Fluorene
Concen: 36.296 ng/ul
RT: 15.540 min Scan# 2117
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31


Tgt Ion:166 Resp: 786623
Ion Ratio Lower Upper
166 100
165 98.3 77.7 116.5
167 13.6 11.1 16.7





#62
4-Chlorophenyl-phenylether
Concen: 36.988 ng/ul
RT: 15.534 min Scan# 2116
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Instrument :
BNA_N
ClientSampleId :
SLCS473




Tgt Ion:204 Resp: 36244
Ion Ratio Lower Upper
204 100
206 32.3 26.7 40.1
141 70.3 50.2 75.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#63
4-Nitroaniline
Concen: 60.190 ng/ul
RT: 15.545 min Scan# 2118
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Tgt Ion:138 Resp: 199129
Ion Ratio Lower Upper
138 100
92 49.2 41.0 61.6
108 82.2 65.3 97.9





#66
4,6-Dinitro-2-methylphenol
Concen: 37.816 ng/ul
RT: 15.610 min Scan# 2129
Delta R.T. 0.000 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Instrument :
BNA_N
ClientSampleId :
SLCS473




Tgt Ion:198 Resp: 126912
Ion Ratio Lower Upper
198 100
182 4.2 3.0 4.4
77 27.3 20.1 30.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#67
N-Nitrosodiphenylamine
Concen: 35.155 ng/ul
RT: 15.745 min Scan# 2152
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Tgt Ion:169 Resp: 697168
Ion Ratio Lower Upper
169 100
168 63.9 52.4 78.6
167 34.7 27.5 41.3





#68
4-Bromophenyl-phenylether
Concen: 34.910 ng/ul
RT: 16.428 min Scan# 21196
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Instrument :
BNA_N
ClientSampleId :
SLCS473




Tgt Ion:248 Resp: 21196
Ion Ratio Lower Upper
248 100
250 91.4 77.8 116.8
141 88.2 62.6 93.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#69
Hexachlorobenzene
Concen: 34.843 ng/ul
RT: 16.545 min Scan# 2288
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Tgt Ion:284 Resp: 237962
Ion Ratio Lower Upper
284 100
142 44.6 27.8 41.8#
249 35.5 25.7 38.5





#70
Atrazine
Concen: 37.990 ng/ul
RT: 16.692 min Scan# 21
Delta R.T. -0.012 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Instrument :
BNA_N
ClientSampleId :
SLCS473




Tgt Ion:200 Resp: 240339
Ion Ratio Lower Upper
200 100
173 27.5 20.5 30.7
215 46.3 38.9 58.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#71
Pentachlorophenol
Concen: 38.063 ng/ul
RT: 16.881 min Scan# 2345
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Tgt Ion:266 Resp: 157759
Ion Ratio Lower Upper
266 100
264 64.6 50.2 75.2
268 65.5 49.3 73.9





#72
Phenanthrene
Concen: 35.023 ng/u1
RT: 17.275 min Scan# 2412
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Instrument :
BNA_N
ClientSampleId :
SLCS473




Tgt Ion:178 Resp: 1265008
Ion Ratio Lower Upper
178 100
179 15.3 12.2 18.4
176 19.0 15.2 22.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#74
Anthracene
Concen: 35.654 ng/u1
RT: 17.363 min Scan# 2427
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Tgt Ion:178 Resp: 1286422
Ion Ratio Lower Upper
178 100
179 14.8 13.0 19.4
176 17.5 15.5 23.3





#75
1,2,3,4-Tetrachlorobenzene
Concen: 33.251 ng/uL
RT: 13.293 min Scan# 11
Delta R.T. -0.012 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Instrument :
BNA_N
ClientSampleId :
SLCS473




Tgt Ion:216 Resp: 331294
Ion Ratio Lower Upper
216 100
214 75.2 63.2 94.8
218 46.7 37.9 56.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#76
Pentachlorobenzene
Concen: 34.401 ng/uL
RT: 14.816 min Scan# 1994
Delta R.T. 0.000 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Tgt Ion:250 Resp: 298586
Ion Ratio Lower Upper
250 100
248 60.2 53.0 79.4
252 64.7 53.4 80.2





#77
 Carbazole
 Concen: 37.148 ng/ul
 RT: 17.628 min Scan# 2472
 Delta R.T. -0.006 min
 Lab File: BN019609.D
 Acq: 30 Apr 2022 02:31

Instrument :
 BNA_N
 ClientSampleId :
 SLCS473






Tgt Ion:167 Resp: 1218478


Ion	Ratio	Lower	Upper
167	100		
166	21.5	16.1	24.1
139	14.0	10.2	15.2

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 05/02/2022
 Supervised By :Yogesh Patel 05/05/2022






#78
 Di-n-butylphthalate
 Concen: 40.717 ng/ul
 RT: 18.204 min Scan# 2570
 Delta R.T. -0.012 min
 Lab File: BN019609.D
 Acq: 30 Apr 2022 02:31



Tgt Ion:149 Resp: 1428692


Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.1	10.7
104	5.4	4.1	6.1





#80
Fluoranthene
Concen: 35.723 ng/u1
RT: 19.286 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Instrument :
BNA_N
ClientSampleId :
SLCS473




Tgt Ion:202 Resp: 145233
Ion Ratio Lower Upper
202 100
101 11.9 8.9 13.3
100 9.5 6.7 10.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#82
Pyrene
Concen: 35.312 ng/u1
RT: 19.651 min Scan# 2816
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Tgt Ion:202 Resp: 1495385
Ion Ratio Lower Upper
202 100
101 13.6 9.7 14.5
100 11.1 8.0 12.0





#83
Butylbenzylphthalate
Concen: 41.939 ng/ul
RT: 20.557 min Scan# 2970
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Instrument :
BNA_N
ClientSampleId :
SLCS473




Tgt Ion:149 Resp: 65696
Ion Ratio Lower Upper
149 100
91 71.3 58.8 88.2
206 19.9 18.1 27.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#84
3,3'-Dichlorobenzidine
Concen: 36.408 ng/ul
RT: 21.328 min Scan# 3101
Delta R.T. -0.012 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Tgt Ion:252 Resp: 438631
Ion Ratio Lower Upper
252 100
254 66.5 50.8 76.2
126 14.4 8.2 12.4





#85
Benzo(a)anthracene
Concen: 36.696 ng/u1
RT: 21.398 min Scan# 3113
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Instrument :
BNA_N
ClientSampleId :
SLCS473



Tgt Ion:228 Resp: 140992
Ion Ratio Lower Upper
228 100
229 20.9 16.3 24.5
226 29.2 23.2 34.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 42.209 ng/u1
RT: 21.339 min Scan# 3103
Delta R.T. -0.012 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31


Tgt Ion:149 Resp: 1014515
Ion Ratio Lower Upper
149 100
167 26.2 22.5 33.7
279 3.4 3.5 5.3#





#87
Chrysene
Concen: 35.493 ng/u1
RT: 21.451 min Scan# 3122
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Instrument :
BNA_N
ClientSampleId :
SLCS473



Tgt Ion:228 Resp: 1339534
Ion Ratio Lower Upper
228 100
226 31.7 24.2 36.2
229 20.8 16.2 24.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#89
Di-n-octyl phthalate
Concen: 41.539 ng/u1
RT: 22.251 min Scan# 3258
Delta R.T. -0.017 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31


Tgt Ion:149 Resp: 1741452





#90
Benzo(b)fluoranthene
Concen: 38.320 ng/u1
RT: 23.051 min Scan# 31
Delta R.T. -0.012 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Instrument :
BNA_N
ClientSampleId :
SLCS473



Tgt Ion:252 Resp: 151810
Ion Ratio Lower Upper
252 100
253 22.0 16.5 24.7
125 11.5 7.8 11.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#91
Benzo(k)fluoranthene
Concen: 36.916 ng/u1
RT: 23.104 min Scan# 3403
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31


Tgt Ion:252 Resp: 1326583
Ion Ratio Lower Upper
252 100
253 21.5 18.2 27.4
125 11.5 8.2 12.4





#93
Benzo(a)pyrene
Concen: 38.247 ng/ul
RT: 23.663 min Scan# 3499
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Instrument :
BNA_N
ClientSampleId :
SLCS473




Tgt Ion:252 Resp: 1438890
Ion Ratio Lower Upper
252 100
253 20.9 18.7 28.1
125 13.0 9.2 13.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 37.635 ng/ul
RT: 26.192 min Scan# 3928
Delta R.T. -0.012 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Tgt Ion:276 Resp: 1622570
Ion Ratio Lower Upper
276 100
138 30.2 19.5 29.3#
277 24.4 21.4 32.0





#95
Dibenzo(a,h)anthracene
Concen: 37.424 ng/ul
RT: 26.204 min Scan# 3931
Delta R.T. -0.017 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31




Instrument :
BNA_N
ClientSampleId :
SLCS473




Tgt Ion:278 Resp: 1379618
Ion Ratio Lower Upper
278 100
139 19.0 13.8 20.8
279 23.2 18.3 27.5

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 05/02/2022
Supervised By :Yogesh Patel 05/05/2022



#96
Benzo(g,h,i)perylene
Concen: 36.902 ng/ul
RT: 26.939 min Scan# 4055
Delta R.T. -0.006 min
Lab File: BN019609.D
Acq: 30 Apr 2022 02:31



Tgt Ion:276 Resp: 1399389
Ion Ratio Lower Upper
276 100
138 25.1 17.0 25.6
277 25.1 17.8 26.6

