


Data Path : Z:\HPCHEM1\BNA\_N\Data\BN050118\  
Data File : BN001025.D  
Acq On : 01 May 2018 12:08  
Operator : JU/SJ  
Sample : DFTPP55  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP55

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA\_N\METHODS\SOM-EPA-BN042418MA.M  
Title : SVOA CALIBRATION  
Last Update : Thu May 03 08:25:51 2018



AutoFind: Scans 2459, 2460, 2461; Background Corrected with Scan 2454

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 34.7      | 51813   | PASS             |
| 68          | 69           | 0.00         | 2            | 1.8       | 1134    | PASS             |
| 69          | 198          | 0.00         | 100          | 42.6      | 63741   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.4       | 231     | PASS             |
| 127         | 198          | 10           | 80           | 50.4      | 75378   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 149520  | PASS             |
| 199         | 198          | 5            | 9            | 6.9       | 10243   | PASS             |
| 275         | 198          | 10           | 60           | 25.2      | 37624   | PASS             |
| 365         | 198          | 1            | 100          | 3.2       | 4814    | PASS             |
| 441         | 443          | 0.01         | 100          | 78.0      | 18841   | PASS             |
| 442         | 198          | 50           | 100          | 80.7      | 120725  | PASS             |
| 443         | 442          | 15           | 24           | 20.0      | 24162   | PASS             |

| m/z    | Abundance |
|--------|-----------|
| 38.00  | 553.0     |
| 39.00  | 3088.0    |
| 40.00  | 509.0     |
| 44.00  | 1071.0    |
| 50.00  | 10897.0   |
| 51.10  | 42128.0   |
| 52.10  | 2324.0    |
| 55.00  | 377.0     |
| 56.00  | 1363.0    |
| 57.00  | 3035.0    |
| 61.00  | 628.0     |
| 62.00  | 692.0     |
| 63.00  | 1820.0    |
| 65.10  | 1178.0    |
| 68.10  | 930.0     |
| 69.00  | 52016.0   |
| 73.00  | 540.0     |
| 74.00  | 4968.0    |
| 75.00  | 8770.0    |
| 76.10  | 3010.0    |
| 77.10  | 60208.0   |
| 78.10  | 4366.0    |
| 79.00  | 4023.0    |
| 80.00  | 2849.0    |
| 81.00  | 4033.0    |
| 82.00  | 991.0     |
| 83.10  | 1097.0    |
| 85.00  | 927.0     |
| 86.00  | 1083.0    |
| 87.00  | 566.0     |
| 91.00  | 975.0     |
| 92.00  | 1187.0    |
| 93.00  | 6566.0    |
| 94.00  | 419.0     |
| 96.00  | 334.0     |
| 98.00  | 4840.0    |
| 99.00  | 4337.0    |
| 100.00 | 503.0     |
| 101.00 | 2260.0    |
| 103.00 | 821.0     |
| 104.00 | 1475.0    |
| 105.00 | 1413.0    |
| 106.10 | 405.0     |
| 107.00 | 17488.0   |
| 108.00 | 2593.0    |
| 109.00 | 638.0     |
| 110.00 | 31304.0   |
| 111.00 | 5089.0    |
| 112.00 | 666.0     |
| 116.00 | 881.0     |
| 117.00 | 13502.0   |
| 117.90 | 997.0     |
| 122.00 | 1160.0    |
| 123.00 | 1532.0    |
| 124.00 | 816.0     |
| 125.00 | 652.0     |
| 127.00 | 56544.0   |
| 128.00 | 4617.0    |
| 129.00 | 21272.0   |
| 130.00 | 1727.0    |
| 131.00 | 365.0     |
| 134.00 | 600.0     |
| 135.00 | 1780.0    |
| 136.00 | 769.0     |
| 137.00 | 1070.0    |
| 141.00 | 2701.0    |
| 142.00 | 972.0     |
| 143.00 | 575.0     |
| 146.00 | 578.0     |
| 147.00 | 1579.0    |
| 148.00 | 3264.0    |
| 149.00 | 871.0     |
| 151.10 | 355.0     |
| 153.00 | 886.0     |
| 154.00 | 552.0     |
| 155.00 | 1653.0    |
| 156.00 | 2384.0    |
| 157.00 | 560.0     |
| 158.00 | 440.0     |
| 159.00 | 380.0     |
| 160.00 | 925.0     |
| 161.00 | 1306.0    |
| 162.00 | 340.0     |
| 165.00 | 1032.0    |
| 166.00 | 768.0     |
| 167.00 | 5118.0    |
| 168.00 | 3448.0    |
| 169.00 | 608.0     |
| 171.90 | 407.0     |
| 173.00 | 548.0     |
| 174.00 | 1212.0    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP55

|        |          |
|--------|----------|
| 175.00 | 1978.0   |
| 176.00 | 638.0    |
| 177.00 | 972.0    |
| 179.00 | 3895.0   |
| 180.00 | 2833.0   |
| 181.00 | 1372.0   |
| 185.00 | 1951.0   |
| 186.10 | 14319.0  |
| 187.00 | 4159.0   |
| 188.00 | 561.0    |
| 189.00 | 859.0    |
| 191.00 | 597.0    |
| 192.00 | 1324.0   |
| 193.00 | 1342.0   |
| 196.10 | 2773.0   |
| 198.00 | 103280.0 |
| 199.00 | 7856.0   |
| 200.00 | 575.0    |
| 201.50 | 377.0    |
| 203.00 | 636.0    |
| 204.10 | 3708.0   |
| 205.10 | 6122.0   |
| 206.10 | 26736.0  |
| 207.10 | 4203.0   |
| 208.00 | 860.0    |
| 210.00 | 456.0    |
| 211.00 | 932.0    |
| 216.00 | 591.0    |
| 217.00 | 6586.0   |
| 218.00 | 999.0    |
| 221.00 | 3588.0   |
| 223.00 | 1550.0   |
| 224.10 | 14101.0  |
| 225.10 | 3547.0   |
| 226.10 | 374.0    |
| 227.00 | 5576.0   |
| 228.00 | 911.0    |
| 229.00 | 1193.0   |
| 231.00 | 553.0    |
| 234.10 | 305.0    |
| 235.00 | 403.0    |
| 236.00 | 320.0    |
| 237.00 | 623.0    |
| 242.00 | 658.0    |
| 243.10 | 775.0    |
| 244.10 | 10366.0  |
| 245.00 | 1503.0   |
| 246.00 | 2230.0   |
| 247.00 | 516.0    |
| 249.00 | 381.0    |
| 255.00 | 53872.0  |
| 256.00 | 8076.0   |
| 257.00 | 627.0    |
| 258.00 | 2830.0   |
| 259.00 | 501.0    |
| 265.00 | 1154.0   |
| 273.10 | 1712.0   |
| 274.00 | 4415.0   |
| 275.00 | 24360.0  |
| 276.10 | 3438.0   |
| 277.00 | 2035.0   |
| 278.00 | 361.0    |
| 285.10 | 376.0    |
| 292.90 | 361.0    |
| 296.00 | 6037.0   |
| 297.00 | 967.0    |
| 303.00 | 734.0    |
| 315.00 | 771.0    |
| 316.00 | 353.0    |
| 323.00 | 1933.0   |
| 324.00 | 431.0    |
| 326.90 | 368.0    |
| 334.00 | 1445.0   |
| 345.90 | 388.0    |
| 352.00 | 572.0    |
| 353.00 | 460.0    |
| 354.00 | 660.0    |
| 365.00 | 2713.0   |
| 366.00 | 523.0    |
| 372.10 | 1023.0   |
| 402.00 | 550.0    |
| 403.00 | 644.0    |
| 421.00 | 468.0    |
| 422.10 | 409.0    |
| 423.10 | 3554.0   |
| 424.00 | 1161.0   |
| 441.10 | 9309.0   |
| 442.10 | 61608.0  |
| 443.10 | 12864.0  |
| 444.10 | 1216.0   |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP55

| m/z    | Abundance |
|--------|-----------|
| 38.00  | 778.0     |
| 39.10  | 4445.0    |
| 40.00  | 572.0     |
| 44.00  | 977.0     |
| 49.00  | 573.0     |
| 50.00  | 15794.0   |
| 51.10  | 61136.0   |
| 52.10  | 3233.0    |
| 55.00  | 466.0     |
| 56.00  | 1784.0    |
| 57.00  | 4566.0    |
| 61.00  | 755.0     |
| 62.10  | 983.0     |
| 63.00  | 2865.0    |
| 64.00  | 435.0     |
| 65.00  | 1549.0    |
| 68.00  | 1345.0    |
| 69.00  | 74816.0   |
| 70.00  | 368.0     |
| 73.00  | 752.0     |
| 74.00  | 7423.0    |
| 75.00  | 12439.0   |
| 76.10  | 4513.0    |
| 77.10  | 89520.0   |
| 78.10  | 6530.0    |
| 79.00  | 5537.0    |
| 80.00  | 4562.0    |
| 81.00  | 6117.0    |
| 82.00  | 1621.0    |
| 83.00  | 1511.0    |
| 85.00  | 1300.0    |
| 86.00  | 1707.0    |
| 86.90  | 787.0     |
| 87.90  | 413.0     |
| 91.00  | 1509.0    |
| 92.00  | 1533.0    |
| 93.00  | 9585.0    |
| 94.00  | 629.0     |
| 96.10  | 493.0     |
| 98.00  | 7494.0    |
| 99.00  | 6368.0    |
| 100.00 | 599.0     |
| 101.00 | 3350.0    |
| 103.00 | 1155.0    |
| 103.90 | 2218.0    |
| 105.00 | 2092.0    |
| 106.10 | 754.0     |
| 107.00 | 26616.0   |
| 108.00 | 4273.0    |
| 109.20 | 857.0     |
| 110.00 | 49224.0   |
| 111.00 | 7273.0    |
| 112.00 | 873.0     |
| 113.10 | 320.0     |
| 116.00 | 1381.0    |
| 117.00 | 20624.0   |
| 118.00 | 1499.0    |
| 119.90 | 342.0     |
| 122.00 | 1568.0    |
| 123.00 | 2411.0    |
| 124.00 | 1136.0    |
| 125.00 | 1087.0    |
| 127.00 | 88464.0   |
| 128.00 | 6828.0    |
| 129.00 | 32872.0   |
| 130.00 | 2946.0    |
| 131.00 | 609.0     |
| 132.00 | 316.0     |
| 134.00 | 945.0     |
| 135.00 | 2846.0    |
| 136.00 | 1170.0    |
| 137.00 | 1503.0    |
| 137.90 | 342.0     |
| 140.00 | 421.0     |
| 141.00 | 4424.0    |
| 142.10 | 1380.0    |
| 143.10 | 982.0     |
| 146.00 | 815.0     |
| 147.00 | 2151.0    |
| 148.00 | 4915.0    |
| 149.00 | 1205.0    |
| 150.00 | 321.0     |
| 151.10 | 606.0     |
| 151.90 | 376.0     |
| 153.00 | 1458.0    |
| 154.00 | 906.0     |
| 155.00 | 2537.0    |
| 156.10 | 3585.0    |
| 157.00 | 777.0     |
| 158.00 | 774.0     |
| 158.90 | 586.0     |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP55

|        |          |
|--------|----------|
| 160.00 | 1289.0   |
| 161.00 | 2144.0   |
| 162.00 | 607.0    |
| 165.00 | 1733.0   |
| 166.10 | 1237.0   |
| 167.10 | 8644.0   |
| 168.00 | 5001.0   |
| 169.00 | 775.0    |
| 170.00 | 357.0    |
| 171.00 | 396.0    |
| 172.00 | 696.0    |
| 173.00 | 999.0    |
| 174.00 | 1955.0   |
| 175.00 | 3220.0   |
| 176.00 | 863.0    |
| 177.00 | 1596.0   |
| 178.00 | 500.0    |
| 179.00 | 6492.0   |
| 180.00 | 4625.0   |
| 181.00 | 1977.0   |
| 182.00 | 354.0    |
| 184.00 | 565.0    |
| 185.10 | 3165.0   |
| 186.10 | 24192.0  |
| 187.00 | 7274.0   |
| 188.10 | 753.0    |
| 189.00 | 1380.0   |
| 191.00 | 726.0    |
| 192.00 | 1977.0   |
| 193.00 | 2146.0   |
| 194.00 | 539.0    |
| 194.90 | 316.0    |
| 196.00 | 5101.0   |
| 198.00 | 173760.0 |
| 199.00 | 11662.0  |
| 200.00 | 887.0    |
| 201.60 | 679.0    |
| 203.10 | 1160.0   |
| 204.10 | 5946.0   |
| 205.10 | 10410.0  |
| 206.10 | 45456.0  |
| 207.10 | 6903.0   |
| 208.00 | 1558.0   |
| 209.00 | 539.0    |
| 210.00 | 616.0    |
| 211.10 | 1877.0   |
| 211.80 | 385.0    |
| 215.10 | 497.0    |
| 216.00 | 926.0    |
| 217.00 | 11972.0  |
| 218.00 | 1678.0   |
| 221.10 | 6262.0   |
| 223.00 | 2312.0   |
| 224.10 | 23920.0  |
| 225.10 | 6027.0   |
| 226.10 | 661.0    |
| 227.00 | 9977.0   |
| 228.00 | 1339.0   |
| 229.00 | 2288.0   |
| 231.00 | 818.0    |
| 234.00 | 704.0    |
| 235.00 | 709.0    |
| 236.00 | 481.0    |
| 237.00 | 827.0    |
| 239.00 | 488.0    |
| 240.90 | 554.0    |
| 242.10 | 1372.0   |
| 243.10 | 1337.0   |
| 244.10 | 18536.0  |
| 245.10 | 2534.0   |
| 246.00 | 3603.0   |
| 247.00 | 794.0    |
| 249.00 | 671.0    |
| 253.10 | 457.0    |
| 255.00 | 94880.0  |
| 256.00 | 14029.0  |
| 257.00 | 1090.0   |
| 258.00 | 5083.0   |
| 259.00 | 818.0    |
| 265.00 | 1857.0   |
| 266.00 | 318.0    |
| 273.00 | 2814.0   |
| 274.10 | 7935.0   |
| 275.10 | 41456.0  |
| 276.10 | 5783.0   |
| 277.00 | 3743.0   |
| 278.00 | 615.0    |
| 283.00 | 484.0    |
| 285.10 | 630.0    |
| 293.00 | 900.0    |
| 295.10 | 358.0    |
| 296.00 | 11579.0  |
| 297.00 | 1630.0   |
| 303.00 | 1319.0   |
| 304.00 | 301.0    |

Instrument :

BNA\_N

ClientSampleId :

DFTPP55

|        |          |
|--------|----------|
| 314.00 | 631.0    |
| 315.00 | 1322.0   |
| 316.10 | 674.0    |
| 321.00 | 352.0    |
| 323.10 | 3705.0   |
| 324.00 | 781.0    |
| 327.00 | 786.0    |
| 328.00 | 326.0    |
| 333.00 | 357.0    |
| 334.10 | 2182.0   |
| 335.00 | 655.0    |
| 341.00 | 350.0    |
| 346.00 | 913.0    |
| 352.00 | 1230.0   |
| 353.10 | 739.0    |
| 354.10 | 1231.0   |
| 365.00 | 5295.0   |
| 366.00 | 773.0    |
| 371.00 | 312.0    |
| 372.10 | 1801.0   |
| 373.00 | 525.0    |
| 383.00 | 566.0    |
| 402.00 | 867.0    |
| 403.10 | 1030.0   |
| 404.00 | 436.0    |
| 421.00 | 950.0    |
| 422.10 | 1169.0   |
| 423.10 | 7335.0   |
| 424.10 | 1729.0   |
| 441.10 | 20376.0  |
| 442.10 | 128536.0 |
| 443.10 | 26312.0  |
| 444.10 | 2260.0   |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP55

| m/z    | Abundance |
|--------|-----------|
| 38.00  | 680.0     |
| 39.10  | 3417.0    |
| 40.00  | 624.0     |
| 44.00  | 1097.0    |
| 49.00  | 574.0     |
| 50.10  | 13512.0   |
| 51.10  | 52176.0   |
| 52.00  | 2557.0    |
| 55.00  | 450.0     |
| 56.00  | 1558.0    |
| 57.00  | 3763.0    |
| 61.00  | 812.0     |
| 62.00  | 882.0     |
| 63.00  | 2320.0    |
| 64.10  | 381.0     |
| 65.10  | 1349.0    |
| 68.20  | 1127.0    |
| 69.00  | 64392.0   |
| 70.00  | 327.0     |
| 73.00  | 636.0     |
| 74.00  | 6734.0    |
| 75.00  | 10492.0   |
| 76.10  | 4122.0    |
| 77.10  | 79416.0   |
| 78.10  | 5554.0    |
| 79.00  | 5180.0    |
| 80.00  | 3837.0    |
| 81.00  | 5367.0    |
| 82.00  | 1374.0    |
| 83.00  | 1201.0    |
| 85.00  | 1228.0    |
| 86.00  | 1641.0    |
| 87.00  | 766.0     |
| 91.00  | 1273.0    |
| 92.00  | 1447.0    |
| 93.00  | 8379.0    |
| 94.00  | 629.0     |
| 96.10  | 397.0     |
| 98.00  | 6933.0    |
| 99.00  | 5605.0    |
| 100.00 | 546.0     |
| 101.00 | 3164.0    |
| 103.00 | 1001.0    |
| 104.00 | 2025.0    |
| 105.00 | 1790.0    |
| 106.00 | 622.0     |
| 107.00 | 23312.0   |
| 108.00 | 3648.0    |
| 109.10 | 758.0     |
| 110.00 | 44032.0   |
| 111.00 | 6738.0    |
| 112.00 | 718.0     |
| 116.00 | 1316.0    |
| 117.00 | 18592.0   |
| 118.00 | 1286.0    |
| 119.90 | 308.0     |
| 122.00 | 1775.0    |
| 123.00 | 2305.0    |
| 123.90 | 891.0     |
| 125.00 | 990.0     |
| 127.00 | 81128.0   |
| 128.00 | 5989.0    |
| 129.00 | 29688.0   |
| 130.00 | 2382.0    |
| 131.00 | 523.0     |
| 132.00 | 314.0     |
| 134.00 | 820.0     |
| 135.00 | 2574.0    |
| 136.00 | 943.0     |
| 137.10 | 1695.0    |
| 138.00 | 346.0     |
| 140.00 | 332.0     |
| 141.00 | 3961.0    |
| 142.00 | 1286.0    |
| 142.90 | 979.0     |
| 146.00 | 634.0     |
| 147.00 | 2153.0    |
| 148.00 | 5010.0    |
| 149.00 | 980.0     |
| 150.00 | 303.0     |
| 150.90 | 507.0     |
| 151.80 | 370.0     |
| 153.00 | 1353.0    |
| 154.00 | 842.0     |
| 155.00 | 2154.0    |
| 156.00 | 3475.0    |
| 157.10 | 655.0     |
| 158.00 | 776.0     |
| 159.00 | 678.0     |
| 160.10 | 1238.0    |
| 161.00 | 1857.0    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP55

|        |          |
|--------|----------|
| 162.00 | 544.0    |
| 165.00 | 1440.0   |
| 166.00 | 1224.0   |
| 167.00 | 8220.0   |
| 168.00 | 4616.0   |
| 169.00 | 891.0    |
| 170.90 | 388.0    |
| 172.00 | 774.0    |
| 173.00 | 1106.0   |
| 174.00 | 1748.0   |
| 175.10 | 3252.0   |
| 175.90 | 973.0    |
| 177.00 | 1518.0   |
| 178.00 | 444.0    |
| 179.00 | 6032.0   |
| 180.00 | 4561.0   |
| 181.00 | 2131.0   |
| 182.10 | 386.0    |
| 183.90 | 419.0    |
| 185.00 | 3002.0   |
| 186.10 | 23848.0  |
| 187.00 | 6759.0   |
| 188.10 | 609.0    |
| 189.00 | 1391.0   |
| 191.00 | 757.0    |
| 192.00 | 2187.0   |
| 193.00 | 2213.0   |
| 194.00 | 434.0    |
| 196.10 | 4965.0   |
| 198.00 | 171520.0 |
| 199.00 | 11213.0  |
| 200.00 | 873.0    |
| 201.40 | 789.0    |
| 203.00 | 1315.0   |
| 204.00 | 5889.0   |
| 205.00 | 10174.0  |
| 206.10 | 43824.0  |
| 207.10 | 6291.0   |
| 208.00 | 1601.0   |
| 209.00 | 466.0    |
| 210.10 | 640.0    |
| 211.10 | 1733.0   |
| 211.90 | 334.0    |
| 214.90 | 596.0    |
| 216.00 | 1008.0   |
| 217.00 | 11447.0  |
| 218.10 | 1340.0   |
| 221.10 | 6987.0   |
| 221.70 | 1732.0   |
| 223.00 | 2676.0   |
| 224.10 | 24120.0  |
| 225.10 | 6057.0   |
| 226.10 | 760.0    |
| 227.00 | 9988.0   |
| 228.00 | 1570.0   |
| 229.00 | 2063.0   |
| 230.00 | 316.0    |
| 231.00 | 962.0    |
| 234.00 | 572.0    |
| 235.00 | 875.0    |
| 236.00 | 455.0    |
| 237.10 | 950.0    |
| 239.00 | 412.0    |
| 240.00 | 400.0    |
| 241.00 | 703.0    |
| 242.00 | 1395.0   |
| 243.10 | 1482.0   |
| 244.10 | 19504.0  |
| 245.10 | 2592.0   |
| 246.00 | 3814.0   |
| 247.00 | 802.0    |
| 249.00 | 608.0    |
| 253.10 | 537.0    |
| 255.10 | 99128.0  |
| 256.10 | 14179.0  |
| 257.00 | 1056.0   |
| 258.00 | 5110.0   |
| 259.00 | 1007.0   |
| 265.00 | 2059.0   |
| 266.00 | 367.0    |
| 272.10 | 330.0    |
| 273.00 | 3227.0   |
| 274.10 | 8187.0   |
| 275.10 | 47056.0  |
| 276.10 | 6086.0   |
| 277.00 | 4044.0   |
| 278.00 | 715.0    |
| 283.00 | 433.0    |
| 284.10 | 307.0    |
| 285.10 | 789.0    |
| 293.00 | 1053.0   |
| 296.00 | 12851.0  |
| 297.10 | 1638.0   |
| 303.10 | 1524.0   |
| 304.00 | 438.0    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP55



|        |          |
|--------|----------|
| 314.00 | 717.0    |
| 315.00 | 1578.0   |
| 316.10 | 838.0    |
| 321.00 | 484.0    |
| 323.10 | 4283.0   |
| 324.10 | 770.0    |
| 327.00 | 845.0    |
| 328.00 | 390.0    |
| 333.10 | 503.0    |
| 334.10 | 3015.0   |
| 335.00 | 737.0    |
| 341.00 | 530.0    |
| 346.00 | 1034.0   |
| 352.00 | 1316.0   |
| 353.10 | 868.0    |
| 354.00 | 1580.0   |
| 355.00 | 322.0    |
| 365.00 | 6434.0   |
| 366.00 | 1034.0   |
| 371.00 | 366.0    |
| 372.00 | 2507.0   |
| 373.00 | 568.0    |
| 383.00 | 631.0    |
| 390.10 | 307.0    |
| 402.10 | 879.0    |
| 403.10 | 1538.0   |
| 404.00 | 507.0    |
| 421.00 | 1299.0   |
| 422.00 | 1258.0   |
| 423.10 | 9271.0   |
| 424.10 | 2101.0   |
| 441.10 | 26840.0  |
| 442.10 | 172032.0 |
| 443.10 | 33312.0  |
| 444.10 | 3149.0   |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP55

Scan 2454 (17.422 min): BN001025.D  
DFTPP55

| m/z    | Abundance |
|--------|-----------|
| 40.00  | 456.0     |
| 44.00  | 1128.0    |
| 207.00 | 612.0     |

**Instrument :**  
BNA\_N  
**ClientSampleId :**  
DFTPP55