




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050121\
Data File : BN014509.D
Acq On : 01 May 2021 10:28
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 02 05:04:16 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun May 02 05:03:56 2021
Response via : Initial Calibration



TIC: BN014509.D\data.ms

(70) Pentachlorophenol (C)




16.663min (0.000) 24593.16 ng

response 399705

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	64.52
264.00	61.60	61.63
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050121\
Data File : BN014509.D
Acq On : 01 May 2021 10:28
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 02 05:04:16 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun May 02 05:03:56 2021
Response via : Initial Calibration



TIC: BN014509.D\data.ms

(77) Benzidine

19.269min (0.000) 0.00 ng

response 2188407

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.98
183.00	13.20	11.43
0.00	0.00	0.00