



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050121\
 Data File : BN014509.D
 Acq On : 01 May 2021 10:28
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN043021.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Apr 30 17:46:00 2021






AutoFind: Scans 2445, 2446, 2447; Background Corrected with Scan 2438

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	25.5	80067	PASS
68	69	0.00	2	1.7	1723	PASS
69	198	0.00	100	32.8	102763	PASS
70	69	0.00	2	0.6	604	PASS
127	198	10	80	45.9	143960	PASS
197	198	0.00	2	0.2	615	PASS
198	198	100	100	100.0	313600	PASS
199	198	5	9	6.8	21325	PASS
275	198	10	60	26.2	82229	PASS
365	198	1	100	3.2	9902	PASS
441	198	0.01	100	14.1	44237	PASS
442	442	50	100	100.0	291307	PASS
443	442	15	24	17.5	51107	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050121\
Data File : BN014509.D
Acq On : 01 May 2021 10:28
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: May 02 05:04:16 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun May 02 05:03:56 2021
Response via : Initial Calibration



TIC: BN014509.D\data.ms

(70) Pentachlorophenol (C)

16.663min (0.000) 24593.16 ng




response 399705

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	64.52
264.00	61.60	61.63
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050121\
 Data File : BN014509.D
 Acq On : 01 May 2021 10:28
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Quant Time: May 02 05:04:16 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun May 02 05:03:56 2021
 Response via : Initial Calibration



TIC: BN014509.D\data.ms

(77) Benzidine

19.269min (0.000) 0.00 ng

response 2188407

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.98
183.00	13.20	11.43
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
5/1/2021	BNA_N	<u>BN014509.D</u>
Compound Name	Response	Retention Time
DDT	1069386	20.516
DDD	3135	20.127
DDE	1092	19.563
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
4227	1073613	0.39