


Data Path : Z:\HPCHEM1\BNA_N\Data\BN050218\
 Data File : BN001060.D
 Acq On : 02 May 2018 09:06
 Operator : JU/SJ
 Sample : DFTPP56
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP56

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN042418MA.M
 Title : SVOA CALIBRATION
 Last Update : Thu May 03 08:25:51 2018



AutoFind: Scans 2460, 2461, 2462; Background Corrected with Scan 2454

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.8	65842	PASS
68	69	0.00	2	1.8	1463	PASS
69	198	0.00	100	40.8	81954	PASS
70	69	0.00	2	0.6	493	PASS
127	198	10	80	49.0	98285	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	200661	PASS
199	198	5	9	6.7	13518	PASS
275	198	10	60	26.3	52770	PASS
365	198	1	100	3.6	7270	PASS
441	443	0.01	100	81.5	30541	PASS
442	198	50	100	96.4	193514	PASS
443	442	15	24	19.4	37472	PASS

m/z	Abundance
38.00	944.0
39.00	5719.0
40.00	620.0
41.00	359.0
44.00	1065.0
49.00	744.0
50.00	19320.0
51.10	77096.0
52.00	4095.0
55.00	625.0
56.00	2283.0
57.00	5171.0
61.00	984.0
62.00	1322.0
63.00	3553.0
64.00	484.0
65.00	1707.0
68.10	1750.0
69.00	94184.0
70.00	553.0
73.00	971.0
74.00	10032.0
75.00	15646.0
76.00	5670.0
77.10	112048.0
78.10	8121.0
79.00	7259.0
80.00	5222.0
81.00	7656.0
82.00	1990.0
83.00	1853.0
85.00	1870.0
86.00	2070.0
87.00	1064.0
87.90	402.0
91.00	1982.0
92.00	2063.0
93.00	11922.0
94.00	900.0
95.10	420.0
96.10	591.0
98.00	9430.0
99.00	7766.0
100.00	706.0
101.00	4366.0
102.00	329.0
103.00	1393.0
104.00	2679.0
105.00	2622.0
106.00	869.0
107.00	33232.0
108.00	5459.0
109.10	1153.0
110.00	60632.0
111.00	9304.0
112.00	1107.0
112.90	328.0
116.00	1789.0
117.00	24960.0
118.00	1969.0
120.00	367.0
122.00	2103.0
123.00	3183.0
123.90	1384.0
125.00	1325.0
127.00	108456.0
128.00	8478.0
129.00	41416.0
130.00	3580.0
131.00	730.0
132.00	344.0
134.00	1249.0
135.00	3532.0
136.00	1458.0
137.00	1816.0
138.00	418.0
140.00	489.0
141.00	5332.0
142.00	1641.0
143.00	1184.0
144.00	316.0
145.00	329.0
146.00	1107.0
147.00	2803.0
148.00	5739.0
149.00	1300.0
149.90	392.0
151.20	664.0
151.80	546.0
153.00	1659.0
154.00	1361.0

Instrument :
BNA_N
ClientSampleId :
DFTPP56

155.00	2940.0
156.10	4268.0
157.10	975.0
158.00	1024.0
159.00	753.0
160.00	1685.0
161.00	2509.0
162.00	785.0
165.00	2010.0
166.00	1470.0
167.10	10419.0
168.00	5377.0
169.00	1059.0
170.00	306.0
171.00	522.0
172.00	1064.0
173.00	1205.0
174.00	2249.0
175.00	4394.0
176.00	1162.0
177.00	1998.0
178.00	600.0
179.00	7323.0
180.00	5241.0
181.00	2591.0
182.00	379.0
182.80	324.0
184.00	569.0
185.00	4040.0
186.00	28456.0
187.00	8371.0
188.00	876.0
189.00	1755.0
191.00	785.0
192.00	2346.0
193.10	2809.0
194.00	662.0
195.20	369.0
196.00	6290.0
198.00	207488.0
199.00	14372.0
200.00	1286.0
201.40	826.0
203.00	1485.0
204.00	6935.0
205.10	12229.0
206.10	53936.0
207.10	7703.0
208.00	1969.0
209.00	678.0
210.10	772.0
211.00	2056.0
215.00	596.0
216.10	1152.0
217.00	13865.0
218.00	1763.0
221.10	8262.0
221.90	2041.0
223.00	2977.0
224.10	28320.0
225.10	7355.0
226.10	907.0
227.00	11830.0
228.00	1971.0
229.00	2571.0
230.00	345.0
231.10	1000.0
234.00	809.0
235.00	926.0
236.00	584.0
237.00	1147.0
239.00	457.0
240.00	378.0
241.00	676.0
242.10	1452.0
243.10	1759.0
244.10	21584.0
245.10	2818.0
246.00	4029.0
247.00	895.0
249.00	851.0
253.10	502.0
255.00	111856.0
256.10	16728.0
257.10	1265.0
258.00	6336.0
259.00	1096.0
265.10	2450.0
266.00	304.0
271.10	329.0
272.00	341.0
273.00	2869.0
274.10	9366.0
275.10	49624.0
276.10	6596.0

Instrument :
BNA_N
ClientSampleId :
DFTPP56

277.00	4568.0
278.00	693.0
283.00	441.0
284.00	392.0
285.00	726.0
293.10	911.0
295.10	344.0
296.00	13733.0
297.00	2106.0
303.10	1632.0
304.00	453.0
314.00	711.0
315.00	1491.0
316.00	799.0
321.20	303.0
322.00	313.0
323.10	4222.0
324.00	843.0
327.00	845.0
328.10	521.0
332.10	315.0
333.10	430.0
334.10	2735.0
335.10	662.0
341.10	527.0
346.10	976.0
352.10	1303.0
353.00	976.0
354.00	1474.0
355.00	335.0
365.00	6271.0
366.00	821.0
371.10	367.0
372.00	2357.0
373.00	647.0
383.00	687.0
390.10	308.0
402.10	936.0
403.00	1290.0
404.10	510.0
421.00	1257.0
422.10	1287.0
423.10	8437.0
424.10	1911.0
441.10	24024.0
442.10	151680.0
443.10	30320.0
444.10	3065.0

Instrument :
BNA_N
ClientSampleId :
DFTPP56

m/z	Abundance
37.00	334.0
38.00	910.0
39.00	5259.0
40.00	630.0
44.00	961.0
49.10	659.0
50.00	19488.0
51.10	77176.0
52.10	4043.0
55.00	489.0
56.00	2396.0
57.00	5751.0
61.00	1074.0
62.00	1259.0
63.00	3344.0
64.10	431.0
65.10	1748.0
66.90	325.0
68.00	1717.0
69.00	97664.0
70.00	571.0
73.10	903.0
74.00	9751.0
75.00	15211.0
76.10	5433.0
77.10	114896.0
78.10	8194.0
79.00	7746.0
80.00	5723.0
81.00	8254.0
82.00	2039.0
83.00	2074.0
85.00	2096.0
86.00	2302.0
87.00	966.0
87.90	391.0
89.00	302.0
91.00	1691.0
92.00	2061.0
93.00	12761.0
94.00	965.0
95.10	342.0
96.10	586.0
98.00	9456.0
99.00	7763.0
100.00	655.0
101.00	4407.0
103.00	1673.0
104.00	2684.0
105.00	2562.0
106.10	933.0
107.00	34808.0
108.00	5731.0
109.10	1115.0
110.00	65488.0
111.00	9123.0
112.00	1421.0
113.00	372.0
116.00	1797.0
117.00	26976.0
118.00	2236.0
119.00	372.0
120.00	413.0
122.00	2140.0
123.00	3542.0
124.00	1667.0
125.00	1561.0
127.00	118000.0
128.00	9114.0
129.00	45184.0
130.00	3903.0
131.00	825.0
132.00	422.0
134.00	1453.0
135.00	3846.0
136.00	1432.0
137.10	1937.0
138.00	438.0
140.00	617.0
141.00	5838.0
142.00	1778.0
142.90	1436.0
144.00	368.0
145.10	313.0
146.00	1219.0
147.00	3189.0
148.00	6889.0
149.00	1499.0
150.00	446.0
151.10	676.0
151.80	710.0

Instrument :
BNA_N
ClientSampleId :
DFTPP56

153.00	1728.0
154.00	1467.0
155.00	3204.0
156.00	4793.0
157.00	930.0
158.00	1007.0
159.00	886.0
160.00	1740.0
161.00	2710.0
162.00	745.0
164.10	414.0
165.00	2258.0
166.10	1903.0
167.00	11925.0
168.00	5878.0
169.00	916.0
170.10	437.0
171.10	605.0
172.00	1240.0
173.00	1338.0
174.00	2645.0
175.10	4781.0
176.00	1249.0
177.00	2192.0
178.00	827.0
179.00	8722.0
180.00	6237.0
181.00	2670.0
182.00	415.0
184.00	827.0
185.00	4219.0
186.10	33560.0
187.00	9617.0
188.10	1022.0
189.00	2269.0
190.00	312.0
191.00	1066.0
192.10	2934.0
193.00	3112.0
194.00	748.0
195.20	363.0
196.10	7290.0
198.00	240640.0
199.00	15941.0
200.00	1128.0
201.50	941.0
203.00	1765.0
204.10	8320.0
205.10	14579.0
206.10	62616.0
207.10	8101.0
208.00	2393.0
209.00	755.0
210.20	924.0
211.00	2244.0
215.00	743.0
216.00	1369.0
217.00	15904.0
218.00	2216.0
221.10	9600.0
223.00	3841.0
224.10	33976.0
225.10	9050.0
226.00	895.0
227.00	14413.0
228.00	1825.0
229.00	3303.0
230.10	385.0
231.10	1239.0
234.00	904.0
235.00	1049.0
236.00	777.0
237.10	1065.0
239.00	638.0
240.00	475.0
241.10	735.0
242.00	1763.0
243.10	2199.0
244.10	26464.0
245.10	3711.0
246.00	5351.0
247.00	1063.0
249.00	981.0
253.00	671.0
255.10	139456.0
256.10	19536.0
257.00	1545.0
258.00	7869.0
259.00	1161.0
265.00	3160.0
266.00	391.0
272.10	440.0
273.10	4096.0
274.10	11300.0
275.10	63680.0

Instrument :
BNA_N
ClientSampleId :
DFTPP56

276.10	8356.0
277.10	5350.0
278.00	1015.0
282.90	582.0
284.00	360.0
285.10	878.0
293.00	1117.0
294.10	305.0
296.00	17544.0
297.00	2510.0
302.00	379.0
303.00	2291.0
304.10	622.0
314.10	959.0
315.00	1916.0
316.10	966.0
321.10	467.0
323.10	5918.0
324.10	1132.0
327.00	1196.0
328.00	520.0
332.00	388.0
333.10	463.0
334.10	3887.0
335.10	1035.0
341.00	800.0
346.00	1408.0
352.00	1651.0
353.10	1321.0
354.10	2218.0
355.00	345.0
365.00	8450.0
366.00	1454.0
371.00	612.0
372.10	3592.0
373.00	762.0
383.00	961.0
390.00	395.0
391.00	382.0
402.00	1287.0
403.00	2033.0
404.10	681.0
421.00	1796.0
422.10	1808.0
423.10	13071.0
424.10	2556.0
425.10	301.0
441.10	37088.0
442.10	232960.0
443.10	45544.0
444.10	4556.0

Instrument :
BNA_N
ClientSampleId :
DFTPP56

m/z	Abundance
38.00	442.0
39.00	2913.0
40.00	599.0
44.00	1041.0
49.00	380.0
50.00	10965.0
51.10	43256.0
52.00	2302.0
55.00	417.0
56.00	1393.0
57.00	3060.0
61.00	642.0
62.00	754.0
63.00	1993.0
65.00	1111.0
68.10	922.0
69.00	54016.0
70.00	356.0
73.00	729.0
74.00	5513.0
75.00	8835.0
76.10	3049.0
77.10	64200.0
78.00	4361.0
79.00	4205.0
80.00	3013.0
81.00	4352.0
82.00	1229.0
83.00	1153.0
84.90	1087.0
86.00	1322.0
87.00	622.0
91.00	1103.0
92.00	1207.0
93.00	7243.0
93.90	587.0
96.10	369.0
98.00	5268.0
99.00	4812.0
100.10	423.0
101.00	2619.0
103.00	869.0
104.00	1715.0
105.00	1549.0
106.10	550.0
107.00	19768.0
108.00	3064.0
109.00	574.0
110.00	37472.0
111.00	5590.0
112.00	747.0
116.00	1023.0
117.00	15546.0
118.00	1193.0
122.00	1120.0
123.00	1970.0
124.00	987.0
125.10	1013.0
127.00	68400.0
128.00	5345.0
129.00	26320.0
130.00	2236.0
131.00	475.0
133.90	821.0
135.00	2452.0
136.00	859.0
137.00	1236.0
141.00	3513.0
142.00	1140.0
143.00	875.0
146.00	654.0
147.00	1961.0
148.00	4295.0
149.00	927.0
151.10	454.0
153.00	1164.0
154.00	819.0
155.00	2076.0
156.00	3147.0
157.00	768.0
158.00	634.0
159.00	436.0
160.00	1210.0
161.00	1796.0
162.00	508.0
165.00	1418.0
166.10	1086.0
167.00	6971.0
168.00	3890.0
169.00	773.0
172.00	645.0

Instrument :
BNA_N
ClientSampleId :
DFTPP56

173.00	819.0
174.00	1469.0
175.00	2806.0
176.10	888.0
177.00	1280.0
178.00	471.0
179.00	5640.0
180.00	3871.0
181.00	1787.0
183.90	411.0
185.10	2654.0
186.00	20480.0
187.10	5958.0
188.00	563.0
189.00	1289.0
191.00	540.0
192.00	2002.0
193.00	2140.0
194.00	442.0
196.00	4342.0
198.00	153856.0
199.00	10241.0
200.10	981.0
201.50	737.0
203.00	1119.0
204.10	5518.0
205.10	9116.0
206.10	40272.0
207.10	5961.0
208.10	1485.0
209.00	481.0
210.10	691.0
211.10	1555.0
215.10	439.0
216.00	878.0
217.00	10709.0
218.00	1431.0
221.00	6331.0
221.90	1403.0
223.00	2433.0
224.10	22136.0
225.10	5619.0
226.10	748.0
227.00	9597.0
228.00	1442.0
229.00	2027.0
231.00	985.0
234.00	594.0
235.00	703.0
236.00	461.0
237.00	723.0
239.00	443.0
240.00	484.0
241.00	603.0
242.00	1234.0
243.10	1292.0
244.10	18520.0
245.00	2400.0
246.00	3846.0
247.00	756.0
249.00	690.0
253.00	460.0
255.00	96040.0
256.00	13869.0
257.00	1133.0
258.00	5452.0
259.00	877.0
265.00	2116.0
266.00	343.0
273.00	3000.0
274.00	8360.0
275.10	45008.0
276.10	5931.0
277.00	3772.0
278.00	661.0
283.00	429.0
285.10	630.0
293.00	886.0
296.00	12521.0
297.00	1552.0
303.00	1552.0
304.00	402.0
314.00	731.0
315.00	1385.0
316.10	791.0
321.00	364.0
322.10	320.0
323.10	4483.0
324.10	908.0
327.00	907.0
328.00	452.0
333.10	543.0
334.10	2844.0
335.00	789.0
341.00	565.0

Instrument :

BNA_N

ClientSampleId :

DFTPP56

346.00	938.0
352.10	1384.0
353.00	833.0
354.10	1563.0
355.00	375.0
365.00	7089.0
366.00	870.0
371.00	340.0
372.10	2660.0
373.10	598.0
383.00	736.0
390.00	356.0
402.00	1244.0
403.10	1662.0
404.00	617.0
421.00	1364.0
422.10	1642.0
423.10	10581.0
424.10	2235.0
441.10	30512.0
442.10	195904.0
443.10	36552.0
444.10	3176.0

Instrument :
BNA_N
ClientSampleId :
DFTPP56

Scan 2454 (17.422 min): BN001060.D
DFTPP56

m/z	Abundance
40.00	448.0
44.00	1275.0
207.10	646.0

Instrument :
BNA_N
ClientSampleId :
DFTPP56