

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050222\
 Data File : BN019639.D
 Acq On : 03 May 2022 02:54
 Operator : CG/JU
 Sample : PB144499BS
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS499

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 05/05/2022
 Supervised By :Yogesh Patel 05/09/2022

Quant Time: May 03 04:06:04 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN042722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 03 03:31:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.869	152	157228	20.000	ng/u1	0.00
20) Naphthalene-d8	10.663	136	661937	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.493	164	397113	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.228	188	798451	20.000	ng/u1	0.00
79) Chrysene-d12	21.410	240	738373	20.000	ng/u1	0.00
88) Perylene-d12	23.763	264	750574	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.340	96	24714	5.932	ng/uL	0.00
4) Pyridine-d5	3.746	84	311234	29.500	ng/u1	0.00
7) Phenol-d5	7.034	99	411451	30.198	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.199	67	254204	32.165	ng/u1	0.00
11) 2-Chlorophenol-d4	7.405	132	340552	31.511	ng/u1	0.00
15) 4-Methylphenol-d8	8.575	113	339008	32.479	ng/u1	0.00
21) Nitrobenzene-d5	9.022	128	166528	34.505	ng/u1	0.00
24) 2-Nitrophenol-d4	9.746	143	171388	36.034	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.281	165	306820	32.226	ng/u1	0.00
31) 4-Chloroaniline-d4	10.793	131	427428	29.574	ng/u1	0.00
46) Dimethylphthalate-d6	13.898	166	927432	34.002	ng/u1	0.00
49) Acenaphthylene-d8	14.181	160	1123210	30.457	ng/u1	0.00
54) 4-Nitrophenol-d4	14.675	143	187193	35.288	ng/u1	0.00
60) Fluorene-d10	15.481	176	788261	32.738	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.592	200	133621	32.781	ng/u1	0.00
73) Anthracene-d10	17.328	188	1183082	31.463	ng/u1	0.00
81) Pyrene-d10	19.616	212	1313554	30.943	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.610	264	1271188	33.131	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.375	88	47784	11.246	ng/uL#	92
5) Pyridine	3.764	79	339736	30.559	ng/u1	94
6) Benzaldehyde	7.011	77	266307	38.135	ng/u1	89
8) Phenol	7.058	94	442070	32.243	ng/u1	96
10) Bis(2-Chloroethyl)ether	7.293	93	337028	31.681	ng/u1	98
12) 2-Chlorophenol	7.434	128	359045	32.713	ng/u1	99
13) 2-Methylphenol	8.305	108	324715	31.986	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.405	45	514615	33.659	ng/u1	97
16) Acetophenone	8.687	105	515523	32.913	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.675	70	253727	34.806	ng/u1	91
18) 4-Methylphenol	8.634	108	358438	32.977	ng/u1	98
19) Hexachloroethane	8.952	117	144409	32.457	ng/u1	88
22) Nitrobenzene	9.063	77	383747	33.724	ng/u1	93
23) Isophorone	9.593	82	700475	34.516	ng/u1	98
25) 2-Nitrophenol	9.775	139	188948	35.180	ng/u1	100
26) 2,4-Dimethylphenol	9.840	107	377533	31.555	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.075	93	430961	32.293	ng/u1	98
29) 2,4-Dichlorophenol	10.304	162	318260	33.341	ng/u1	99
30) Naphthalene	10.710	128	1109229	31.289	ng/u1	100
32) 4-Chloroaniline	10.816	127	422406	29.078	ng/u1	99
33) Hexachlorobutadiene	11.010	225	179176	31.288	ng/u1	98
34) Caprolactam	11.569	113	106127m	39.312	ng/u1	
35) 4-Chloro-3-methylphenol	11.940	107	357725	35.406	ng/u1	99
36) 2-Methylnaphthalene	12.322	142	740812	32.470	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050222\
 Data File : BN019639.D
 Acq On : 03 May 2022 02:54
 Operator : CG/JU
 Sample : PB144499BS
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS499


Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 05/05/2022
 Supervised By :Yogesh Patel 05/09/2022

Quant Time: May 03 04:06:04 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN042722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 03 03:31:44 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.540	142	761675	33.347	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.687	216	343413	31.166	ng/ul	97
40) Hexachlorocyclopentadiene	12.675	237	177909	26.240	ng/ul	95
41) 2,4,6-Trichlorophenol	12.922	196	234715	34.147	ng/ul	92
42) 2,4,5-Trichlorophenol	12.993	196	262020	35.108	ng/ul	97
43) 1,1'-Biphenyl	13.328	154	968493	30.846	ng/ul	99
44) 2-Chloronaphthalene	13.369	162	743902	31.011	ng/ul	98
45) 2-Nitroaniline	13.563	65	227788	39.571	ng/ul	96
47) Dimethylphthalate	13.951	163	906228	33.457	ng/ul	99
48) 2,6-Dinitrotoluene	14.063	165	184831	39.556	ng/ul	97
50) Acenaphthylene	14.210	152	1216936	32.119	ng/ul	99
51) 3-Nitroaniline	14.387	138	196469	41.021	ng/ul	97
52) Acenaphthene	14.551	153	785248	31.667	ng/ul	97
53) 2,4-Dinitrophenol	14.587	184	89927	34.695	ng/ul	86
55) 4-Nitrophenol	14.692	109	160214	38.009	ng/ul	98
56) Dibenzofuran	14.887	168	1090392	31.479	ng/ul	96
57) 2,4-Dinitrotoluene	14.845	165	273856	40.305	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.116	232	201943	37.096	ng/ul	94
59) Diethylphthalate	15.316	149	935786	35.125	ng/ul	97
61) Fluorene	15.534	166	849585	31.595	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.534	204	389069	32.001	ng/ul	93
63) 4-Nitroaniline	15.545	138	211272	51.469	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.610	198	133656	32.238	ng/ul	95
67) N-Nitrosodiphenylamine	15.739	169	749218	30.581	ng/ul	98
68) 4-Bromophenyl-phenylether	16.422	248	236419	31.519	ng/ul#	89
69) Hexachlorobenzene	16.539	284	262477	31.110	ng/ul#	88
70) Atrazine	16.692	200	262809	33.628	ng/ul	97
71) Pentachlorophenol	16.881	266	169963	33.195	ng/ul	98
72) Phenanthrene	17.269	178	1386942	31.083	ng/ul	98
74) Anthracene	17.363	178	1409743	31.628	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.293	216	345443	28.066	ng/uL	98
76) Pentachlorobenzene	14.810	250	320036	29.847	ng/uL	92
77) Carbazole	17.628	167	1335642	32.962	ng/ul	98
78) Di-n-butylphthalate	18.204	149	1601647	36.950	ng/ul	100
80) Fluoranthene	19.286	202	1624495	32.464	ng/ul	99
82) Pyrene	19.645	202	1648524	31.627	ng/ul#	95
83) Butylbenzylphthalate	20.551	149	713221	36.991	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.327	252	466034	31.428	ng/ul#	97
85) Benzo(a)anthracene	21.398	228	1554135	32.863	ng/ul	96
86) Bis(2-ethylhexyl)phtha...	21.339	149	1074734	36.328	ng/ul	97
87) Chrysene	21.445	228	1515801	32.631	ng/ul	99
89) Di-n-octyl phthalate	22.251	149	1843741	35.882	ng/ul	100
90) Benzo(b)fluoranthene	23.051	252	1650821	33.998	ng/ul	98
91) Benzo(k)fluoranthene	23.098	252	1512307	34.336	ng/ul	97
93) Benzo(a)pyrene	23.657	252	1571309	34.077	ng/ul	95
94) Indeno(1,2,3-cd)pyrene	26.186	276	1781081	33.706	ng/ul	93
95) Dibenzo(a,h)anthracene	26.204	278	1546603	34.230	ng/ul	96
96) Benzo(g,h,i)perylene	26.933	276	1526178	32.836	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 11.246 ng/uL
RT: 3.375 min Scan# 49
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Instrument :
BNA_N
ClientSampleId :
SLCS499




Tgt Ion: 88 Resp: 47784
Ion Ratio Lower Upper
88 100
43 37.9 22.4 33.6
58 80.1 61.0 91.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022




#5
Pyridine
Concen: 30.559 ng/uL
RT: 3.764 min Scan# 115
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Tgt Ion: 79 Resp: 339736
Ion Ratio Lower Upper
79 100
52 87.4 65.0 97.4
51 42.8 31.6 47.4





#6
Benzaldehyde
Concen: 38.135 ng/ul
RT: 7.011 min Scan# 667
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Instrument :
BNA_N
ClientSampleId :
SLCS499




Tgt Ion: 77 Resp: 266307
Ion Ratio Lower Upper
77 100
105 102.8 74.7 112.1
106 98.3 70.2 105.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022




#8
Phenol
Concen: 32.243 ng/ul
RT: 7.058 min Scan# 675
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54



Tgt Ion: 94 Resp: 442070
Ion Ratio Lower Upper
94 100
65 30.9 23.6 35.4
66 41.7 31.2 46.8





#10

Bis(2-Chloroethyl)ether

Concen: 31.681 ng/ul

RT: 7.293 min Scan# 715

Delta R.T. 0.000 min

Lab File: BN019639.D

Acq: 03 May 2022 02:54

Instrument :

BNA_N

ClientSampleId :

SLCS499

Tgt Ion: 93 Resp: 337028

Ion Ratio Lower Upper

93 100

63 81.6 63.4 95.0





95 33.3 26.2 39.2

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 05/05/2022

Supervised By :Yogesh Patel 05/09/2022



#12

2-Chlorophenol

Concen: 32.713 ng/ul

RT: 7.434 min Scan# 739

Delta R.T. 0.000 min

Lab File: BN019639.D

Acq: 03 May 2022 02:54




Tgt Ion:128 Resp: 359045


Ion Ratio Lower Upper

128 100

64 47.2 38.6 57.8

130 31.5 25.0 37.4





#13
2-Methylphenol
Concen: 31.986 ng/ul
RT: 8.305 min Scan# 888
Delta R.T. -0.006 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54

Instrument :

BNA_N

ClientSampleId :





SLCS499

Tgt Ion:108 Resp: 324715
Ion Ratio Lower Upper
108 100
107 89.8 71.5 107.3

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022



#14
2,2'-oxybis(1-Chloropropane)
Concen: 33.659 ng/ul
RT: 8.405 min Scan# 904
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54


Tgt Ion: 45 Resp: 514615
Ion Ratio Lower Upper
45 100
77 12.3 10.9 16.3
79 9.5 8.6 13.0





#16
Acetophenone
Concen: 32.913 ng/ul
RT: 8.687 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Instrument :
BNA_N
ClientSampleId :
SLCS499




Tgt Ion:105 Resp: 51552
Ion Ratio Lower Upper
105 100
77 76.2 63.5 95.3
51 30.0 23.5 35.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022




#17
N-Nitroso-di-n-propylamine
Concen: 34.806 ng/ul
RT: 8.675 min Scan# 950
Delta R.T. -0.006 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Tgt Ion: 70 Resp: 253727
Ion Ratio Lower Upper
70 100
42 67.8 47.8 71.6
101 10.0 7.8 11.6
130 23.1 15.8 23.8





#18
4-Methylphenol
Concen: 32.977 ng/ul
RT: 8.634 min Scan# 943
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Instrument :
BNA_N
ClientSampleId :
SLCS499




Tgt Ion:108 Resp: 358438
Ion Ratio Lower Upper
108 100
107 112.0 88.0 132.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022




#19
Hexachloroethane
Concen: 32.457 ng/ul
RT: 8.952 min Scan# 997
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54



Tgt Ion:117 Resp: 144409
Ion Ratio Lower Upper
117 100
201 89.4 83.2 124.8
199 55.7 50.2 75.4









#22
Nitrobenzene
Concen: 33.724 ng/ul
RT: 9.063 min Scan# 1016
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54

Instrument :
BNA_N
ClientSampleId :
SLCS499

Tgt Ion: 77 Resp: 38374
Ion Ratio Lower Upper
77 100
123 48.5 34.5 51.7
65 14.7 11.1 16.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022



#23
Isophorone
Concen: 34.516 ng/ul
RT: 9.593 min Scan# 1106
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54

Tgt Ion: 82 Resp: 700475
Ion Ratio Lower Upper
82 100
95 7.6 5.8 8.8
138 18.3 13.8 20.8









#25
2-Nitrophenol
Concen: 35.180 ng/ul
RT: 9.775 min Scan# 1137
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54

Instrument :
BNA_N
ClientSampleId :
SLCS499

Tgt Ion:139 Resp: 188948
Ion Ratio Lower Upper
139 100
65 54.5 43.7 65.5
109 32.6 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022



#26
2,4-Dimethylphenol
Concen: 31.555 ng/ul
RT: 9.840 min Scan# 1148
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54


Tgt Ion:107 Resp: 377533
Ion Ratio Lower Upper
107 100
121 50.1 38.3 57.5
122 83.2 66.3 99.5





#27
Bis(2-Chloroethoxy)methane
Concen: 32.293 ng/ul
RT: 10.075 min Scan# 1188
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Instrument :
BNA_N
ClientSampleId :
SLCS499



Tgt Ion: 93 Resp: 43096
Ion Ratio Lower Upper
93 100
95 32.1 25.4 38.0
123 12.0 8.3 12.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022



#29
2,4-Dichlorophenol
Concen: 33.341 ng/ul
RT: 10.304 min Scan# 1227
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54


Tgt Ion:162 Resp: 318260
Ion Ratio Lower Upper
162 100
164 65.7 52.9 79.3
98 38.9 29.8 44.8





#30
Naphthalene
Concen: 31.289 ng/ul
RT: 10.710 min Scan# 1296
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Instrument :
BNA_N
ClientSampleId :
SLCS499



Tgt Ion:128 Resp: 1109229
Ion Ratio Lower Upper
128 100
129 10.8 8.7 13.1
127 13.2 10.6 16.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022



#32
4-Chloroaniline
Concen: 29.078 ng/ul
RT: 10.816 min Scan# 1314
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54


Tgt Ion:127 Resp: 422406
Ion Ratio Lower Upper
127 100
129 32.0 26.1 39.1





#33
Hexachlorobutadiene
Concen: 31.288 ng/ul
RT: 11.010 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Instrument :
BNA_N
ClientSampleId :
SLCS499




Tgt Ion:225 Resp: 179170
Ion Ratio Lower Upper
225 100
223 65.4 51.8 77.6
227 65.1 50.6 76.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022




#34
Caprolactam
Concen: 39.312 ng/ul m
RT: 11.569 min Scan# 1442
Delta R.T. 0.006 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Tgt Ion:113 Resp: 106127
Ion Ratio Lower Upper
113 100
55 182.0 145.5 218.3
56 130.2 106.5 159.7





#35
4-Chloro-3-methylphenol
Concen: 35.406 ng/ul
RT: 11.940 min Scan# 1505
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Instrument :
BNA_N
ClientSampleId :
SLCS499




Tgt Ion:107 Resp: 357725
Ion Ratio Lower Upper
107 100
144 24.8 20.6 31.0
142 78.3 62.2 93.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022




#36
2-Methylnaphthalene
Concen: 32.470 ng/ul
RT: 12.322 min Scan# 1570
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Tgt Ion:142 Resp: 740812
Ion Ratio Lower Upper
142 100
141 90.9 70.6 105.8





#37
1-Methylnaphthalene
Concen: 33.347 ng/ul
RT: 12.540 min Scan# 1607
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Instrument :
BNA_N
ClientSampleId :
SLCS499




Tgt Ion:142 Resp: 76167
Ion Ratio Lower Upper
142 100
141 91.3 70.9 106.3
116 3.8 3.1 4.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 31.166 ng/ul
RT: 12.687 min Scan# 1632
Delta R.T. -0.006 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Tgt Ion:216 Resp: 343413
Ion Ratio Lower Upper
216 100
214 76.9 63.4 95.0
179 23.6 17.8 26.8
108 20.8 15.1 22.7





#40
Hexachlorocyclopentadiene
Concen: 26.240 ng/ul
RT: 12.675 min Scan# 1630
Delta R.T. -0.006 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Instrument :
BNA_N
ClientSampleId :
SLCS499




Tgt Ion:237 Resp: 177909
Ion Ratio Lower Upper
237 100
235 65.2 48.8 73.2
272 13.6 9.3 13.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022




#41
2,4,6-Trichlorophenol
Concen: 34.147 ng/ul
RT: 12.922 min Scan# 1672
Delta R.T. -0.006 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Tgt Ion:196 Resp: 234715
Ion Ratio Lower Upper
196 100
198 90.4 79.8 119.8
200 29.7 25.0 37.6





#42
2,4,5-Trichlorophenol
Concen: 35.108 ng/uI
RT: 12.993 min Scan# 1684
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Instrument :
BNA_N
ClientSampleId :
SLCS499




Tgt Ion:196 Resp: 262020
Ion Ratio Lower Upper
196 100
198 94.1 77.6 116.4
200 29.9 24.4 36.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022




#43
1,1'-Biphenyl
Concen: 30.846 ng/uI
RT: 13.328 min Scan# 1741
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Tgt Ion:154 Resp: 968493
Ion Ratio Lower Upper
154 100
153 39.7 31.5 47.3
76 12.7 10.5 15.7





#44
2-Chloronaphthalene
Concen: 31.011 ng/uI
RT: 13.369 min Scan# 1748
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Instrument :
BNA_N
ClientSampleId :
SLCS499




Tgt Ion:162 Resp: 743902
Ion Ratio Lower Upper
162 100
164 32.1 25.8 38.6
127 41.3 31.4 47.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022




#45
2-Nitroaniline
Concen: 39.571 ng/uI
RT: 13.563 min Scan# 1781
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Tgt Ion: 65 Resp: 227788
Ion Ratio Lower Upper
65 100
92 66.9 54.1 81.1
138 101.9 77.0 115.4





#47
Dimethylphthalate
Concen: 33.457 ng/ul
RT: 13.951 min Scan# 1847
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Instrument :
BNA_N
ClientSampleId :
SLCS499




Tgt Ion:163 Resp: 906228
Ion Ratio Lower Upper
163 100
194 4.0 3.4 5.0
164 9.8 7.7 11.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022




#48
2,6-Dinitrotoluene
Concen: 39.556 ng/ul
RT: 14.063 min Scan# 1866
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Tgt Ion:165 Resp: 184831
Ion Ratio Lower Upper
165 100
89 59.6 50.6 75.8
121 22.5 17.9 26.9





#50
Acenaphthylene
Concen: 32.119 ng/ul
RT: 14.210 min Scan# 1891
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Instrument :
BNA_N
ClientSampleId :
SLCS499




Tgt Ion:152 Resp: 1216930
Ion Ratio Lower Upper
152 100
151 21.0 16.0 24.0
153 13.5 10.7 16.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022




#51
3-Nitroaniline
Concen: 41.021 ng/ul
RT: 14.387 min Scan# 1921
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Tgt Ion:138 Resp: 196469
Ion Ratio Lower Upper
138 100
108 10.8 8.6 12.8
92 112.0 92.9 139.3





#52
Acenaphthene
Concen: 31.667 ng/ul
RT: 14.551 min Scan# 1949
Delta R.T. -0.006 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Instrument :
BNA_N
ClientSampleId :
SLCS499




Tgt Ion:153 Resp: 785248
Ion Ratio Lower Upper
153 100
152 46.5 38.4 57.6
154 84.2 69.7 104.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022




#53
2,4-Dinitrophenol
Concen: 34.695 ng/ul
RT: 14.587 min Scan# 1955
Delta R.T. -0.006 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Tgt Ion:184 Resp: 89927
Ion Ratio Lower Upper
184 100
63 83.6 55.9 83.9
154 74.7 53.1 79.7





#55
4-Nitrophenol
Concen: 38.009 ng/ul
RT: 14.692 min Scan# 1973
Delta R.T. 0.006 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Instrument :
BNA_N
ClientSampleId :
SLCS499



Tgt Ion:109 Resp: 160214
Ion Ratio Lower Upper
109 100
139 128.5 100.1 150.1
65 129.8 104.6 157.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022



#56
Dibenzofuran
Concen: 31.479 ng/ul
RT: 14.887 min Scan# 2006
Delta R.T. -0.006 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54


Tgt Ion:168 Resp: 1090392
Ion Ratio Lower Upper
168 100
139 41.3 31.1 46.7





#57
2,4-Dinitrotoluene
Concen: 40.305 ng/ul
RT: 14.845 min Scan# 1999
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Instrument :
BNA_N
ClientSampleId :
SLCS499




Tgt Ion:165 Resp: 273850
Ion Ratio Lower Upper
165 100
63 48.4 36.3 54.5
182 2.9 2.3 3.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 37.096 ng/ul
RT: 15.116 min Scan# 2045
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54



Tgt Ion:232 Resp: 201943
Ion Ratio Lower Upper
232 100
131 53.1 38.5 57.7
130 2.8 2.2 3.2
166 34.5 25.8 38.8









#59
Diethylphthalate
Concen: 35.125 ng/ul
RT: 15.316 min Scan# 2079
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54

Instrument :
BNA_N
ClientSampleId :
SLCS499

Tgt Ion:149 Resp: 935780
Ion Ratio Lower Upper
149 100
177 20.7 17.8 26.8
150 12.5 10.5 15.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022



#61
Fluorene
Concen: 31.595 ng/ul
RT: 15.534 min Scan# 2116
Delta R.T. -0.006 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54


Tgt Ion:166 Resp: 849585
Ion Ratio Lower Upper
166 100
165 99.1 77.7 116.5
167 13.9 11.1 16.7





#62
4-Chlorophenyl-phenylether
Concen: 32.001 ng/ul
RT: 15.534 min Scan# 2116
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Instrument :
BNA_N
ClientSampleId :
SLCS499




Tgt Ion:204 Resp: 389069
Ion Ratio Lower Upper
204 100
206 32.2 26.7 40.1
141 70.0 50.2 75.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022




#63
4-Nitroaniline
Concen: 51.469 ng/ul
RT: 15.545 min Scan# 2118
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Tgt Ion:138 Resp: 211272
Ion Ratio Lower Upper
138 100
92 51.8 41.0 61.6
108 83.7 65.3 97.9





#66
4,6-Dinitro-2-methylphenol
Concen: 32.238 ng/ul
RT: 15.610 min Scan# 2128
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Instrument :
BNA_N
ClientSampleId :
SLCS499




Tgt Ion:198 Resp: 133650
Ion Ratio Lower Upper
198 100
182 4.4 3.0 4.4
77 28.1 20.1 30.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022




#67
N-Nitrosodiphenylamine
Concen: 30.581 ng/ul
RT: 15.739 min Scan# 2151
Delta R.T. -0.006 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54



Tgt Ion:169 Resp: 749218
Ion Ratio Lower Upper
169 100
168 66.8 52.4 78.6
167 35.7 27.5 41.3






#68
4-Bromophenyl-phenylether
Concen: 31.519 ng/ul
RT: 16.422 min Scan# 2267
Delta R.T. -0.006 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54

Instrument :

BNA_N

ClientSampleId :

SLCS499






Tgt Ion:248 Resp: 236419
Ion Ratio Lower Upper
248 100
250 101.3 77.8 116.8
141 95.9 62.6 93.8


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022




#69
Hexachlorobenzene
Concen: 31.110 ng/ul
RT: 16.539 min Scan# 2287
Delta R.T. -0.006 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Tgt Ion:284 Resp: 262477
Ion Ratio Lower Upper
284 100
142 46.7 27.8 41.8#
249 33.5 25.7 38.5





#70
Atrazine
Concen: 33.628 ng/ul
RT: 16.692 min Scan# 213
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Instrument :
BNA_N
ClientSampleId :
SLCS499




Tgt Ion:200 Resp: 262809
Ion Ratio Lower Upper
200 100
173 28.3 20.5 30.7
215 47.4 38.9 58.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022




#71
Pentachlorophenol
Concen: 33.195 ng/ul
RT: 16.881 min Scan# 2345
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Tgt Ion:266 Resp: 169963
Ion Ratio Lower Upper
266 100
264 64.5 50.2 75.2
268 63.4 49.3 73.9





#72
Phenanthrene
Concen: 31.083 ng/u1
RT: 17.269 min Scan# 2411
Delta R.T. -0.006 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Instrument :
BNA_N
ClientSampleId :
SLCS499



Tgt Ion:178 Resp: 138694
Ion Ratio Lower Upper
178 100
179 16.1 12.2 18.4
176 20.0 15.2 22.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022



#74
Anthracene
Concen: 31.628 ng/u1
RT: 17.363 min Scan# 2427
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54


Tgt Ion:178 Resp: 1409743
Ion Ratio Lower Upper
178 100
179 15.2 13.0 19.4
176 18.6 15.5 23.3





#75
1,2,3,4-Tetrachlorobenzene
Concen: 28.066 ng/uL
RT: 13.293 min Scan# 1735
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Instrument :
BNA_N
ClientSampleId :
SLCS499




Tgt Ion:216 Resp: 34544
Ion Ratio Lower Upper
216 100
214 80.5 63.2 94.8
218 49.0 37.9 56.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022




#76
Pentachlorobenzene
Concen: 29.847 ng/uL
RT: 14.810 min Scan# 1993
Delta R.T. -0.006 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Tgt Ion:250 Resp: 320036
Ion Ratio Lower Upper
250 100
248 61.1 53.0 79.4
252 59.5 53.4 80.2





#77
 Carbazole
 Concen: 32.962 ng/ul
 RT: 17.628 min Scan# 2472
 Delta R.T. 0.000 min
 Lab File: BN019639.D
 Acq: 03 May 2022 02:54

Instrument :
 BNA_N
 ClientSampleId :
 SLCS499






Tgt Ion:167 Resp: 133564


Ion	Ratio	Lower	Upper
167	100		
166	21.0	16.1	24.1
139	13.7	10.2	15.2

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 05/05/2022
 Supervised By :Yogesh Patel 05/09/2022






#78
 Di-n-butylphthalate
 Concen: 36.950 ng/ul
 RT: 18.204 min Scan# 2570
 Delta R.T. 0.000 min
 Lab File: BN019639.D
 Acq: 03 May 2022 02:54



Tgt Ion:149 Resp: 1601647


Ion	Ratio	Lower	Upper
149	100		
150	9.0	7.1	10.7
104	5.3	4.1	6.1





#80
Fluoranthene
Concen: 32.464 ng/u1
RT: 19.286 min Scan# 2754
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Instrument :
BNA_N
ClientSampleId :
SLCS499



Tgt Ion:202 Resp: 162449
Ion Ratio Lower Upper
202 100
101 11.3 8.9 13.3
100 9.0 6.7 10.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022



#82
Pyrene
Concen: 31.627 ng/u1
RT: 19.645 min Scan# 2815
Delta R.T. -0.006 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54


Tgt Ion:202 Resp: 1648524
Ion Ratio Lower Upper
202 100
101 14.2 9.7 14.5
100 12.0 8.0 12.0#





#83
Butylbenzylphthalate
Concen: 36.991 ng/ul
RT: 20.551 min Scan# 2969
Delta R.T. -0.006 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Instrument :
BNA_N
ClientSampleId :
SLCS499




Tgt Ion:149 Resp: 71322
Ion Ratio Lower Upper
149 100
91 71.6 58.8 88.2
206 19.4 18.1 27.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022




#84
3,3'-Dichlorobenzidine
Concen: 31.428 ng/ul
RT: 21.327 min Scan# 3101
Delta R.T. -0.006 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Tgt Ion:252 Resp: 466034
Ion Ratio Lower Upper
252 100
254 65.0 50.8 76.2
126 13.4 8.2 12.4#





#85
Benzo(a)anthracene
Concen: 32.863 ng/ul
RT: 21.398 min Scan# 3113
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Instrument :
BNA_N
ClientSampleId :
SLCS499



Tgt Ion:228 Resp: 1554135
Ion Ratio Lower Upper
228 100
229 19.3 16.3 24.5
226 26.6 23.2 34.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 36.328 ng/ul
RT: 21.339 min Scan# 3103
Delta R.T. -0.006 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54


Tgt Ion:149 Resp: 1074734
Ion Ratio Lower Upper
149 100
167 26.7 22.5 33.7
279 3.6 3.5 5.3





#87
Chrysene
Concen: 32.631 ng/ul
RT: 21.445 min Scan# 3122
Delta R.T. -0.006 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Instrument :
BNA_N
ClientSampleId :
SLCS499



Tgt Ion:228 Resp: 151580
Ion Ratio Lower Upper
228 100
226 29.4 24.2 36.2
229 19.6 16.2 24.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022



#89
Di-n-octyl phthalate
Concen: 35.882 ng/ul
RT: 22.251 min Scan# 3258
Delta R.T. -0.006 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54

Tgt Ion:149 Resp: 1843741






#90
Benzo(b)fluoranthene
Concen: 33.998 ng/u1
RT: 23.051 min Scan# 3394
Delta R.T. -0.006 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54

Instrument :

BNA_N

ClientSampleId :

SLCS499






Tgt Ion:252 Resp: 165082
Ion Ratio Lower Upper
252 100
253 20.9 16.5 24.7
125 11.5 7.8 11.8


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022




#91
Benzo(k)fluoranthene
Concen: 34.336 ng/u1
RT: 23.098 min Scan# 3402
Delta R.T. -0.006 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Tgt Ion:252 Resp: 1512307
Ion Ratio Lower Upper
252 100
253 21.3 18.2 27.4
125 11.7 8.2 12.4





#93
Benzo(a)pyrene
Concen: 34.077 ng/ul
RT: 23.657 min Scan# 3498
Delta R.T. -0.012 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54




Instrument :
BNA_N
ClientSampleId :
SLCS499




Tgt Ion:252 Resp: 1571309
Ion Ratio Lower Upper
252 100
253 20.7 18.7 28.1
125 12.9 9.2 13.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 33.706 ng/ul
RT: 26.186 min Scan# 3927
Delta R.T. -0.006 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54



Tgt Ion:276 Resp: 1781081
Ion Ratio Lower Upper
276 100
138 28.9 19.5 29.3
277 24.1 21.4 32.0






#95
Dibenzo(a,h)anthracene
Concen: 34.230 ng/ul
RT: 26.204 min Scan# 3930
Delta R.T. 0.000 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54

Instrument :

BNA_N

ClientSampleId :




SLCS499




Tgt Ion:278 Resp: 1546600
Ion Ratio Lower Upper
278 100
139 19.2 13.8 20.8
279 24.4 18.3 27.5

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 05/05/2022
Supervised By :Yogesh Patel 05/09/2022



#96
Benzo(g,h,i)perylene
Concen: 32.836 ng/ul
RT: 26.933 min Scan# 4054
Delta R.T. 0.006 min
Lab File: BN019639.D
Acq: 03 May 2022 02:54



Tgt Ion:276 Resp: 1526178
Ion Ratio Lower Upper
276 100
138 24.3 17.0 25.6
277 24.2 17.8 26.6

