

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050321\
 Data File : BN014516.D
 Acq On : 03 May 2021 15:48
 Operator : CG/JU
 Sample : SSTDIC0.2
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SSTDIC0.2

Quant Time: May 03 17:11:45 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN050321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 03 17:11:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.620	152	6501	0.40	ng	0.00
7) Naphthalene-d8	10.391	136	22989	0.40	ng	# 0.00
13) Acenaphthene-d10	14.257	164	11734	0.40	ng	0.00
19) Phenanthrene-d10	17.006	188	23175	0.40	ng	# 0.00
29) Chrysene-d12	21.205	240	18163	0.40	ng	# 0.00
35) Perylene-d12	23.397	264	18827	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.212	112	2622	0.20	ng	0.00
5) Phenol-d6	6.790	99	2541	0.17	ng	0.00
8) Nitrobenzene-d5	8.769	82	2269	0.17	ng	0.00
11) 2-Methylnaphthalene-d10	11.991	152	6439	0.14	ng	0.00
14) 2,4,6-Tribromophenol	15.754	330	540	0.15	ng	0.00
15) 2-Fluorobiphenyl	12.874	172	7876	0.18	ng	0.00
27) Fluoranthene-d10	19.042	212	10960	0.13	ng	0.00
31) Terphenyl-d14	19.652	244	8076	0.19	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.158	88	1739	0.22	ng	# 84
3) n-Nitrosodimethylamine	3.521	42	616	0.19	ng	# 58
6) bis(2-Chloroethyl)ether	7.056	93	3106	0.20	ng	95
9) Naphthalene	10.442	128	11299	0.19	ng	99
10) Hexachlorobutadiene	10.733	225	2161	0.18	ng	# 100
12) 2-Methylnaphthalene	12.066	142	6964	0.18	ng	98
16) Acenaphthylene	13.978	152	7553	0.17	ng	99
17) Acenaphthene	14.321	154	6011	0.18	ng	98
18) Fluorene	15.310	166	7278	0.18	ng	100
20) 4,6-Dinitro-2-methylph...	15.412	198	168	0.10	ng	# 1
21) 4-Bromophenyl-phenylether	16.203	248	1988	0.15	ng	# 84
22) Hexachlorobenzene	16.312	284	2837	0.17	ng	95
23) Atrazine	16.483	200	1284	0.18	ng	96
24) Pentachlorophenol	16.678	266	430	0.13	ng	# 36
25) Phenanthrene	17.043	178	11350	0.17	ng	100
26) Anthracene	17.140	178	8134	0.16	ng	100
28) Fluoranthene	19.071	202	11625	0.16	ng	99
30) Pyrene	19.439	202	11846	0.19	ng	99
32) Benzo(a)anthracene	21.194	228	7470	0.15	ng	99
33) Chrysene	21.237	228	11107	0.19	ng	98
34) Bis(2-ethylhexyl)phtha...	21.130	149	3496	0.18	ng	99
36) Indeno(1,2,3-cd)pyrene	25.592	276	11058	0.16	ng	95
37) Benzo(b)fluoranthene	22.744	252	10353	0.15	ng	# 83
38) Benzo(k)fluoranthene	22.785	252	13978	0.17	ng	# 92
39) Benzo(a)pyrene	23.305	252	11277	0.18	ng	# 46
40) Dibenzo(a,h)anthracene	25.609	278	8811	0.15	ng	# 92
41) Benzo(g,h,i)perylene	26.256	276	11793	0.17	ng	98





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.620 min Scan# 601
 Delta R.T. -0.000 min
 Lab File: BN014516.D
 Acq: 03 May 2021 15:48




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2


Tgt Ion:152 Resp: 6501
 Ion Ratio Lower Upper
 152 100
 150 152.0 123.8 185.8
 115 59.0 47.0 70.4



#2
 1,4-Dioxane
 Concen: 0.22 ng
 RT: 3.158 min Scan# 47
 Delta R.T. 0.000 min
 Lab File: BN014516.D
 Acq: 03 May 2021 15:48

Tgt Ion: 88 Resp: 1739
 Ion Ratio Lower Upper
 88 100
 58 67.9 44.4 66.6#
 43 25.2 15.9 23.9#









#3
n-Nitrosodimethylamine
Concen: 0.19 ng
RT: 3.521 min Scan# 92
Delta R.T. 0.000 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48




Instrument :
BNA_N
ClientSampled :
SSTDIC0.2


Tgt Ion: 42 Resp: 616
Ion Ratio Lower Upper
42 100
74 150.6 175.1 262.7#
44 0.0 7.5 11.3#



#4
2-Fluorophenol
Concen: 0.20 ng
RT: 5.212 min Scan# 302
Delta R.T. -0.008 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48

Tgt Ion: 112 Resp: 2622
Ion Ratio Lower Upper
112 100
64 55.9 41.8 62.6
63 30.5 22.0 33.0









#5
Phenol-d6
Concen: 0.17 ng
RT: 6.790 min Scan# 498
Delta R.T. 0.000 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion: 99 Resp: 2541
Ion Ratio Lower Upper
99 100
42 23.0 13.2 19.8#
71 32.9 26.2 39.2



#6
bis(2-Chloroethyl)ether
Concen: 0.20 ng
RT: 7.056 min Scan# 531
Delta R.T. 0.000 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48

Tgt Ion: 93 Resp: 3106
Ion Ratio Lower Upper
93 100
63 67.9 58.0 87.0
95 31.2 27.0 40.6









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.391 min Scan# 851
Delta R.T. 0.000 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:	136	Resp:	22989
Ion Ratio	Lower	Upper	
136	100		
137	11.3	8.9	13.3
54	7.4	3.8	5.8#
68	5.7	3.8	5.6#



#8
Nitrobenzene-d5
Concen: 0.17 ng
RT: 8.769 min Scan# 723
Delta R.T. 0.000 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48

Tgt Ion:	82	Resp:	2269
Ion Ratio	Lower	Upper	
82	100		
128	45.6	29.5	44.3#
54	52.9	36.9	55.3









#9
Naphthalene
Concen: 0.19 ng
RT: 10.442 min Scan# 855
Delta R.T. 0.000 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:	128	Resp:	11299
Ion Ratio	Lower	Upper	
128	100		
129	11.5	9.0	13.4
127	13.7	10.4	15.6



#10
Hexachlorobutadiene
Concen: 0.18 ng
RT: 10.733 min Scan# 878
Delta R.T. 0.000 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48

Tgt Ion:	225	Resp:	2161
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	63.6	50.9	76.3









#11
2-Methylnaphthalene-d10
Concen: 0.14 ng
RT: 11.991 min Scan# 1048
Delta R.T. 0.000 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:152 Resp: 6439
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.2



#12
2-Methylnaphthalene
Concen: 0.18 ng
RT: 12.066 min Scan# 1065
Delta R.T. 0.004 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48

Tgt Ion:142 Resp: 6964
Ion Ratio Lower Upper
142 100
141 86.5 70.2 105.4
115 35.9 26.7 40.1









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.257 min Scan# 1302
Delta R.T. 0.000 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:164 Resp: 11734
Ion Ratio Lower Upper
164 100
162 100.6 78.6 118.0
160 47.1 35.8 53.8



#14
2,4,6-Tribromophenol
Concen: 0.15 ng
RT: 15.754 min Scan# 1420
Delta R.T. 0.000 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48

Tgt Ion:330 Resp: 540
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 40.2 27.0 40.4









#15
2-Fluorobiphenyl
Concen: 0.18 ng
RT: 12.874 min Scan# 1193
Delta R.T. 0.000 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:	172	Resp:	7876
Ion Ratio	Lower	Upper	
172	100		
171	35.8	27.1	40.7
170	24.9	18.3	27.5



#16
Acenaphthylene
Concen: 0.17 ng
RT: 13.978 min Scan# 1280
Delta R.T. 0.000 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48

Tgt Ion:	152	Resp:	7553
Ion Ratio	Lower	Upper	
152	100		
151	19.9	15.6	23.4
153	13.1	10.5	15.7









#17
Acenaphthene
Concen: 0.18 ng
RT: 14.321 min Scan# 1307
Delta R.T. 0.000 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:154 Resp: 6011
Ion Ratio Lower Upper
154 100
153 115.5 90.2 135.4
152 57.9 46.2 69.2



#18
Fluorene
Concen: 0.18 ng
RT: 15.310 min Scan# 1385
Delta R.T. 0.000 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48

Tgt Ion:166 Resp: 7278
Ion Ratio Lower Upper
166 100
165 98.3 78.4 117.6
167 13.5 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.006 min Scan# 1522
Delta R.T. 0.000 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:188 Resp: 23175
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 8.4 5.2 7.8#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.10 ng
RT: 15.412 min Scan# 1393
Delta R.T. 0.013 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48

Tgt Ion:198 Resp: 168
Ion Ratio Lower Upper
198 100
51 250.0 865.3 1297.9#
105 135.0 93.6 140.4









#21
4-Bromophenyl-phenylether
Concen: 0.15 ng
RT: 16.203 min Scan# 1456
Delta R.T. 0.000 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:248 Resp: 1988
Ion Ratio Lower Upper
248 100
250 96.2 80.5 120.7
141 74.5 37.2 55.8#



#22
Hexachlorobenzene
Concen: 0.17 ng
RT: 16.312 min Scan# 1465
Delta R.T. 0.000 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48

Tgt Ion:284 Resp: 2837
Ion Ratio Lower Upper
284 100
142 36.9 26.6 39.8
249 30.7 23.6 35.4









#23
Atrazine
Concen: 0.18 ng
RT: 16.483 min Scan# 1479
Delta R.T. 0.000 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:200 Resp: 1284
Ion Ratio Lower Upper
200 100
173 29.7 23.0 34.6
215 51.6 38.6 57.8



#24
Pentachlorophenol
Concen: 0.13 ng
RT: 16.678 min Scan# 1495
Delta R.T. 0.012 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48

Tgt Ion:266 Resp: 430
Ion Ratio Lower Upper
266 100
264 58.4 24.2 36.4#
268 73.5 25.2 37.8#









#25
Phenanthrene
Concen: 0.17 ng
RT: 17.043 min Scan# 1525
Delta R.T. 0.000 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:178 Resp: 11350
Ion Ratio Lower Upper
178 100
176 19.0 15.2 22.8
179 15.7 12.4 18.6



#26
Anthracene
Concen: 0.16 ng
RT: 17.140 min Scan# 1533
Delta R.T. 0.000 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48

Tgt Ion:178 Resp: 8134
Ion Ratio Lower Upper
178 100
176 18.2 14.8 22.2
179 15.3 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.13 ng
RT: 19.042 min Scan# 1813
Delta R.T. 0.000 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:212 Resp: 10960
Ion Ratio Lower Upper
212 100
106 13.3 10.1 15.1
104 7.5 5.4 8.2



#28
Fluoranthene
Concen: 0.16 ng
RT: 19.071 min Scan# 1819
Delta R.T. 0.000 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48

Tgt Ion:202 Resp: 11625
Ion Ratio Lower Upper
202 100
101 11.1 8.4 12.6
203 16.6 13.5 20.3





#29

Chrysene-d12

Concen: 0.40 ng

RT: 21.205 min Scan# 2120

Delta R.T. 0.000 min

Lab File: BN014516.D

Acq: 03 May 2021 15:48

Tgt Ion:240 Resp: 18163

Ion Ratio Lower Upper

240 100

120 10.7 5.8 8.8#





236 26.4 20.6 30.8

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.2



#30

Pyrene

Concen: 0.19 ng

RT: 19.439 min Scan# 1893

Delta R.T. 0.005 min

Lab File: BN014516.D

Acq: 03 May 2021 15:48




Tgt Ion:202 Resp: 11846


Ion Ratio Lower Upper

202 100

200 20.6 16.3 24.5

203 17.1 14.0 21.0









#31
Terphenyl-d14
Concen: 0.19 ng
RT: 19.652 min Scan# 1936
Delta R.T. 0.000 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:244 Resp: 8076
Ion Ratio Lower Upper
244 100
212 8.8 6.0 9.0
122 12.9 9.3 13.9



#32
Benzo(a)anthracene
Concen: 0.15 ng
RT: 21.194 min Scan# 2119
Delta R.T. 0.011 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48

Tgt Ion:228 Resp: 7470
Ion Ratio Lower Upper
228 100
226 27.1 21.7 32.5
229 21.1 15.7 23.5









#33
Chrysene
Concen: 0.19 ng
RT: 21.237 min Scan# 2123
Delta R.T. 0.000 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:228	Resp:	11107
Ion Ratio	Lower	Upper
228	100	
226	30.8	23.8 35.6
229	20.3	15.6 23.4



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.18 ng
RT: 21.130 min Scan# 2113
Delta R.T. 0.000 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48

Tgt Ion:149	Resp:	3496
Ion Ratio	Lower	Upper
149	100	
167	26.1	20.5 30.7
279	4.1	2.9 4.3





#35

Perylene-d12

Concen: 0.40 ng

RT: 23.397 min Scan# 2523

Delta R.T. 0.000 min

Lab File: BN014516.D

Acq: 03 May 2021 15:48





Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2

Tgt Ion:264	Resp:	18827
Ion Ratio	Lower	Upper
264	100	
260	25.4	19.5 29.3
265	42.4	35.4 53.0



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.16 ng




RT: 25.592 min Scan# 3164


Delta R.T. 0.003 min

Lab File: BN014516.D

Acq: 03 May 2021 15:48

Tgt Ion:276	Resp:	11058
Ion Ratio	Lower	Upper
276	100	
138	18.9	18.8 28.2
277	24.8	19.8 29.8









#37
Benzo(b)fluoranthene
Concen: 0.15 ng
RT: 22.744 min Scan# 2332
Delta R.T. 0.003 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:252 Resp: 10353
Ion Ratio Lower Upper
252 100
253 29.0 19.6 29.4
125 25.6 10.3 15.5#



#38
Benzo(k)fluoranthene
Concen: 0.17 ng
RT: 22.785 min Scan# 2344
Delta R.T. 0.000 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48

Tgt Ion:252 Resp: 13978
Ion Ratio Lower Upper
252 100
253 28.5 19.7 29.5
125 15.5 9.7 14.5#





#39

Benzo(a)pyrene

Concen: 0.18 ng

RT: 23.305 min Scan# 2496

Delta R.T. 0.003 min

Lab File: BN014516.D

Acq: 03 May 2021 15:48





Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2

Tgt Ion:252	Resp:	11277
Ion Ratio	Lower	Upper
252	100	
253	37.6	21.8 32.6#
125	64.7	14.9 22.3#



#40

Dibenzo(a,h)anthracene

Concen: 0.15 ng




RT: 25.609 min Scan# 3169


Delta R.T. 0.003 min

Lab File: BN014516.D

Acq: 03 May 2021 15:48

Tgt Ion:278	Resp:	8811
Ion Ratio	Lower	Upper
278	100	
139	19.6	13.9 20.9
279	31.1	20.7 31.1#





#41
Benzo(g,h,i)perylene
Concen: 0.17 ng
RT: 26.256 min Scan# 3358
Delta R.T. 0.000 min
Lab File: BN014516.D
Acq: 03 May 2021 15:48

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2

Tgt Ion:	276	Resp:	11793
Ion Ratio	Lower	Upper	
276	100		
277	25.1	19.6	29.4
138	23.7	17.5	26.3

