

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050322\
 Data File : BN019656.D
 Acq On : 03 May 2022 19:37
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 16 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: May 04 07:33:59 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 28 02:56:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.876	152	4596	0.400	ng	0.00
7) Naphthalene-d8	10.658	136	12968	0.400	ng	-0.01
13) Acenaphthene-d10	14.485	164	6584	0.400	ng	-0.01
19) Phenanthrene-d10	17.226	188	12646	0.400	ng	0.00
29) Chrysene-d12	21.414	240	10531	0.400	ng	# 0.00
35) Perylene-d12	23.761	264	9404	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.457	112	4138	0.365	ng	0.00
5) Phenol-d6	7.031	99	4903	0.361	ng	0.00
8) Nitrobenzene-d5	9.014	82	3865	0.419	ng	-0.01
11) 2-Methylnaphthalene-d10	12.246	152	8118	0.403	ng	-0.01
14) 2,4,6-Tribromophenol	15.974	330	765	0.363	ng	0.00
15) 2-Fluorobiphenyl	13.116	172	11507	0.408	ng	-0.01
27) Fluoranthene-d10	19.253	212	13478	0.374	ng	0.00
31) Terphenyl-d14	19.856	244	9708	0.419	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.377	88	2132	0.397	ng	97
3) n-Nitrosodimethylamine	3.687	42	1908	0.455	ng	# 91
6) bis(2-Chloroethyl)ether	7.298	93	5070	0.438	ng	99
9) Naphthalene	10.712	128	14903	0.402	ng	100
10) Hexachlorobutadiene	11.011	225	2877	0.410	ng	# 100
12) 2-Methylnaphthalene	12.322	142	9418	0.404	ng	98
16) Acenaphthylene	14.207	152	11836	0.399	ng	99
17) Acenaphthene	14.549	154	8379	0.395	ng	99
18) Fluorene	15.532	166	10358	0.399	ng	98
20) 4,6-Dinitro-2-methylph...	15.607	198	389	0.321	ng	91
21) 4-Bromophenyl-phenylether	16.426	248	3106	0.401	ng	# 72
22) Hexachlorobenzene	16.538	284	3815	0.422	ng	98
23) Atrazine	16.682	200	1618	0.353	ng	99
24) Pentachlorophenol	16.877	266	1021	0.334	ng	96
25) Phenanthrene	17.267	178	16752	0.396	ng	100
26) Anthracene	17.359	178	13267	0.379	ng	99
28) Fluoranthene	19.287	202	17237	0.377	ng	99
30) Pyrene	19.649	202	17496	0.405	ng	100
32) Benzo(a)anthracene	21.396	228	14269	0.376	ng	99
33) Chrysene	21.449	228	16494	0.397	ng	99
34) Bis(2-ethylhexyl)phtha...	21.333	149	5951	0.343	ng	# 98
36) Indeno(1,2,3-cd)pyrene	26.176	276	17149	0.380	ng	99
37) Benzo(b)fluoranthene	23.048	252	14736	0.384	ng	95
38) Benzo(k)fluoranthene	23.095	252	16125	0.417	ng	# 92
39) Benzo(a)pyrene	23.656	252	11023	0.368	ng	94
40) Dibenzo(a,h)anthracene	26.194	278	13782	0.375	ng	97
41) Benzo(g,h,i)perylene	26.910	276	13992	0.359	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed






#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.876 min Scan# 6
 Delta R.T. -0.007 min
 Lab File: BN019656.D
 Acq: 03 May 2022 19:37

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC







Tgt Ion:152 Resp: 4596
 Ion Ratio Lower Upper
 152 100
 150 152.6 122.6 183.8
 115 60.2 49.0 73.6



#2
 1,4-Dioxane
 Concen: 0.397 ng
 RT: 3.377 min Scan# 41
 Delta R.T. -0.007 min
 Lab File: BN019656.D
 Acq: 03 May 2022 19:37

Tgt Ion: 88 Resp: 2132
 Ion Ratio Lower Upper
 88 100
 43 34.0 25.2 37.8
 58 86.4 67.6 101.4









#3
n-Nitrosodimethylamine
Concen: 0.455 ng
RT: 3.687 min Scan# 84
Delta R.T. -0.014 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 42 Resp: 1908
Ion Ratio Lower Upper
42 100
74 117.3 101.8 152.6
44 6.5 8.4 12.6#



#4
2-Fluorophenol
Concen: 0.365 ng
RT: 5.457 min Scan# 329
Delta R.T. -0.007 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37

Tgt Ion: 112 Resp: 4138
Ion Ratio Lower Upper
112 100
64 59.7 47.3 70.9
63 31.8 24.6 36.8









#5
Phenol-d6
Concen: 0.361 ng
RT: 7.031 min Scan# 54
Delta R.T. -0.007 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 99 Resp: 4903
Ion Ratio Lower Upper
99 100
42 23.4 15.8 23.8
71 34.1 26.5 39.7



#6
bis(2-Chloroethyl)ether
Concen: 0.438 ng
RT: 7.298 min Scan# 584
Delta R.T. -0.007 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37

Tgt Ion: 93 Resp: 5070
Ion Ratio Lower Upper
93 100
63 81.5 65.8 98.8
95 34.0 27.4 41.2









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.658 min Scan# 94
Delta R.T. -0.010 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:136	Resp:	12968
Ion Ratio	Lower	Upper
136	100	
137	11.7	9.1 13.7
54	10.1	7.8 11.6
68	6.7	5.4 8.2



#8
Nitrobenzene-d5
Concen: 0.419 ng
RT: 9.014 min Scan# 790
Delta R.T. -0.010 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37

Tgt Ion: 82	Resp:	3865
Ion Ratio	Lower	Upper
82	100	
128	40.1	32.4 48.6
54	60.6	44.4 66.6









#9
Naphthalene
Concen: 0.402 ng
RT: 10.712 min Scan# 949
Delta R.T. -0.010 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:128 Resp: 14903
Ion Ratio Lower Upper
128 100
129 11.4 9.0 13.4
127 13.3 10.8 16.2



#10
Hexachlorobutadiene
Concen: 0.410 ng
RT: 11.011 min Scan# 977
Delta R.T. -0.010 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37

Tgt Ion:225 Resp: 2877
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.9 50.8 76.2









#11
2-Methylnaphthalene-d10
Concen: 0.403 ng
RT: 12.246 min Scan# 1218
Delta R.T. -0.011 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 8118
Ion Ratio Lower Upper
152 100
151 20.6 16.2 24.4



#12
2-Methylnaphthalene
Concen: 0.404 ng
RT: 12.322 min Scan# 1238
Delta R.T. -0.007 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37

Tgt Ion:142 Resp: 9418
Ion Ratio Lower Upper
142 100
141 87.1 71.0 106.6
115 36.5 29.6 44.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.485 min Scan# 1473
Delta R.T. -0.010 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:164 Resp: 6584
Ion Ratio Lower Upper
164 100
162 103.1 81.8 122.6
160 49.8 38.3 57.5



#14
2,4,6-Tribromophenol
Concen: 0.363 ng
RT: 15.974 min Scan# 1612
Delta R.T. -0.010 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37

Tgt Ion:330 Resp: 765
Ion Ratio Lower Upper
330 100
332 93.9 76.2 114.4
141 49.4 39.7 59.5









#15
2-Fluorobiphenyl
Concen: 0.408 ng
RT: 13.116 min Scan# 11
Delta R.T. -0.010 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:172 Resp: 11507
Ion Ratio Lower Upper
172 100
171 35.7 28.0 42.0
170 24.4 19.0 28.4



#16
Acenaphthylene
Concen: 0.399 ng
RT: 14.207 min Scan# 1446
Delta R.T. -0.010 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37

Tgt Ion:152 Resp: 11836
Ion Ratio Lower Upper
152 100
151 19.9 16.2 24.4
153 13.3 10.5 15.7









#17
Acenaphthene
Concen: 0.395 ng
RT: 14.549 min Scan# 14
Delta R.T. -0.010 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:154 Resp: 8379
Ion Ratio Lower Upper
154 100
153 112.0 89.3 133.9
152 57.6 44.1 66.1



#18
Fluorene
Concen: 0.399 ng
RT: 15.532 min Scan# 1570
Delta R.T. -0.010 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37

Tgt Ion:166 Resp: 10358
Ion Ratio Lower Upper
166 100
165 97.2 79.1 118.7
167 13.4 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.226 min Scan# 1734
Delta R.T. -0.010 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:188 Resp: 12646
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.9 8.6 12.8



#20
4,6-Dinitro-2-methylphenol
Concen: 0.321 ng
RT: 15.607 min Scan# 1577
Delta R.T. 0.000 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37

Tgt Ion:198 Resp: 389
Ion Ratio Lower Upper
198 100
182 25.3 16.9 25.3
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.401 ng
RT: 16.426 min Scan# 1656
Delta R.T. 0.000 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:248 Resp: 3106
Ion Ratio Lower Upper
248 100
250 99.4 75.5 113.3
141 50.8 79.9 119.9#



#22
Hexachlorobenzene
Concen: 0.422 ng
RT: 16.538 min Scan# 1667
Delta R.T. -0.010 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37

Tgt Ion:284 Resp: 3815
Ion Ratio Lower Upper
284 100
142 43.3 33.2 49.8
249 32.8 26.4 39.6









#23
 Atrazine
 Concen: 0.353 ng
 RT: 16.682 min Scan# 1681
 Delta R.T. -0.010 min
 Lab File: BN019656.D
 Acq: 03 May 2022 19:37




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:	200	Resp:	1618
Ion Ratio	Lower	Upper	
200	100		
173	34.9	28.5	42.7
215	45.7	37.0	55.4



#24
 Pentachlorophenol
 Concen: 0.334 ng
 RT: 16.877 min Scan# 1700
 Delta R.T. -0.010 min
 Lab File: BN019656.D
 Acq: 03 May 2022 19:37

Tgt Ion:	266	Resp:	1021
Ion Ratio	Lower	Upper	
266	100		
264	63.3	48.0	72.0
268	64.4	54.2	81.2









#25
Phenanthrene
Concen: 0.396 ng
RT: 17.267 min Scan# 1747
Delta R.T. -0.010 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:178 Resp: 16752
Ion Ratio Lower Upper
178 100
176 19.2 15.4 23.2
179 15.2 12.1 18.1



#26
Anthracene
Concen: 0.379 ng
RT: 17.359 min Scan# 1747
Delta R.T. -0.010 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37

Tgt Ion:178 Resp: 13267
Ion Ratio Lower Upper
178 100
176 18.6 14.3 21.5
179 14.9 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.374 ng
RT: 19.253 min Scan# 2106
Delta R.T. -0.008 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:212 Resp: 13478
Ion Ratio Lower Upper
212 100
106 16.3 13.5 20.3
104 9.2 7.6 11.4



#28
Fluoranthene
Concen: 0.377 ng
RT: 19.287 min Scan# 2114
Delta R.T. -0.004 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37

Tgt Ion:202 Resp: 17237
Ion Ratio Lower Upper
202 100
101 12.5 10.6 15.8
203 17.0 13.6 20.4









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.414 min Scan# 2442
Delta R.T. 0.000 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:240 Resp: 10531
Ion Ratio Lower Upper
240 100
120 13.3 17.4 26.2#
236 27.3 23.0 34.4



#30
Pyrene
Concen: 0.405 ng
RT: 19.649 min Scan# 2200
Delta R.T. -0.004 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37

Tgt Ion:202 Resp: 17496
Ion Ratio Lower Upper
202 100
200 20.9 16.7 25.1
203 17.8 14.2 21.4









#31
Terphenyl-d14
Concen: 0.419 ng
RT: 19.856 min Scan# 21
Delta R.T. -0.008 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:244 Resp: 9708
Ion Ratio Lower Upper
244 100
212 9.3 7.6 11.4
122 0.0 0.0 0.0



#32
Benzo(a)anthracene
Concen: 0.376 ng
RT: 21.396 min Scan# 2440
Delta R.T. -0.009 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37

Tgt Ion:228 Resp: 14269
Ion Ratio Lower Upper
228 100
226 27.1 21.6 32.4
229 19.7 16.5 24.7





#33

Chrysene

Concen: 0.397 ng

RT: 21.449 min Scan# 2446

Delta R.T. 0.000 min

Lab File: BN019656.D

Acq: 03 May 2022 19:37

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:228 Resp: 16494

Ion Ratio Lower Upper

228 100

226 29.7 24.6 36.8

229 20.1 15.9 23.9



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.343 ng

RT: 21.333 min Scan# 2433

Delta R.T. -0.009 min

Lab File: BN019656.D

Acq: 03 May 2022 19:37




Tgt Ion:149 Resp: 5951


Ion Ratio Lower Upper

149 100

167 26.2 21.7 32.5

279 2.7 2.8 4.2#









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.761 min Scan# 3016
Delta R.T. -0.006 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:264 Resp: 9404
Ion Ratio Lower Upper
264 100
260 26.4 21.0 31.6
265 54.4 67.9 101.9#



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.380 ng
RT: 26.176 min Scan# 3842
Delta R.T. -0.011 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37

Tgt Ion:276 Resp: 17149
Ion Ratio Lower Upper
276 100
138 32.8 26.6 40.0
277 24.6 20.0 30.0









#37
Benzo(b)fluoranthene
Concen: 0.384 ng
RT: 23.048 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:252 Resp: 14736
Ion Ratio Lower Upper
252 100
253 24.2 21.2 31.8
125 17.1 15.9 23.9



#38
Benzo(k)fluoranthene
Concen: 0.417 ng
RT: 23.095 min Scan# 2788
Delta R.T. -0.006 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37

Tgt Ion:252 Resp: 16125
Ion Ratio Lower Upper
252 100
253 23.8 21.7 32.5
125 16.6 17.0 25.4#









#39
Benzo(a)pyrene
Concen: 0.368 ng
RT: 23.656 min Scan# 2982
Delta R.T. -0.006 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	252	Resp:	11023
Ion Ratio	Lower	Upper	
252	100		
253	26.3	23.1	34.7
125	21.9	20.4	30.6



#40
Dibenzo(a,h)anthracene
Concen: 0.375 ng
RT: 26.194 min Scan# 3848
Delta R.T. -0.011 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37


Tgt Ion:	278	Resp:	13782
Ion Ratio	Lower	Upper	
278	100		
139	24.6	19.4	29.2
279	25.0	21.8	32.8





#41
Benzo(g,h,i)perylene
Concen: 0.359 ng
RT: 26.910 min Scan# 4093
Delta R.T. -0.017 min
Lab File: BN019656.D
Acq: 03 May 2022 19:37

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:	276	Resp:	13992
Ion	Ratio	Lower	Upper
276	100		
277	24.8	20.2	30.2
138	30.0	23.5	35.3

