

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050322\
 Data File : BN019648.D
 Acq On : 03 May 2022 14:51
 Operator : CG/JU
 Sample : N2622-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SP-100-20220428

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 05/04/2022
 Supervised By : Jagrut Upadhyay 05/05/2022

Quant Time: May 04 07:33:20 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 28 02:56:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.876	152	3994	0.400	ng	0.00
7) Naphthalene-d8	10.658	136	10891	0.400	ng	-0.01
13) Acenaphthene-d10	14.484	164	5303	0.400	ng	-0.01
19) Phenanthrene-d10	17.225	188	10478	0.400	ng	-0.01
29) Chrysene-d12	21.413	240	9091	0.400	ng	# 0.00
35) Perylene-d12	23.761	264	7533	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.456	112	1171	0.119	ng	0.00
5) Phenol-d6	7.038	99	761	0.065	ng	0.00
8) Nitrobenzene-d5	9.014	82	2488m	0.321	ng	-0.01
11) 2-Methylnaphthalene-d10	12.249	152	5372	0.318	ng	0.00
14) 2,4,6-Tribromophenol	15.974	330	473	0.279	ng	-0.01
15) 2-Fluorobiphenyl	13.116	172	7236	0.319	ng	-0.01
27) Fluoranthene-d10	19.253	212	12241	0.410	ng	0.00
31) Terphenyl-d14	19.855	244	8149	0.408	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.376	88	20764m	4.448	ng	Qvalue
6) bis(2-Chloroethyl)ether	7.298	93	320	0.032	ng	98
34) Bis(2-ethylhexyl)phtha...	21.333	149	815m	0.054	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050322\
Data File : BN019648.D
Acq On : 03 May 2022 14:51
Operator : CG/JU
Sample : N2622-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SP-100-20220428

Quant Time: May 04 07:33:20 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 28 02:56:24 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 05/04/2022
Supervised By :Jagrut Upadhyay 05/05/2022

